

DATAMASTER DMT MAINTENANCE LOG

Brattleboro USP

Department/Agency

100161

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
9/2/11	DataMaster Ext. Std. Solution Lot # 11-41-100 EtOH_{aq} Conc: 0.101 Range: 0.096 - 0.106 Preparation Date: 06/29/2011 Expires: 06/29/2012 OPENED: 9-2-11 [For _____ Place Label Here] Installed Per Procedure	Kirk Kimball Prog. Chief VDHL Steven Harnois
	[For RPC or Simulator Solution Change Place Label Here]	TMK
9/29/11	Updated Software	A. Bole
10/25/11	DataMaster Ext. Std. Solution Lot # 11-41-100 EtOH_{aq} Conc: 0.101 Range: 0.096 - 0.106 Preparation Date: 06/29/2011 Expires: 06/29/2012 OPENED: 10/25/11 [For RPC _____ ce Label Here]	RPC TMK
2/9/12	DataMaster Simulator Solution Lot # 12-32-100 EtOH_{aq} Conc: 0.100 Range: 0.095 - 0.105 Preparation Date: 10/26/2011 Expires: 10/26/2012 OPENED: [For RPC _____ e Label Here]	RPC TMK

DATAMASTER DMT MAINTENANCE LOG

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
	[For RPC or Simulator Solution Change Place Label Here]	
8/10/12	APM Performed	R. Driscoll
10/25/12	DataMaster Simulator Solution Lot # 12-77-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 05/08/2012 Expires: 05/08/2013 OPENED: [For RPC or Simulator Solution Change Place Label Here]	D.F.B.
2/5/13	DataMaster Simulator Solution Lot # 12-80-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 07/5/2012 Expires: 7/5/2013 OPENED: 2/5/13 RCD Detector Voltage approaching -300V. Removed from service to optimize detector voltage, recalibrate and certify. Install protocol performed RCD	R. Driscoll
7/18/13 RCD	DataMaster Simulator Solution Lot # 12-87-100 EtOHaq Conc: 0.101 Range: 0.096 - 0.106 Preparation Date: 10/31/2012 Expires: 10/31/2013 OPENED: 06/16/13 D.B. [For RPC or Simulator Solution Change Place Label Here]	R. Driscoll DMT TECH
6/16/13	D.B. completed RPC	

DATAMASTER DMT MAINTENANCE LOG

Bottleboro USP

Department/Agency

100161

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
7/10/13	Completed APR [For RPC or Simulator Solution Change Place Label Here]	R. Driscoll
09/06/13	DataMaster Simulator Solution Lot # 12-90-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 12/18/2012 Expires: 12/18/2013 OPENED: 09/06/2013 1350hrs Simulation Change DB	Dan Bennett
7/1/14	[For RPC or Simulator Solution Change Place Label Here] RPC Completed by Dan Bennette 7/21/14 RCD	R. Driscoll DMT TECH
7/21/14	Completed APR Protocol [For RPC or Simulator Solution Change Place Label Here]	R. Driscoll DMT TECH

DATAMASTER DMT MAINTENANCE LOG

Brattleboro VSP
Department/Agency

100161
DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
7/27/15	Completed APM protocol, downloaded [For RPC or Simulator Solution Change Place Label Here] data to USB drive, uploaded new software, and installed new communication PCB RCD 7/27/15	R. Driscoll DMT Tech
10/11/15	[For RPC or Simulator Solution Change Place Label Here] DataMaster Simulator Solution Lot # 15-05-100 EtOHaq Conc: 0.099 g/210L Range: 0.094 - 0.104 g/210L Preparation Date: 2/19/2015 Expires: 2/19/2016 OPENED: 10/11/15 J. Waitkus	John Waitkus
2-1-16	[For RPC or Simulator Solution Change Place Label Here] DataMaster Simulator Solution 2-1-16 Lot # 15-13-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 09/15/2015 Expires: 09/15/2016 OPENED: 2-1-16 John Waitkus	John Waitkus Test/RPC
6-3-16	[For RPC or Simulator Solution Change Place Label Here] DataMaster Simulator Solution Lot # 15-13-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 09/15/2015 Expires: 09/15/2016 OPENED: 6-3-16 John Waitkus	John Waitkus RPC

DATAMASTER DMT MAINTENANCE LOG

Westminster VSP

Department/Agency

100161
DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
7/18/16	DataMaster External Standard Solution Lot # 16-05-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 4/19/2016 Expires: 4/19/2017 <u>7/18/16 RCD</u> [For RPC or] Label Here]	R. Driscoll DMT Tech.
	Instrument moved from Brattleboro VSP to Westminster VSP. Installation and APM protocols completed.	
10/3/16	DataMaster External Standard Solution Lot # 16-04-100 EtOHaq Conc: 0.098 Range: 0.093 - 0.103 Preparation Date: 3/25/2016 Expires: 3/25/2017 OPENED: <u>KMD 10/3/16</u> [For F] Place Label Here] RPC	K. Deslauriers
2/28/17	DataMaster External Standard Solution Lot # 16-04-100 EtOHaq Conc: 0.098 Range: 0.093 - 0.103 Preparation Date: 3/25/2016 Expires: 3/25/2017 OPENED: <u>KMD 2/28/17</u> [For] Place Label Here] RPC	K. Deslauriers
4/21/17	DMT Simulator Solution Lot # 17-16-100 EtOHaq Certified Conc: 0.101 Range: 0.096 - 0.106 Preparation Date: 04/03/2017 Expires: 04/03/2018 OPENED: <u>4/21/17 RCD</u> [For RPC or] Label Here] Install post simulator update.	R. Driscoll VFL

DATAMASTER DMT MAINTENANCE LOG

Westminster VSP

Department/Agency

100161

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
6/29/17	DMT Simulator Solution Lot # 17-16-100 EtOHaq Certified Conc: 0.101 Range: 0.096 - 0.106 Preparation Date: 04/03/2017 Expires: 04/03/2018 OPENED: <u>06/29/17 JWS</u> [For RPC e Label Here] RPC	John W. White TPR.
10/17/17	DMT Simulator Solution Lot # 17-19-100 EtOHaq Certified Conc: 0.101 Range: 0.096 - 0.106 Preparation Date: 05/30/2017 Expires: 05/30/2018 OPENED: <u>10/17/17</u> [For RPC e Label Here]	Anthony S. TPR.
10/24/17	<u>Installed version 2.08 PIC chip.</u> [For RPC or Simulator Solution Change Place Label Here]	R. Triscoll VFL
2/28/18	[For RPC or Simulator Solution Change Place Label Here] DMT Simulator Solution Lot # 17-19-100 EtOHaq Certified Conc: 0.101 Range: 0.096 - 0.106 Preparation Date: 05/30/2017 Expires: 05/30/2018 OPENED: <u>2/28/18 568</u>	Anthony S. TPR.

DATAMASTER DMT MAINTENANCE LOG

Westminster VSP
Department/Agency

100161
DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
4/2/18	Completed APM protocol [For RPC or Simulator Solution Change Place Label Here]	R. Driscoll VFL
06/27/18	DMT External Std Solution Lot: 18-04-100 EtOH Certified Conc. 0.101 Range: 0.096 - 0.106 Prep Date: 2/28/18 Exp Date: 02/28/19 Opened: 06-27-18 JW 522	John Witek
10/19/18	DMT External Std Solution Lot: 18-12-100 EtOH Certified Conc. 0.100 Range: 0.095 - 0.105 Prep Date: 7/11/18 Exp Date: 07/11/19 Opened: 10/19/18 Ashlu White	Ashlu White
2/26/19	DMT External Std Solution Lot: 18-12-100 EtOH Certified Conc. 0.100 Range: 0.095 - 0.105 Prep Date: 7/11/18 Exp Date: 07/11/19 Opened: 2/26/19 [Label Here]	Ashlu White Barman Clerk

DATAMASTER DMT MAINTENANCE LOG

Department/Agency

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
3/14/19 JTV	Completed APM protocol [For RPC or Simulator Solution Change Place Label Here]	J. Dukette SOFI
06-11-19	[For RPC or Simulator Solution Change Place Label Here] DMT External Std Solution Lot: 19-04-100 EtOH Certified Conc. 0.101 Range: 0.096 - 0.106 Prep Date: 2/28/19 Exp Date: 02/28/20 Opened: 06-11-19 John Waitkus	RPC John Waitkus TPR.
10-25-19	DMT External Std Solution [For RPC or Simulator Solution Change Place Label Here] Lot: 19-04-100 EtOH Certified Conc. 0.101 Range: 0.096 - 0.106 Prep Date: 2/28/19 Exp Date: 02/28/20 Opened: 10/25/19	Ashley White Barracks Clerk
2/27/20	DMT External Std Solution [For RPC or Simulator Solution Change Place Label Here] Lot: 19-13-100 EtOH Certified Conc. 0.100 Range: 0.095 - 0.105 Prep Date: 9/23/19 Exp Date: 09/23/20 Opened: 2/27/20	A. Soule Trooper

DATAMASTER DMT MAINTENANCE LOG

Department/Agency

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
3/10/2020	Completed APM protocol [For RPC or Simulator Solution Change Place Label Here]	S Stratton (VFL)
06-02-20	DMT External Std Solution Lot: 19-13-100 EtOH Certified Conc. 0.100 Range: 0.095 - 0.105 Prep Date: 9/23/19 Exp Date: 09/23/20 Opened: 06-02-20 FAIL to PRINT showed good test	John Waitekus <i>John Waitekus</i>
10-30-20	DMT External Std Solution Lot: 20-04-100 EtOH Certified Conc. 0.101 Range: 0.096 - 0.106 Prep Date: 2/21/20 Exp Date: 02/21/21 Opened: 10-30-20 ge Place Label Here]	Ashlee White Barracks Clerk <i>Ashlee White</i>
2-26-20	DMT External Std Solution Lot: 20-11-100 EtOH Certified Conc. 0.100 Range: 0.095 - 0.105 Prep Date: 7/9/20 Exp Date: 07/09/22 Opened: 2-26-21 AAW Place Label Here]	Ashlee White - Barracks Clerk <i>Ashlee White</i>

DATAMASTER DMT MAINTENANCE LOG

Department/Agency

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
3/8/21	Completed ARM protocol [For RPC or Simulator Solution Change Place Label Here]	J. Dickett (VFL)
06/24/21	DMT External Std Solution Lot: 21-04-100 EtOH Certified Conc. 0.100 Range: 0.095 - 0.105 Prep Date: 2/12/21 Exp Date: 02/12/23 Opened: RPC	A. White 6/24/21
10/26/21	DMT External Std Solution Lot: 21-09-100 EtOH Certified Conc. 0.099 Range: 0.094 - 0.104 Prep Date: 6/4/21 Exp Date: 06/04/23 Opened: RPC 10-26-21	Ashley White - Clerk
2/24/22	DMT External Std Solution [F Lot: 21-09-100 EtOH Certified Conc. 0.099 Range: 0.094 - 0.104 Prep Date: 6/4/21 Exp Date: 06/04/23 Opened: 2-24-22 aaw	Ashley White Barack's Clerk

DATAMASTER DMT MAINTENANCE LOG

Department/Agency

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
3/2/22	Removed operator use log - no longer necessary. Contact VFL for records as necessary. [For RPC or Simulator Solution Change Place Label Here] Completed APM protocol	S. Stratton (VFL)
6/30/22	DMT External Std Solution RPC 6/30/22 Lot: 22-06-100 EtOH Certified Conc. 0.100 Range: 0.095 - 0.105 Prep Date: 3/7/22 Exp Date: 03/07/24 Opened: 6/30/22 Ashlee White [nge Place Label Here]	Ashlee White Barracks Clerk
10/30/22	DMT External Std Solution RPC 10/30/22 Lot: 22-06-100 EtOH Certified Conc. 0.100 Range: 0.095 - 0.105 Prep Date: 3/7/22 Exp Date: 03/07/24 Opened: 10/30/22 [nge Place Label Here]	ant Trooper
1/29/23	DMT External Std Solution RPC Lot: 22-12-100 EtOH Certified Conc. 0.101 Range: 0.096 - 0.106 Prep Date: 5/31/22 Exp Date: 05/31/24 Opened: 1/29/23 G. Mills [nge Place Label Here]	G. Mills Trooper

DATAMASTER DMT MAINTENANCE LOG

Department/Agency _____

DataMaster DMT Serial Number _____

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
3/1/23	Completed ALM [For RPC or Simulator Solution Change Place Label Here]	J. Dukette (VFL)
6/26/23	DMT External Std Solution Lot: 22-12-100 EtOH Certified Conc. 0.101 Range: 0.096 - 0.106 Prep Date: 5/31/22 Exp Date: 05/31/24 Opened: 6/26/23 A. White	A. White Baracks Clerk
11/4/23	DMT External Std Solution Lot: 23-07-100 EtOH Certified Conc. 0.100 Prep Date: 6/20/23 Exp Date: 06/20/25 Opened: 11-4-23	Ashley White Clerk
2/4/24	DMT External Std Solution Lot: 23-07-100 EtOH Certified Conc. 0.100 Prep Date: 6/20/23 Exp Date: 06/20/25 Opened: 2/4/24 546	TRAC MMS 546 TRACER

DATAMASTER DMT MAINTENANCE LOG

Department/Agency _____

DataMaster DMT Serial Number _____

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
2/27/24	Completed APM protocol [For RPC or Simulator Solution Change Place Label Here]	S. Stratton (VFL)
5/11/24	DMT External Std Solution Lot: 23-08-100 EtOH Certified Conc. 0.099 Prep Date: 7/19/23 Exp Date: 07/19/25 Opened: 5/11/24 <i>[Signature]</i> 546 [Change Place Label Here]	G. Mills 546 <i>[Signature]</i>
6/24/24	DMT Quality Control Solution Lot: 24-02-100 EtOH Certified Conc. 0.099 Prep Date: 4/9/24 Exp Date: 04/09/26 Opened: 6/24/24 <i>[Signature]</i> <i>RPC</i> [Change Place Label Here]	A. White Barracks Clerk
10/25/24	DMT Quality Control Solution Lot: 24-02-100 EtOH Certified Conc. 0.099 Prep Date: 4/9/24 Exp Date: 04/09/26 Opened: 10/25/24 <i>[Signature]</i> <i>RPC</i> [Change Place Label Here]	A. White Barracks Clerk

DATAMASTER DMT MAINTENANCE LOG

Department/Agency _____

DataMaster DMT Serial Number _____

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
1/17/25	<i>Solution Change</i> DMT Quality Control Solution <i>Change Place Label Here]</i> Lot: 24-17-100 EtOH Certified Conc. 0.100 Prep Date: 10/21/24 Exp Date: 10/21/26 Opened: <i>1/17/25</i> <i>AW</i>	<i>Ashlee White - Clerk</i>
2/26/25	DMT Quality Control Solution <i>RPC</i> Lot: 24-17-100 EtOH Certified Conc. 0.100 Prep Date: 10/21/24 Exp Date: 10/21/26 Opened: <i>2/26/25 Ashlee White</i> <i>Change Place Label Here]</i>	<i>Ashlee White - Clerk</i>
	[For RPC or Simulator Solution Change Place Label Here]	
	[For RPC or Simulator Solution Change Place Label Here]	