

STATE OF VERMONT  
DEPARTMENT OF PUBLIC SAFETY  
VERMONT FORENSIC LABORATORY  
P.O. BOX 47  
WATERBURY, VERMONT 05676-0047



TEL: 802-244-8788  
FAX: 802-241-5557  
E-MAIL: [dps.forensiclab@vermont.gov](mailto:dps.forensiclab@vermont.gov)  
<http://www.vfl.vermont.gov/>

---

---

**MEMORANDUM**

---

---

**TO:** DMT QA DOCUMENTS  
**FROM:** AMANDA BOLDUC, TOXICOLOGY SECTION SUPERVISOR  
**SUBJECT:** APPROVAL  
**DATE:** SEPTEMBER 11, 2024

A handwritten signature in blue ink, appearing to be "AB", is written over the "DATE:" line of the memorandum.

This memo is to authorize the refurbished DMT as fit for purpose as the evidential breath testing instrument in the State of Vermont, following the procedure outlined in TOX\_P200\_DMT Manual\_version 5. DMT instruments were refurbished by the manufacturer, Intoximeters, Inc., and new analytical and computer components were installed. Subsequent to refurbishment, three instruments were tested to develop and validate the calibration method. The DMT instrument is deemed fit for use as all acceptance criteria of the validation have been met with one exception: the in-simulator stability of the certified reference material testing is ongoing. In-simulator solution life will follow the limits defined in TOX\_P200 and TOX\_P500 until the stability study is complete.