

## DPS - DMT

---

**From:** Bolduc, Amanda  
**Sent:** Friday, January 15, 2016 9:15 AM  
**To:** Driscoll, Rob; DPS - DMT  
**Cc:** Dukette, Jeffery  
**Subject:** COMM LOG

FYI

*Amanda Bolduc, MFS*

Alcohol Program Supervisor  
Vermont Forensic Laboratory  
(802) 241-5300 (ph)  
(802) 585-0504 (cell)

---

**From:** Dukette, Jeffery  
**Sent:** Friday, January 15, 2016 9:14 AM  
**To:** Keefe, Michael  
**Cc:** Bolduc, Amanda  
**Subject:** DataMaster DMT

Hi Michael,

Rob is on the road today so I'm writing to follow up on the DMT. We noticed that the simulator solution failed high last night for one of the tests and you performed a solution change protocol early this morning. It appears that the last test run back in November also had a sim. out of range error. I wanted to see if you could run a couple tests for us for look at. We would like you to run a diagnostic test and then perform an accuracy and precision test. Until we are able to take a further look at it, the instrument will need to be put out of service. I spoke with the dept. earlier and heard that you are out of the office for the weekend. I will contact the dept. again to put the instrument out of service for the time being. Please let us know when you have had the chance to run these tests. Feel free to contact us with any questions.

Thanks,

*Jeff Dukette*  
Forensic Chemist  
Vermont Forensic Laboratory  
45 State Drive  
Waterbury, VT 05671  
(802) 241-5275  
[Jeffery.Dukette@vermont.gov](mailto:Jeffery.Dukette@vermont.gov)