

9/20/11 - 12/15/23

DATAMASTER DMT MAINTENANCE LOG

Springfield Pd

102606

Department/Agency

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
9/20/11	DataMaster Ext. Std. Solution Lot # 11-41-100 EtOH_{aq} Conc: 0.101 Range: 0.096 - 0.106 Preparation Date: 06/29/2011 Expires: 06/29/2012 OPENED: 9/20/2011 Instal Per Protocol	Steve Hannaford Program Chief 9-20-11
10/24/11	DataMaster Ext. Std. Solution Lot # 11-41-100 EtOH_{aq} Conc: 0.101 Range: 0.096 - 0.106 Preparation Date: 06/29/2011 Expires: 06/29/2012 OPENED:	Shirley Williams
11/7/11	Software update installed	W. L. L. C. Prog. Chief. N.D.H.L.
2/14/12	[For RPC or Simulator Solution Change Place Label Here] DataMaster Simulator Solution Lot # 12-32-100 EtOH_{aq} Conc: 0.100 Range: 0.095 - 0.105 Preparation Date: 10/26/2011 Expires: 10/26/2012 OPENED: 02-14-12 RPC	CAN 8/9/11
	[For RPC or Simulator Solution Change Place Label Here]	
3-11-12	Time change for DST, one hour ahead.	CAN

DATAMASTER DMT MAINTENANCE LOG

SPRINGFIELD PD

102606

Department/Agency

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
6/22/12	<u>DataMaster Simulator Solution</u> Lot # 12-73-100 EtOH _{aq} Conc: 0.098 Range: 0.093 - 0.103 Preparation Date: 04/27/2012 Expires: 04/27/2013 [For RPC] OPENED: 06-22-12	
	Attempted RPC. Solution average out of range. Will attempt 2nd solution change.	CAN
6/22/12	<u>DataMaster Simulator Solution</u> Lot # 12-45-100 EtOH _{aq} Conc: 0.100 Range: 0.095 - 0.105 Preparation Date: 12/12/2011 Expires: 12/12/2012 [For RPC] OPENED: 06-22-12 June RPC	CAN
6/26/12	[For RPC or Simulator Solution Change Place Label Here] Placed DMT out of service. Sim out of range.	
6/28/12	DMT REMOVED FOR SERVICE [For RPC or Simulator Solution Change Place Label Here] AT VFL.	TRISHA CONTI ALC PROG. SUPERVISOR VFL

DATAMASTER DMT MAINTENANCE LOG

SPRINGFIELD PD

Department/Agency

100144 *

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
7/3/12	INSTRUMENT # 100144 [For RPC or Simulator Solution Change Place Label Here] INSTALLED. <u>DataMaster Simulator Solution</u> Lot # 12-45-100 EtOH_{aq} Conc: 0.100 Range: 0.095 - 0.105 Preparation Date: 12/12/2011 Expires: 12/12/2012 OPENED: 7/3/2012 TLC	TEXNA CONTI ALC PROG. SUPERVISOR VFL
10/20/12	<u>DataMaster Simulator Solution</u> Lot # 12-73-100 EtOH_{aq} Conc: 0.098 Range: 0.093 - 0.103 Preparation Date: 04/27/2012 Expires: 04/27/2013 OPENED: 10-20-12 [For] Place Label Here]	
10/20/12	RFI test voltage read .0008 no RFI detected aborted test contacted	
10/23/12	Reran RPC protocol. RFI TEST [For RPC or Simulator Solution Change Place Label Here] Voltage stable @ .006 when keying main radio console. RPC passed	Robert Driscoll DMT TECH.
2-13-13	<u>DataMaster Simulator Solution</u> Lot # 12-77-100 EtOH_{aq} Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 05/08/2012 Expires: 05/08/2013 OPENED: 02-13-13 [For RP] Place Label Here]	CAN
	Feb. RPC all good	

DATAMASTER DMT MAINTENANCE LOG

SPD

Department/Agency

100144

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
6/12/13	performed ARM protocol [For RPC or Simulator Solution Change Place Label Here]	R. Driscoll DMT TECH
6/21/13	DataMaster Simulator Solution Lot # 13-03-100 EtOHaq Conc: 0.098 Range: 0.093 - 0.103 Preparation Date: 04/03/2013 Expires: 04/03/2014 OPENED: 6-21-13 [For RPC (Label Here)] June RPC	Aul
7/11/13	[For RPC or Simulator Solution Change Place Label Here] Adjusted RFI sensitivity	R. Driscoll DMT TECH
10/31/13	DataMaster Simulator Solution Lot # 13-03-100 EtOHaq Conc: 0.098 Range: 0.093 - 0.103 Preparation Date: 04/03/2013 Expires: 04/03/2014 OPENED: 10/31/13 2:37 PM [For RPC or Simulator Solution Change Place Label Here] October RPC	Gregory J. Molgano Sergeant

DATAMASTER DMT MAINTENANCE LOG

Springfield Police
Department/Agency

100144
DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
11/5/10	[For RPC or Simulator Solution Change Place Label Here] Failed to print DMT supervisor reset DMT but still no printing. Reconnected printer USB and rebooted DMT. Printer now working properly.	R. Driscoll
	[For RPC or Simulator Solution Change Place Label Here] DataMaster Simulator Solution Lot # 13-19-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 10/25/2013 Expires: 10/25/2014 OPENED: 2-13-14 AUC February '14 RPC 2x trouble w/ sim head that's on jar, inadvertently aborted test once	A. Remond
6/12/14	Completed ARM Protocol [For RPC or Simulator Solution Change Place Label Here] Also adjusted clock ahead 5 mins. to correct time.	R. Driscoll DMT Tech.
6/17/14	[For RPC or Simulator Solution Change Place Label Here] DataMaster Simulator Solution Lot # 13-19-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 10/25/2013 Expires: 10/25/2014 OPENED: RPC Sol. Change	Molano ES3M

DATAMASTER DMT MAINTENANCE LOG

Department/Agency



DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
10/12/14	DataMaster Simulator Solution Lot # 14-04-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 5/27/2014 Expires: 5/27/2015 [For RP _____ _____ice Label Here] OPENED: <u>RPC/Sol. Change</u>	Molgano <i>[Signature]</i>
2/11/15	DataMaster Simulator Solution Lot # 14-04-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 5/27/2014 Expires: 5/27/2015 [_____ _____nge Place Label Here] OPENED: <u>RPC/Sol. Change</u>	Molgano <i>[Signature]</i>
5/22/15	DataMaster Simulator Solution Lot # 15-01-100 EtOHaq Conc: 0.098 g/210L Range: 0.093 - 0.103 g/210L Preparation Date: 1/21/2015 Expires: 1/21/2016 [_____ _____ange Place Label Here] OPENED: <u>Solution Change</u>	Molgano <i>[Signature]</i>
6/26/15	DataMaster Simulator Solution Lot # 15-01-100 EtOHaq Conc: 0.098 g/210L Range: 0.093 - 0.103 g/210L Preparation Date: 1/21/2015 Expires: 1/21/2016 [_____ _____ange Place Label Here] OPENED: <u>RPC/Sol. Change</u>	Molgano <i>[Signature]</i>

DATAMASTER DMT MAINTENANCE LOG

Department/Agency

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
	[For RPC or Simulator Solution Change Place Label Here]	
7/25/15	Changed time. DMT was about 11 minutes slow. Time synced to dispatch console	
8/12/15	Download data to USB, uploaded new software, installed new com-pcs and completed APM protocol	R. Driscoll DMT TECH
10/18/15	DataMaster Simulator Solution Lot # 15-13-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 09/15/2015 Expires: 09/15/2016 OPENED: _____ [For RPC (Label Here)] RPC / Solution Change	
2/5/16	DataMaster Simulator Solution Lot # 15-12-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 07/28/2015 Expires: 07/28/2016 OPENED: _____ [For RPC (Label Here)] RPC / Solution Change	

DATAMASTER DMT MAINTENANCE LOG

Department/Agency

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
6/23/16	DataMaster External Standard Solution Lot # 15-25-100 EtOHaq Conc: 0.100 Range: 0.095 - 0.105 Preparation Date: 11/30/2015 Expires: 11/30/2016 [For RPC or Simulator Solution Change Place Label Here] OPENED: RPC / Solution Change	AWL
8/24/16	Completed APM Protocol [For RPC or Simulator Solution Change Place Label Here]	R. Driscoll DMT Tech.
10/23/16	DataMaster External Standard Solution Lot # 15-25-100 EtOHaq Conc: 0.100 Range: 0.095 - 0.105 Preparation Date: 11/30/2015 Expires: 11/30/2016 [For RPC or Simulator Solution Change Place Label Here] OPENED: 10/23/16 833M RPC / Solution Change	833M
2/20/17	DataMaster External Standard Solution Lot # 16-05-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 4/19/2016 Expires: 4/19/2017 [For RPC or Simulator Solution Change Place Label Here] OPENED: RPC / Solution Change	833M

DATAMASTER DMT MAINTENANCE LOG

Springfield PD
Department/Agency

1000000000
DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
4/10/17	DMT Simulator Solution Lot # 17-16-100 EtOHaq Certified Conc: 0.101 Range: 0.096 - 0.106 Preparation Date: 04/03/2017 Expires: 04/03/2018 OPENED: 4/10/17 <i>RC</i> [For RPC or Simulator Solution Change Place Label Here] <i>Install post simulator update</i>	<i>R. Driscoll</i> VFL
6/28/17	DMT Simulator Solution Lot # 17-16-100 EtOHaq Certified Conc: 0.101 Range: 0.096 - 0.106 Preparation Date: 04/03/2017 Expires: 04/03/2018 OPENED: 6/28/17 <i>ESM</i> [For RPC or Simulator Solution Change Place Label Here] <i>RPC / Solution Change</i>	<i>[Signature]</i>
10/22/17	DMT Simulator Solution Lot # 17-19-100 EtOHaq Certified Conc: 0.101 Range: 0.096 - 0.106 Preparation Date: 05/30/2017 Expires: 05/30/2018 OPENED: 10-22-17 <i>AK</i> [For RPC or Simulator Solution Change Place Label Here] <i>RPC / Solution Change</i>	<i>[Signature]</i>
10/24/17	<i>Installed version 2.08 PIC chip</i> [For RPC or Simulator Solution Change Place Label Here]	<i>R. Driscoll</i> VFL

DATAMASTER DMT MAINTENANCE LOG

Springfield PD
Department/Agency

09100144
DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
2/20/18	DMT Simulator Solution Lot # 17-19-100 EtOHaq Certified Conc: 0.101 Range: 0.096 - 0.106 Preparation Date: 05/30/2017 Expires: 05/30/2018 OPENED: 2-20-18 AWC [For RPC or Simulator Solution Change Place Label Here] RPC / Solution Change	AWC
4/2/18	Completed APC protocol. [For RPC or Simulator Solution Change Place Label Here]	R. Driscoll VFL
6/27/18	DMT Simulator Solution Lot # 17-21-100 EtOHaq Certified Conc: 0.100 Range: 0.095 - 0.105 Preparation Date: 09/20/2017 Expires: 09/20/2018 OPENED: 6/27/18 JGM [For RPC or Simulator Solution Change Place Label Here] RPC / Solution Change	JGM
10/6/18	DMT External Std Solution Lot: 18-11-100 EtOH Certified Conc. 0.101 Range: 0.096 - 0.106 Prep Date: 6/26/18 Exp Date: 06/26/19 Opened: 10/6/18 JGM [For RPC or Simulator Solution Change Place Label Here] RPC / Solution Change	JGM

DATAMASTER DMT MAINTENANCE LOG

Springfield P.D.
Department/Agency

100148
DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
	[For RPC or Simulator Solution Change Place Label Here]	
10/27/18	Communication error with printer - restarted DataMaster which cleared Problem	(Aul)
2/20/19	DMT External Std Solution Lot: 18-11-100 EtOH Certified Conc. 0.101 Range: 0.096 - 0.106 Prep Date: 6/26/18 Exp Date: 06/26/19 Opened: <u>2-20-19</u> Aul [F] ge Place Label Here] RPC & Solution change Passed all tests	(Aul)
2/27/19	Completed APM protocol [For RPC or Simulator Solution Change Place Label Here]	J Dukette (UFL)
6/5/19	DMT External Std Solution Lot: 19-01-100 EtOH Certified Conc. 0.101 Range: 0.096 - 0.106 Prep Date: 1/15/19 Exp Date: 01/15/20 Opened: <u>6/5/19</u> JSB [F] ge Place Label Here] RPC / Solution Change	(JSB)

DATAMASTER DMT MAINTENANCE LOG

Springfield P.D.
Department/Agency

100144
DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
10/29/19	DMT External Std Solution [For RPC Lot: 19-01-100 EtOH Certified Conc. 0.101 se Label Here] Range: 0.096 - 0.106 Prep Date: 1/15/19 Exp Date: 01/15/20 Opened: 10-29-19 RPC / Solution Change	Solution: AWL RPC: 83m
2/13/2020	Completed APM protocol [For RPC or Simulator Solution Change Place Label Here]	J. Dickette (VFL)
2/19/20	DMT External Std Solution Lot: 20-03-100 EtOH Certified Conc. 0.100 Range: 0.095 - 0.105 Prep Date: 1/29/20 Exp Date: 01/29/21 Opened: RPC/Solution Change	83m
6/16/20	DMT External Std Solution Lot: 19-13-100 EtOH Certified Conc. 0.100 [For RPC Range: 0.095 - 0.105 se Label Here] Prep Date: 9/23/19 Exp Date: 09/23/20 Opened: 6-16-20 <u>AWL</u> RPC / Solution change	AWL
10/24/2020	DMT External Std Solution Lot: 20-11-100 EtOH Certified Conc. 0.100 [For RPC Range: 0.095 - 0.105 se Label Here] Prep Date: 7/9/20 Exp Date: 07/09/22 Opened: 10-24-2020 <u>AWL</u> Solution change / RPC	AWL

DATAMASTER DMT MAINTENANCE LOG

Springfield Police
Department/Agency

100144
DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
2/10/21	DMT External Std Solution Lot: 20-11-100 EtOH Certified Conc. 0.100 Range: 0.095 - 0.105 Prep Date: 7/9/20 Exp Date: 07/09/22 Opened: 2-10-21 (Aut) [For RPC or Simulator Solution Change Place Label Here]	(Aut) SPD
2/10/21	RPC	(Aut) SPD
2/10/21	Completed APM protocol [For RPC or Simulator Solution Change Place Label Here]	S. Stratton (VFL)
6/7/21	DMT External Std Solution Lot: 21-06-100 EtOH Certified Conc. 0.100 Range: 0.095 - 0.105 Prep Date: 3/25/21 Exp Date: 03/25/23 Opened: 6/7/21 JS3M June RPC	JS3M Molganio
10/12/21	DMT External Std Solution Lot: 21-06-100 EtOH Certified Conc. 0.100 Range: 0.095 - 0.105 Prep Date: 3/25/21 Exp Date: 03/25/23 Opened: 10/12/21 JS3M October RPC Change Place Label Here]	JS3M Molganio
2/9/22	Completed APM protocol [For RPC or Simulator Solution Change Place Label Here]	J. Dukette (VFL)

DATAMASTER DMT MAINTENANCE LOG

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
2.24.22	RPC [For Range: 0.094 - 0.104 Prep Date: 10/8/21 Opened: 2.24.22 Raw DMT External Std Solution Lot: 21-12-100 EtOH Certified Conc. 0.099 Exp Date: 10/08/23 Place Label Here]	Ashlee White VSP Barracks Clerk
6/4/22	RPC Solution Change Range: 0.094 - 0.104 Prep Date: 10/8/21 Opened: 6/4/22 JEF DMT External Std Solution Lot: 21-12-100 EtOH Certified Conc. 0.099 Exp Date: 10/08/23 Place Label Here]	Fitzgerald SPD
10/3/22	RPC Range: 0.095 - 0.105 Prep Date: 1/26/22 Opened: 10/3/22 JEF DMT External Std Solution Lot: 22-04-100 EtOH Certified Conc. 0.100 Exp Date: 01/26/24 Place Label Here]	Fitzgerald SPD
1/1/23	BSC Range: 0.098 Prep Date: 10/5/22 Opened: 1/1/23 RPC DMT External Std Solution Lot: 22-14-100 EtOH Certified Conc. 0.098 Exp Date: 10/05/24 Place Label Here]	Clark SPD

DATAMASTER DMT MAINTENANCE LOG

Department/Agency

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
2/6/23	Completed APM protocol [For RPC or Simulator Solution Change Place Label Here]	S. Stratton (VFL)
3/8/23	DMT External Std Solution Lot: 22-14-100 EtOH Certified Conc. 0.098 Prep Date: 10/5/22 Exp Date: 10/05/24 Opened: 3/8/23	Clark SPD
6/17/23	DMT External Std Solution Lot: 23-01-100 EtOH Certified Conc. 0.100 Prep Date: 1/25/23 Exp Date: 01/25/25 Opened: 6/17/23 JEF	RPC FITZGERALD SPD
9/9/23	DMT External Std Solution Lot: 23-01-100 EtOH Certified Conc. 0.100 Prep Date: 1/25/23 Exp Date: 01/25/25 Opened:	RPC SSC RPC SPD

SPRINGFIELD

100144

DataMaster DMT Serial Number

ALC 803 REV 03 08/2011