DATAMASTER DMT 100157

VERMONT DEPARTMENT OF HEALTH LABORATORY

PERFORMANCE TESTING RESULTS

August 2010

Table of Contents

Section 1: Results	1
Section 2: NPAS Tests	5
Section 3: Power-Up Test	6
Section 4: Linearity	11
Section 5: Interference	43
Section 6: Low Level Accuracy	51
Section 7: Breath Volume	53
Section 8: Software Testing	55
Section 9: Error Reports	57
Section 10: Live Subject Tests	83
Section 11: Communication Notes	85

Section 1: Results


100157 Linearity
GC Certified Value
0.450
0.400
0.350
0.300
0.250
0.200
0.150
0.100
0.050
0.000
0.000
0.050
0.100
0.150
0.200
0.250
0.300
0.350
0.400
0.450
DMT Result
y = 0.9642x + 0.0004
R2 = 0.9997

Operator	Date	Lot #	GC Value	DMT Result	Std Dev	% Recovery	
SH	9/10/2010	11-05-100	0.102	0.102	0.0005	100.0	Outside 5%
SH	9/13/2010	10-31-160	0.160	0.157	0.0004	98.1	Outside 10%
		10-37-080	0.079	0.078	0.0004	98.7	
		10-34-020	0.020	0.018	0.0004	90.0	
		10-32-400	0.373	0.362	0.0006	97.1	
SH	9/15/2010	10-31-160	0.160	0.154	0.0003	96.3	
		10-37-080	0.079	0.077	0.0004	97.5	
		10-32-400	0.373	0.358	0.0007	96.0	
		10-34-020	0.020	0.017	0.0004	85.0	depleted
SH	9/17/2010	10-34-020	0.020	0.019	0.0004	95.0	
SH	9/21/2010	10-37-080	0.079	0.076	0.0005	96.2	
		10-32-400	0.373	0.357	0.0009	95.7	
		11-06-160	0.162	0.158	0.0007	97.5	
		10-34-020	0.020	0.018	0.0004	90.0	
SH	9/21/2010	10-34-020	0.020	0.019	0.0005	95.0	
		11-06-160	0.162	0.159	0.0005	98.1	
		10-32-400	0.373	0.357	0.0008	95.7	
		10-37-080	0.079	0.078	0.0006	98.7	
SH	9/22/2010	10-37-080	0.079	0.076	0.0003	96.2	
		10-32-400	0.373	0.356	0.0008	95.4	
		11-06-160	0.162	0.157	0.0004	96.9	
		10-34-020	0.020	0.019	0.0005	95.0	
SH	9/22/2010	10-37-080	0.079	0.077	0.0006	97.5	
		10-34-020	0.020	0.021	0.0010	105.0	
		11-07-400	0.412	0.404	0.0005	98.1	
		11-06-160	0.162	0.156	0.0007	96.3	
				Average	0.0005	96.2	
				Std Dev	0.0002	3.63	
				Range	0.0000 - 0.0010	90.0 - 105.0	



Operator	Date	Concentration	Lot Number	DMT Result	Apparent EtOH	N = 10
SH	09/13/10	.01 Acetone in .08 EtOH	10-33-08A	9 OUT OF 10	0.08	
SH	09/13/10	.01 Acetone in .04 EtOH	10-35-04A1	INTERFERENCE	0.04	
SH	09/13/10	.02 Acetone in .04 EtOH	10-35-04A2	INTERFERENCE	0.04	
SH	09/13/10	.04 MeOH in water	11-08-04M	INTERFERENCE	0.02	
SH	09/10/10	.04 MeOH in .08 EtOH	10-10-08M4	INTERFERENCE	0.1	
SH	09/13/10	.04 IPA in water	10-26-04I	INTERFERENCE	0.02	
SH	09/13/10	.04 IPA in .08 EtOH	09-30-08I	INTERFERENCE	0.1	

Section 2: NPAS Tests

Section 3: Power-Up Tests

CALIBRATION REPORT

DataMaster DMT: 100157
Calibration Date: 09/10/2010
Calibrated by: STEVEN E HARNOIS
Lot: 11-03-100

VERMONT
DEPARTMENT OF HEALTH

```
Ca   =  0.102000
CAL  =  0.997341     0.800 <= CAL < 1.200
b1   =  0.000460     0.000 <=  b1 < 0.004
b2   =  0.003993     0.001 <=  b2 < 0.012
b3   =  0.000021     0.000 <=  b3 < 0.004
Xq   =  0.108668     0.050 <=  Xq < 0.200
a21  =  1.179005     1.050 <= a21 < 1.300
a31  =  0.418157     0.300 <= a31 < 0.800
```

Performed by x _____ Date 09/10/2010

ACCURACY AND PRECISION TEST

STATE OF VERMONT

DataMaster DMT: 100157

Date: 09/10/2010
Time: 08:33:15

SUPERVISOR NAME: SEH CAL CHECK

SOLUTION LOT #: 11-05-100
SOLUTION CONCENTRATION: 0.102

BLANK TEST	0.000	08:34
CALIBRATION CHECK	PASSED	08:34
SIMULATOR VAPOR	0.102	08:34
SIMULATOR VAPOR	0.102	08:35
SIMULATOR VAPOR	0.102	08:36
SIMULATOR VAPOR	0.103	08:37
SIMULATOR VAPOR	0.103	08:38
SIMULATOR VAPOR	0.102	08:39
SIMULATOR VAPOR	0.102	08:40
SIMULATOR VAPOR	0.103	08:41
SIMULATOR VAPOR	0.103	08:42
SIMULATOR VAPOR	0.102	08:43
BLANK TEST	0.000	08:44

Average = 0.102
Std Dev = 0.0005



CERTIFICATION REPORT

DataMaster DMT: 100157
Calibration Date: 09/10/2010
Certification Date: 09/10/2010
Certified by: STEVEN E HARNOIS

VERMONT
DEPARTMENT OF HEALTH

Diagnostic Results

VERSIONS
DMT 1.05
PIC 2.06
Modem 1.06
Questions 1.00
Reports 1.02

TEMPERATURES
Sample Chamber = 49.26°C
Breath Tube = 46.96°C

SETTINGS
Lamp Voltage = 2.10 V
Cooler Voltage = 1.79 V
Bias Voltage = 80.3 V
Chopper Freq = 538.1 Hz

PUMP INFO
Flow Rate = 5.497 L/M

DETECTOR INFO
PUMP ON OFF
MAX(V) -0.0791 -0.0746
MIN(V) -0.0807 -0.0760

FILTER INFO
Filter 1 -0.075 Zero = true
Filter 2 0.133 Zero = true
Filter 3 1.369 Zero = true

CALIBRATION CHECK
Xq = 0.1079 0.65%

Options

Printer: yes
Number of Copies: 3
Number of Supervisor Tests: 10
Tolerance Check: yes
Data Collection: yes
Units: g/210L
Simulator Check: yes
Uses Wet Bath Simulator: yes
Simulator Nominal: 0.102
Digital Simulator: none
Simulator Before: yes
Simulator Between: yes
Simulator After: no
Number of Subject Tests: 2
AskQuestions: yes
Query Refusal: yes
Alcohol Display: yes
Show Two Digits: no
Volume Display: yes
Number of Calibration Tests: 1
Calibrate with Wet Bath: yes
Calibration Nominal: 0.100

Linearity Check Results

True Value	Reported Average	Std Dev
0.020 g/210L	0.018 g/210L	0.000
Lot # 10-34-020		
0.079 g/210L	0.077 g/210L	0.001
Lot # 10-37-080		
0.160 g/210L	0.156 g/210L	0.000
Lot # 10-31-160		
0.373 g/210L	0.363 g/210L	0.002
Lot # 10-32-400		

$R^2 = 1.0000$



Acetone Interference Test

Lot # 10-33-08A

Interference Detected

Mouth Alcohol Test

Mouth Alcohol Detected



RF Interference Test

RFI detected



Sample Acceptance Test

Passed

CERTIFICATION PASSED

Performed by _____ **Date** 09/10/2010

Reviewed by _____ **Date** _____

Section 4: Linearity

DIAGNOSTIC RESULT

DataMaster DMT: 100157
Location:
Calibration Date: 09/10/2010
Certification Date: 09/10/2010
Installation Date: 09/10/2010
Test Date: 09/13/2010
Time: 13:57

VERMONT
DEPARTMENT OF HEALTH

```
VERSIONS
DMT 1.05
PIC 2.06
Modem 1.06
Questions 1.00
Reports 1.02

TEMPERATURES
Sample Chamber = 48.80°C
Breath Tube    = 48.10°C

SETTINGS
Lamp Voltage   = 2.10 V
Cooler Voltage = 1.79 V
Bias Voltage   = 80.4 V
Chopper Freq   = 545.3 Hz

PUMP INFO
Flow Rate = 5.259 L/M

DETECTOR INFO
PUMP     ON        OFF
MAX(V) -0.0571   -0.0602
MIN(V) -0.0669   -0.0624

FILTER INFO
Filter 1 -0.059  Zero = true
Filter 2  0.150  Zero = true
Filter 3  1.410  Zero = true

CALIBRATION CHECK
Xq = 0.1081  0.49%
```

ACCURACY AND PRECISION TEST

STATE OF VERMONT

DataMaster DMT: 100157

Date: 09/13/2010
Time: 11:33:05

SUPERVISOR NAME: STEVEN E HARNOIS

SOLUTION LOT #: 10-31-160
SOLUTION CONCENTRATION: 0.160

BLANK TEST	0.000	11:34
CALIBRATION CHECK	PASSED	11:34
SIMULATOR VAPOR	0.156	11:34
SIMULATOR VAPOR	0.157	11:35
SIMULATOR VAPOR	0.157	11:36
SIMULATOR VAPOR	0.157	11:37
SIMULATOR VAPOR	0.157	11:38
SIMULATOR VAPOR	0.157	11:39
SIMULATOR VAPOR	0.157	11:40
SIMULATOR VAPOR	0.157	11:41
SIMULATOR VAPOR	0.158	11:42
SIMULATOR VAPOR	0.157	11:43
BLANK TEST	0.000	11:44

Average = 0.157
Std Dev = 0.0004



ACCURACY AND PRECISION TEST

STATE OF VERMONT

DataMaster DMT: 100157

Date: 09/13/2010
Time: 11:47:49

SUPERVISOR NAME: STEVEN E HARNOIS

SOLUTION LOT #: 10-37-080
SOLUTION CONCENTRATION: 0.079

BLANK TEST	0.000	11:48
CALIBRATION CHECK	PASSED	11:48
SIMULATOR VAPOR	0.078	11:49
SIMULATOR VAPOR	0.078	11:50
SIMULATOR VAPOR	0.078	11:51
SIMULATOR VAPOR	0.079	11:52
SIMULATOR VAPOR	0.079	11:53
SIMULATOR VAPOR	0.079	11:54
SIMULATOR VAPOR	0.079	11:55
SIMULATOR VAPOR	0.079	11:56
SIMULATOR VAPOR	0.079	11:57
SIMULATOR VAPOR	0.079	11:58
BLANK TEST	0.000	11:58

Average = 0.078
Std Dev = 0.0004



ACCURACY AND PRECISION TEST

STATE OF VERMONT

DataMaster DMT: 100157

Date: 09/13/2010
Time: 12:21:27

SUPERVISOR NAME: STEVEN E HARNOIS

SOLUTION LOT #: 10-34-020
SOLUTION CONCENTRATION: 0.020

BLANK TEST	0.000	12:22
CALIBRATION CHECK	PASSED	12:22
SIMULATOR VAPOR	0.017	12:22
SIMULATOR VAPOR	0.018	12:23
SIMULATOR VAPOR	0.019	12:24
SIMULATOR VAPOR	0.018	12:25
SIMULATOR VAPOR	0.018	12:26
SIMULATOR VAPOR	0.018	12:27
SIMULATOR VAPOR	0.018	12:28
SIMULATOR VAPOR	0.018	12:29
SIMULATOR VAPOR	0.018	12:30
SIMULATOR VAPOR	0.018	12:31
BLANK TEST	0.000	12:32

Average out of range!

Average = 0.018
Std Dev = 0.0004



ACCURACY AND PRECISION TEST

STATE OF VERMONT

DataMaster DMT: 100157

Date: 09/13/2010
Time: 12:37:49

SUPERVISOR NAME: STEVEN E HARNOIS

SOLUTION LOT #: 10-32-400
SOLUTION CONCENTRATION: 0.373

BLANK TEST	0.000	12:38
CALIBRATION CHECK	PASSED	12:38
SIMULATOR VAPOR	0.363	12:39
SIMULATOR VAPOR	0.364	12:40
SIMULATOR VAPOR	0.363	12:41
SIMULATOR VAPOR	0.362	12:42
SIMULATOR VAPOR	0.362	12:43
SIMULATOR VAPOR	0.362	12:44
SIMULATOR VAPOR	0.362	12:45
SIMULATOR VAPOR	0.362	12:46
SIMULATOR VAPOR	0.362	12:47
SIMULATOR VAPOR	0.362	12:48
BLANK TEST	0.000	12:49

Average = 0.362
Std Dev = 0.0006



DIAGNOSTIC RESULT

DataMaster DMT: 100157
Location:
Calibration Date: 09/10/2010
Certification Date: 09/10/2010
Installation Date: 09/10/2010
Test Date: 09/15/2010
Time: 05:53

VERMONT
DEPARTMENT OF HEALTH

```
VERSIONS
DMT 1.05
PIC 2.06
Modem 1.06
Questions 1.00
Reports 1.02

TEMPERATURES
Sample Chamber = 49.03°C
Breath Tube    = 44.19°C

SETTINGS
Lamp Voltage   = 2.10 V
Cooler Voltage = 1.79 V
Bias Voltage   = 80.3 V
Chopper Freq   = 544.1 Hz

PUMP INFO
Flow Rate = 5.348 L/M

DETECTOR INFO
PUMP     ON        OFF
MAX(V) -0.1163   -0.1126
MIN(V) -0.1177   -0.1150

FILTER INFO
Filter 1 -0.113  Zero = true
Filter 2  0.096  Zero = true
Filter 3  1.347  Zero = true

CALIBRATION CHECK
Xq = 0.1082  0.41%
```

ACCURACY AND PRECISION TEST

STATE OF VERMONT

DataMaster DMT: 100157

Date: 09/15/2010
Time: 06:26:59

SUPERVISOR NAME: STEVEN E HARNOIS

SOLUTION LOT #: 10-31-160
SOLUTION CONCENTRATION: 0.160

BLANK TEST	0.000	06:27
CALIBRATION CHECK	PASSED	06:28
Xq = 0.1078 (0.73%)		
SIMULATOR VAPOR	0.154	06:28
X[1] = 0.1540 (0.0011) (-0.0003)		
SIMULATOR VAPOR	0.154	06:29
X[1] = 0.1541 (0.0003) (-0.0017)		
SIMULATOR VAPOR	0.154	06:30
X[1] = 0.1540 (0.0019) (-0.0011)		
SIMULATOR VAPOR	0.154	06:31
X[1] = 0.1540 (0.0019) (-0.0021)		
SIMULATOR VAPOR	0.154	06:32
X[1] = 0.1546 (0.0019) (-0.0024)		
SIMULATOR VAPOR	0.155	06:33
X[1] = 0.1550 (0.0019) (-0.0022)		
SIMULATOR VAPOR	0.154	06:34
X[1] = 0.1549 (0.0018) (-0.0022)		
SIMULATOR VAPOR	0.154	06:35
X[1] = 0.1545 (0.0016) (-0.0023)		
SIMULATOR VAPOR	0.154	06:36
X[1] = 0.1546 (0.0019) (-0.0022)		
SIMULATOR VAPOR	0.154	06:37
X[1] = 0.1546 (0.0017) (-0.0020)		
BLANK TEST	0.000	06:38

Average = 0.154
Std Dev = 0.0003



ACCURACY AND PRECISION TEST

STATE OF VERMONT

DataMaster DMT: 100157

Date: 09/15/2010
Time: 14:11:02

SUPERVISOR NAME: STEVEN E HARNOIS

SOLUTION LOT #: 10-37-080
SOLUTION CONCENTRATION: 0.079

BLANK TEST	0.000	14:11
CALIBRATION CHECK	PASSED	14:12
Xq = 0.1086 (0.00%)		
SIMULATOR VAPOR	0.077	14:12
X[1] = 0.0777 (0.0002) (0.0007)		
SIMULATOR VAPOR	0.077	14:13
X[1] = 0.0779 (-0.0003) (-0.0002)		
SIMULATOR VAPOR	0.078	14:14
X[1] = 0.0781 (-0.0006) (-0.0004)		
SIMULATOR VAPOR	0.078	14:15
X[1] = 0.0782 (-0.0006) (0.0000)		
SIMULATOR VAPOR	0.078	14:16
X[1] = 0.0784 (-0.0002) (-0.0002)		
SIMULATOR VAPOR	0.078	14:17
X[1] = 0.0783 (-0.0010) (-0.0006)		
SIMULATOR VAPOR	0.078	14:18
X[1] = 0.0783 (-0.0007) (-0.0001)		
SIMULATOR VAPOR	0.077	14:19
X[1] = 0.0778 (-0.0015) (-0.0013)		
SIMULATOR VAPOR	0.078	14:20
X[1] = 0.0780 (-0.0012) (-0.0002)		
SIMULATOR VAPOR	0.078	14:21
X[1] = 0.0781 (-0.0008) (-0.0007)		
BLANK TEST	0.000	14:22

Average = 0.077
Std Dev = 0.0004



ACCURACY AND PRECISION TEST

STATE OF VERMONT

DataMaster DMT: 100157

Date: 09/15/2010
Time: 13:34:31

SUPERVISOR NAME: STEVEN E HARNOIS

SOLUTION LOT #: 10-32-400
SOLUTION CONCENTRATION: 0.373

BLANK TEST		0.000	13:35
CALIBRATION CHECK		PASSED	13:35
Xq = 0.1084 (0.16%)			
SIMULATOR VAPOR		0.360	13:36
X[1] = 0.3602	(0.0049)	(-0.0007)	
SIMULATOR VAPOR		0.359	13:37
X[1] = 0.3596	(0.0057)	(-0.0014)	
SIMULATOR VAPOR		0.359	13:38
X[1] = 0.3591	(0.0056)	(-0.0009)	
SIMULATOR VAPOR		0.358	13:39
X[1] = 0.3587	(0.0052)	(-0.0018)	
SIMULATOR VAPOR		0.358	13:40
X[1] = 0.3584	(0.0052)	(-0.0017)	
SIMULATOR VAPOR		0.360	13:41
X[1] = 0.3603	(0.0060)	(-0.0010)	
SIMULATOR VAPOR		0.358	13:42
X[1] = 0.3588	(0.0047)	(-0.0013)	
SIMULATOR VAPOR		0.359	13:43
X[1] = 0.3595	(0.0057)	(-0.0009)	
SIMULATOR VAPOR		0.359	13:44
X[1] = 0.3597	(0.0053)	(-0.0007)	
SIMULATOR VAPOR		0.359	13:45
X[1] = 0.3592	(0.0044)	(-0.0011)	
BLANK TEST		0.000	13:46

Average = 0.358
Std Dev = 0.0007



ACCURACY AND PRECISION TEST

STATE OF VERMONT

DataMaster DMT: 100157

Date: 09/15/2010
Time: 06:54:56

SUPERVISOR NAME: STEVEN E HARNOIS

SOLUTION LOT #: 10-34-020
SOLUTION CONCENTRATION: 0.020

BLANK TEST	0.000	06:55
CALIBRATION CHECK	PASSED	06:55
Xq = 0.1084 (0.16%)		
SIMULATOR VAPOR	0.017	06:56
X[1] = 0.0175 (-0.0001) (0.0000)		
SIMULATOR VAPOR	0.018	06:57
X[1] = 0.0182 (0.0004) (-0.0005)		
SIMULATOR VAPOR	0.018	06:58
X[1] = 0.0183 (-0.0006) (-0.0002)		
SIMULATOR VAPOR	0.017	06:59
X[1] = 0.0179 (-0.0008) (0.0002)		
SIMULATOR VAPOR	0.018	07:00
X[1] = 0.0187 (0.0001) (0.0001)		
SIMULATOR VAPOR	0.018	07:01
X[1] = 0.0180 (-0.0005) (0.0000)		
SIMULATOR VAPOR	0.017	07:02
X[1] = 0.0179 (-0.0010) (-0.0001)		
SIMULATOR VAPOR	0.018	07:03
X[1] = 0.0180 (-0.0022) (0.0006)		
SIMULATOR VAPOR	0.018	07:04
X[1] = 0.0183 (0.0001) (-0.0001)		
SIMULATOR VAPOR	0.018	07:05
X[1] = 0.0181 (-0.0020) (-0.0001)		
BLANK TEST	0.000	07:05

Average out of range!

Average = 0.017
Std Dev = 0.0004

Solution Depleted Replace + Rerun
SH 9/17/10



ACCURACY AND PRECISION TEST

STATE OF VERMONT

DataMaster DMT: 100157

Date: 09/17/2010
Time: 06:36:10

SUPERVISOR NAME: SEH REPLACMENT

SOLUTION LOT #: 10-34-020
SOLUTION CONCENTRATION: 0.020

BLANK TEST	0.000	06:37
CALIBRATION CHECK	PASSED	06:37
SIMULATOR VAPOR	0.019	06:37
SIMULATOR VAPOR	0.019	06:38
SIMULATOR VAPOR	0.019	06:39
SIMULATOR VAPOR	0.019	06:40
SIMULATOR VAPOR	0.019	06:41
SIMULATOR VAPOR	0.019	06:42
SIMULATOR VAPOR	0.020	06:43
SIMULATOR VAPOR	0.020	06:44
SIMULATOR VAPOR	0.020	06:45
SIMULATOR VAPOR	0.019	06:46
BLANK TEST	0.000	06:47

Average = 0.019
Std Dev = 0.0004



DIAGNOSTIC RESULT

DataMaster DMT: 100157
Location:
Calibration Date: 09/10/2010
Certification Date: 09/10/2010
Installation Date: 09/10/2010
Test Date: 09/21/2010
Time: 08:39

VERMONT
DEPARTMENT OF HEALTH

```
VERSIONS
DMT 1.05
PIC 2.06
Modem 1.06
Questions 1.00
Reports 1.02

TEMPERATURES
Sample Chamber = 48.97°C
Breath Tube    = 47.12°C

SETTINGS
Lamp Voltage   = 2.10 V
Cooler Voltage = 1.79 V
Bias Voltage   = 80.3 V
Chopper Freq   = 547.4 Hz

PUMP INFO
Flow Rate = 5.418 L/M

DETECTOR INFO
PUMP     ON        OFF
MAX(V) -0.1266   -0.1236
MIN(V) -0.1280   -0.1248

FILTER INFO
Filter 1 -0.124   Zero = true
Filter 2  0.085   Zero = true
Filter 3  1.346   Zero = true

CALIBRATION CHECK
Xq = 0.1081  0.49%
```

ACCURACY AND PRECISION TEST

STATE OF VERMONT

DataMaster DMT: 100157

Date: 09/21/2010
Time: 08:42:00

SUPERVISOR NAME: STEVEN E HARNOIS

SOLUTION LOT #: 10-37-080
SOLUTION CONCENTRATION: 0.079

BLANK TEST	0.000	08:42
CALIBRATION CHECK	PASSED	08:43
SIMULATOR VAPOR	0.076	08:43
SIMULATOR VAPOR	0.077	08:44
SIMULATOR VAPOR	0.076	08:45
SIMULATOR VAPOR	0.077	08:46
SIMULATOR VAPOR	0.077	08:47
SIMULATOR VAPOR	0.076	08:48
SIMULATOR VAPOR	0.077	08:49
SIMULATOR VAPOR	0.077	08:50
SIMULATOR VAPOR	0.076	08:51
SIMULATOR VAPOR	0.076	08:52
BLANK TEST	0.000	08:52

Average = 0.076
Std Dev = 0.0005



ACCURACY AND PRECISION TEST

STATE OF VERMONT

DataMaster DMT: 100157

Date: 09/21/2010
Time: 08:54:53

SUPERVISOR NAME: STEVEN E HARNOIS

SOLUTION LOT #: 10-32-400
SOLUTION CONCENTRATION: 0.373

BLANK TEST	0.000	08:55
CALIBRATION CHECK	PASSED	08:55
SIMULATOR VAPOR	0.359	08:56
SIMULATOR VAPOR	0.358	08:57
SIMULATOR VAPOR	0.358	08:58
SIMULATOR VAPOR	0.357	08:59
SIMULATOR VAPOR	0.356	09:00
SIMULATOR VAPOR	0.357	09:01
SIMULATOR VAPOR	0.357	09:02
SIMULATOR VAPOR	0.357	09:03
SIMULATOR VAPOR	0.359	09:04
SIMULATOR VAPOR	0.358	09:05
BLANK TEST	0.000	09:06

Average = 0.357
Std Dev = 0.0009



ACCURACY AND PRECISION TEST

STATE OF VERMONT

DataMaster DMT: 100157

Date: 09/21/2010
Time: 09:11:50

SUPERVISOR NAME: STEVEN E HARNOIS

SOLUTION LOT #: 11-06-160
SOLUTION CONCENTRATION: 0.162

BLANK TEST	0.000	09:12
CALIBRATION CHECK	PASSED	09:12
SIMULATOR VAPOR	0.159	09:13
SIMULATOR VAPOR	0.158	09:14
SIMULATOR VAPOR	0.159	09:15
SIMULATOR VAPOR	0.158	09:16
SIMULATOR VAPOR	0.159	09:17
SIMULATOR VAPOR	0.160	09:18
SIMULATOR VAPOR	0.160	09:19
SIMULATOR VAPOR	0.159	09:20
SIMULATOR VAPOR	0.159	09:21
SIMULATOR VAPOR	0.158	09:22
BLANK TEST	0.000	09:22

Average = 0.158
Std Dev = 0.0007



ACCURACY AND PRECISION TEST

STATE OF VERMONT

DataMaster DMT: 100157

Date: 09/21/2010
Time: 09:33:51

SUPERVISOR NAME: STEVEN E HARNOIS

SOLUTION LOT #: 10-34-020
SOLUTION CONCENTRATION: 0.020

BLANK TEST	0.000	09:34
CALIBRATION CHECK	PASSED	09:34
SIMULATOR VAPOR	0.018	09:35
SIMULATOR VAPOR	0.019	09:36
SIMULATOR VAPOR	0.018	09:37
SIMULATOR VAPOR	0.019	09:38
SIMULATOR VAPOR	0.019	09:39
SIMULATOR VAPOR	0.019	09:40
SIMULATOR VAPOR	0.019	09:41
SIMULATOR VAPOR	0.019	09:42
SIMULATOR VAPOR	0.019	09:42
SIMULATOR VAPOR	0.019	09:43
BLANK TEST	0.000	09:44

Average out of range!

Average = 0.018
Std Dev = 0.0004



DIAGNOSTIC RESULT

DataMaster DMT: 100157
Location:
Calibration Date: 09/10/2010
Certification Date: 09/10/2010
Installation Date: 09/10/2010
Test Date: 09/21/2010
Time: 13:23

VERMONT
DEPARTMENT OF HEALTH

```
VERSIONS
DMT 1.05
PIC 2.06
Modem 1.06
Questions 1.00
Reports 1.02

TEMPERATURES
Sample Chamber = 49.03°C
Breath Tube    = 48.01°C

SETTINGS
Lamp Voltage   = 2.10 V
Cooler Voltage = 1.79 V
Bias Voltage   = 80.4 V
Chopper Freq   = 547.3 Hz

PUMP INFO
Flow Rate = 5.359 L/M

DETECTOR INFO
PUMP     ON        OFF
MAX(V) -0.1142   -0.1112
MIN(V) -0.1163   -0.1136

FILTER INFO
Filter 1 -0.110  Zero = true
Filter 2  0.099  Zero = true
Filter 3  1.365  Zero = true

CALIBRATION CHECK
Xq = 0.1083  0.32%
```

ACCURACY AND PRECISION TEST

STATE OF VERMONT

DataMaster DMT: 100157

Date: 09/21/2010
Time: 13:25:42

SUPERVISOR NAME: STEVEN E HARNOIS

SOLUTION LOT #: 10-34-020
SOLUTION CONCENTRATION: 0.020

BLANK TEST	0.000	13:26
CALIBRATION CHECK	PASSED	13:26
SIMULATOR VAPOR	0.020	13:27
SIMULATOR VAPOR	0.020	13:28
SIMULATOR VAPOR	0.019	13:29
SIMULATOR VAPOR	0.019	13:29
SIMULATOR VAPOR	0.019	13:30
SIMULATOR VAPOR	0.020	13:31
SIMULATOR VAPOR	0.019	13:32
SIMULATOR VAPOR	0.019	13:33
SIMULATOR VAPOR	0.019	13:34
SIMULATOR VAPOR	0.020	13:35
BLANK TEST	0.000	13:36

Average = 0.019
Std Dev = 0.0005



ACCURACY AND PRECISION TEST

STATE OF VERMONT

DataMaster DMT: 100157

Date: 09/21/2010
Time: 13:39:51

SUPERVISOR NAME: STEVEN E HARNOIS

SOLUTION LOT #: 11-06-160
SOLUTION CONCENTRATION: 0.162

BLANK TEST	0.000	13:40
CALIBRATION CHECK	PASSED	13:40
SIMULATOR VAPOR	0.160	13:41
SIMULATOR VAPOR	0.159	13:42
SIMULATOR VAPOR	0.159	13:43
SIMULATOR VAPOR	0.159	13:44
SIMULATOR VAPOR	0.159	13:45
SIMULATOR VAPOR	0.159	13:46
SIMULATOR VAPOR	0.160	13:47
SIMULATOR VAPOR	0.159	13:48
SIMULATOR VAPOR	0.159	13:49
SIMULATOR VAPOR	0.158	13:50
BLANK TEST	0.000	13:51

Average = 0.159
Std Dev = 0.0005



ACCURACY AND PRECISION TEST

STATE OF VERMONT

DataMaster DMT: 100157

Date: 09/21/2010
Time: 13:52:56

SUPERVISOR NAME: STEVEN E HARNOIS

SOLUTION LOT #: 10-32-400
SOLUTION CONCENTRATION: 0.373

BLANK TEST	0.000	13:53
CALIBRATION CHECK	PASSED	13:53
SIMULATOR VAPOR	0.358	13:54
SIMULATOR VAPOR	0.359	13:55
SIMULATOR VAPOR	0.357	13:56
SIMULATOR VAPOR	0.358	13:57
SIMULATOR VAPOR	0.356	13:59
SIMULATOR VAPOR	0.357	13:59
SIMULATOR VAPOR	0.357	14:01
SIMULATOR VAPOR	0.357	14:02
SIMULATOR VAPOR	0.357	14:03
SIMULATOR VAPOR	0.357	14:04
BLANK TEST	0.000	14:04

Average = 0.357
Std Dev = 0.0008



ACCURACY AND PRECISION TEST

STATE OF VERMONT

DataMaster DMT: 100157

Date: 09/21/2010
Time: 14:07:26

SUPERVISOR NAME: STEVEN E HARNOIS

SOLUTION LOT #: 10-37-080
SOLUTION CONCENTRATION: 0.079

BLANK TEST	0.000	14:08
CALIBRATION CHECK	PASSED	14:08
SIMULATOR VAPOR	0.079	14:08
SIMULATOR VAPOR	0.079	14:09
SIMULATOR VAPOR	0.078	14:10
SIMULATOR VAPOR	0.078	14:11
SIMULATOR VAPOR	0.078	14:12
SIMULATOR VAPOR	0.078	14:13
SIMULATOR VAPOR	0.078	14:14
SIMULATOR VAPOR	0.077	14:15
SIMULATOR VAPOR	0.077	14:16
SIMULATOR VAPOR	0.078	14:17
BLANK TEST	0.000	14:18

Average = 0.078
Std Dev = 0.0006



DIAGNOSTIC RESULT

DataMaster DMT: 100157
Location:
Calibration Date: 09/10/2010
Certification Date: 09/10/2010
Installation Date: 09/10/2010
Test Date: 09/22/2010
Time: 05:54

VERMONT
DEPARTMENT OF HEALTH

```
VERSIONS
DMT 1.05
PIC 2.06
Modem 1.06
Questions 1.00
Reports 1.02

TEMPERATURES
Sample Chamber = 48.74°C
Breath Tube    = 47.95°C

SETTINGS
Lamp Voltage   = 2.10 V
Cooler Voltage = 1.79 V
Bias Voltage   = 80.3 V
Chopper Freq   = 547.3 Hz

PUMP INFO
Flow Rate = 5.264 L/M

DETECTOR INFO
PUMP     ON        OFF
MAX(V) -0.1191   -0.1169
MIN(V) -0.1224   -0.1187

FILTER INFO
Filter 1 -0.117  Zero = true
Filter 2  0.094  Zero = true
Filter 3  1.359  Zero = true

CALIBRATION CHECK
Xq = 0.1081  0.49%
```

ACCURACY AND PRECISION TEST

STATE OF VERMONT

DataMaster DMT: 100157

Date: 09/22/2010
Time: 05:59:05

SUPERVISOR NAME: STEVEN E HARNOIS

SOLUTION LOT #: 10-37-080
SOLUTION CONCENTRATION: 0.079

BLANK TEST	0.000	06:00
CALIBRATION CHECK	PASSED	06:00
SIMULATOR VAPOR	0.076	06:00
SIMULATOR VAPOR	0.077	06:01
SIMULATOR VAPOR	0.077	06:02
SIMULATOR VAPOR	0.077	06:03
SIMULATOR VAPOR	0.077	06:04
SIMULATOR VAPOR	0.077	06:05
SIMULATOR VAPOR	0.077	06:06
SIMULATOR VAPOR	0.077	06:07
SIMULATOR VAPOR	0.077	06:08
SIMULATOR VAPOR	0.077	06:09
BLANK TEST	0.000	06:10

Average = 0.076
Std Dev = 0.0003



ACCURACY AND PRECISION TEST

STATE OF VERMONT

DataMaster DMT: 100157

Date: 09/22/2010
Time: 06:14:33

SUPERVISOR NAME: STEVEN E HARNOIS

SOLUTION LOT #: 10-32-400
SOLUTION CONCENTRATION: 0.373

BLANK TEST	0.000	06:15
CALIBRATION CHECK	PASSED	06:15
SIMULATOR VAPOR	0.357	06:16
SIMULATOR VAPOR	0.355	06:17
SIMULATOR VAPOR	0.355	06:18
SIMULATOR VAPOR	0.356	06:19
SIMULATOR VAPOR	0.355	06:20
SIMULATOR VAPOR	0.356	06:21
SIMULATOR VAPOR	0.356	06:22
SIMULATOR VAPOR	0.356	06:23
SIMULATOR VAPOR	0.357	06:24
SIMULATOR VAPOR	0.357	06:25
BLANK TEST	0.000	06:26

Average = 0.356
Std Dev = 0.0008



ACCURACY AND PRECISION TEST

STATE OF VERMONT

DataMaster DMT: 100157

Date: 09/22/2010
Time: 06:27:46

SUPERVISOR NAME: STEVEN E HARNOIS

SOLUTION LOT #: 11-06-160
SOLUTION CONCENTRATION: 0.162

BLANK TEST	0.000	06:28
CALIBRATION CHECK	PASSED	06:28
SIMULATOR VAPOR	0.158	06:29
SIMULATOR VAPOR	0.158	06:30
SIMULATOR VAPOR	0.157	06:31
SIMULATOR VAPOR	0.157	06:32
SIMULATOR VAPOR	0.157	06:33
SIMULATOR VAPOR	0.157	06:34
SIMULATOR VAPOR	0.157	06:35
SIMULATOR VAPOR	0.157	06:36
SIMULATOR VAPOR	0.157	06:37
SIMULATOR VAPOR	0.157	06:38
BLANK TEST	0.000	06:38

Average = 0.157
Std Dev = 0.0004



ACCURACY AND PRECISION TEST

STATE OF VERMONT

DataMaster DMT: 100157

Date: 09/22/2010
Time: 06:54:21

SUPERVISOR NAME: STEVEN E HARNOIS

SOLUTION LOT #: 10-34-020
SOLUTION CONCENTRATION: 0.020

BLANK TEST	0.000	06:55
CALIBRATION CHECK	PASSED	06:55
SIMULATOR VAPOR	0.020	06:55
SIMULATOR VAPOR	0.019	06:56
SIMULATOR VAPOR	0.020	06:57
SIMULATOR VAPOR	0.020	06:58
SIMULATOR VAPOR	0.019	06:59
SIMULATOR VAPOR	0.020	07:00
SIMULATOR VAPOR	0.019	07:01
SIMULATOR VAPOR	0.020	07:02
SIMULATOR VAPOR	0.019	07:03
SIMULATOR VAPOR	0.019	07:04
BLANK TEST	0.000	07:05

Average = 0.019
Std Dev = 0.0005



ACCURACY AND PRECISION TEST

STATE OF VERMONT

DataMaster DMT: 100157

Date: 09/22/2010
Time: 09:40:23

SUPERVISOR NAME: STEVEN E HARNOIS

SOLUTION LOT #: 10-37-080
SOLUTION CONCENTRATION: 0.079

BLANK TEST	0.000	09:41
CALIBRATION CHECK	PASSED	09:41
SIMULATOR VAPOR	0.076	09:41
SIMULATOR VAPOR	0.077	09:42
SIMULATOR VAPOR	0.078	09:43
SIMULATOR VAPOR	0.077	09:44
SIMULATOR VAPOR	0.078	09:45
SIMULATOR VAPOR	0.078	09:46
SIMULATOR VAPOR	0.077	09:47
SIMULATOR VAPOR	0.077	09:48
SIMULATOR VAPOR	0.077	09:49
SIMULATOR VAPOR	0.077	09:50
BLANK TEST	0.000	09:51

Average = 0.077
Std Dev = 0.0006



ACCURACY AND PRECISION TEST

STATE OF VERMONT

DataMaster DMT: 100157

Date: 09/22/2010
Time: 09:17:47

SUPERVISOR NAME: STEVEN E HARNOIS

SOLUTION LOT #: 10-34-020
SOLUTION CONCENTRATION: 0.020

BLANK TEST	0.000	09:18
CALIBRATION CHECK	PASSED	09:18
SIMULATOR VAPOR	0.023	09:19
SIMULATOR VAPOR	0.022	09:20
SIMULATOR VAPOR	0.022	09:21
SIMULATOR VAPOR	0.022	09:22
SIMULATOR VAPOR	0.021	09:23
SIMULATOR VAPOR	0.021	09:24
SIMULATOR VAPOR	0.020	09:25
SIMULATOR VAPOR	0.020	09:25
SIMULATOR VAPOR	0.020	09:26
SIMULATOR VAPOR	0.021	09:27
BLANK TEST	0.000	09:28

Average = 0.021
Std Dev = 0.0010



ACCURACY AND PRECISION TEST

STATE OF VERMONT

DataMaster DMT: 100157

Date: 09/22/2010
Time: 08:49:41

SUPERVISOR NAME: STEVEN E HARNOIS

SOLUTION LOT #: 11-07-400
SOLUTION CONCENTRATION: 0.412

BLANK TEST	0.000	08:50
CALIBRATION CHECK	PASSED	08:50
SIMULATOR VAPOR	0.405	08:51
SIMULATOR VAPOR	0.404	08:52
SIMULATOR VAPOR	0.404	08:53
SIMULATOR VAPOR	0.404	08:54
SIMULATOR VAPOR	0.405	08:55
SIMULATOR VAPOR	0.404	08:56
SIMULATOR VAPOR	0.404	08:57
SIMULATOR VAPOR	0.403	08:58
SIMULATOR VAPOR	0.404	08:59
SIMULATOR VAPOR	0.404	09:00
BLANK TEST	0.000	09:01

Average = 0.404
Std Dev = 0.0005



ACCURACY AND PRECISION TEST

STATE OF VERMONT

DataMaster DMT: 100157

Date: 09/22/2010
Time: 08:18:12

SUPERVISOR NAME: STEVEN E HARNOIS

SOLUTION LOT #: 11-06-160
SOLUTION CONCENTRATION: 0.162

BLANK TEST	0.000	08:19
CALIBRATION CHECK	PASSED	08:19
SIMULATOR VAPOR	0.155	08:19
SIMULATOR VAPOR	0.155	08:20
SIMULATOR VAPOR	0.156	08:21
SIMULATOR VAPOR	0.157	08:22
SIMULATOR VAPOR	0.157	08:23
SIMULATOR VAPOR	0.157	08:24
SIMULATOR VAPOR	0.156	08:25
SIMULATOR VAPOR	0.156	08:26
SIMULATOR VAPOR	0.157	08:27
SIMULATOR VAPOR	0.156	08:28
BLANK TEST	0.000	08:29

Average = 0.156
Std Dev = 0.0007



DIAGNOSTIC RESULT

DataMaster DMT: 100157
Location:
Calibration Date: 09/10/2010
Certification Date: 09/10/2010
Installation Date: 09/10/2010
Test Date: 09/22/2010
Time: 08:16

VERMONT
DEPARTMENT OF HEALTH

```
VERSIONS
DMT 1.05
PIC 2.06
Modem 1.06
Questions 1.00
Reports 1.02

TEMPERATURES
Sample Chamber = 48.91°C
Breath Tube    = 48.10°C

SETTINGS
Lamp Voltage   = 2.10 V
Cooler Voltage = 1.79 V
Bias Voltage   = 80.4 V
Chopper Freq   = 546.9 Hz

PUMP INFO
Flow Rate = 5.323 L/M

DETECTOR INFO
PUMP     ON        OFF
MAX(V) -0.1079   -0.1047
MIN(V) -0.1098   -0.1073

FILTER INFO
Filter 1 -0.107  Zero = true
Filter 2  0.105  Zero = true
Filter 3  1.374  Zero = true

CALIBRATION CHECK
Xq = 0.1084  0.16%
```

Section 5: Interference

ACCURACY AND PRECISION TEST

STATE OF VERMONT

DataMaster DMT: 100157

Date: 09/10/2010
Time: 09:07:09

SUPERVISOR NAME: SEH 0.04MEOH0.08ETOHINH2O

SOLUTION LOT #: 10-10-08M4
SOLUTION CONCENTRATION: 0.080

BLANK TEST	0.000	09:08
CALIBRATION CHECK	PASSED	09:08
SIMULATOR VAPOR	INTERFERENCE	09:08
SIMULATOR VAPOR	INTERFERENCE	09:09
SIMULATOR VAPOR	INTERFERENCE	09:10
SIMULATOR VAPOR	INTERFERENCE	09:11
SIMULATOR VAPOR	INTERFERENCE	09:12
SIMULATOR VAPOR	INTERFERENCE	09:13
SIMULATOR VAPOR	INTERFERENCE	09:14
SIMULATOR VAPOR	INTERFERENCE	09:15
SIMULATOR VAPOR	INTERFERENCE	09:16
SIMULATOR VAPOR	INTERFERENCE	09:17
BLANK TEST	0.000	09:18

Average out of range!



ACCURACY AND PRECISION TEST

STATE OF VERMONT

DataMaster DMT: 100157

Date: 09/13/2010
Time: 07:46:46

SUPERVISOR NAME: SEH 08ETOH01ACEINH2O

SOLUTION LOT #: 10-33-08A
SOLUTION CONCENTRATION: 0.080

BLANK TEST	0.000	07:47
CALIBRATION CHECK	PASSED	07:47
Xq = 0.1090 (0.32%)		
SIMULATOR VAPOR	INTERFERENCE	07:48
X[1] = 0.0831 (-0.0054) (0.0006)		
SIMULATOR VAPOR	0.082	07:49
X[1] = 0.0826 (-0.0049) (-0.0003)		
SIMULATOR VAPOR	INTERFERENCE	07:50
X[1] = 0.0833 (-0.0056) (-0.0003)		
SIMULATOR VAPOR	INTERFERENCE	07:51
X[1] = 0.0829 (-0.0062) (-0.0007)		
SIMULATOR VAPOR	INTERFERENCE	07:52
X[1] = 0.0826 (-0.0067) (-0.0008)		
SIMULATOR VAPOR	INTERFERENCE	07:53
X[1] = 0.0832 (-0.0058) (-0.0011)		
SIMULATOR VAPOR	INTERFERENCE	07:54
X[1] = 0.0832 (-0.0058) (-0.0002)		
SIMULATOR VAPOR	INTERFERENCE	07:55
X[1] = 0.0825 (-0.0067) (-0.0011)		
SIMULATOR VAPOR	INTERFERENCE	07:56
X[1] = 0.0825 (-0.0059) (-0.0006)		
SIMULATOR VAPOR	INTERFERENCE	07:57
X[1] = 0.0829 (-0.0056) (-0.0006)		
BLANK TEST	0.000	07:57



ACCURACY AND PRECISION TEST

STATE OF VERMONT

DataMaster DMT: 100157

Date: 09/13/2010
Time: 08:13:32

SUPERVISOR NAME: SEH 04MEOHINH2O

SOLUTION LOT #: 11-08-04M
SOLUTION CONCENTRATION: 0.040

BLANK TEST	0.000	08:14
CALIBRATION CHECK	PASSED	08:14
Xq = 0.1090 (0.32%)		
SIMULATOR VAPOR	INTERFERENCE	08:14
X[1] = 0.0191	(-0.0057) (-0.0051)	
SIMULATOR VAPOR	INTERFERENCE	08:15
X[1] = 0.0183	(-0.0072) (-0.0069)	
SIMULATOR VAPOR	INTERFERENCE	08:16
X[1] = 0.0180	(-0.0070) (-0.0062)	
SIMULATOR VAPOR	INTERFERENCE	08:17
X[1] = 0.0188	(-0.0073) (-0.0071)	
SIMULATOR VAPOR	INTERFERENCE	08:18
X[1] = 0.0181	(-0.0072) (-0.0068)	
SIMULATOR VAPOR	INTERFERENCE	08:19
X[1] = 0.0184	(-0.0079) (-0.0075)	
SIMULATOR VAPOR	INTERFERENCE	08:20
X[1] = 0.0185	(-0.0068) (-0.0065)	
SIMULATOR VAPOR	INTERFERENCE	08:21
X[1] = 0.0180	(-0.0073) (-0.0068)	
SIMULATOR VAPOR	INTERFERENCE	08:22
X[1] = 0.0187	(-0.0061) (-0.0062)	
SIMULATOR VAPOR	INTERFERENCE	08:23
X[1] = 0.0183	(-0.0079) (-0.0073)	
BLANK TEST	0.000	08:24

Average out of range!



ACCURACY AND PRECISION TEST

STATE OF VERMONT

DataMaster DMT: 100157

Date: 09/13/2010
Time: 08:27:16

SUPERVISOR NAME: SEH 04IPAINH20

SOLUTION LOT #: 10-26-04I
SOLUTION CONCENTRATION: 0.040

BLANK TEST	0.000	08:28
CALIBRATION CHECK	PASSED	08:28
Xq = 0.1083 (0.32%)		
SIMULATOR VAPOR	INTERFERENCE	08:28
X[1] = 0.0199 (-0.0223) (0.0000)		
SIMULATOR VAPOR	INTERFERENCE	08:29
X[1] = 0.0198 (-0.0224) (0.0011)		
SIMULATOR VAPOR	INTERFERENCE	08:30
X[1] = 0.0204 (-0.0208) (0.0005)		
SIMULATOR VAPOR	INTERFERENCE	08:31
X[1] = 0.0201 (-0.0223) (0.0009)		
SIMULATOR VAPOR	INTERFERENCE	08:32
X[1] = 0.0198 (-0.0227) (0.0008)		
SIMULATOR VAPOR	INTERFERENCE	08:33
X[1] = 0.0198 (-0.0225) (0.0007)		
SIMULATOR VAPOR	INTERFERENCE	08:34
X[1] = 0.0202 (-0.0222) (0.0011)		
SIMULATOR VAPOR	INTERFERENCE	08:35
X[1] = 0.0197 (-0.0229) (-0.0002)		
SIMULATOR VAPOR	INTERFERENCE	08:36
X[1] = 0.0198 (-0.0228) (0.0012)		
SIMULATOR VAPOR	INTERFERENCE	08:37
X[1] = 0.0201 (-0.0224) (0.0013)		
BLANK TEST	0.000	08:38

Average out of range!



ACCURACY AND PRECISION TEST

STATE OF VERMONT

DataMaster DMT: 100157

Date: 09/13/2010
Time: 08:42:04

SUPERVISOR NAME: SEH 04ETOH01ACEINH2O

SOLUTION LOT #: 10-35-04A1
SOLUTION CONCENTRATION: 0.040

BLANK TEST	0.000	08:43
CALIBRATION CHECK	PASSED	08:43
Xq = 0.1085 (0.08%)		
SIMULATOR VAPOR	INTERFERENCE	08:43
X[1] = 0.0433 (-0.0109) (0.0014)		
SIMULATOR VAPOR	INTERFERENCE	08:44
X[1] = 0.0436 (-0.0108) (0.0005)		
SIMULATOR VAPOR	INTERFERENCE	08:45
X[1] = 0.0440 (-0.0093) (0.0009)		
SIMULATOR VAPOR	INTERFERENCE	08:46
X[1] = 0.0436 (-0.0105) (0.0003)		
SIMULATOR VAPOR	INTERFERENCE	08:47
X[1] = 0.0447 (-0.0084) (0.0004)		
SIMULATOR VAPOR	INTERFERENCE	08:48
X[1] = 0.0435 (-0.0108) (0.0001)		
SIMULATOR VAPOR	INTERFERENCE	08:49
X[1] = 0.0430 (-0.0118) (-0.0004)		
SIMULATOR VAPOR	INTERFERENCE	08:50
X[1] = 0.0436 (-0.0108) (0.0001)		
SIMULATOR VAPOR	INTERFERENCE	08:51
X[1] = 0.0435 (-0.0104) (0.0001)		
SIMULATOR VAPOR	INTERFERENCE	08:52
X[1] = 0.0434 (-0.0108) (0.0004)		
BLANK TEST	0.000	08:52

Average out of range!



ACCURACY AND PRECISION TEST

STATE OF VERMONT

DataMaster DMT: 100157

Date: 09/13/2010
Time: 10:22:12

SUPERVISOR NAME: SEH 04ETOH02ACETONE

SOLUTION LOT #: 10-35-04A2
SOLUTION CONCENTRATION: 0.040

BLANK TEST	0.000	10:23
CALIBRATION CHECK	PASSED	10:23
Xq = 0.1083 (0.32%)		
SIMULATOR VAPOR	INTERFERENCE	10:23
X[1] = 0.0364 (-0.0164) (0.0004)		
SIMULATOR VAPOR	INTERFERENCE	10:24
X[1] = 0.0363 (-0.0160) (-0.0002)		
SIMULATOR VAPOR	INTERFERENCE	10:25
X[1] = 0.0364 (-0.0162) (-0.0001)		
SIMULATOR VAPOR	INTERFERENCE	10:26
X[1] = 0.0363 (-0.0166) (-0.0010)		
SIMULATOR VAPOR	INTERFERENCE	10:27
X[1] = 0.0363 (-0.0168) (-0.0012)		
SIMULATOR VAPOR	INTERFERENCE	10:28
X[1] = 0.0367 (-0.0166) (-0.0002)		
SIMULATOR VAPOR	INTERFERENCE	10:29
X[1] = 0.0365 (-0.0165) (-0.0002)		
SIMULATOR VAPOR	INTERFERENCE	10:30
X[1] = 0.0364 (-0.0168) (-0.0008)		
SIMULATOR VAPOR	INTERFERENCE	10:31
X[1] = 0.0365 (-0.0165) (-0.0005)		
SIMULATOR VAPOR	INTERFERENCE	10:32
X[1] = 0.0364 (-0.0166) (-0.0007)		
BLANK TEST	0.000	10:33

Average out of range!



ACCURACY AND PRECISION TEST

STATE OF VERMONT

DataMaster DMT: 100157

Date: 09/13/2010
Time: 10:39:24

SUPERVISOR NAME: SEH 04IPA08ETOHINH2O

SOLUTION LOT #: 09-30-08I
SOLUTION CONCENTRATION: 0.080

BLANK TEST	0.000	10:40
CALIBRATION CHECK	PASSED	10:40
Xq = 0.1083 (0.32%)		
SIMULATOR VAPOR	INTERFERENCE	10:40
X[1] = 0.0923 (-0.0225) (0.0024)		
SIMULATOR VAPOR	INTERFERENCE	10:41
X[1] = 0.0929 (-0.0199) (0.0018)		
SIMULATOR VAPOR	INTERFERENCE	10:42
X[1] = 0.0926 (-0.0216) (0.0019)		
SIMULATOR VAPOR	INTERFERENCE	10:43
X[1] = 0.0931 (-0.0216) (0.0017)		
SIMULATOR VAPOR	INTERFERENCE	10:44
X[1] = 0.0927 (-0.0223) (0.0016)		
SIMULATOR VAPOR	INTERFERENCE	10:45
X[1] = 0.0921 (-0.0224) (0.0012)		
SIMULATOR VAPOR	INTERFERENCE	10:46
X[1] = 0.0925 (-0.0226) (0.0016)		
SIMULATOR VAPOR	INTERFERENCE	10:47
X[1] = 0.0929 (-0.0219) (0.0025)		
SIMULATOR VAPOR	INTERFERENCE	10:48
X[1] = 0.0930 (-0.0220) (0.0016)		
SIMULATOR VAPOR	INTERFERENCE	10:49
X[1] = 0.0928 (-0.0210) (0.0016)		
BLANK TEST	0.000	10:50

Average out of range!



Section 6: Low Level Accuracy

This Instrument not tested for this specification

Section 7: Breath Volume

This Instrument not tested for this specification

Section 8: Software Testing

This Instrument not tested for this specification

Section 9: Error Reports

Vermont Department of Health Laboratory
195 Colchester Avenue, P.O. Box 1125, Burlington VT 05402-1125
DataMaster Technical Services (ph) 802-863-7641; (fax) 802-863-7632

DATAMASTER TECHNICAL SUPPORT INQUIRY

Agency/Site: VDHL **DataMaster Serial #:** 100157

Date: 8/27/2010 Requestor: Amanda Bolduc

Instrument Supervisor: _____ Phone #: _____

Taken By: Harnois

Problem: Will no call MEOH Interference

Laboratory Response: Talked to NPAS advised to replace filter wheel.

Work Performed:

9/7/2010 Replaced Filter Wheel SH

☐ On Site ☒ In-house

Performed By: Steven Harnois SH Date: 9/7/2010

ALC 626 Rev 1 06/07

March/April 2010

~~October 2009~~

~~Third~~ 4th Round Testing Protocol for DMT Instruments

Power Up Procedure:

a) Open unit and inspect connections. Ensure all tubing is secure. Check that all cables are properly seated and tied down or strapped together as necessary. Check that all screws are in place and snug; this will require the removal of the side pieces to allow the technician access. Correct as needed.

b) Manually turn the filter wheel. The filter wheel should turn smoothly and should not rub or wobble. Correct as needed.

c) Power on the instrument. Allow approximately one hour for it to come to temperature.

d) Access the SETUP menu. On the MISC tab, type "State of Vermont" in the NAME field. In the EXPLORE box, highlight STYLUS and press the OK button to activate the stylus settings. Recalibrate the stylus. Once complete, exit the SETUP menu by pressing the OK button to save the settings. If you exit the SETUP menu using the CANCEL button, none of the changes will be saved.

e) Enter the TECH screen. Scroll through all of the filter settings to ensure the filter wheel is activating properly.

f) Optimize the detector voltage
-Adjust the cooler voltage setting to 1.70Vdc +/-0.10.
-Manually manipulate the detector position to get the detector voltage as low as possible.
-Adjust the Lamp voltage setting so that the detector voltage displayed is approximately +0.000 +/- 0.100.
-Note the initial noise of the detector voltage once unit has had a minute to equilibrate. Noise tolerance at NPAS is +/-0.003 on a detector.
-Save the voltage settings
-Per VDHL protocol (P-Alc-116 Power-Up SOP), the detector voltage must read 0.000V with a tolerance of +/- 0.100V. Correct as needed.

g) While still in the TECH screen, press SET in the RF Sensitivity box to automatically set the RF sensitivity level. Press SAVE to save the settings. Confirm the setting by keying a radio near the instrument. Correct as needed.

h) Using the large syringe, push 1.5L of air into the breath tube. Monitor the volume of air displayed on the TECH screen. The instrument should read 1.5L +/- 0.2L. If the instrument is out of range, adjust the distance between the "T" fittings on the breath exhaust tubing. Retest until the volume reads appropriately.

i) Once all the TECH settings are correct, re-save every setting. Exit the TECH screen.

j) Close up the instrument using locking screws on the top four holes. Attach the stylus to the DMT using a locking screw in the center right screw hole nearest the simulator tower. Attach the keyboard to the top of the instrument using sticky-back velcro. Wrap a small piece of velcro around the top of the stylus and affix the other side to the front of the instrument so the stylus can rest on the front of the unit.

k) Calibrate the instrument following the Calibration Protocol using Calibration Solution (0.100 EtOH). During Calibration and Certification, it may be necessary to turn off the simulator temperature monitoring if using Guth 2100 Simulators as these do not have BNC connectors for temperature monitoring. To do this, open the SETUP menu. In the Simulator box, select NONE as the type of

DRAFT

100157

simulator. Press OK to exit. Document the room temperature and humidity level at the time of calibration in the communication notebook.

l) Certify the instrument following the Certification Protocol.

m) Preliminary interference testing. Run the following concentrations as an Accuracy and Precision Test with n=10. The instrument must report Interference at least 80% of the time.

0.01% Acetone in 0.08% EtOH
0.04% MeOH in Water
0.04% Isopropanol in Water

The instrument is now ready for performance testing. If any of the above steps fail or there is **any** problem with the instrument, the DataMaster Team should be consulted as to the action required. It may be necessary to return an instrument to the manufacture for refurbishment prior to use by VDHL. All adjustments and repairs to an instrument should be documented in the instrument's testing notebook located in the testing binder. All reports generated during the Power Up Procedure should be filed in the power up section of that instrument's testing binder. If the instrument has passed, it should now be Installed to VDHL.

Unit failed to recognize MeOH in H_2O

RFI

100157

Performance Testing

a) Linearity Test: Four concentrations of EtOH solutions will be used, 0.020, 0.080, 0.160, and 0.400. Each solution will be run using the Accuracy and Precision test (n=10). All four concentrations will be run consecutively and will count as one linearity test. Five linearity tests will be run on each instrument. Linearity tests should not be completed back to back. Testing should allow for inter- and intra- day performance.

b) Low Level Performance: A solution of 0.040EtOH will be tested using the Accuracy and Precision test (n=10) once

c) Interference Test: The following concentrations will be tested using the Accuracy and Precision test (n=10). The instrument should report Interference at least 80% of the time.

0.01% Acetone in 0.04% EtOH
0.02% Acetone in 0.04% EtOH
0.04% MeOH in 0.08% EtOH
0.04% Isopropanol in 0.08% EtOH

d) Limited Volume Testing:

Target Volume	Flow (lpm)	Time (sec.)	1	2	3	4	5	6	7	Ave.	% s.d.
1.5 L	8.2	11									
	15	6									
	22.5	4									
2.0 L	10	12									
	15	8									
	24	5									

Pressure: _____ Flow rate +/- _____ time +/- _____

R

CALIBRATION REPORT

DataMaster DMT: 100157
Calibration Date: 08/24/2010
Calibrated by: AMANDA BOLDUC
Lot: 11-03-100

VERMONT
DEPARTMENT OF HEALTH

```
Ca  =  0.102000
CAL =  1.014461    0.800 <= CAL < 1.200
b1  =  0.000397    0.000 <=  b1 < 0.004
b2  =  0.005584    0.001 <=  b2 < 0.012
b3  =  0.000000    0.000 <=  b3 < 0.004
Xq  =  0.099267    0.050 <=  Xq < 0.200
a21 =  1.165331    1.050 <= a21 < 1.300
a31 =  0.540924    0.300 <= a31 < 0.800
```

Performed by AB Bolduc Date 08/24/2010

ACCURACY AND PRECISION TEST

STATE OF VERMONT

DataMaster DMT: 100157

Date: 08/24/2010
Time: 14:40:54

SUPERVISOR NAME: ALB CAL CHECK

SOLUTION LOT #: 11-02-100
SOLUTION CONCENTRATION: 0.100

BLANK TEST	0.000	14:41
CALIBRATION CHECK	PASSED	14:41
SIMULATOR VAPOR	0.102	14:42
SIMULATOR VAPOR	0.102	14:43
SIMULATOR VAPOR	0.101	14:44
SIMULATOR VAPOR	0.102	14:45
SIMULATOR VAPOR	0.102	14:46
SIMULATOR VAPOR	0.102	14:47
SIMULATOR VAPOR	0.101	14:48
SIMULATOR VAPOR	0.102	14:49
SIMULATOR VAPOR	0.102	14:50
SIMULATOR VAPOR	0.102	14:51
BLANK TEST	0.000	14:51

Average = 0.101
Std Dev = 0.0004



CERTIFICATION REPORT

DataMaster DMT: 100157
Calibration Date: 08/24/2010
Certification Date: 08/26/2010
Certified by: AMANDA BOLDUC

VERMONT
DEPARTMENT OF HEALTH

Diagnostic Results

VERSIONS
DMT 1.05
PIC 2.06
Modem 1.06
Questions 1.00
Reports 1.02

TEMPERATURES
Sample Chamber = 49.03°C
Breath Tube = 45.39°C

SETTINGS
Lamp Voltage = 2.18 V
Cooler Voltage = 1.74 V
Bias Voltage = 80.7 V
Chopper Freq = 536.7 Hz

PUMP INFO
Flow Rate = 5.456 L/M

DETECTOR INFO
PUMP ON OFF
MAX(V) 0.0069 0.0114
MIN(V) 0.0053 0.0100

FILTER INFO
Filter 1 0.010 Zero = true
Filter 2 0.329 Zero = true
Filter 3 0.637 Zero = true

CALIBRATION CHECK
Xq = 0.1006 1.41%

Options

Printer: yes
Number of Copies: 3
Number of Supervisor Tests: 10
Tolerance Check: yes
Data Collection: yes
Units: g/210L
Simulator Check: yes
Uses Wet Bath Simulator: yes
Simulator Nominal: 0.102
Digital Simulator: none
Simulator Before: yes
Simulator Between: yes
Simulator After: no
Number of Subject Tests: 2
AskQuestions: yes
Query Refusal: yes
Alcohol Display: yes
Show Two Digits: no
Volume Display: yes
Number of Calibration Tests: 1
Calibrate with Wet Bath: yes
Calibration Nominal: 0.100

Linearity Check Results

True Value	Reported Average	Std Dev
0.020 g/210L Lot # 10-34-020	0.019 g/210L	0.001
0.079 g/210L Lot # 10-37-080	0.081 g/210L	0.000
0.160 g/210L Lot # 10-31-160	0.161 g/210L	0.001
0.373 g/210L Lot # 10-32-400	0.384 g/210L	0.001

$R^2 = 0.9999$



Acetone Interference Test

Lot # 10-33-08A

Interference Detected

Mouth Alcohol Test

Mouth Alcohol Detected



RF Interference Test

RFI detected



Sample Acceptance Test

Passed

CERTIFICATION PASSED

Performed by _____ **Date** 08/26/2010

Reviewed by _____ **Date** _____

ACCURACY AND PRECISION TEST

STATE OF VERMONT

DataMaster DMT: 100157

Date: 08/26/2010
Time: 09:41:44

SUPERVISOR NAME: ALB .01ACETONE IN .08ETOH

SOLUTION LOT #: 10-33-08A
SOLUTION CONCENTRATION: 0.080

BLANK TEST	0.000	09:42
CALIBRATION CHECK	PASSED	09:42
SIMULATOR VAPOR	INTERFERENCE	09:43
SIMULATOR VAPOR	INTERFERENCE	09:44
SIMULATOR VAPOR	INTERFERENCE	09:45
SIMULATOR VAPOR	INTERFERENCE	09:46
SIMULATOR VAPOR	0.089	09:47
SIMULATOR VAPOR	INTERFERENCE	09:48
SIMULATOR VAPOR	INTERFERENCE	09:49
SIMULATOR VAPOR	INTERFERENCE	09:50
SIMULATOR VAPOR	INTERFERENCE	09:51
SIMULATOR VAPOR	INTERFERENCE	09:52
BLANK TEST	0.000	09:53

Average out of range!



ACCURACY AND PRECISION TEST

STATE OF VERMONT

DataMaster DMT: 100157

Date: 08/26/2010
Time: 10:05:59

SUPERVISOR NAME: ALB .04IPA IN H20

SOLUTION LOT #: 10-26-04I
SOLUTION CONCENTRATION: 0.040

BLANK TEST	0.000	10:06
CALIBRATION CHECK	PASSED	10:07
SIMULATOR VAPOR	INTERFERENCE	10:07
SIMULATOR VAPOR	INTERFERENCE	10:08
SIMULATOR VAPOR	INTERFERENCE	10:09
SIMULATOR VAPOR	INTERFERENCE	10:10
SIMULATOR VAPOR	INTERFERENCE	10:11
SIMULATOR VAPOR	INTERFERENCE	10:12
SIMULATOR VAPOR	INTERFERENCE	10:13
SIMULATOR VAPOR	INTERFERENCE	10:14
SIMULATOR VAPOR	INTERFERENCE	10:15
SIMULATOR VAPOR	INTERFERENCE	10:16
BLANK TEST	0.000	10:16

Average out of range!



ACCURACY AND PRECISION TEST

STATE OF VERMONT

DataMaster DMT: 100157

Date: 08/26/2010
Time: 10:22:06

SUPERVISOR NAME: ALB .01ACETONE IN .08ETOH

SOLUTION LOT #: 10-35-04A1
SOLUTION CONCENTRATION: 0.080

BLANK TEST	0.000	10:23
CALIBRATION CHECK	PASSED	10:23
SIMULATOR VAPOR	INTERFERENCE	10:23
SIMULATOR VAPOR	INTERFERENCE	10:24
SIMULATOR VAPOR	INTERFERENCE	10:25
SIMULATOR VAPOR	INTERFERENCE	10:26
SIMULATOR VAPOR	INTERFERENCE	10:27
SIMULATOR VAPOR	INTERFERENCE	10:28
SIMULATOR VAPOR	INTERFERENCE	10:29
SIMULATOR VAPOR	INTERFERENCE	10:30
SIMULATOR VAPOR	INTERFERENCE	10:31
SIMULATOR VAPOR	INTERFERENCE	10:32
BLANK TEST	0.000	10:33

Average out of range!



ACCURACY AND PRECISION TEST

STATE OF VERMONT

DataMaster DMT: 100157

Date: 08/26/2010
Time: 11:11:36

SUPERVISOR NAME: ALB .02ACETONE IN .04ETOH

SOLUTION LOT #: 10-35-04A2
SOLUTION CONCENTRATION: 0.040

BLANK TEST	0.000	11:12
CALIBRATION CHECK	PASSED	11:12
SIMULATOR VAPOR	INTERFERENCE	11:13
SIMULATOR VAPOR	INTERFERENCE	11:13
SIMULATOR VAPOR	INTERFERENCE	11:14
SIMULATOR VAPOR	INTERFERENCE	11:15
SIMULATOR VAPOR	INTERFERENCE	11:16
SIMULATOR VAPOR	INTERFERENCE	11:17
SIMULATOR VAPOR	INTERFERENCE	11:18
SIMULATOR VAPOR	INTERFERENCE	11:19
SIMULATOR VAPOR	INTERFERENCE	11:20
SIMULATOR VAPOR	INTERFERENCE	11:21
BLANK TEST	0.000	11:22

Average out of range!



ACCURACY AND PRECISION TEST

STATE OF VERMONT

DataMaster DMT: 100157

Date: 08/26/2010
Time: 11:43:52

SUPERVISOR NAME: ALB .04MEOH IN .08ETOH

SOLUTION LOT #: 10-10-08M4
SOLUTION CONCENTRATION: 0.080

BLANK TEST	0.000	11:44
CALIBRATION CHECK	PASSED	11:44
SIMULATOR VAPOR	INTERFERENCE	11:45
SIMULATOR VAPOR	INTERFERENCE	11:46
SIMULATOR VAPOR	INTERFERENCE	11:47
SIMULATOR VAPOR	INTERFERENCE	11:48
SIMULATOR VAPOR	INTERFERENCE	11:49
SIMULATOR VAPOR	INTERFERENCE	11:50
SIMULATOR VAPOR	INTERFERENCE	11:51
SIMULATOR VAPOR	INTERFERENCE	11:52
SIMULATOR VAPOR	INTERFERENCE	11:53
SIMULATOR VAPOR	INTERFERENCE	11:54
BLANK TEST	0.000	11:55

Average out of range!



ACCURACY AND PRECISION TEST

STATE OF VERMONT

DataMaster DMT: 100157

Date: 08/26/2010
Time: 12:14:44

SUPERVISOR NAME: ALB .04IPA IN .08ETOH

SOLUTION LOT #: 09-30-08I
SOLUTION CONCENTRATION: 0.079

BLANK TEST	0.000	12:15
CALIBRATION CHECK	PASSED	12:15
SIMULATOR VAPOR	INTERFERENCE	12:16
SIMULATOR VAPOR	INTERFERENCE	12:17
SIMULATOR VAPOR	INTERFERENCE	12:18
SIMULATOR VAPOR	INTERFERENCE	12:19
SIMULATOR VAPOR	INTERFERENCE	12:20
SIMULATOR VAPOR	INTERFERENCE	12:21
SIMULATOR VAPOR	INTERFERENCE	12:22
SIMULATOR VAPOR	INTERFERENCE	12:23
SIMULATOR VAPOR	INTERFERENCE	12:24
SIMULATOR VAPOR	INTERFERENCE	12:25
BLANK TEST	0.000	12:25

Average out of range!



ACCURACY AND PRECISION TEST

STATE OF VERMONT

DataMaster DMT: 100157

Date: 08/26/2010
Time: 13:48:30

SUPERVISOR NAME: ALB .04MEOH IN H2O

SOLUTION LOT #: 11-08-04M
SOLUTION CONCENTRATION: 0.040

BLANK TEST	0.000	13:49
CALIBRATION CHECK	PASSED	13:49
SIMULATOR VAPOR	0.019	13:49
SIMULATOR VAPOR	0.020	13:50
SIMULATOR VAPOR	0.021	13:51
SIMULATOR VAPOR	0.021	13:52
SIMULATOR VAPOR	0.022	13:53
SIMULATOR VAPOR	0.021	13:54
SIMULATOR VAPOR	0.021	13:55
SIMULATOR VAPOR	0.022	13:56
SIMULATOR VAPOR	0.021	13:57
SIMULATOR VAPOR	0.021	13:58
BLANK TEST	0.000	13:59

Average out of range!

Average = 0.020
Std Dev = 0.0008



ACCURACY AND PRECISION TEST

STATE OF VERMONT

DataMaster DMT: 100157

Date: 08/26/2010
Time: 13:48:30

SUPERVISOR NAME: ALB .04MEOH IN H2O

SOLUTION LOT #: 11-08-04M
SOLUTION CONCENTRATION: 0.040

BLANK TEST	0.000	13:49
CALIBRATION CHECK	PASSED	13:49
Xq = 0.1002 (0.97%)		
SIMULATOR VAPOR	0.019	13:49
X[1] = 0.0193 (-0.0006) (-0.0029)		
SIMULATOR VAPOR	0.020	13:50
X[1] = 0.0206 (0.0001) (-0.0031)		
SIMULATOR VAPOR	0.021	13:51
X[1] = 0.0219 (-0.0002) (-0.0027)		
SIMULATOR VAPOR	0.021	13:52
X[1] = 0.0210 (-0.0007) (-0.0033)		
SIMULATOR VAPOR	0.022	13:53
X[1] = 0.0220 (-0.0006) (-0.0029)		
SIMULATOR VAPOR	0.021	13:54
X[1] = 0.0218 (-0.0002) (-0.0027)		
SIMULATOR VAPOR	0.021	13:55
X[1] = 0.0217 (-0.0002) (-0.0030)		
SIMULATOR VAPOR	0.022	13:56
X[1] = 0.0220 (-0.0006) (-0.0030)		
SIMULATOR VAPOR	0.021	13:57
X[1] = 0.0215 (-0.0012) (-0.0037)		
SIMULATOR VAPOR	0.021	13:58
X[1] = 0.0214 (-0.0016) (-0.0033)		
BLANK TEST	0.000	13:59

Average out of range!

Average = 0.020
Std Dev = 0.0008



National Patent
Analytical Systems, Inc.

Certificate of Calibration

100157
Instrument Serial Number

0938U
Standard Reference Number

The above instrument has been calibrated and is hereby certified for use in determining alcohol breath concentration when used in accordance with the operating guidelines by a properly trained operator.

6.28.10
Date

Technician

DIAGNOSTIC RESULT

DataMaster DMT: 100157
Location: ADDISON COUNTY SO
Calibration Date: 06/28/2010
Certification Date: 06/28/2010
Installation Date: 06/28/2010
Test Date: 06/28/2010
Time: 16:09

VERMONT
DEPARTMENT OF HEALTH

```
VERSIONS
DMT 1.05
PIC 2.06
Modem 1.06
Questions 1.00
Reports 1.02

TEMPERATURES
Sample Chamber = 48.86°C
Breath Tube    = 46.42°C

SETTINGS
Lamp Voltage   = 2.18 V
Cooler Voltage = 1.71 V
Bias Voltage   = 80.4 V
Chopper Freq   = 535.9 Hz

PUMP INFO
Flow Rate = 5.321 L/M

DETECTOR INFO
PUMP     ON       OFF
MAX(V)   0.5110   0.5108
MIN(V)   0.5085   0.5083

FILTER INFO
Filter 1   0.511   Zero = true
Filter 2   0.774   Zero = true
Filter 3   1.099   Zero = true

CALIBRATION CHECK
Xq = 0.1046   0.58%
```

DIAGNOSTIC RESULT

DataMaster DMT: 100157
Location:
Calibration Date: 06/28/2010
Certification Date:
Installation Date:
Test Date: 06/28/2010
Time: 13:37

VERMONT
DEPARTMENT OF HEALTH

```
VERSIONS
DMT 1.05
PIC 2.06
Modem 1.06
Questions 1.00
Reports 1.02

TEMPERATURES
Sample Chamber = 48.91°C
Breath Tube    = 46.53°C

SETTINGS
Lamp Voltage   = 2.18 V
Cooler Voltage = 1.71 V
Bias Voltage   = 80.4 V
Chopper Freq   = 536.1 Hz

PUMP INFO
Flow Rate = 5.323 L/M

DETECTOR INFO
PUMP     ON       OFF
MAX(V)   0.4937   0.4949
MIN(V)   0.4902   0.4933

FILTER INFO
Filter 1   0.495   Zero = true
Filter 2   0.762   Zero = true
Filter 3   1.074   Zero = true

CALIBRATION CHECK
Xq = 0.1060   0.75%
```



Check-In Confirmation Test

STATE OF VERMONT

DataMaster DMT: 100157
ADDISON COUNTY SO
Certification Date: 06/28/2010
Installation Date: 06/28/2010

Date: 06/28/2010
Time: 15:58:09

Subject Name: T T T
Date of Birth: 09/09/1975 Age: 34
Gender: M
License #:

Test Operator Name: T T T
Agency: ADDISON COUNTY SO

BLANK TEST	0.000	15:59
CALIBRATION CHECK	PASSED	15:59
Xq = 0.1058 (0.58%)		
SIMULATOR VAPOR	0.100	15:59
X[1] = 0.1007 (-0.0009) (0.0018)		
BLANK TEST	0.000	16:00
SUBJECT SAMPLE (Vol=1.8L)	0.098	16:00
X[1] = 0.0985 (-0.0021) (0.0006)		

All alcohol values reported in g/210L

Sim. Vapor Concentration	0.098
Sim. Vapor Range	0.093 - 0.103

INSTALLATION REPORT

DataMaster DMT: 100157
Location: ADDISON COUNTY SO
Calibration Date: 06/28/2010
Certification Date: 06/28/2010
Installation Date: 06/28/2010
Installed By: T T T

VERMONT
DEPARTMENT OF HEALTH

Site meets specification in doc P-ALC-210

Diagnostic Results

VERSIONS
DMT 1.05
PIC 2.06
Modem 1.06
Questions 1.00
Reports 1.02

TEMPERATURES
Sample Chamber = 48.74°C
Breath Tube = 47.06°C

SETTINGS
Lamp Voltage = 2.18 V
Cooler Voltage = 1.71 V
Bias Voltage = 80.4 V
Chopper Freq = 535.8 Hz

PUMP INFO
Flow Rate = 5.375 L/M

DETECTOR INFO
PUMP ON OFF
MAX(V) 0.5032 0.5043
MIN(V) 0.5004 0.5026

FILTER INFO
Filter 1 0.505 Zero = true
Filter 2 0.771 Zero = true
Filter 3 1.090 Zero = true

CALIBRATION CHECK
Xq = 0.1054 0.17%

INSTALLATION PASSED

Accuracy and Precision Check

Concentration = 0.098 g/210L
Lot # = 87
Average = 0.099 g/210L
Std Dev = 0.000

RFI Interference Test

RFI detected

Sample Acceptance Test

Passed



Performed by _____ **Date** 06/28/2010

Reviewed by _____ **Date** _____

Linearity Check Results

True Value	Reported Average	Std Dev
0.041 g/210L	0.041 g/210L	0.001
Lot # 76		
0.099 g/210L	0.098 g/210L	0.000
Lot # 67		
0.097 g/210L	0.098 g/210L	0.000
Lot # 98765		
0.430 g/210L	0.438 g/210L	0.001
Lot # YT45		

$R^2 = 1.0000$



Acetone Interference Test

Lot # 8765

Interference Detected

Mouth Alcohol Test

Mouth Alcohol Detected



RF Interference Test

RFI detected

Sample Acceptance Test

Passed



CERTIFICATION PASSED

Performed by _____ **Date** 06/28/2010

Reviewed by _____ **Date** _____

CERTIFICATION REPORT

DataMaster DMT: 100157
Calibration Date: 06/28/2010
Certification Date: 06/28/2010
Certified by: T T T

VERMONT
DEPARTMENT OF HEALTH

Diagnostic Results

VERSIONS
DMT 1.05
PIC 2.06
Modem 1.06
Questions 1.00
Reports 1.02

TEMPERATURES
Sample Chamber = 48.91°C
Breath Tube = 46.61°C

SETTINGS
Lamp Voltage = 2.18 V
Cooler Voltage = 1.71 V
Bias Voltage = 80.3 V
Chopper Freq = 536.1 Hz

PUMP INFO
Flow Rate = 5.313 L/M

DETECTOR INFO
PUMP ON OFF
MAX(V) 0.4955 0.4955
MIN(V) 0.4921 0.4945

FILTER INFO
Filter 1 0.496 Zero = true
Filter 2 0.762 Zero = true
Filter 3 1.077 Zero = true

CALIBRATION CHECK
Xq = 0.1055 0.33%

Options

Printer: yes
Number of Copies: 1
Number of Supervisor Tests: 10
Tolerance Check: no
Data Collection: yes
Units: g/210L
Simulator Check: no
Uses Wet Bath Simulator: yes
Simulator Nominal: 0.097
Digital Simulator: none
Simulator Before: yes
Simulator Between: yes
Simulator After: yes
Number of Subject Tests: 2
AskQuestions: yes
Query Refusal: yes
Alcohol Display: yes
Show Two Digits: no
Volume Display: yes
Number of Calibration Tests: 2
Calibrate with Wet Bath: yes
Calibration Nominal: 0.100

CALIBRATION REPORT

DataMaster DMT: 100157
Calibration Date: 06/28/2010
Calibrated by: T T T
Lot: 76

VERMONT
DEPARTMENT OF HEALTH

```
Ca  =  0.097000
CAL =  0.989619    0.800 <= CAL < 1.200
b1  =  0.001368    0.000 <=  b1 < 0.004
b2  =  0.005893    0.001 <=  b2 < 0.012
b3  =  0.000240    0.000 <=  b3 < 0.004
Xq  =  0.105236    0.050 <=  Xq < 0.200
a21 =  1.165906    1.050 <= a21 < 1.300
a31 =  0.549310    0.300 <= a31 < 0.800
```

Performed by _____ Date 06/28/2010

Section 10: Live Subject Tests

This Instrument not tested for this specification

Section 11: Communication Notes

March/April 2010

~~October 2009~~

100157

~~Third~~ 4th Round Testing Protocol for DMT Instruments

Power Up Procedure:

a) Open unit and inspect connections. Ensure all tubing is secure. Check that all cables are properly seated and tied down or strapped together as necessary. Check that all screws are in place and snug; this will require the removal of the side pieces to allow the technician access. Correct as needed. 8/23/10

b) Manually turn the filter wheel. The filter wheel should turn smoothly and should not rub or wobble. Correct as needed. ALb

c) Power on the instrument. Allow approximately one hour for it to come to temperature.

d) Access the SETUP menu. On the MISC tab, type "State of Vermont" in the NAME field. In the EXPLORE box, highlight STYLUS and press the OK button to activate the stylus settings. Recalibrate the stylus. Once complete, exit the SETUP menu by pressing the OK button to save the settings. If you exit the SETUP menu using the CANCEL button, none of the changes will be saved.

e) Enter the TECH screen. Scroll through all of the filter settings to ensure the filter wheel is activating properly.

f) Optimize the detector voltage
-Adjust the cooler voltage setting to 1.70Vdc +/-0.10.
-Manually manipulate the detector position to get the detector voltage as low as possible.
-Adjust the Lamp voltage setting so that the detector voltage displayed is approximately +0.000 +/- 0.100. SA
-Note the initial noise of the detector voltage once unit has had a minute to equilibrate. Noise tolerance at NPAS is +/-0.003 on a detector. 9/10/10
-Save the voltage settings
-Per VDHL protocol (P-Alc-116 Power-Up SOP), the detector voltage must read 0.000V with a tolerance of +/- 0.100V. Correct as needed.

g) While still in the TECH screen, press SET in the RF Sensitivity box to automatically set the RF sensitivity level. Press SAVE to save the settings. Confirm the setting by keying a radio near the instrument. Correct as needed.

h) Using the large syringe, push 1.5L of air into the breath tube. Monitor the volume of air displayed on the TECH screen. The instrument should read 1.5L +/- 0.2L. If the instrument is out of range, adjust the distance between the "T" fittings on the breath exhaust tubing. Retest until the volume reads appropriately. SA

i) Once all the TECH settings are correct, re-save every setting. Exit the TECH screen.

j) Close up the instrument using locking screws on the top four holes. Attach the stylus to the DMT using a locking screw in the center right screw hole nearest the simulator tower. Attach the keyboard to the top of the instrument using sticky-back velcro. Wrap a small piece of velcro around the top of the stylus and affix the other side to the front of the instrument so the stylus can rest on the front of the unit. 9/10/10

k) Calibrate the instrument following the Calibration Protocol using Calibration Solution (0.100 EtOH). During Calibration and Certification, it may be necessary to turn off the simulator temperature monitoring if using Guth 2100 Simulators as these do not have BNC connectors for temperature monitoring. To do this, open the SETUP menu. In the Simulator box, select NONE as the type of

9/10/10 SH

100157

simulator. Press OK to exit. Document the room temperature and humidity level at the time of calibration in the communication notebook.

l) Certify the instrument following the Certification Protocol.

m) Preliminary interference testing. Run the following concentrations as an Accuracy and Precision Test with n=10. The instrument must report Interference at least 80% of the time.

SH 9/13/10

0.01% Acetone in 0.08% EtOH ✓ SH 9/13/10
0.04% MeOH in Water ✓ SH 9/13/10
0.04% Isopropanol in Water ✓ SH 9/13/10

The instrument is now ready for performance testing. If any of the above steps fail or there is **any** problem with the instrument, the DataMaster Team should be consulted as to the action required. It may be necessary to return an instrument to the manufacture for refurbishment prior to use by VDHL. All adjustments and repairs to an instrument should be documented in the instrument's testing notebook located in the testing binder. All reports generated during the Power Up Procedure should be filed in the power up section of that instrument's testing binder. If the instrument has passed, it should now be Installed to VDHL.

RFI

100157

Performance Testing

a) Linearity Test: Four concentrations of EtOH solutions will be used, 0.020, 0.080, 0.160, and 0.400. Each solution will be run using the Accuracy and Precision test (n=10). All four concentrations will be run consecutively and will count as one linearity test. Five linearity tests will be run on each instrument. Linearity tests should not be completed back to back. Testing should allow for inter- and intra- day performance.

b) ~~Low Level Performance: A solution of 0.040EtOH will be tested using the Accuracy and Precision test~~ (n=10) once

c) Interference Test: The following concentrations will be tested using the Accuracy and Precision test (n=10). The instrument should report Interference at least 80% of the time.

SH 9/13/10

0.01% Acetone in 0.04% EtOH ✓ SH 9/13/10
0.02% Acetone in 0.04% EtOH ✓ SH 9/13/10
0.04% MeOH in 0.08% EtOH ✓ SH 9/10/10
0.04% Isopropanol in 0.08% EtOH ✓ SH 9/13/10

d) Limited Volume Testing:

Target Volume	Flow (lpm)	Time (sec.)	1	2	3	4	5	6	7	Ave.	% s.d.
1.5 L	8.2	11									
	15	6									
	22.5	4									
2.0 L	10	12									
	15	8									
	24	5									

Pressure: _____ Flow rate +/- _____ time +/- _____

RF