DATAMASTER DMT 100140

VERMONT DEPARTMENT OF HEALTH LABORATORY

PERFORMANCE TESTING RESULTS

August 2010

Table of Contents

Section 1: Results	1
Section 2: NPAS Tests	5
Section 3: Power-Up Test	7
Section 4: Linearity	12
Section 5: Interference	37
Section 6: Low Level Accuracy	46
Section 7: Breath Volume	48
Section 8: Software Testing	50
Section 9: Error Reports	52
Section 10: Live Subject Tests	53
Section 11: Communication Notes	55

Section 1: Results


100140Linearity
GC Certified Value
0.450
0.400
0.350
0.300
0.250
0.200
0.150
0.100
0.050
0.000
0.000
0.050
0.100
0.150
0.200
0.250
0.300
0.350
0.400
0.450
DMT Result
y = 0.9959x - 0.0025
R2 = 0.9999

DMT Serial # 100140

Operator	Date	Lot #	GC Value	DMT Result	Std Dev	% Recovery	
ALB	4/26/2010	10-19-080	0.081	0.077	0.0004	95.1	Outside 5%
DMR	4/26/2010	10-20-160	0.159	0.156	0.0003	98.1	Outside 10%
ALB	4/26/2010	10-14-400	0.387	0.384	0.0008	99.2	
DMR	4/26/2010	10-21-020	0.019	0.018	0.0000	94.7	
ALB	4/26/2010	10-21-020	0.019	0.017	0.0003	89.5	
		10-19-080	0.081	0.076	0.0006	93.8	
		10-20-160	0.159	0.154	0.0004	96.9	
ALB	4/27/2010	10-21-020	0.019	0.017	0.0004	89.5	
		10-19-080	0.081	0.076	0.0004	93.8	
		10-20-160	0.159	0.154	0.0000	96.9	
		10-14-400	0.387	0.382	0.0011	98.7	
ALB	5/11/2010	10-32-400	0.373	0.370	0.0010	99.2	
		10-20-160	0.160	0.158	0.0004	98.8	
		10-30-080	0.081	0.080	0.0004	98.8	
		10-21-020	0.019	0.018	0.0003	94.7	
DMR	5/11/2010	10-21-020	0.019	0.018	0.0004	94.7	
		10-30-080	0.081	0.079	0.0005	97.5	
		10-31-160	0.160	0.158	0.0004	98.8	
		10-32-400	0.373	0.371	0.0008	99.5	
				Average	0.0005	96.2	
				Std Dev	0.0003	3.08	
				Range	0.0000 - 0.0011	89.5 - 99.5	



DMT Serial # 100140

Operator	Date	Concentration	Lot Number	DMT Result	Apparent EtOH	N = 10
ALB	04/26/10	.01 Acetone in .08 EtOH	10-24-08A	9 OUT OF 10	0.08	
SH	04/21/10	.01 Acetone in .04 EtOH	09-27-04A1	8 OUT OF 10	0.04	
SH	04/21/10	.02 Acetone in .04 EtOH	09-31-04A2	INTERFERENCE	0.05	
DMR	04/26/10	.04 MeOH in water	10-25-04M	INTERFERENCE	0.02	
DMR	04/26/10	.04 MeOH in water	10-25-04M	INTERFERENCE	0.02	
SH	04/21/10	.04 MeOH in .08 EtOH	10-10-08M4	INTERFERENCE	0.1	
DMR	04/26/10	.04 IPA in water	10-26-04I	INTERFERENCE	0.02	
SH	04/21/10	.04 IPA in .08 EtOH	09-30-08I	INTERFERENCE	0.1	

Section 2: NPAS Tests

National Patent
Analytical Systems, Inc.

Certificate of Calibration

100140

Instrument Serial Number

09380

Standard Reference Number

The above instrument has been calibrated and is hereby certified for use in determining alcohol breath concentration when used in accordance with the operating guidelines by a properly trained operator.

3-23-10

Date

Charles E Townsend

Technician

Section 3: Power-Up Tests

CALIBRATION REPORT

DataMaster DMT: 100140
Calibration Date: 04/20/2010
Calibrated by: STEVEN E HARNOIS
Lot: 10-22-100

VERMONT
DEPARTMENT OF HEALTH

Ca =	0.101000		
CAL =	0.983461	0.800 <= CAL < 1.200	
b1 =	0.000000	0.000 <= b1 < 0.004	
b2 =	0.005406	0.001 <= b2 < 0.012	
b3 =	0.000000	0.000 <= b3 < 0.004	
Xq =	0.076683	0.050 <= Xq < 0.200	
a21 =	1.192462	1.050 <= a21 < 1.300	
a31 =	0.403027	0.300 <= a31 < 0.800	

Cal Done 4/20/10
Cert Almost Done 4/20/10
Unit Powered Down
(My Mistake)
Cert Redone 4/21/10
SH

Performed by _____ **Date** 04/20/2010

ACCURACY AND PRECISION TEST

STATE OF VERMONT

DataMaster DMT: 100140

Date: 04/20/2010
Time: 10:51:52

SUPERVISOR NAME: SEH TEST CAL CHECK

SOLUTION LOT #: 10-23-100
SOLUTION CONCENTRATION: 0.100

BLANK TEST	0.000	10:52
CALIBRATION CHECK	PASSED	10:52
SIMULATOR VAPOR	0.099	10:53
SIMULATOR VAPOR	0.100	10:54
SIMULATOR VAPOR	0.100	10:55
SIMULATOR VAPOR	0.100	10:56
SIMULATOR VAPOR	0.100	10:57
SIMULATOR VAPOR	0.100	10:58
SIMULATOR VAPOR	0.101	10:59
SIMULATOR VAPOR	0.100	11:00
SIMULATOR VAPOR	0.100	11:01
SIMULATOR VAPOR	0.100	11:02
BLANK TEST	0.000	11:02

Average = 0.099
Std Dev = 0.0004



CERTIFICATION REPORT

DataMaster DMT: 100140
Calibration Date: 04/20/2010
Certification Date: 04/21/2010
Certified by: STEVEN E HARNOIS

VERMONT
DEPARTMENT OF HEALTH

Diagnostic Results

VERSIONS
DMT 1.04
PIC 2.05
Modem 1.06
Questions 1.00
Reports 1.02

TEMPERATURES
Sample Chamber = 48.97°C
Breath Tube = 47.50°C

SETTINGS
Lamp Voltage = 1.52 V
Cooler Voltage = 1.65 V
Bias Voltage = 80.4 V
Chopper Freq = 537.5 Hz

PUMP INFO
Flow Rate = 5.477 L/M

DETECTOR INFO

PUMP	ON	OFF
MAX(V)	0.0760	0.0793
MIN(V)	0.0746	0.0780

FILTER INFO
Filter 1 0.078 Zero = true
Filter 2 0.430 Zero = true
Filter 3 1.031 Zero = true

CALIBRATION CHECK
Xq = 0.0759 0.91%

Options

Printer: yes
Number of Copies: 1
Number of Supervisor Tests: 10
Tolerance Check: yes
Data Collection: yes
Units: g/210L
Simulator Check: yes
Uses Wet Bath Simulator: yes
Simulator Nominal: 0.101
Digital Simulator: none
Simulator Before: yes
Simulator Between: yes
Simulator After: no
Number of Subject Tests: 2
AskQuestions: yes
Query Refusal: yes
Alcohol Display: yes
Show Two Digits: no
Volume Display: yes
Number of Calibration Tests: 1
Calibrate with Wet Bath: yes
Calibration Nominal: 0.100

Linearity Check Results

True Value	Reported Average	Std Dev
0.019 g/210L Lot # 10-21-020	0.018 g/210L	0.001
0.081 g/210L Lot # 10-19-080	0.077 g/210L	0.000
0.159 g/210L Lot # 10-20-160	0.157 g/210L	0.001
0.387 g/210L Lot # 10-14-400	0.395 g/210L	0.001

$R^2 = 0.9998$



Acetone Interference Test

Lot # 10-24-08A

Interference Detected

Mouth Alcohol Test

Mouth Alcohol Detected



RF Interference Test

RFI detected

Sample Acceptance Test

Passed



CERTIFICATION PASSED

Performed by _____ **Date** 04/21/2010

Reviewed by _____ **Date** 4/21/10

Section 4: Linearity

DIAGNOSTIC RESULT

DataMaster DMT: 100140
Location:
Calibration Date: 04/20/2010
Certification Date: 04/21/2010
Installation Date: 04/21/2010
Test Date: 04/26/2010
Time: 10:02

VERMONT
DEPARTMENT OF HEALTH

```
VERSIONS
DMT 1.04
PIC 2.05
Modem 1.06
Questions 1.00
Reports 1.02

TEMPERATURES
Sample Chamber = 48.68°C
Breath Tube    = 48.11°C

SETTINGS
Lamp Voltage   = 1.52 V
Cooler Voltage = 1.65 V
Bias Voltage   = 80.4 V
Chopper Freq   = 539.3 Hz

PUMP INFO
Flow Rate = 5.424 L/M

DETECTOR INFO
PUMP     ON       OFF
MAX(V)   0.0898   0.0931
MIN(V)   0.0886   0.0917

FILTER INFO
Filter 1  0.092  Zero = true
Filter 2  0.446  Zero = true
Filter 3  1.051  Zero = true

CALIBRATION CHECK
Xq = 0.0765  0.11%
```

ACCURACY AND PRECISION TEST

STATE OF VERMONT

DataMaster DMT: 100140

Date: 04/26/2010
Time: 10:18:27

SUPERVISOR NAME: A L B

SOLUTION LOT #: 10-19-080
SOLUTION CONCENTRATION: 0.081

BLANK TEST	0.000	10:19
CALIBRATION CHECK	PASSED	10:19
SIMULATOR VAPOR	0.077	10:19
SIMULATOR VAPOR	0.077	10:20
SIMULATOR VAPOR	0.077	10:21
SIMULATOR VAPOR	0.077	10:22
SIMULATOR VAPOR	0.077	10:23
SIMULATOR VAPOR	0.078	10:24
SIMULATOR VAPOR	0.077	10:25
SIMULATOR VAPOR	0.077	10:26
SIMULATOR VAPOR	0.078	10:27
SIMULATOR VAPOR	0.077	10:28
BLANK TEST	0.000	10:29

Average = 0.077
Std Dev = 0.0004



ACCURACY AND PRECISION TEST

STATE OF VERMONT

DataMaster DMT: 100140

Date: 04/26/2010
Time: 10:32:33

SUPERVISOR NAME: DARCY M RICHARDSON

SOLUTION LOT #: 10-20-160
SOLUTION CONCENTRATION: 0.159

BLANK TEST	0.000	10:33
CALIBRATION CHECK	PASSED	10:33
SIMULATOR VAPOR	0.156	10:33
SIMULATOR VAPOR	0.156	10:34
SIMULATOR VAPOR	0.157	10:35
SIMULATOR VAPOR	0.156	10:36
SIMULATOR VAPOR	0.156	10:37
SIMULATOR VAPOR	0.156	10:38
SIMULATOR VAPOR	0.156	10:39
SIMULATOR VAPOR	0.156	10:40
SIMULATOR VAPOR	0.156	10:41
SIMULATOR VAPOR	0.156	10:42
BLANK TEST	0.000	10:43

Average = 0.156
Std Dev = 0.0003



ACCURACY AND PRECISION TEST

STATE OF VERMONT

DataMaster DMT: 100140

Date: 04/26/2010
Time: 10:48:00

SUPERVISOR NAME: A L B

SOLUTION LOT #: 10-14-400
SOLUTION CONCENTRATION: 0.387

BLANK TEST	0.000	10:48
CALIBRATION CHECK	PASSED	10:49
SIMULATOR VAPOR	0.386	10:49
SIMULATOR VAPOR	0.386	10:50
SIMULATOR VAPOR	0.385	10:51
SIMULATOR VAPOR	0.385	10:52
SIMULATOR VAPOR	0.384	10:53
SIMULATOR VAPOR	0.385	10:54
SIMULATOR VAPOR	0.384	10:55
SIMULATOR VAPOR	0.384	10:56
SIMULATOR VAPOR	0.384	10:57
SIMULATOR VAPOR	0.384	10:58
BLANK TEST	0.000	10:59

Average = 0.384
Std Dev = 0.0008



ACCURACY AND PRECISION TEST

STATE OF VERMONT

DataMaster DMT: 100140

Date: 04/26/2010
Time: 11:07:07

SUPERVISOR NAME: DARCY M RICHARDSON

SOLUTION LOT #: 10-21-020
SOLUTION CONCENTRATION: 0.019

BLANK TEST	0.000	11:08
CALIBRATION CHECK	PASSED	11:08
SIMULATOR VAPOR	0.018	11:08
SIMULATOR VAPOR	0.018	11:09
SIMULATOR VAPOR	0.018	11:10
SIMULATOR VAPOR	0.018	11:11
SIMULATOR VAPOR	0.018	11:12
SIMULATOR VAPOR	0.018	11:13
SIMULATOR VAPOR	0.018	11:14
SIMULATOR VAPOR	0.018	11:15
SIMULATOR VAPOR	0.018	11:16
SIMULATOR VAPOR	0.018	11:17
BLANK TEST	0.000	11:17

Average = 0.018
Std Dev = 0.0000



ACCURACY AND PRECISION TEST

STATE OF VERMONT

DataMaster DMT: 100140

Date: 04/26/2010
Time: 13:28:03

SUPERVISOR NAME: A L B

SOLUTION LOT #: 10-21-020
SOLUTION CONCENTRATION: 0.019

BLANK TEST	0.000	13:28
CALIBRATION CHECK	PASSED	13:29
SIMULATOR VAPOR	0.018	13:29
SIMULATOR VAPOR	0.017	13:30
SIMULATOR VAPOR	0.018	13:31
SIMULATOR VAPOR	0.018	13:32
SIMULATOR VAPOR	0.018	13:33
SIMULATOR VAPOR	0.018	13:34
SIMULATOR VAPOR	0.018	13:35
SIMULATOR VAPOR	0.018	13:36
SIMULATOR VAPOR	0.018	13:37
SIMULATOR VAPOR	0.018	13:38
BLANK TEST	0.000	13:38

Average out of range!

Average = 0.017
Std Dev = 0.0003



ACCURACY AND PRECISION TEST

STATE OF VERMONT

DataMaster DMT: 100140

Date: 04/26/2010
Time: 13:47:31

SUPERVISOR NAME: A L B

SOLUTION LOT #: 10-19-080
SOLUTION CONCENTRATION: 0.081

BLANK TEST	0.000	13:48
CALIBRATION CHECK	PASSED	13:48
SIMULATOR VAPOR	0.075	13:48
SIMULATOR VAPOR	0.076	13:49
SIMULATOR VAPOR	0.076	13:50
SIMULATOR VAPOR	0.077	13:51
SIMULATOR VAPOR	0.076	13:52
SIMULATOR VAPOR	0.077	13:53
SIMULATOR VAPOR	0.077	13:54
SIMULATOR VAPOR	0.077	13:55
SIMULATOR VAPOR	0.076	13:56
SIMULATOR VAPOR	0.076	13:57
BLANK TEST	0.000	13:58

Average = 0.076
Std Dev = 0.0006



ACCURACY AND PRECISION TEST

STATE OF VERMONT

DataMaster DMT: 100140

Date: 04/26/2010
Time: 14:00:46

SUPERVISOR NAME: A L B

SOLUTION LOT #: 10-20-160
SOLUTION CONCENTRATION: 0.159

BLANK TEST	0.000	14:01
CALIBRATION CHECK	PASSED	14:01
SIMULATOR VAPOR	0.153	14:02
SIMULATOR VAPOR	0.154	14:03
SIMULATOR VAPOR	0.154	14:04
SIMULATOR VAPOR	0.154	14:05
SIMULATOR VAPOR	0.154	14:06
SIMULATOR VAPOR	0.154	14:07
SIMULATOR VAPOR	0.155	14:08
SIMULATOR VAPOR	0.154	14:09
SIMULATOR VAPOR	0.154	14:10
SIMULATOR VAPOR	0.154	14:11
BLANK TEST	0.000	14:11

Average = 0.154
Std Dev = 0.0004



ACCURACY AND PRECISION TEST

STATE OF VERMONT

DataMaster DMT: 100140

Date: 04/26/2010
Time: 14:13:40

SUPERVISOR NAME: DARCY M RICHARDSON

SOLUTION LOT #: 10-14-400
SOLUTION CONCENTRATION: 0.387

BLANK TEST	0.000	14:14
CALIBRATION CHECK	PASSED	14:14
SIMULATOR VAPOR	0.384	14:15
SIMULATOR VAPOR	0.384	14:16
SIMULATOR VAPOR	0.384	14:17
SIMULATOR VAPOR	0.382	14:18
SIMULATOR VAPOR	0.382	14:19
SIMULATOR VAPOR	0.384	14:20
SIMULATOR VAPOR	0.383	14:21
SIMULATOR VAPOR	0.383	14:22
SIMULATOR VAPOR	0.383	14:23
SIMULATOR VAPOR	0.382	14:24
BLANK TEST	0.000	14:25

Average = 0.383
Std Dev = 0.0008



DIAGNOSTIC RESULT

DataMaster DMT: 100140
Location:
Calibration Date: 04/20/2010
Certification Date: 04/21/2010
Installation Date: 04/21/2010
Test Date: 04/27/2010
Time: 07:33

VERMONT
DEPARTMENT OF HEALTH

```
VERSIONS
DMT 1.04
PIC 2.05
Modem 1.06
Questions 1.00
Reports 1.02

TEMPERATURES
Sample Chamber = 49.09°C
Breath Tube    = 48.07°C

SETTINGS
Lamp Voltage   = 1.52 V
Cooler Voltage = 1.65 V
Bias Voltage   = 80.4 V
Chopper Freq   = 539.1 Hz

PUMP INFO
Flow Rate = 5.446 L/M

DETECTOR INFO
PUMP     ON       OFF
MAX(V)   0.0915   0.0941
MIN(V)   0.0896   0.0927

FILTER INFO
Filter 1  0.093  Zero = true
Filter 2  0.445  Zero = true
Filter 3  1.049  Zero = true

CALIBRATION CHECK
Xq = 0.0765  0.23%
```

ACCURACY AND PRECISION TEST

STATE OF VERMONT

DataMaster DMT: 100140

Date: 04/27/2010
Time: 07:35:00

SUPERVISOR NAME: A L B

SOLUTION LOT #: 10-21-020
SOLUTION CONCENTRATION: 0.019

BLANK TEST	0.000	07:35
CALIBRATION CHECK	PASSED	07:36
SIMULATOR VAPOR	0.017	07:36
SIMULATOR VAPOR	0.017	07:37
SIMULATOR VAPOR	0.018	07:38
SIMULATOR VAPOR	0.018	07:39
SIMULATOR VAPOR	0.018	07:40
SIMULATOR VAPOR	0.018	07:41
SIMULATOR VAPOR	0.018	07:42
SIMULATOR VAPOR	0.018	07:43
SIMULATOR VAPOR	0.018	07:44
SIMULATOR VAPOR	0.018	07:45
BLANK TEST	0.000	07:45

Average out of range!

Average = 0.017
Std Dev = 0.0004



ACCURACY AND PRECISION TEST

STATE OF VERMONT

DataMaster DMT: 100140

Date: 04/27/2010
Time: 07:48:42

SUPERVISOR NAME: A L B

SOLUTION LOT #: 10-19-080
SOLUTION CONCENTRATION: 0.081

BLANK TEST	0.000	07:49
CALIBRATION CHECK	PASSED	07:49
SIMULATOR VAPOR	0.077	07:50
SIMULATOR VAPOR	0.076	07:51
SIMULATOR VAPOR	0.076	07:52
SIMULATOR VAPOR	0.076	07:53
SIMULATOR VAPOR	0.077	07:53
SIMULATOR VAPOR	0.076	07:54
SIMULATOR VAPOR	0.076	07:55
SIMULATOR VAPOR	0.076	07:56
SIMULATOR VAPOR	0.077	07:57
SIMULATOR VAPOR	0.076	07:58
BLANK TEST	0.000	07:59

Average = 0.076
Std Dev = 0.0004



ACCURACY AND PRECISION TEST

STATE OF VERMONT

DataMaster DMT: 100140

Date: 04/27/2010
Time: 08:02:34

SUPERVISOR NAME: A L B

SOLUTION LOT #: 10-20-160
SOLUTION CONCENTRATION: 0.159

BLANK TEST	0.000	08:03
CALIBRATION CHECK	PASSED	08:03
SIMULATOR VAPOR	0.154	08:04
SIMULATOR VAPOR	0.154	08:05
SIMULATOR VAPOR	0.154	08:06
SIMULATOR VAPOR	0.154	08:07
SIMULATOR VAPOR	0.154	08:08
SIMULATOR VAPOR	0.154	08:08
SIMULATOR VAPOR	0.154	08:09
SIMULATOR VAPOR	0.154	08:10
SIMULATOR VAPOR	0.154	08:11
SIMULATOR VAPOR	0.154	08:12
BLANK TEST	0.000	08:13

Average = 0.154
Std Dev = 0.0000



ACCURACY AND PRECISION TEST

STATE OF VERMONT

DataMaster DMT: 100140

Date: 04/27/2010
Time: 08:15:07

SUPERVISOR NAME: A L B

SOLUTION LOT #: 10-14-400
SOLUTION CONCENTRATION: 0.387

BLANK TEST	0.000	08:16
CALIBRATION CHECK	PASSED	08:16
SIMULATOR VAPOR	0.384	08:16
SIMULATOR VAPOR	0.384	08:17
SIMULATOR VAPOR	0.382	08:18
SIMULATOR VAPOR	0.382	08:19
SIMULATOR VAPOR	0.383	08:20
SIMULATOR VAPOR	0.381	08:21
SIMULATOR VAPOR	0.382	08:22
SIMULATOR VAPOR	0.382	08:23
SIMULATOR VAPOR	0.384	08:24
SIMULATOR VAPOR	0.381	08:25
BLANK TEST	0.000	08:26

Average = 0.382
Std Dev = 0.0011



DIAGNOSTIC RESULT

DataMaster DMT: 100140
Location:
Calibration Date: 04/20/2010
Certification Date: 04/21/2010
Installation Date: 04/21/2010
Test Date: 05/11/2010
Time: 10:30

VERMONT
DEPARTMENT OF HEALTH

```
VERSIONS
DMT 1.04
PIC 2.05
Modem 1.06
Questions 1.00
Reports 1.02

TEMPERATURES
Sample Chamber = 48.68°C
Breath Tube    = 48.06°C

SETTINGS
Lamp Voltage   = 1.52 V
Cooler Voltage = 1.65 V
Bias Voltage   = 80.4 V
Chopper Freq   = 541.5 Hz

PUMP INFO
Flow Rate = 5.644 L/M

DETECTOR INFO
PUMP     ON       OFF
MAX(V)   0.0994   0.1022
MIN(V)   0.0976   0.1006

FILTER INFO
Filter 1  0.101  Zero = true
Filter 2  0.454  Zero = true
Filter 3  1.061  Zero = true

CALIBRATION CHECK
Xq = 0.0759  0.91%
```

ACCURACY AND PRECISION TEST

STATE OF VERMONT

DataMaster DMT: 100140

Date: 05/11/2010
Time: 10:34:11

SUPERVISOR NAME: AMANDA BOLDUC

SOLUTION LOT #: 10-32-400
SOLUTION CONCENTRATION: 0.373

BLANK TEST	0.000	10:35
CALIBRATION CHECK	PASSED	10:35
SIMULATOR VAPOR	0.369	10:35
SIMULATOR VAPOR	0.370	10:36
SIMULATOR VAPOR	0.369	10:37
SIMULATOR VAPOR	0.370	10:38
SIMULATOR VAPOR	0.371	10:39
SIMULATOR VAPOR	0.370	10:40
SIMULATOR VAPOR	0.369	10:41
SIMULATOR VAPOR	0.371	10:42
SIMULATOR VAPOR	0.371	10:43
SIMULATOR VAPOR	0.372	10:45
BLANK TEST	0.000	10:45

Average = 0.370
Std Dev = 0.0010



ACCURACY AND PRECISION TEST

STATE OF VERMONT

DataMaster DMT: 100140

Date: 05/11/2010
Time: 11:01:09

SUPERVISOR NAME: AMANDA BOLDUC

SOLUTION LOT #: 10-30-080
SOLUTION CONCENTRATION: 0.081

BLANK TEST	0.000	11:02
CALIBRATION CHECK	PASSED	11:02
SIMULATOR VAPOR	0.081	11:02
SIMULATOR VAPOR	0.080	11:03
SIMULATOR VAPOR	0.080	11:04
SIMULATOR VAPOR	0.080	11:05
SIMULATOR VAPOR	0.080	11:06
SIMULATOR VAPOR	0.080	11:07
SIMULATOR VAPOR	0.080	11:08
SIMULATOR VAPOR	0.081	11:09
SIMULATOR VAPOR	0.080	11:10
SIMULATOR VAPOR	0.080	11:11
BLANK TEST	0.000	11:12

Average = 0.080
Std Dev = 0.0004



ACCURACY AND PRECISION TEST

STATE OF VERMONT

DataMaster DMT: 100140

Date: 05/11/2010
Time: 10:48:28

SUPERVISOR NAME: AMANDA BOLDUC

SOLUTION LOT #: 10-~~21~~ 20-160 10/22/10 [illegible]
SOLUTION CONCENTRATION: 0.160

BLANK TEST	0.000	10:49
CALIBRATION CHECK	PASSED	10:49
SIMULATOR VAPOR	0.158	10:49
SIMULATOR VAPOR	0.159	10:50
SIMULATOR VAPOR	0.158	10:51
SIMULATOR VAPOR	0.159	10:52
SIMULATOR VAPOR	0.159	10:53
SIMULATOR VAPOR	0.159	10:54
SIMULATOR VAPOR	0.159	10:55
SIMULATOR VAPOR	0.159	10:56
SIMULATOR VAPOR	0.158	10:57
SIMULATOR VAPOR	0.159	10:58
BLANK TEST	0.000	10:59

Average = 0.158
Std Dev = 0.0004



ACCURACY AND PRECISION TEST

STATE OF VERMONT

DataMaster DMT: 100140

Date: 05/11/2010
Time: 11:14:19

SUPERVISOR NAME: AMANDA BOLDUC

SOLUTION LOT #: 10-21-020
SOLUTION CONCENTRATION: 0.019

BLANK TEST	0.000	11:15
CALIBRATION CHECK	PASSED	11:15
SIMULATOR VAPOR	0.018	11:15
SIMULATOR VAPOR	0.018	11:16
SIMULATOR VAPOR	0.018	11:17
SIMULATOR VAPOR	0.019	11:18
SIMULATOR VAPOR	0.018	11:19
SIMULATOR VAPOR	0.018	11:20
SIMULATOR VAPOR	0.018	11:21
SIMULATOR VAPOR	0.018	11:22
SIMULATOR VAPOR	0.018	11:23
SIMULATOR VAPOR	0.018	11:24
BLANK TEST	0.000	11:25

Average = 0.018
Std Dev = 0.0003



DIAGNOSTIC RESULT

DataMaster DMT: 100140
Location:
Calibration Date: 04/20/2010
Certification Date: 04/21/2010
Installation Date: 04/21/2010
Test Date: 05/11/2010
Time: 17:21

VERMONT
DEPARTMENT OF HEALTH

```
VERSIONS
DMT 1.04
PIC 2.05
Modem 1.06
Questions 1.00
Reports 1.02

TEMPERATURES
Sample Chamber = 48.74°C
Breath Tube    = 46.93°C

SETTINGS
Lamp Voltage   = 1.52 V
Cooler Voltage = 1.66 V
Bias Voltage   = 80.4 V
Chopper Freq   = 540.6 Hz

PUMP INFO
Flow Rate = 5.527 L/M

DETECTOR INFO
PUMP     ON       OFF
MAX(V)   0.1059   0.1061
MIN(V)   0.1016   0.1045

FILTER INFO
Filter 1  0.106  Zero = true
Filter 2  0.460  Zero = true
Filter 3  1.065  Zero = true

CALIBRATION CHECK
Xq = 0.0765  0.11%
```

ACCURACY AND PRECISION TEST

STATE OF VERMONT

DataMaster DMT: 100140

Date: 05/11/2010
Time: 15:53:58

SUPERVISOR NAME: DARCY M RICHARDSON

SOLUTION LOT #: 10-21-020
SOLUTION CONCENTRATION: 0.019

BLANK TEST	0.000	15:54
CALIBRATION CHECK	PASSED	15:55
SIMULATOR VAPOR	0.018	15:55
SIMULATOR VAPOR	0.018	15:56
SIMULATOR VAPOR	0.019	15:57
SIMULATOR VAPOR	0.018	15:58
SIMULATOR VAPOR	0.018	15:59
SIMULATOR VAPOR	0.018	16:00
SIMULATOR VAPOR	0.018	16:01
SIMULATOR VAPOR	0.018	16:02
SIMULATOR VAPOR	0.019	16:03
SIMULATOR VAPOR	0.018	16:04
BLANK TEST	0.000	16:04

Average = 0.018
Std Dev = 0.0004



ACCURACY AND PRECISION TEST

STATE OF VERMONT

DataMaster DMT: 100140

Date: 05/11/2010
Time: 16:11:05

SUPERVISOR NAME: DARCY M RICHARDSON

SOLUTION LOT #: 10-30-080
SOLUTION CONCENTRATION: 0.081

BLANK TEST	0.000	16:12
CALIBRATION CHECK	PASSED	16:12
SIMULATOR VAPOR	0.079	16:12
SIMULATOR VAPOR	0.079	16:13
SIMULATOR VAPOR	0.080	16:14
SIMULATOR VAPOR	0.080	16:15
SIMULATOR VAPOR	0.080	16:16
SIMULATOR VAPOR	0.080	16:17
SIMULATOR VAPOR	0.080	16:18
SIMULATOR VAPOR	0.080	16:19
SIMULATOR VAPOR	0.081	16:20
SIMULATOR VAPOR	0.080	16:21
BLANK TEST	0.000	16:22

Average = 0.079
Std Dev = 0.0005



ACCURACY AND PRECISION TEST

STATE OF VERMONT

DataMaster DMT: 100140

Date: 05/11/2010
Time: 16:26:51

SUPERVISOR NAME: DARCY M RICHARDSON

SOLUTION LOT #: 10-31-160
SOLUTION CONCENTRATION: 0.160

BLANK TEST	0.000	16:27
CALIBRATION CHECK	PASSED	16:27
SIMULATOR VAPOR	0.158	16:28
SIMULATOR VAPOR	0.158	16:29
SIMULATOR VAPOR	0.158	16:30
SIMULATOR VAPOR	0.159	16:31
SIMULATOR VAPOR	0.159	16:32
SIMULATOR VAPOR	0.159	16:33
SIMULATOR VAPOR	0.159	16:34
SIMULATOR VAPOR	0.159	16:35
SIMULATOR VAPOR	0.159	16:36
SIMULATOR VAPOR	0.159	16:37
BLANK TEST	0.000	16:37

Average = 0.158
Std Dev = 0.0004



ACCURACY AND PRECISION TEST

STATE OF VERMONT

DataMaster DMT: 100140

Date: 05/11/2010
Time: 16:45:16

SUPERVISOR NAME: DARCY M RICHARDSON

SOLUTION LOT #: 10-32-400
SOLUTION CONCENTRATION: 0.373

BLANK TEST	0.000	16:46
CALIBRATION CHECK	PASSED	16:46
SIMULATOR VAPOR	0.370	16:46
SIMULATOR VAPOR	0.370	16:48
SIMULATOR VAPOR	0.372	16:49
SIMULATOR VAPOR	0.372	16:50
SIMULATOR VAPOR	0.372	16:51
SIMULATOR VAPOR	0.372	16:52
SIMULATOR VAPOR	0.371	16:53
SIMULATOR VAPOR	0.372	16:54
SIMULATOR VAPOR	0.371	16:55
SIMULATOR VAPOR	0.372	16:56
BLANK TEST	0.000	16:56

Average = 0.371
Std Dev = 0.0008



Section 5: Interference

ACCURACY AND PRECISION TEST

STATE OF VERMONT

DataMaster DMT: 100140

Date: 04/21/2010
Time: 08:55:58

SUPERVISOR NAME: SEH .02ACETONE .04ETOH

SOLUTION LOT #: 09-31-04A2
SOLUTION CONCENTRATION: 0.040

BLANK TEST	0.000	08:56
CALIBRATION CHECK	PASSED	08:57
SIMULATOR VAPOR	INTERFERENCE	08:57
SIMULATOR VAPOR	INTERFERENCE	08:58
SIMULATOR VAPOR	INTERFERENCE	08:59
SIMULATOR VAPOR	INTERFERENCE	09:00
SIMULATOR VAPOR	INTERFERENCE	09:01
SIMULATOR VAPOR	INTERFERENCE	09:02
SIMULATOR VAPOR	INTERFERENCE	09:03
SIMULATOR VAPOR	INTERFERENCE	09:04
SIMULATOR VAPOR	INTERFERENCE	09:05
SIMULATOR VAPOR	INTERFERENCE	09:06
BLANK TEST	0.000	09:06

Average out of range!



ACCURACY AND PRECISION TEST

STATE OF VERMONT

DataMaster DMT: 100140

Date: 04/21/2010
Time: 09:09:13

SUPERVISOR NAME: SEH 04ETOH 01 ACETONE

SOLUTION LOT #: 09-27-04A1
SOLUTION CONCENTRATION: 0.040

BLANK TEST	0.000	09:10
CALIBRATION CHECK	PASSED	09:10
SIMULATOR VAPOR	INTERFERENCE	09:10
SIMULATOR VAPOR	INTERFERENCE	09:11
SIMULATOR VAPOR	INTERFERENCE	09:12
SIMULATOR VAPOR	0.042	09:13
SIMULATOR VAPOR	INTERFERENCE	09:14
SIMULATOR VAPOR	0.042	09:15
SIMULATOR VAPOR	INTERFERENCE	09:16
SIMULATOR VAPOR	INTERFERENCE	09:17
SIMULATOR VAPOR	INTERFERENCE	09:18
SIMULATOR VAPOR	INTERFERENCE	09:19
BLANK TEST	0.000	09:20



ACCURACY AND PRECISION TEST

STATE OF VERMONT

DataMaster DMT: 100140

Date: 04/21/2010
Time: 09:26:20

SUPERVISOR NAME: SEH TEST 04IPA 08ETOH

SOLUTION LOT #: 09-30-08I
SOLUTION CONCENTRATION: 0.079

BLANK TEST	0.000	09:27
CALIBRATION CHECK	PASSED	09:27
SIMULATOR VAPOR	INTERFERENCE	09:27
SIMULATOR VAPOR	INTERFERENCE	09:28
SIMULATOR VAPOR	INTERFERENCE	09:29
SIMULATOR VAPOR	INTERFERENCE	09:30
SIMULATOR VAPOR	INTERFERENCE	09:31
SIMULATOR VAPOR	INTERFERENCE	09:32
SIMULATOR VAPOR	INTERFERENCE	09:33
SIMULATOR VAPOR	INTERFERENCE	09:34
SIMULATOR VAPOR	INTERFERENCE	09:35
SIMULATOR VAPOR	INTERFERENCE	09:36
BLANK TEST	0.000	09:37

Average out of range!



ACCURACY AND PRECISION TEST

STATE OF VERMONT

DataMaster DMT: 100140

Date: 04/21/2010
Time: 09:40:20

SUPERVISOR NAME: SEH TEST 08ETOH 04MEOH

SOLUTION LOT #: 10-10-084
SOLUTION CONCENTRATION: 0.080

BLANK TEST	0.000	09:41
CALIBRATION CHECK	PASSED	09:41
SIMULATOR VAPOR	INTERFERENCE	09:41
SIMULATOR VAPOR	INTERFERENCE	09:42
SIMULATOR VAPOR	INTERFERENCE	09:43
SIMULATOR VAPOR	INTERFERENCE	09:44
SIMULATOR VAPOR	INTERFERENCE	09:45
SIMULATOR VAPOR	INTERFERENCE	09:46
SIMULATOR VAPOR	INTERFERENCE	09:47
SIMULATOR VAPOR	INTERFERENCE	09:48
SIMULATOR VAPOR	INTERFERENCE	09:49
SIMULATOR VAPOR	INTERFERENCE	09:50
BLANK TEST	0.000	09:51

Average out of range!



ACCURACY AND PRECISION TEST

STATE OF VERMONT

DataMaster DMT: 100140

Date: 04/26/2010
Time: 14:45:29

SUPERVISOR NAME: ALB .01ACETONE IN 08ETOH

SOLUTION LOT #: 10-24-08A
SOLUTION CONCENTRATION: 0.081

BLANK TEST	0.000	14:46
CALIBRATION CHECK	PASSED	14:46
SIMULATOR VAPOR	INTERFERENCE	14:46
SIMULATOR VAPOR	INTERFERENCE	14:47
SIMULATOR VAPOR	INTERFERENCE	14:48
SIMULATOR VAPOR	INTERFERENCE	14:49
SIMULATOR VAPOR	INTERFERENCE	14:50
SIMULATOR VAPOR	0.084	14:51
SIMULATOR VAPOR	INTERFERENCE	14:52
SIMULATOR VAPOR	INTERFERENCE	14:53
SIMULATOR VAPOR	INTERFERENCE	14:54
SIMULATOR VAPOR	INTERFERENCE	14:55
BLANK TEST	0.000	14:56



ACCURACY AND PRECISION TEST

STATE OF VERMONT

DataMaster DMT: 100140

Date: 04/26/2010
Time: 15:57:56

SUPERVISOR NAME: DMR 04 IPA IN H20

SOLUTION LOT #: 10-26-04I
SOLUTION CONCENTRATION: 0.040

BLANK TEST	0.000	15:58
CALIBRATION CHECK	PASSED	15:58
SIMULATOR VAPOR	INTERFERENCE	15:59
SIMULATOR VAPOR	INTERFERENCE	16:00
SIMULATOR VAPOR	INTERFERENCE	16:01
SIMULATOR VAPOR	INTERFERENCE	16:02
SIMULATOR VAPOR	INTERFERENCE	16:03
SIMULATOR VAPOR	INTERFERENCE	16:04
SIMULATOR VAPOR	INTERFERENCE	16:05
SIMULATOR VAPOR	INTERFERENCE	16:06
SIMULATOR VAPOR	INTERFERENCE	16:07
SIMULATOR VAPOR	INTERFERENCE	16:08
BLANK TEST	0.000	16:08

Average out of range!



ACCURACY AND PRECISION TEST

STATE OF VERMONT

DataMaster DMT: 100140

Date: 04/26/2010
Time: 16:12:52

SUPERVISOR NAME: DMR 04 MEOH IN H20

SOLUTION LOT #: 10-25-04M
SOLUTION CONCENTRATION: 0.040

BLANK TEST	0.000	16:13
CALIBRATION CHECK	PASSED	16:13
SIMULATOR VAPOR	INTERFERENCE	16:14
SIMULATOR VAPOR	INTERFERENCE	16:15
SIMULATOR VAPOR	INTERFERENCE	16:16
SIMULATOR VAPOR	INTERFERENCE	16:17
SIMULATOR VAPOR	INTERFERENCE	16:18
SIMULATOR VAPOR	INTERFERENCE	16:19
SIMULATOR VAPOR	INTERFERENCE	16:20
SIMULATOR VAPOR	INTERFERENCE	16:21
SIMULATOR VAPOR	INTERFERENCE	16:22
SIMULATOR VAPOR	INTERFERENCE	16:22
BLANK TEST	0.000	16:23

Average out of range!



ACCURACY AND PRECISION TEST

STATE OF VERMONT

DataMaster DMT: 100140

Date: 04/26/2010
Time: 16:38:18

SUPERVISOR NAME: DMR 04 MEOH IN H20

SOLUTION LOT #: 10-25-04M
SOLUTION CONCENTRATION: 0.040

BLANK TEST	0.000	16:39
CALIBRATION CHECK	PASSED	16:39
SIMULATOR VAPOR	INTERFERENCE	16:39
SIMULATOR VAPOR	INTERFERENCE	16:40
SIMULATOR VAPOR	INTERFERENCE	16:41
SIMULATOR VAPOR	INTERFERENCE	16:42
SIMULATOR VAPOR	INTERFERENCE	16:43
SIMULATOR VAPOR	INTERFERENCE	16:44
SIMULATOR VAPOR	INTERFERENCE	16:45
SIMULATOR VAPOR	INTERFERENCE	16:46
SIMULATOR VAPOR	INTERFERENCE	16:47
SIMULATOR VAPOR	INTERFERENCE	16:48
BLANK TEST	0.000	16:49

Average out of range!



Section 6: Low Level Accuracy

This Instrument not tested for this specification

Section 7: Breath Volume

DMT Serial #100140

Operator	Date
ALB	5/11/2010

Target Volume	Set Flow (lpm)	Time (sec)	1	2	3	4	5	6	7	Average	% s.d.
1.5 L	8.2	11	1.86	1.97	1.94	1.97	1.97	1.84	1.9	1.92143	0.055205245
	15	6	1.89	1.83	1.72	1.8	1.59	1.73	1.86	1.77429	0.102771128
	22.5	4	1.74	1.73	1.8	1.59	1.73	1.86	1.94	1.77	0.111355287
2.5 L	10	15	2.38	2.39	2.47	2.44	2.39	2.43	2.48	2.42571	0.040355563
	15	10	2.31	2.19	2.22	2.25	2.3	2.23	2.24	2.24857	0.042983939
	25	6	2.4	2.52	2.5	2.26	2.36	2.48	2.36	2.41143	0.093706736

Section 8: Software Testing

This Instrument not tested for this specification

Section 9: Error Reports

Section 10: Live Subject Tests

This Instrument not tested for this specification

Section 11: Communication Notes

100140

March/April 2010

~~October 2009~~

~~Third~~ 4th Round Testing Protocol for DMT Instruments

Power Up Procedure:

SH a) Open unit and inspect connections. Ensure all tubing is secure. Check that all cables are properly seated and tied down or strapped together as necessary. Check that all screws are in place and snug; this will require the removal of the side pieces to allow the technician access. Correct as needed.

SH b) Manually turn the filter wheel. The filter wheel should turn smoothly and should not rub or wobble. Correct as needed.

SH c) Power on the instrument. Allow approximately one hour for it to come to temperature.

SH d) Access the SETUP menu. On the MISC tab, type "State of Vermont" in the NAME field. In the EXPLORE box, highlight STYLUS and press the OK button to activate the stylus settings. Recalibrate the stylus. Once complete, exit the SETUP menu by pressing the OK button to save the settings. If you exit the SETUP menu using the CANCEL button, none of the changes will be saved.

SH 4/13/10 e) Enter the TECH screen. Scroll through all of the filter settings to ensure the filter wheel is activating properly.

f) Optimize the detector voltage
-Adjust the cooler voltage setting to 1.70Vdc +/-0.10.
-Manually manipulate the detector position to get the detector voltage as low as possible.
-Adjust the Lamp voltage setting so that the detector voltage displayed is approximately +0.000 +/- 0.100.
SH 4/20/10 -Note the initial noise of the detector voltage once unit has had a minute to equilibrate. Noise tolerance at NPAS is +/-0.003 on a detector.
-Save the voltage settings
-Per VDHL protocol (P-Alc-116 Power-Up SOP), the detector voltage must read 0.000V with a tolerance of +/- 0.100V. Correct as needed.

SH 4/13/10 g) While still in the TECH screen, press SET in the RF Sensitivity box to automatically set the RF sensitivity level. Press SAVE to save the settings. Confirm the setting by keying a radio near the instrument. Correct as needed.

SH h) Using the large syringe, push 1.5L of air into the breath tube. Monitor the volume of air displayed on the TECH screen. The instrument should read 1.5L +/- 0.2L. If the instrument is out of range, adjust the distance between the "T" fittings on the breath exhaust tubing. Retest until the volume reads appropriately.

SH 4/20/10 i) Once all the TECH settings are correct, re-save every setting. Exit the TECH screen.

SH 4/20/10 j) Close up the instrument using locking screws on the top four holes. Attach the stylus to the DMT using a locking screw in the center right screw hole nearest the simulator tower. Attach the keyboard to the top of the instrument using sticky-back velcro. Wrap a small piece of velcro around the top of the stylus and affix the other side to the front of the instrument so the stylus can rest on the front of the unit. Needs To Be Done 4/22/10

SH k) Calibrate the instrument following the Calibration Protocol using Calibration Solution (0.100 EtOH). During Calibration and Certification, it may be necessary to turn off the simulator temperature monitoring if using Guth 2100 Simulators as these do not have BNC connectors for temperature monitoring. To do this, open the SETUP menu. In the Simulator box, select NONE as the type of

SH 4/21/10 simulator. Press OK to exit. Document the room temperature and humidity level at the time of calibration in the communication notebook.

l) Certify the instrument following the Certification Protocol.

m) Preliminary interference testing. Run the following concentrations as an Accuracy and Precision Test with n=10. The instrument must report Interference at least 80% of the time.

needs to Be done 4/22/10

0.01% Acetone in 0.08% EtOH 4/26/10
0.04% MeOH in Water 4/26/10
0.04% Isopropanol in Water 4/26/10

The instrument is now ready for performance testing. If any of the above steps fail or there is **any** problem with the instrument, the DataMaster Team should be consulted as to the action required. It may be necessary to return an instrument to the manufacture for refurbishment prior to use by VDHL. All adjustments and repairs to an instrument should be documented in the instrument's testing notebook located in the testing binder. All reports generated during the Power Up Procedure should be filed in the power up section of that instrument's testing binder. If the instrument has passed, it should now be Installed to VDHL.

DRAFT

Performance Testing

a) Linearity Test: Four concentrations of EtOH solutions will be used, 0.020, 0.080, 0.160, and 0.400. Each solution will be run using the Accuracy and Precision test (n=10). All four concentrations will be run consecutively and will count as one linearity test. Five linearity tests will be run on each instrument. Linearity tests should not be completed back to back. Testing should allow for inter- and intra- day performance.

b) Low Level Performance: A solution of 0.040EtOH will be tested using the Accuracy and Precision test (n=10) once

c) Interference Test: The following concentrations will be tested using the Accuracy and Precision test (n=10). The instrument should report Interference at least 80% of the time.

SH 4/21/10

0.01% Acetone in 0.04% EtOH
0.02% Acetone in 0.04% EtOH
0.04% MeOH in 0.08% EtOH
0.04% Isopropanol in 0.08% EtOH

d) Limited Volume Testing:

Target Volume	Flow (lpm)	Time (sec.)	1	2	3	4	5	6	7	Ave.	% s.d.
1.5 L	8.2	11	1.86	1.97	1.94	1.97	1.97	1.84	1.90		
	15	6	1.89	1.83	1.72	1.65	1.64	1.85	1.74		
	22.5	4	1.74	1.73	1.80	1.59	1.73	1.86	1.94		
2.0 L	10	12	2.38	2.39	2.47	2.44	2.39	2.43	2.48		
	15	8	2.31	2.19	2.22	2.25	2.30	2.23	2.24		
	24	5	2.40	2.52	2.50	2.26	2.36	2.48	2.36		

Pressure: _____ Flow rate +/- _____ time +/- _____

DRAFT