

DMT Software Upgrade  
Work/Fixes Still Needed  
December 30<sup>th</sup>, 2008

1. When performing a stand-alone Accuracy and Precision Check, the “true value” of the Simulator Vapor stored in the Set-up menu should not change.
2. When performing a stand-alone Accuracy and Precision Check, the reported average should be to 3 decimal places.
3. The Cal Factors are no longer accessible in the reports menu ??
4. The Cal Records are no longer accessible in the reports menu ??