

- RFI Interference error message
 - Generated when possibly interfering RFI is detected
- RFI testing sequence added to the following protocols
 - Certification
 - Installation
 - Routine Performance Check
 - Annual Preventative Maintenance check
 - During each of the above protocols, prior to the sample acceptance test the instrument should perform and RFI interference test
 - The instrument will prompt operator to administer RFI to unit
 - Unit will report RFI detected on screen and print RFI detection passed on reports
- Anything detected less than a 0.004g/210L should be reported as 0.000
- Unit should NOT abort when screen is touched during a test sequence