SECTION 6

ERROR REPORTS

ACCURACY & PRECISION REPORT

STATE OF VERMONT

DataMater DMT: 100170

Date: 08/10/2011
Time: 14:12:56

SUPERVISOR NAME:
KIRK L KIMBALL

SOLUTION LOT #: 12-02-080
SOLUTION CONCENTRATION: 0.080

BLANK TEST	0.000	14:13
CALIBRATION CHECK	PASSED	14:13

Simulator Timed Out KLK 8-10-11

ACCURACY & PRECISION REPORT

STATE OF VERMONT

DataMater DMT: 100170

Date: 08/10/2011
Time: 14:35:08

SUPERVISOR NAME:
KIRK L KIMBALL

SOLUTION LOT #: 12-03-160
SOLUTION CONCENTRATION: 0.161

BLANK TEST	0.000	14:36
CALIBRATION CHECK	PASSED	14:36

Simulator Temperature Too Low

Not Connected
mcc
8-10-11

ACCURACY & PRECISION REPORT

STATE OF VERMONT

DataMater DMT: 100170

Date: 08/10/2011
Time: 14:37:31

SUPERVISOR NAME:
KIRK L KIMBALL

SOLUTION LOT #: 12-03-160
SOLUTION CONCENTRATION: 0.161

BLANK TEST	0.000	14:38
CALIBRATION CHECK	PASSED	14:38

Simulator Timed Out

ACCURACY & PRECISION REPORT

STATE OF VERMONT

DataMater DMT: 100170

Date: 08/10/2011
Time: 14:56:10

SUPERVISOR NAME:
KIRK L KIMBALL

SOLUTION LOT #: 11-34-400
SOLUTION CONCENTRATION: 0.400

BLANK TEST	0.000	14:57
CALIBRATION CHECK	PASSED	14:57

Simulator Timed Out

ACCURACY & PRECISION REPORT

STATE OF VERMONT

DataMater DMT: 100170

Date: 08/10/2011
Time: 16:07:48

SUPERVISOR NAME:
KIRK L KIMBALL

SOLUTION LOT #: 12-04-080A1
SOLUTION CONCENTRATION: 0.080

BLANK TEST	0.000	16:08
CALIBRATION CHECK	PASSED	16:08

Simulator Timed Out ✓ KLK 8-10-11

ACCURACY & PRECISION REPORT

STATE OF VERMONT

DataMater DMT: 100170

Date: 08/11/2011
Time: 09:22:09

SUPERVISOR NAME:
KIRK L KIMBALL

SOLUTION LOT #: 11-33-080
SOLUTION CONCENTRATION: 0.080

BLANK TEST	0.000	09:23
CALIBRATION CHECK	PASSED	09:23

Simulator Timed Out

ACCURACY & PRECISION REPORT

STATE OF VERMONT

DataMater DMT: 100170

Date: 08/11/2011
Time: 09:40:13

SUPERVISOR NAME:
KIRK L KIMBALL

SOLUTION LOT #: 12-03-160
SOLUTION CONCENTRATION: 0.161

BLANK TEST	0.000	09:41
CALIBRATION CHECK	PASSED	09:41

Simulator Timed Out KLKC 8-11-11

ACCURACY & PRECISION REPORT

STATE OF VERMONT

DataMater DMT: 100170

Date: 08/11/2011
Time: 10:13:39

SUPERVISOR NAME:
KIRK L KIMBALL

SOLUTION LOT #: 11-34-400
SOLUTION CONCENTRATION: 0.400

BLANK TEST	0.000	10:14
CALIBRATION CHECK	PASSED	10:14

Simulator Timed Out

ACCURACY & PRECISION REPORT

STATE OF VERMONT

DataMater DMT: 100170

Date: 08/11/2011
Time: 10:16:04

SUPERVISOR NAME:
KIRK L KIMBALL

SOLUTION LOT #: 11-34-400
SOLUTION CONCENTRATION: 0.400

BLANK TEST	0.000	10:16
CALIBRATION CHECK	PASSED	10:17

Simulator Timed Out

ACCURACY & PRECISION REPORT

STATE OF VERMONT

DataMater DMT: 100170

Date: 08/11/2011
Time: 10:18:27

SUPERVISOR NAME:
KIRK L KIMBALL

SOLUTION LOT #: 11-34-400
SOLUTION CONCENTRATION: 0.400

BLANK TEST	0.000	10:19
CALIBRATION CHECK	PASSED	10:19

Simulator Timed Out

ACCURACY & PRECISION REPORT

STATE OF VERMONT

DataMater DMT: 100170

Date: 08/11/2011
Time: 10:20:56

SUPERVISOR NAME:
KIRK L KIMBALL

SOLUTION LOT #: 11-34-400
SOLUTION CONCENTRATION: 0.400

BLANK TEST	0.000	10:21
CALIBRATION CHECK	PASSED	10:21

Simulator Timed Out

ACCURACY & PRECISION REPORT

STATE OF VERMONT

DataMater DMT: 100170

Date: 08/11/2011
Time: 11:02:03

SUPERVISOR NAME:
KIRK L KIMBALL

SOLUTION LOT #: 12-05-040M
SOLUTION CONCENTRATION: 0.040

BLANK TEST	0.000	11:02
CALIBRATION CHECK	PASSED	11:03

Simulator Timed Out

ACCURACY & PRECISION REPORT

STATE OF VERMONT

DataMater DMT: 100170

Date: 08/11/2011
Time: 11:04:41

SUPERVISOR NAME:
KIRK L KIMBALL

SOLUTION LOT #: 12-05-040M
SOLUTION CONCENTRATION: 0.040

BLANK TEST	0.000	11:05
CALIBRATION CHECK	PASSED	11:05

Simulator Timed Out

ACCURACY & PRECISION REPORT

STATE OF VERMONT

DataMater DMT: 100170

Date: 08/11/2011
Time: 13:03:22

SUPERVISOR NAME:
KIRK L KIMBALL

SOLUTION LOT #: 11-33-080
SOLUTION CONCENTRATION: 0.080

BLANK TEST	0.000	13:04
CALIBRATION CHECK	PASSED	13:04

Simulator Timed Out

L3 KLK

ACCURACY & PRECISION REPORT

STATE OF VERMONT

DataMater DMT: 100170

Date: 08/11/2011
Time: 13:51:14

SUPERVISOR NAME:
KIRK L KIMBALL

SOLUTION LOT #: 12-03-160
SOLUTION CONCENTRATION: 0.161

BLANK TEST	0.000	13:52
CALIBRATION CHECK	PASSED	13:52

Simulator Timed Out

ACCURACY & PRECISION REPORT

STATE OF VERMONT

DataMater DMT: 100170

Date: 08/11/2011
Time: 13:54:15

SUPERVISOR NAME:
KIRK L KIMBALL

SOLUTION LOT #: 12-03-160
SOLUTION CONCENTRATION: 0.161

BLANK TEST	0.000	13:55
CALIBRATION CHECK	PASSED	13:55

Simulator Timed Out

ACCURACY & PRECISION REPORT

STATE OF VERMONT

DataMater DMT: 100170

Date: 08/11/2011
Time: 14:20:52

SUPERVISOR NAME:
KIRK L KIMBALL

SOLUTION LOT #: 11-34-400
SOLUTION CONCENTRATION: 0.400

BLANK TEST	0.000	14:21
CALIBRATION CHECK	PASSED	14:21

Simulator Timed Out

ACCURACY & PRECISION REPORT

STATE OF VERMONT

DataMater DMT: 100170

Date: 08/11/2011
Time: 14:16:42

SUPERVISOR NAME:
KIRK L KIMBALL

SOLUTION LOT #: 11-34-400
SOLUTION CONCENTRATION: 0.400

BLANK TEST	0.000	14:17
CALIBRATION CHECK	PASSED	14:17

Simulator Timed Out

stopped due to soln. being used 10x
8-11-11

ACCURACY & PRECISION REPORT

STATE OF VERMONT

DataMater DMT: 100170

Date: 08/12/2011
Time: 08:30:38

SUPERVISOR NAME:
KIRK L KIMBALL

SOLUTION LOT #: 12-01-040A2
SOLUTION CONCENTRATION: 0.040

BLANK TEST	0.000	08:31
CALIBRATION CHECK	PASSED	08:31

Simulator Timed Out

ACCURACY & PRECISION REPORT

STATE OF VERMONT

DataMater DMT: 100170

Date: 08/12/2011
Time: 11:49:51

SUPERVISOR NAME:
KIRK L KIMBALL

SOLUTION LOT #: 11-38-020
SOLUTION CONCENTRATION: 0.020

BLANK TEST	0.000	11:50
CALIBRATION CHECK	PASSED	11:50
SIMULATOR VAPOR 33.6°C	0.022	11:51
SIMULATOR VAPOR 33.6°C	0.021	11:52
SIMULATOR VAPOR 33.7°C	0.021	11:53
SIMULATOR VAPOR 33.6°C	0.020	11:54
SIMULATOR VAPOR 33.6°C	0.020	11:55
SIMULATOR VAPOR 33.6°C	0.020	11:56
SIMULATOR VAPOR 33.7°C	0.020	11:57
SIMULATOR VAPOR 33.6°C	0.020	11:58

Simulator Temperature Too Low

simulator temp w/NIST Thermometer was 34.0°C
KLK 8-12-11

Signal reading on DMT not set correctly. KLK 8-12-11



ACCURACY & PRECISION REPORT

STATE OF VERMONT

DataMater DMT: 100170

Date: 08/12/2011
Time: 14:05:41

SUPERVISOR NAME:
KIRK L KIMBALL

SOLUTION LOT #: 11-33-080
SOLUTION CONCENTRATION: 0.080

BLANK TEST	0.000	14:06
CALIBRATION CHECK	PASSED	14:06

Simulator Timed Out

ACCURACY & PRECISION REPORT

STATE OF VERMONT

DataMater DMT: 100170

Date: 08/12/2011
Time: 14:34:23

SUPERVISOR NAME:
KIRK L KIMBALL

SOLUTION LOT #: 12-03-160
SOLUTION CONCENTRATION: 0.161

BLANK TEST	0.000	14:35
CALIBRATION CHECK	PASSED	14:35

Simulator Timed Out

ACCURACY & PRECISION REPORT

STATE OF VERMONT

DataMater DMT: 100170

Date: 08/12/2011
Time: 14:38:35

SUPERVISOR NAME:
KIRK L KIMBALL

SOLUTION LOT #: 12-03-160
SOLUTION CONCENTRATION: 0.161

BLANK TEST	0.000	14:39
CALIBRATION CHECK	PASSED	14:39

Simulator Timed Out

ACCURACY & PRECISION REPORT

STATE OF VERMONT

DataMater DMT: 100170

Date: 08/15/2011
Time: 09:29:50

SUPERVISOR NAME:
KIRK L KIMBALL

SOLUTION LOT #: 11-33-080I
SOLUTION CONCENTRATION: 0.080

BLANK TEST	0.000	09:30
CALIBRATION CHECK	PASSED	09:30

Simulator Timed Out