DataMaster DMT Monthly Update Check-list

Department/Agency: Middlebury PD DM serial #: 123206

Date: 10 / 10 / 12

	Provided	None this month
Operators Logbook	☐	☒
Maintenance Logbook	☒	☐
Breath Summary Report	☒	☐
Status Summary Report	☒	☐
Calibration Report	☐	☒
Certification Report	☐	☒
Installation Report	☐	☒
Annual Preventative Maintenance	☐	☒
Routine Performance Check	☒	☐
Simulator Solution Change	☐	☒
Other Reports	☐	☒

List other report type and date:

REPORT FOR JUNE 2012

ROUTINE PERFORMANCE CHECK REPORT

DataMaster DMT: 123206
Location: Middlebury PD
Calibration Date: 01/03/2012
Certification Date: 01/04/2012
Installation Date: 01/12/2012
RPC Date: 06/27/2012
Supervisor Name: KEVIN A EMILIO

Diagnostic Results

```
VERSIONS
DMT: 1.01
PIC: 2.05
Modem: 2.2
Questions: 2.0

TEMPERATURES

Sample Chamber = 48.9°C
Breath Tube    = 43.7°C
Digital Sim    = 34.0°C

SETTINGS
Lamp Voltage   = 1.73 V
Cooler Voltage = 1.78 V
Bias Voltage   = 80 V
Chopper Freq   = 541 Hz

PUMP INFO
Flow Rate = 4.334 L/M

DETECTOR INFO
PUMP     ON        OFF
MAX(V)   0.1508    0.1541
MIN(V)   0.1488    0.1522

FILTER INFO
Filter 1   0.153  Zero = true
Filter 2   0.515  Zero = true
Filter 3   1.503  Zero = true

CALIBRATION CHECK
Xq  = 0.081  0.21%
```

Routine Performance Check Passed

Accuracy and Precision Check

Concentration	= 0.102 g/210L
Lot #	= 12-13-100
Range	= 0.097 - 0.107
Average	= 0.100 g/210L
Std Dev	= 0.0000

RF Detection Test

Passed

Sample Acceptance Test

Passed



Performed by KEE Date 06/27/2012

Reviewed by _____ Date _____

Log On Record History

Test Date	Test Time	Name
06/27/2012	14:22:55	Supervisor

Status Record History

Status Date	Status Time	Test Type	Status Message
06/27/2012	14:53:38	RPC Record	User Abort

Breath Record History

Test Date	Test Time	Agency	Operator Name	Alcohol Val. 1	Alcohol Val. 2	Sim. Conc. 1	Sim. Conc. 2	Status Message
06/09/2012	02:50:34	MIDDLEBURY PD	VEGAR BOE	0.142		0.097		

DATAMASTER DMT MAINTENANCE LOG

Middlebury PD
Department/Agency

123206
DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
1-12-12	**DataMaster Simulator Solution** Lot # 12-45-100 $EtOH_{aq}$ Conc: 0.100 Range: 0.095 - 0.105 Preparation Date: 12/12/2011 Expires: 12/12/2012 [For RF ... ace Label Here] OPENED: 1-12-12 WM Install Per Protocal SJH 1/12/12	Kirk Kimball, VDAC Prog Chief Steven Harnois
2/27/12	**DataMaster Simulator Solution** Lot # 12-45-100 $EtOH_{aq}$ Conc: 0.100 Range: 0.095 - 0.105 Preparation Date: 12/12/2011 Expires: 12/12/2012 RPC Conducted [... hange Place Label Here] OPENED: 2/27/12	KEVIN Emilio
6/27/12	**DataMaster Simulator Solution** Lot # 12-13-100 $EtOH_{aq}$ Conc: 0.102 Range: 0.097 - 0.107 Preparation Date: 09/09/2011 Expires: 09/09/2012 RPC [F ... ge Place Label Here] OPENED: 6-27-12	KEVIN EMILIO
	[For RPC or Simulator Solution Change Place Label Here]	

ALC 803 REV 03 08/2011

DataMaster DMT Monthly Update Check-list

Department/Agency: Middlebury DM serial #: 123206

Date: 10 / 10 / 12

	Provided	None this month
Operators Logbook	☐	☒
Maintenance Logbook	☐	☒
Breath Summary Report	☒	☐
Status Summary Report	☒	☐
Calibration Report	☐	☒
Certification Report	☐	☒
Installation Report	☐	☒
Annual Preventative Maintenance	☐	☒
Routine Performance Check	☐	☒
Simulator Solution Change	☐	☒
Other Reports	☐	☒

List other report type and date:

REPORT FOR JULY 2012

Breath Record History

Test Date	Test Time	Agency	Operator Name	Alcohol Val. 1	Alcohol Val. 2	Sim. Conc. 1	Sim. Conc. 2	Status Message
07/01/2012	13:31:20	middlebury	JAMIELEE M KILEY	REFUSED		0.101		
07/01/2012	17:28:24	middlebury	JAMIELEE M KILEY			0.101		RESPONSE TIMEOUT
07/01/2012	17:33:37	middlebury	JAMIELEE M KILEY	0.165	0.174	0.101	0.100	
07/06/2012	02:23:09	MIDDLEBURY PD	NATHAN N HAYES	0.179	0.181	0.097	0.098	
07/06/2012	23:09:34	MIDDLEBURY PD	VEGAR BOE	0.143		0.100		
07/09/2012	21:38:46	MIDDLEBURY PD	NEIL F MOGERLEY	0.080		0.100		
07/27/2012	20:50:37	middlebury	NEIL F MOGERLEY	0.084		0.100		

Status Record History

Status Date	Status Time	Test Type	Status Message
07/01/2012	17:31:38	DUI Subject Record	RESPONSE TIMEOUT

Log On Record History

Test Date	Test Time	Name

DataMaster DMT Monthly Update Check-list

Department/Agency: Middlebury PD DM serial #: 123206

Date: 10 / 10 / 12

	Provided	None this month
Operators Logbook	☐	☒
Maintenance Logbook	☐	☒
Breath Summary Report	☒	☐
Status Summary Report	☒	☐
Calibration Report	☐	☒
Certification Report	☐	☒
Installation Report	☐	☒
Annual Preventative Maintenance	☐	☒
Routine Performance Check	☐	☒
Simulator Solution Change	☐	☒
Other Reports	☐	☒

List other report type and date:

REPORT FOR AUGUST 2012

Status Record History

Status Date	Status Time	Test Type	Status Message

Breath Record History

Test Date	Test Time	Agency	Operator Name	Alcohol Val. 1	Alcohol Val. 2	Sim. Conc. 1	Sim. Conc. 2	Status Message
08/18/2012	02:13:24	MIDDLEBURY PD	WILLIAM J AUSTIN	0.094		0.099	0.099	
08/19/2012	23:06:25	MIDDLEBURY PD	JAMIELEE M KILEY	0.142		0.098		

Log On Record History

Test Date	Test Time	Name
08/29/2012	10:05:12	Amanda Bolduc

DataMaster DMT Monthly Update Check-list

Department/Agency: MIDDLEBURY PD DM serial #: 123206

Date: 10 / 10 / 12

	Provided	None this month
Operators Logbook	☐	☒
Maintenance Logbook	☐	☒
Breath Summary Report	☒	☐
Status Summary Report	☒	☐
Calibration Report	☐	☒
Certification Report	☐	☒
Installation Report	☐	☒
Annual Preventative Maintenance	☐	☒
Routine Performance Check	☐	☒
Simulator Solution Change	☐	☒
Other Reports	☐	☒

List other report type and date:

REPORT FOR SEPT. 2012

Breath Record History

Test Date	Test Time	Agency	Operator Name	Alcohol Val. 1	Alcohol Val. 2	Sim. Conc. 1	Sim. Conc. 2	Status Message
09/01/2012	23:30:11	MIDDLEBURY PD	WILLIAM J AUSTIN	0.161		0.097		
09/02/2012	01:41:11	MIDDLEBURY PD	VEGAR BOE	0.197	0.200	0.098	0.097	
09/13/2012	01:24:15	MIDDLEBURY PD	WILLIAM J AUSTIN	0.092		0.097		
09/29/2012	02:31:42	MIDDLEBURY PD	VEGAR BOE	0.138	0.139	0.097	0.097	

Status Record History

Status Date	Status Time	Test Type	Status Message

Log On Record History

Test Date	Test Time	Name