Breath Record History

Test Date	Test Time	Agency	Operator Name	Alcohol Val. 1	Alcohol Val. 2	Sim. Conc. 1	Sim. Conc. 2	Status Message
09/02/2012	21:33:15	NEW HAVEN VSP	PETER M DEMPSEY	REFUSED		0.097		
09/02/2012	21:52:09	NEW HAVEN VSP	PETER M DEMPSEY	0.106	0.108	0.098	0.098	
09/04/2012	22:41:59	NEW HAVEN VSP	EDEN R NEARY	REFUSED		0.096		
09/04/2012	23:01:30	NEW HAVEN VSP	EDEN R NEARY	0.079		0.098		RESPONSE TIMEOUT
09/07/2012	23:34:57	NEW HAVEN VSP	STEPHEN E MCNAMARA	0.146		0.098		
09/13/2012	00:33:20	NEW HAVEN VSP	EVAN S DOXSEE			0.097		RESPONSE TIMEOUT
09/13/2012	00:52:10	NEW HAVEN VSP	EVAN S DOXSEE	0.060	0.061	0.097	0.098	
09/13/2012	03:57:57	VERGENNES PD	ADAM J ONEILL	0.156	0.150	0.098	0.097	
09/15/2012	00:07:49	NEW HAVEN VSP	JOSEPH T SZAREJKO	0.152		0.098		
09/15/2012	04:49:53	NEW HAVEN VSP	JOSEPH T SZAREJKO	0.115		0.097		
09/16/2012	01:47:17	NEW HAVEN VSP	JOSEPH T SZAREJKO	0.098	0.095	0.097	0.097	
09/17/2012	10:56:20	NEW HAVEN VSP	PETER M DEMPSEY	0.087		0.097		
09/22/2012	02:36:22	NEW HAVEN VSP	JOSEPH T SZAREJKO	REFUSED		0.095		
09/22/2012	02:57:21	NEW HAVEN VSP	JOSEPH T SZAREJKO	0.172	INCOMPLETE	0.095	0.095	
09/28/2012	02:39:48	NEW HAVEN VSP	EVAN S DOXSEE	REFUSED		0.096		
10/03/2012	22:44:48	NEW HAVEN VSP	EVAN S DOXSEE	0.134		0.096		RESPONSE TIMEOUT
10/05/2012	21:10:13	NEW HAVEN VSP	TIMOTHY P HANLEY	0.131	0.138	0.095	0.095	
10/12/2012	22:25:32	NEW HAVEN VSP	TIMOTHY P HANLEY	REFUSED		0.095		
10/13/2012	23:47:43	NEW HAVEN VSP	KAITLYN ARMSTRONG	0.089	0.089	0.095	0.095	
10/14/2012	02:04:11	MIDDLEBURY PD	NEIL F MOGERLEY			0.094		Simulator Out of Range
10/14/2012	02:07:02	MIDDLEBURY PD	NEIL F MOGERLEY	0.155	INCOMPLETE	0.095	0.095	
10/14/2012	23:19:31	NEW HAVEN VSP	JOSEPH T SZAREJKO			0.094		Simulator Out of Range
10/14/2012	23:41:00	NEW HAVEN VSP	JOSEPH T SZAREJKO	REFUSED		0.096		

Status Record History

Status Date	Status Time	Test Type	Status Message
09/04/2012	23:05:36	DUI Subject Record	RESPONSE TIMEOUT
09/13/2012	00:35:46	DUI Subject Record	RESPONSE TIMEOUT
10/03/2012	22:49:43	DUI Subject Record	RESPONSE TIMEOUT
10/14/2012	02:05:41	DUI Subject Record	Simulator Out of Range
10/14/2012	23:21:02	DUI Subject Record	Simulator Out of Range