DataMaster DMT Monthly Update Check-list

Department/Agency: Middlebury PD DM serial #: 950160

Date: 01 / 12 / 12

	Provided	None this month
Operators Logbook	☐	☒
Maintenance Logbook	☐	☒
Breath Summary Report	☒	☐
Status Summary Report	☒	☐
Calibration Report	☐	☒
Certification Report	☐	☒
Installation Report	☐	☒
Annual Preventative Maintenance	☐	☒
Routine Performance Check	☐	☒
Simulator Solution Change	☐	☒
Other Reports	☐	☒

List other report type and date:

REPORT FOR 12/11

Status Record Summary

X = Interference, R = Refusal, V = Invalid, I = Incomplete

Test Date	Test Time	Test Type	Status Message
05/01/2010	21:34	Breath Test	RESPONSE TIMEOUT
06/05/2010	03:23	Breath Test	RESPONSE TIMEOUT
07/16/2010	05:02	Breath Test	STANDARD OUT OF RANGE
07/28/2010	04:06	Breath Test	RESPONSE TIMEOUT
09/11/2010	03:10	Breath Test	RESPONSE TIMEOUT
09/19/2010	01:25	Breath Test	STANDARD OUT OF RANGE
09/19/2010	01:27	Breath Test	STANDARD OUT OF RANGE
09/22/2010	01:50	Breath Test	STANDARD OUT OF RANGE
09/22/2010	01:52	Breath Test	STANDARD OUT OF RANGE
10/08/2010	03:20	Breath Test	RESPONSE TIMEOUT
11/06/2010	03:57	Breath Test	RESPONSE TIMEOUT
11/13/2010	04:03	Breath Test	RESPONSE TIMEOUT
11/14/2010	15:32	Breath Test	RESPONSE TIMEOUT
11/14/2010	15:44	Breath Test	RESPONSE TIMEOUT
11/25/2010	01:51	Breath Test	STANDARD OUT OF RANGE
01/26/2011	02:50	Breath Test	RESPONSE TIMEOUT
02/05/2011	02:08	Breath Test	RESPONSE TIMEOUT
02/26/2011	01:51	Breath Test	RESPONSE TIMEOUT
07/04/2011	20:11	Breath Test	RESPONSE TIMEOUT
11/04/2011	23:56	Breath Test	RESPONSE TIMEOUT

Breath Record Summary

X = Interference, R = Refusal, V = Invalid, I = Incomplete

Test Date	Test Time	Agency	Operator Name	Alcohol Val. 1	Alcohol Val. 2	Sim. Conc. 1	Sim. Conc. 2	Status Message
08/09/2011	02:53	MIDDLEBURY PD	PAUL F SOKOLOWSKI	0.197	0.182	0.098	0.096	
08/12/2011	02:59	MIDDLEBURY PD	CHRISTOPHER L MASON	0.110		0.098		
08/22/2011	12:16	VDHL	STEVEN E BARNOIS	0.000		0.097		
09/21/2011	03:36	MIDDLEBURY PD	CHRISTOPHER L MASON	0.121	0.117	0.096	0.093	
09/30/2011	02:25	MIDDLEBURY PD	WILLIAM J AUSTIN	0.160		0.095	0.094	
10/01/2011	02:41	MIDDLEBURY PD	CHRISTOPHER L MASON	R		0.095		
10/01/2011	03:21	MIDDLEBURY PD	CHRISTOPHER L MASON	0.110	0.110	0.095	0.095	
10/15/2011	02:29	MIDDLEBURY PD	CHRISTOPHER L MASON	0.133	0.132	0.094	0.094	
10/16/2011	18:45	MIDDLEBURY PD	NEIL F MOGERLEY	R		0.093		
10/16/2011	19:04	MIDDLEBURY PD	NEIL M MOGERLEY	0.107	0.104	0.094	0.093	
10/29/2011	02:20	MIDDLEBURY PD	CHRISTOPHER L MASON	0.081	0.078	0.101	0.100	
10/29/2011	16:37	MIDDLEBURY PD	KEVIN A EMILIO	0.194		0.100		
10/30/2011	02:10	MIDDLEBURY PD	PAUL F SOKOLOWSKI	0.132	I	0.101	0.101	
10/30/2011	07:44	MIDDLEBURY PD	CHRISTOPHER L MASON	0.119	0.118	0.100	0.097	
10/31/2011	24:44	MIDDLEBURY PD	KEVIN A EMILIO	0.000		0.101		
11/04/2011	23:36	MIDDLEBURY PD	KRIS S BOWDISH	R		0.100		
11/04/2011	23:56	MIDDLEBURY PD	KRIS S BOWDISH	I	0.125	0.099		RESPONSE TIMEOUT
11/05/2011	03:42	MIDDLEBURY PD	CHRISTOPHER L MASON	0.105		0.101		
11/19/2011	02:05	MIDDLEBURY PD	CHRISTOPHER L MASON	0.083	0.087	0.100	0.100	
11/26/2011	03:24	MIDDLEBURY PD	CHRISTOPHER L MASON	0.113	0.112	0.100	0.100	
12/02/2011	02:53	MIDDLEBURY PD	WILLIAM J AUSTIN	0.202		0.100		
12/10/2011	02:37	MIDDLEBURY PD	CHRISTOPHER L MASON	0.103	0.109	0.099	0.099	
12/18/2011	00:54	MIDDLEBURY PD	CHRISTOPHER L MASON	0.072		0.100		
12/20/2011	20:52	MIDDLEBURY PD	NEIL F MOGERLEY	0.134	0.132	0.099	0.098	
12/24/2011	02:34	ADDISON COUNTY SO	OSCAR J GARDNER	0.128	0.128	0.099	0.098	

DMT -- 100137-000101

Breath Record Summary

X = Interference, R = Refusal, V = Invalid, I = Incomplete

Test Date	Test Time	Agency	Operator Name	Alcohol Val. 1	Alcohol Val. 2	Sim. Conc. 1	Sim. Conc. 2	Status Message
12/16/2010	02:34	MIDDLEBURY PD	CHRISTOPHER L MASON	0.167	0.172	0.099	0.099	
12/23/2010	00:17	MIDDLEBURY PD	CHRISTOPHER L MASON	0.102	0.102	0.100	0.100	
01/01/2011	03:22	MIDDLEBURY PD	NEIL F MOGERLEY	0.123		0.099		
01/01/2011	18:20	MIDDLEBURY PD	KRIS S BOWDISH	0.089	0.086	0.098	0.099	
01/14/2011	02:39	MIDDLEBURY PD	CHRISTOPHER L MASON	0.086	0.084	0.098	0.097	
01/15/2011	02:35	MIDDLEBURY PD	CHRISTOPHER L MASON	0.163		0.098		
01/20/2011	00:20	MIDDLEBURY PD	CHRISTOPHER L MASON	0.133		0.098		
01/21/2011	02:14	MIDDLEBURY PD	KRIS S BOWDISH	0.148		0.098		
01/22/2011	02:03	MIDDLEBURY PD	CHRISTOPHER L MASON	0.174		0.099		
01/26/2011	02:50	MIDDLEBURY PD	CHRISTOPHER L MASON	0.101		0.098		RESPONSE TIMEOUT
01/26/2011	02:56	MIDDLEBURY PD	CHRISTOPHER L MASON	0.099		0.098		
02/05/2011	01:43	MANCHESTER PD	CHRISTOPHER L MASON	R		0.099		
02/05/2011	02:08	MIDDLEBURY PD	CHRISTOPHER L MASON	0.075		0.097		RESPONSE TIMEOUT
02/13/2011	01:34	MIDDLEBURY PD	CHRISTOPHER L MASON	0.089		0.099		
02/17/2011	22:16	MIDDLEBURY PD	NATHAN N HAYES	0.067		0.099		
02/18/2011	02:50	MIDDLEBURY PD	CHRISTOPHER L MASON	0.176		0.100		
02/19/2011	02:43	MIDDLEBURY PD	CHRISTOPHER L MASON	0.120		0.098		
02/26/2011	01:51	MIDDLEBURY PD	CHRISTOPHER L MASON	I	I	0.100		RESPONSE TIMEOUT
03/06/2011	02:52	MIDDLEBURY PD	CHRISTOPHER L MASON	0.123		0.100		
03/12/2011	02:19	MIDDLEBURY PD	CHRISTOPHER L MASON	0.169		0.098		
03/16/2011	02:50	MIDDLEBURY PD	CHRISTOPHER L MASON	0.110		0.098		
03/18/2011	00:49	MIDDLEBURY PD	CHRISTOPHER L MASON	0.063		0.099		
03/18/2011	03:58	MIDDLEBURY PD	KRIS S BOWDISH	0.183	0.176	0.099	0.097	
03/20/2011	03:06	MIDDLEBURY PD	KRIS S BOWDISH	0.085	0.089	0.098	0.099	
04/01/2011	09:36	NEW HAVEN VSP	PETER M DEMPSEY	0.088		0.097		
04/01/2011	17:50	MIDDLEBURY PD	PAUL F SOKOLOWSKI	0.188	0.188	0.096	0.095	
04/05/2011	11:48	NEW HAVEN VSP	ANDREW W LEISE	0.184	0.174	0.096	0.095	
04/08/2011	03:11	MIDDLEBURY PD	CHRISTOPHER L MASON	0.194		0.097		
04/10/2011	04:10	MIDDLEBURY PD	CHRISTOPHER L MASON	0.135	0.133	0.096	0.095	
04/30/2011	03:36	MIDDLEBURY PD	CHRISTOPHER L MASON	0.100	0.096	0.096	0.096	
05/05/2011	01:54	MIDDLEBURY PD	PAUL F SOKOLOWSKI	0.128		0.096		
05/05/2011	02:15	MIDDLEBURY PD	CHRISTOPHER L MASON	0.102	0.098	0.094	0.093	
05/14/2011	02:19	MIDDLEBURY PD	CHRISTOPHER L MASON	R		0.096		
05/14/2011	02:37	MIDDLEBURY PD	CHRISTOPHER L MASON	0.195		0.097		
05/16/2011	20:34	MIDDLEBURY PD	WILLIAM J AUSTIN	0.227		0.096		
05/21/2011	20:57	MIDDLEBURY PD	WILLIAM J AUSTIN	0.088		0.095		
05/22/2011	03:49	MIDDLEBURY PD	PAUL F SOKOLOWSKI	0.093		0.097		
05/27/2011	14:25	MIDDLEBURY PD	PETER D NEWTON	0.202		0.096	0.095	
06/06/2011	18:40	MIDDLEBURY PD	WILLIAM J AUSTIN	0.000	0.000	0.096	0.097	
06/06/2011	20:53	MIDDLEBURY PD	KEVIN A EMILIO	0.157		0.095		
06/08/2011	02:26	MIDDLEBURY PD	CHRISTOPHER L MASON	0.145	0.153	0.095	0.094	
06/11/2011	03:22	MIDDLEBURY PD	CHRISTOPHER L MASON	I	I	0.099		
06/18/2011	03:56	MIDDLEBURY PD	CHRISTOPHER L MASON	0.118		0.100		
06/25/2011	03:34	MIDDLEBURY PD	CHRISTOPHER L MASON	0.105		0.099		
07/03/2011	03:07	MIDDLEBURY PD	CHRISTOPHER L MASON	0.116		0.100		
07/03/2011	22:23	NEW HAVEN VSP	TIMOTHY P HANLEY	0.105		0.099		
07/04/2011	20:11	NEW HAVEN VSP	STEPHEN E MCNAMARA	0.099		0.099		RESPONSE TIMEOUT
07/09/2011	22:23	MIDDLEBURY PD	KRIS S BOWDISH	0.176	0.180	0.099	0.099	
07/16/2011	01:22	MIDDLEBURY PD	CHRISTOPHER L MASON	0.194		0.098		
07/17/2011	03:29	MIDDLEBURY PD	CHRISTOPHER L MASON	0.084	0.085	0.099	0.099	
07/22/2011	15:51	MIDDLEBURY PD	NEIL F MOGERLEY	0.208	0.208	0.097	0.098	
07/22/2011	20:14	ADDISON COUNTY SO	OSCAR J GARDNER	0.181	0.178	0.097	0.098	
07/30/2011	03:20	MIDDLEBURY PD	CHRISTOPHER L MASON	0.080	0.069	0.098	0.098	
07/31/2011	02:19	MIDDLEBURY PD	PAUL F SOKOLOWSKI	0.216	0.212	0.098	0.098	

DMT -- 100137-000102

Breath Record Summary

X = Interference, R = Refusal, V = Invalid, I = Incomplete

Test Date	Test Time	Agency	Operator Name	Alcohol Val. 1	Alcohol Val. 2	Sim. Conc. 1	Sim. Conc. 2	Status Message
04/29/2010	04:04	MIDDLEBURY PD	CHRISTOPHER L MASON	0.000		0.102		
05/01/2010	03:37	NEW HAVEN VSP	TIMOTHY P HANLEY	0.107	0.087	0.100	0.099	
05/01/2010	21:34	MIDDLEBURY PD	KRIS S BOWDISH	0.201		0.099		RESPONSE TIMEOUT
05/02/2010	22:09	MIDDLEBURY PD	PETER D NEWTON	0.181		0.101		
05/09/2010	16:21	MIDDLEBURY PD	PETER D NEWTON	0.085		0.100	0.100	
05/25/2010	18:22	MIDDLEBURY PD	MICHELE L MAGEE	0.068	0.067	0.100	0.100	
05/30/2010	03:23	MIDDLEBURY PD	CHRISTOPHER L MASON	0.167		0.100		
06/02/2010	03:39	MIDDLEBURY PD	CHRISTOPHER L MASON	0.161		0.099		
06/05/2010	03:23	MIDDLEBURY PD	CHRISTOPHER L MASON	0.163		0.100		RESPONSE TIMEOUT
06/05/2010	03:30	MIDDLEBURY PD	CHRISTOPHER L MASON	0.166		0.100		
07/06/2010	01:53	MIDDLEBURY PD	KEVIN A EMILIO	0.134		0.100		
07/16/2010	03:12	MIDDLEBURY PD	PETER D NEWTON	0.107		0.100		
07/16/2010	03:59	MIDDLEBURY PD	CHRISTOPHER L MASON	0.177		0.099		
07/16/2010	05:02	NEW HAVEN VSP	PHILIP WAGNER			0.097		STANDARD OUT OF RANGE
07/18/2010	01:03	MIDDLEBURY PD	PETER D NEWTON	0.131		0.100		
07/23/2010	00:55	MIDDLEBURY PD	PAUL F SOKOLOWSKI	0.114	0.120	0.100	0.099	
07/28/2010	03:47	MIDDLEBURY PD	CHRISTOPHER L MASON	V		0.099		
07/28/2010	04:06	MIDDLEBURY PD	CHRISTOPHER L MASON	0.212		0.100		RESPONSE TIMEOUT
08/18/2010	19:58	NEW HAVEN VSP	PHILIP WAGNER	0.224		0.098		
08/21/2010	01:47	NEW HAVEN VSP	CJ CAMPBELL	0.159		0.098		
08/21/2010	02:25	MIDDLEBURY PD	CHRISTOPHER L MASON	0.100	0.103	0.099	0.098	
08/28/2010	02:53	MIDDLEBURY PD	CHRISTOPHER L MASON	0.125	0.119	0.098	0.098	
09/11/2010	03:10	MIDDLEBURY PD	CHRISTOPHER L MASON	0.082		0.099		RESPONSE TIMEOUT
09/19/2010	01:25	MIDDLEBURY PD	CHRISTOPHER L MASON			0.096		STANDARD OUT OF RANGE
09/19/2010	01:27	MIDDLEBURY PD	CHRISTOPHER L MASON			0.097		STANDARD OUT OF RANGE
09/22/2010	01:50	MIDDLEBURY PD	CHRISTOPHER L MASON			0.097		STANDARD OUT OF RANGE
09/22/2010	01:52	MIDDLEBURY PD	CHRISTOPHER L MASON			0.097		STANDARD OUT OF RANGE
09/26/2010	00:43	MIDDLEBURY PD	CHRISTOPHER L MASON	0.073		0.102		
10/02/2010	03:13	MIDDLEBURY PD	CHRISTOPHER L MASON	0.136		0.102		
10/06/2010	02:59	MIDDLEBURY PD	CHRISTOPHER L MASON	0.132		0.102		
10/06/2010	03:20	MIDDLEBURY PD	CHRISTOPHER L MASON	0.133	I	0.101	0.100	RESPONSE TIMEOUT
10/16/2010	03:14	MIDDLEBURY PD	PAUL F SOKOLOWSKI	0.092		0.101		
10/21/2010	14:13	BRISTOL PD	KEVIN E GIBBS	0.190	0.184	0.100	0.100	
10/23/2010	01:06	MIDDLEBURY PD	CHRISTOPHER L MASON	0.235		0.100		
10/23/2010	04:22	MIDDLEBURY PD	CHRISTOPHER L MASON	0.134		0.100	0.098	
10/24/2010	02:01	MIDDLEBURY PD	CHRISTOPHER L MASON	0.109		0.101		
10/30/2010	02:57	MIDDLEBURY PD	PAUL F SOKOLOWSKI	0.200	0.208	0.099	0.098	
10/30/2010	03:26	MIDDLEBURY PD	MORGAN B BRUNELL	0.108		0.098		
11/05/2010	03:47	MIDDLEBURY PD	CHRISTOPHER L MASON	0.124	0.129	0.102	0.101	
11/06/2010	03:57	MIDDLEBURY PD	CHRISTOPHER L MASON	0.148		0.102		RESPONSE TIMEOUT
11/07/2010	02:21	MIDDLEBURY PD	PAUL F SOKOLOWSKI	0.162	0.163	0.102	0.101	
11/11/2010	23:41	MIDDLEBURY PD	PETER D NEWTON	0.149	0.147	0.101	0.101	
11/12/2010	02:15	MIDDLEBURY PD	CHRISTOPHER L MASON	0.085	0.090	0.100	0.101	
11/13/2010	04:03	MIDDLEBURY PD	PAUL F SOKOLOWSKI	0.090		0.100		RESPONSE TIMEOUT
11/14/2010	02:39	MIDDLEBURY PD	CHRISTOPHER L MASON	0.106	0.109	0.102	0.101	
11/14/2010	15:32	MIDDLEBURY PD	NEIL F MOGERLEY	I	I	0.101		RESPONSE TIMEOUT
11/14/2010	15:44	MIDDLEBURY PD	NEIL F MOGERLEY	I	I	0.100		RESPONSE TIMEOUT
11/14/2010	16:33	MIDDLEBURY PD	NEIL F MOGERLEY	I	R	0.100		
11/19/2010	02:43	MIDDLEBURY PD	CHRISTOPHER L MASON	0.137		0.101		
11/20/2010	02:08	MIDDLEBURY PD	PAUL F SOKOLOWSKI	0.125		0.101		
11/21/2010	02:49	MIDDLEBURY PD	CHRISTOPHER L MASON	0.073		0.101		
11/22/2010	03:15	MIDDLEBURY PD	PAUL F SOKOLOWSKI	0.111		0.101		
11/25/2010	01:51	MIDDLEBURY PD	CHRISTOPHER L MASON	0.154		0.100	0.096	STANDARD OUT OF RANGE
12/10/2010	02:59	MIDDLEBURY PD	PAUL F SOKOLOWSKI	0.126		0.100		

DMT -- 100137-000103