DATAMASTER DMT MAINTENANCE LOG

VERGENNES PD.
Department/Agency

122106
DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
3/1/13	[For RPC Label Here] **DataMaster Simulator Solution** Lot # 12-80-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 07/5/2012 Expires: 7/5/2013 OPENED: 3/1/2013	B. Newton DMT Sup.
6/26/13	[For RPC Label Here] OPENED: **DataMaster Simulator Solution** Lot # 12-88-100 EtOHaq Conc: 0.098 Range: 0.093 - 0.103 Preparation Date: 11/9/2012 Expires: 11/9/2013 OPENED: 6/26/2013	B. Newton DMT Sup.
10/17/13	[For RPC ce Label Here] **DataMaster Simulator Solution** Lot # 12-89-100 EtOHaq Conc: 0.097 Range: 0.092 - 0.102 Preparation Date: 11/21/2012 Expires: 11/21/2013 OPENED: 10/17/2013	B. Newton DMT Sup.
3/5/14	[For RPC or abel Here] **DataMaster Simulator Solution** Lot # 13-16-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 9/26/2013 Expires: 09/26/2014 OPENED: 3/5/14 RCD Installation into new building	R. Driscoll DMT TECH

ALC 803 REV 03 08/2011

DATAMASTER DMT MAINTENANCE LOG

VERGENNES P.D.
Department/Agency

122106
DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
5/09/14	**DataMaster Simulator Solution** Lot # 13-16-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 9/26/2013 Expires: 09/26/2014 OPENED: 05/09/14 [For RF ... ace Label Here]	Adam O'Neill Police Officer
7/4/14	**DataMaster Simulator Solution** Lot # 14-01-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 1/9/2014 Expires: 1/9/2015 OPENED: 7/4/2014 [For RP ... ace Label Here]	BRENT NEWTON DMT SUPERVISOR
10/27/14	**DataMaster Simulator Solution** Lot # 14-05-100 EtOHaq Conc: 0.098 Range: 0.093 - 0.103 Preparation Date: 6/19/2014 Expires: 6/19/2015 OPENED: 10/27/14 [For RP(... ce Label Here]	PATRICK GREENSLEY DMT SUPERVISOR
	[For RPC or Simulator Solution Change Place Label Here]	

ALC 803 REV 03 08/2011