DataMaster DMT Monthly Update Check-list

Department/Agency: Manchester PD DM serial #: 100165

Supervisor: D. Steere Date: 02/01/13

	Provided	None this month
Operators Logbook	☑	☐
Maintenance Logbook	☑	☐
Breath Test History Report	☑	☐
Status History Report	☑	☐
Log On History Report	☑	☐
Calibration Report	☐	☑
Certification Report	☐	☑
Installation Report	☐	☑
Annual Preventative Maintenance	☐	☑
Routine Performance Check	☐	☑
Simulator Solution Change	☐	☑
Other Reports	☑	☐

List other report type and date:

Check in report.

DATAMASTER DMT OPERATOR USE LOG

Manchester Pd
Department/Agency

100165
DataMaster DMT Serial Number

Date	DMT Operator's Name	Subject's Name	Simulator Vapor 1	Simulator Vapor 2
10/11/11	Steven Harnois	Install	.100	
11-1-11	Thomas Crowe	Logan Spencer	.099 DS	.098 DS
11-4-11	Tienken	Devon Heiser	.100	.099
11-13-11	Schmidt (4")	Kileen Deets	.098	
11-20-11	Hess	Aaron	.098	
12-9-11	Godfrey	Guryan, Sheppard A	.096 (60R) DS	
12/20/11	STEERE	Test Printer		
12/27/11	Hatch 447	Refused	.097	—
1/14/11 (1/14/12) DS	Godfrey	Kevin Benasick	.100 DS	
1-14-14 (1/14/12) DS	Godfrey	Walters, Patrick	.099	
1-17-12	Tienken	Williams Richard	.099	
1-28-12	FRASIER	Call-Lie Big, Claudia	.102	
2.11.12	Hatch 447	Anthony Turnbo	.102	
02/14/12	Daniel Steere	Test/RPC 0.102	.102	
02/25/12	FRASIER	Slocum Daniel	.101	
2-26-12	Godfrey	Aldrich, James A	.101	
3-10-12	Godfrey	Durrschmidt, Chris	.099	.099
3-18-12	CARROCCIO	GUSTAFSON, Hope	~~.098~~ .101	
3-22-12	J. Epstein	Denish Wolff	.100 DS	
3-22-12	Schmidt (4")	Adrian Sweeney	.099	.099

ALC 603 Rev 3 08/11

DATAMASTER DMT OPERATOR USE LOG

Manchester P.D
Department/Agency

DataMaster DMT Serial Number

Date	DMT Operator's Name	Subject's Name	Simulator Vapor 1	Simulator Vapor 2
4/3/12	L. Hatch 1447	David Roberts	.101	
4/3/12	J. Blanchard	Tanya Parker	JB .098	
4/26/12	C. Mason	Gary Fifer	.096	
4/29/12	C. Mason	Carolyn Hadden	.101	
5/5/12	L. Hatch 447	Charles W. Wise	.101	
5-11-12	D. Tienken	Daniel Flint	.096	
5-18-12	GODFREY	Barden, Carson	.097	
5/24/12	J ROGERS	Conte, Christopher	.095	
5/24/12	J ROGERS	Conte, Christopher	.095	
5/24/12	J Rogers	Conte, Christopher	.096	
5/27/12	Oliver, T	O'Brien, ME	Out of Range .094	.095 DD
5/28/12	D. Steere	Test Simulator Solution Change	0.102	
6-2-12	Godfrey	Garay, Sarah H	.102	
6/3/12	Mosar	Weber, Brett	.101	
6/8/12	A. Boldue	APM	0.099	
6/19/12	D. Steere	RPC	—	
06/24/12	C. WATROUS	PETER ROWE	.102	.101
7/5/12	L. Hatch	Allen Erickson	.101	.101
7/8/12	SCHMIDT	Katelyn Barcellos	time out .099	.099
7/8/12	SCHMIDT	Katelyn Barcellos	.101	

ALC 603 Rev 3 08/11

DATAMASTER DMT OPERATOR USE LOG

Manchester PD
Department/Agency

100165
DataMaster DMT Serial Number

Date	DMT Operator's Name	Subject's Name	Simulator Vapor 1	Simulator Vapor 2
07/12/12	WATROUS - M249	JOHN MARINO	.101	N/A
7/13/12	Mason - M248	Ray White Jr	.100	N/A
8/7/12	D. Tienken	Carol Lancaster	.098	.098
8/10/12	L. HATCH	Austen Adams	.100	.100
8-13-12	GODFREY	Boyd, Ian S	.100	.099
8-14-12	Thomas	Alpert, David	.099	.098
8/17/12	Blanchard	Melissa Loughlin	.099	—
8/24/12	Tienken	Gavin Wilcox	.099	.098
9/02/12	Steere (Check In Test)	Test	.095	—
9/02/12	Steere (DUI Test)	Test	.097	—
9/08/12	Blanchard	Robert W. Swinarton	.098	—
09/23/12	Thomas	Jaron C Gardner	.098	
10/13/12	Steere	RPC	.100	
10/24/12	GODFREY	Alfred F. Gray	.100	
11/16/12	Walker	Jesse D Wadsworth	.100	
11/24/12	Zimmer	William Whitman	.100 burned .100	
11/30	Howard BC51	Barbosa Jorge	.100	
12/24	BURNETT	DONALD BLOOD	.100	REFUSED
12/30	HATCH 447	Scott Freeman	.098	
1/4/13	Blanchard	Kathleen McHenry	.098	

DATAMASTER DMT OPERATOR USE LOG

Department/Agency

DataMaster DMT Serial Number

Date	DMT Operator's Name	Subject's Name	Simulator Vapor 1	Simulator Vapor 2
01/07/13	J.D. Bossonny	Harry McDermott	.097	
1/9/12	J. Blanchard	Kathleen McHenry	.100	
1/12/13	A. CARROCCIO	KRISTOPHER DESRAVILES	.062	
1/12/13	~~Kayla Bothomus~~ D. Wright	Kayla Bothomus	.144	
1/21/13	HATCH	Abigail Lanilko	.096	
01/30/13	Steere	Shannon O'brian	.098	

DATAMASTER DMT MAINTENANCE LOG

Manchester Pd
Department/Agency

100165
DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
10/11/11	**DataMaster Simulator Solution** Lot # 12-13-100 $EtOH_{aq}$ Conc: 0.102 Range: 0.097 - 0.107 Preparation Date: 09/09/2011 Expires: 09/09/2012 OPENED: 10/11/11 [...ange Place Label Here] Install Per Protocall [signature] 10/11/11	Steven Harnois
12-9-11	out of Range [For RPC or Simulator Solution Change Place Label Here]	
12-11-11	**DataMaster Ext. Std. Solution** Lot # 11-41-100 $EtOH_{aq}$ Conc: 0.101 Range: 0.096 - 0.106 Preparation Date: 06/29/2011 Expires: 06/29/2012 OPENED: 12-11-11 Solution Chgd Solution OK Pump ERROR	[signature] out of Serv.
12/20/11	Printer problem, Attempted Reprint Last Ticket [For RPC or Simulator Solution Change Place Label Here] would not print, Turned off + on Instrument. Printed last ticket + prior from 9/27/11/Reset Time back 1 hour.	~~Would not Print.~~ ~~Turned off Instrument~~ ~~+ Turned on~~ [signature]
1/10/12	[For RPC or Simulator Solution Change Place Label Here] Updated Software	Steven Harnois

ALC 803 REV 03 08/2011

DATAMASTER DMT MAINTENANCE LOG

Manchester PD
Department/Agency

100165
DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
02/14/12	**DataMaster Simulator Solution** Lot # 12-38-100 $EtOH_{aq}$ Conc: 0.101 Range: 0.096 - 0.106 Preparation Date: 11/17/2011 Expires: 11/17/2012 [For RPC ce Label Here] OPENED: 02/14/12 DS.	Ofc. Daniel Steere
	[For RPC or Simulator Solution Change Place Label Here]	
04/01/12	Accuracy & Precision Check Reset Clock ahead 1hr.	Ofc. Daniel Steere
	DataMaster Simulator Solution Lot # 12-64-100 $EtOH_{aq}$ Conc: 0.101 Range: 0.096 - 0.106 Preparation Date: 02/17/2012 Expires: 02/17/2013 [For RI ace Label Here] OPENED: 05/28/12 DS	
05/28/12	Accuracy + Precision Check Solution Change	Daniel Steere
6/8/12	Performed APM, software update [For RPC or Simulator Solution Change Place Label Here] and lock change	SBR

ALC-803 REV 03 08/2011

DATAMASTER DMT MAINTENANCE LOG

Manchester PD
Department/Agency

100165
DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
06/19/12	[For RPC **DataMaster Simulator Solution** Lot # 12-64-100 $EtOH_{aq}$ Conc: 0.101 Range: 0.096 - 0.106 Preparation Date: 02/17/2012 Expires: 02/17/2013 Label Here] OPENED: 06/19/12 D.S. RPC	Daniel Steere
	[For RPC or Simulator Solution Change Place Label Here]	
09/02/12	Check-in + Dui Tests	Daniel Steere
10/13/12	[For RI **DataMaster Simulator Solution** Lot # 12-72-100 EtOHaq Conc: 0.100 Range: 0.095 - 0.105 Preparation Date: 04/24/2012 Expires: 04/24/2013 ace Label Here] OPENED: 10/13/12 D.S. R.P.C.	Daniel
	[For RPC or Simulator Solution Change Place Label Here]	

ALC 803 REV 03 08/2011

Breath Record History

Test Date	Test Time	Agency	Operator Name	Alcohol Val. 1	Alcohol Val. 2	Sim. Conc. 1	Sim. Conc. 2	Status Message
01/30/2013	18:12:38	MANCHESTER PD	DANIEL STEERE	0.000		0.098		
01/21/2013	01:09:16	SHAFTSBURY VSP	LEWIS A HATCH IV	0.066		0.096		
01/12/2013	06:30:46	WINHALL PD	DOUGLAS W WRIGHT	0.139	0.144	0.097	0.095	
01/12/2013	06:08:01	WINHALL PD	ANTHONY H CARROCCIO	0.062		0.097		
01/08/2013	20:21:32	MANCHESTER PD	JAMES W BLANCHARD	0.027		0.098		
01/07/2013	01:06:18	MANCHESTER PD	JAMES D BOSSONG	0.230		0.097	0.097	
01/04/2013	21:10:57	MANCHESTER PD	JAMES W BLANCHARD	0.066		0.098		

Status Record History

Status Date	Status Time	Test Type	Status Message

Log On Record History

Test Date	Test Time	Name



CHECK-IN CONFIRMATION TEST

STATE OF VERMONT

DataMater DMT: 100165
Location: Manchester PD
Certification Date: 09/27/2011
Installation Date: 10/11/2011

Date: 01/30/2013
Time: 18:12:38

SUBJECT NAME:
SHANNON L OBRIAN
DOB: 03/20/1972 AGE: 40
GENDER: M
LICENSE #: 991755519 NY

TEST OPERATOR NAME:
DANIEL STEERE
AGENCY: MANCHESTER PD

BLANK TEST	0.000	18:13
CALIBRATION CHECK	PASSED	18:13
SIMULATOR VAPOR 34.1°C	0.098	18:13
SUBJECT SAMPLE (Vol=4.9L)	0.000	18:14
BLANK TEST	0.000	18:15

All alcohol values reported in g/210L
Sim. Vapor Concentration 0.100
Sim. Vapor Range 0.095 - 0.105
Lot #: 12-72-100