

Vermont Department of Health
Division of Health Surveillance
Public Health Laboratory
195 Colchester Avenue - P.O. Box 1125
Burlington, VT 05402-1125

[phone] 802-863-7335
[fax] 802-863-7632
[or] 800-660-9997

Agency of Human Services

COUNTY: Bennington
Page 1 of 1

August 18, 2011

Dear colleague:


In a review of our records, we realized that some documents associated with DataMaster DMT Serial Number 100155 were omitted in your monthly update.

We apologize for any inconvenience this may cause your office.

If you have questions or suggestions regarding this process, please don't hesitate to contact me at (802) 863-7570.

Thank you.

Sincerely,



Mary Celotti
Laboratory Director

Enclosures:

DataMaster DMT S.N. 100155 (Bennington County Sheriff's Dept)
Simulator Solution Change Report, 11/12/2010
Simulator Solution Change Report, 11/19/2010
Accuracy and Precision Test, 01/05/2011



SIMULATOR SOLUTION CHANGE REPORT


DataMaster DMT: 100155
Location: BENNINGTON COUNTY SO
Date: 11/12/2010
Performed by: JOEL R HOWARD



Accuracy and Precision Check

Concentration = 0.098 g/210L
Lot # = 11-04-100
Range = 0.093 - 0.103
Average = 0.092 g/210L
Std Dev = 0.000

Simulator Temperature: 34.1°C

Performed by 

Date 11/12/2010

SIMULATOR SOLUTION CHANGE REPORT

DataMaster DMT: 100155
Location: BENNINGTON COUNTY SO
Date: 11/19/2010
Performed by: JOEL R HOWARD



Accuracy and Precision Check

Concentration = 0.102 g/210L
Lot # = 11-05-100
Range = 0.096 - 0.106
Average = 0.097 g/210L
Std Dev = 0.001

Simulator Temperature: 34.1°C

Performed by

Joel

Date

11/19/2010



ACCURACY AND PRECISION TEST

STATE OF VERMONT

DataMaster DMT: 100155

Date: 01/05/2011

Time: 14:46:24

SUPERVISOR NAME: ALB SOLUTION TEST

SOLUTION LOT #: 11-26-100

SOLUTION CONCENTRATION: 0.098

BLANK TEST	0.000	14:47
CALIBRATION CHECK	PASSED	14:47
SIMULATOR VAPOR	0.098	14:47

USER ABORT

