



Vermont Department of Health Laboratory

195 Colchester Avenue, P.O. Box 1125, Burlington VT 05402-1125
DataMaster Technical Services (ph) 802-863-7641; (fax) 802-863-7632

DATAMASTER TECHNICAL SUPPORT INQUIRY

Agency/Site: Bennington PD _____ DataMaster Serial #: 100146 _____

Date: 5/23 /2011 Requestor: Lauzon, Chris _____

Instrument Supervisor: Lauzon, Chris _____ Phone #: _____

Taken By: Harnois, Steve _____

Problem: Interference reported

Laboratory Response: Email back to DataMaster supervisor asking what type of interference, Radio, suspect simulator? Talked to supervisor while he ran a check in test.

Work Performed: _____

5/24/2011 - Asked Supervisor to run a check in test while on the phone with him, the interference happened during the Simulator solution vapor sample. Asked supervisor to perform a simulator solution change. SH

5/25/2011 - Talked to Chris the DataMaster Supervisor he indicated that the DataMaster Passed the simulator solution change Procedure. SH

5/31/2011 - Received passing Simulator Solution Change Report and Check in test, that both passed. ALB

7/5/2011 - reviewed breath record summary Fax of 6/30/2011 with no simulator Interference reports SH

Attachments 21 pages

On Site In-house Corrected by Phone call / E-mail

SH 7/9/11

Performed By: Steven Harnois and ALB _____

Date: 7/5/2011

Final Review: A. Bolduc _____

Date: 7/5 /2011

Reviewed ALB 7-5-11

Breath Record Summary

(05/25/2011-06/30/2011)

X = Interference, R = Refusal, V = Invalid, I = Incomplete

Test Date	Test Time	Agency	Operator Name	Alcohol Val. 1	Alcohol Val. 2	Sim. Conc. 1	Sim. Conc. 2	Status Message
05/25/2011	10:59	BENNINGTON PD	LAUZON C E	0.000		0.102		
05/27/2011	22:29	BRANDON PD	DAVID J FADEN	0.177		0.102		
05/30/2011	01:33	SHAFTSBURY VSP	JUSTIN W WALKER	0.131		0.102		
06/05/2011	02:17	SHAFTSBURY VSP	WAYNE M GODFREY	0.076		0.102		
06/11/2011	03:16	BENNINGTON PD	JOSHUA M STEW	R		0.102		
06/12/2011	05:03	BENNINGTON PD	JAMES A GUILLEY	I	0.123	0.102		
06/16/2011	02:43	BRANDON PD	ROSCOE E HARRINGTON	0.130	I	0.100	0.100	
06/18/2011	04:39	SHAFTSBURY VSP	JUSTIN W WALKER	I	I	0.100	0.100	

Handwritten signature/initials



BENNINGTON POLICE DEPARTMENT
118 SOUTH STREET
BENNINGTON, VERMONT 05201

Paul J. Doucette
CHIEF OF POLICE

FAX TRANSMITTAL COVER SHEET

TO: Steve
FAX #: 863-7632
FROM: Chris
DATE: 6/30/11

MESSAGE:

NUMBER OF PAGES: 2
(Including cover sheet)

This message is intended for the use of the individual or entity to which it is addressed and may contain information that is privileged, confidential, and exempt from disclosure under applicable law. If the reader of this message is not the intended recipient or the employee agent responsible for delivering the message to the intended recipient, you are hereby notified that the dissemination, distribution or copying of this communication is strictly prohibited. If you have received this communication in error, please notify us immediately by telephone and return all portions of the communication to us at the above address via the U.S. Postal Service.

Bolduc, Amanda

From: Chris Lauzon [clauzon@dps.state.vt.us]
Sent: Saturday, June 04, 2011 7:38 AM
To: Bolduc, Amanda
Subject: RE: Interference
will do...thanks chris

*Corporal Christopher "Pappy" Lauzon
Bennington Police Department
118 South Street
Bennington, Vermont 05201
802-442-1030 voice
802-442-1067 fax*

From: Bolduc, Amanda [mailto:Amanda.Bolduc@ahs.state.vt.us]
Sent: Wed 6/1/2011 7:57 AM
To: Chris Lauzon
Subject: RE: Interference

I received the reports today. Please contact us right away if you see interference detected again. Thanks

ABolduc

Amanda Bolduc, MFS
PH Chemist III
Vermont Department of Health Laboratory
Alcohol and Toxicology Program
195 Colchester Ave
Burlington, VT 05402
(802) 585-6707 (mobile)
(802) 863-7412 (ph)
(802) 863-7632 (fax)
amanda.bolduc@ahs.state.vt.us
PLEASE NOTE new phone and email as of 10/13/10

From: Chris Lauzon [mailto:clauzon@dps.state.vt.us]
Sent: Friday, May 27, 2011 7:36 AM
To: Bolduc, Amanda
Subject: RE: Interference

no, they went out via mail yesterday. chris

*Corporal Christopher "Pappy" Lauzon
Bennington Police Department
118 South Street
Bennington, Vermont 05201
802-442-1030 voice
802-442-1067 fax*

6/7/2011

From: Bolduc, Amanda [mailto:Amanda.Bolduc@ahs.state.vt.us]
Sent: Thu 5/26/2011 7:54 AM
To: Chris Lauzon
Subject: RE: Interference

did you fax them, too?

ABolduc

Amanda Bolduc, MFS
PH Chemist III
Vermont Department of Health Laboratory
Alcohol and Toxicology Program
195 Colchester Ave
Burlington, VT 05402
(802) 585-6707 (mobile)
(802) 863-7412 (ph)
(802) 863-7632 (fax)
amanda.bolduc@ahs.state.vt.us
PLEASE NOTE new phone and email as of 10/13/10

From: Chris Lauzon [mailto:clauzon@dps.state.vt.us]
Sent: Thursday, May 26, 2011 7:44 AM
To: Bolduc, Amanda
Subject: RE: Interference

amanda, i will mail them out today. chris

Corporal Christopher "Pappy" Lauzon
Bennington Police Department
118 South Street
Bennington, Vermont 05201
802-442-1030 voice
802-442-1067 fax

From: Bolduc, Amanda [mailto:Amanda.Bolduc@ahs.state.vt.us]
Sent: Wed 5/25/2011 12:23 PM
To: Chris Lauzon
Subject: RE: Interference

We would prefer not. If the only alternative is snail mail, I guess a fax would be okay, as long as you also mail them in, too. We prefer color copies of tickets.
thanks

ABolduc

Amanda Bolduc, MFS
PH Chemist III

6/7/2011

Vermont Department of Health Laboratory
Alcohol and Toxicology Program
195 Colchester Ave
Burlington, VT 05402
(802) 585-6707 (mobile)
(802) 863-7412 (ph)
(802) 863-7632 (fax)
amanda.bolduc@ahs.state.vt.us
PLEASE NOTE new phone and email as of 10/13/10

From: Chris Lauzon [mailto:clauzon@dps.state.vt.us]
Sent: Wednesday, May 25, 2011 11:31 AM
To: Bolduc, Amanda
Subject: RE: Interference

amanda, ok to fax?

Corporal Christopher "Pappy" Lauzon
Bennington Police Department
118 South Street
Bennington, Vermont 05201
802-442-1030 voice
802-442-1067 fax

From: Bolduc, Amanda [mailto:Amanda.Bolduc@ahs.state.vt.us]
Sent: Wed 5/25/2011 9:05 AM
To: Harnois, Steven; Chris Lauzon
Subject: Interference

Chris,
Would it be possible for you to scan and email Steve and/or I copies of these tickets?

ABolduc

Amanda Bolduc, MFS
PH Chemist III
Vermont Department of Health Laboratory
Alcohol and Toxicology Program
195 Colchester Ave
Burlington, VT 05402
(802) 585-6707 (mobile)
(802) 863-7412 (ph)
(802) 863-7632 (fax)
amanda.bolduc@ahs.state.vt.us
PLEASE NOTE new phone and email as of 10/13/10

6/7/2011

From: Harnois, Steven
Sent: Monday, May 23, 2011 1:26 PM
To: 'Chris Lauzon'
Cc: Kimball, Kirk; Bolduc, Amanda; Merrill, Steven; Harnois, Steven
Subject: RE: [SPAM?] FW: Revised Proper Sample Acceptance test (pg 17 supervisor manual D-Alc-011 Addendum A)

Chris,

Thanks for the notification, unfortunately, there is a lot instances of interference that can happen, is this a subject breath, Radio Frequency Interference, when does this happen?

Steven Harnois
State of Vermont
Agency Of Human Services
Department of Health
Public Health Electronics Technician
195 Colchester Avenue
Burlington, Vermont 05401
Phone (802) 863 7641
Fax (802) 863 7632
E Mail: NOTE NEW EMAIL Steven.Harnois@ahs.state.vt.us

From: Chris Lauzon [mailto:clauzon@dps.state.vt.us]
Sent: Monday, May 23, 2011 11:51 AM
To: Harnois, Steven
Subject: RE: [SPAM?] FW: Revised Proper Sample Acceptance test (pg 17 supervisor manual D-Alc-011 Addendum A)

stevo, i keep getting "interference" messages on the dmt...chris

Corporal Christopher "Pappy" Lauzon
Bennington Police Department
118 South Street
Bennington, Vermont 05201
802-442-1030 voice
802-442-1067 fax

From: Harnois, Steven [mailto:Steven.Harnois@ahs.state.vt.us]
Sent: Mon 5/23/2011 8:56 AM
Subject: [SPAM?] FW: Revised Proper Sample Acceptance test (pg 17 supervisor manual D-Alc-011 Addendum A)

DataMaster Supervisors,

Please see the two Attachments & Note that June is almost here and June is time to do an RPC at your site, when you have completed your RPC, *We will only except A COLOR scan of the RPC report OR Send a COLOR Copy to the lab by U.S.P.S.* , Please No Faxes, Faxes are not Color and therefore we cannot tell the black from the blue line.

6/7/2011

Thanks, Steve

Steven Harnois
State of Vermont
Agency Of Human Services
Department of Health
Public Health Electronics Technician
195 Colchester Avenue
Burlington, Vermont 05401
Phone (802) 863 7641
Fax (802) 863 7632
E Mail: NOTE NEW EMAIL Steven.Harnois@ahs.state.vt.us

From: Kimball, Kirk
Sent: Friday, May 20, 2011 2:24 PM
To: Harnois, Steven
Cc: Bolduc, Amanda; 'Schurr, Stuart'
Subject: Revised Proper Sample Acceptance test (pg 17 supervisor manual D-Alc-011 Addendum A)

Steve,

Here is the letter and D-Alc-011 Addendum A files that need to go to all DMT DataMaster Supervisors in the state.

Kirk L. Kimball
Organic Chemistry Program Chief
VT. Dept. of Health
Burlington, VT 05402
(802) 951-1293
kirk.kimball@ahs.state.vt.us

6/7/2011

RECEIVED

MAY 31 2011

SIMULATOR SOLUTION CHANGE REPORT

Vermont Department of
Health Laboratory

DataMaster DMT: 100146
Location: BENNINGTON PD
Date: 05/25/2011
Performed by: CHRIS E LAUZON



VERMONT

DEPARTMENT OF HEALTH

Accuracy and Precision Check

Concentration = 0.103 g/210L
Lot # = 11-25-100
Range = 0.098 - 0.108
Average = 0.102 g/210L
Std Dev = 0.000

Simulator Temperature: 34.0°C

Performed by ChE

Date 05/25/2011

ChE
6-25-11



Check-In Confirmation Test

STATE OF VERMONT

DataMaster DMT: 100146
BENNINGTON PD
Certification Date: 01/20/2011
Installation Date: 02/18/2011


Date: 05/25/2011
Time: 10:59:10

Subject Name: TEST T TEST
Date of Birth: 05/25/2011 Age: 0
Guardian Present: N
Gender: M
License #: 123 VT

Test Operator Name: LAUZON C E
Agency: BENNINGTON PD

BLANK TEST	0.000	11:00
CALIBRATION CHECK	PASSED	11:00
SIMULATOR VAPOR	0.102	11:00
BLANK TEST	0.000	11:01
SUBJECT SAMPLE (Vol=2.7L)	0.000	11:01

All alcohol values reported in g/210L
Sim. Vapor Concentration 0.103
Sim. Vapor Range 0.098 - 0.108
Simulator Temp 34.0°C



RECEIVED

MAY 31 2011

Vermont Department of
Health Laboratory



DUI Subject Test

STATE OF VERMONT

DataMaster DMT: 100146
BENNINGTON PD
Certification Date: 01/20/2011
Installation Date: 02/18/2011

Date: 05/21/2011
Time: 02:48:32

Subject Name: LEEANN LYKSTAD
Date of Birth: 09/29/1981 Age: 29
Gender: F
License #: 81917892 VT

Case #: 11BN03394
Time of Operation: 01:19
Location of Incident: MORGAN STREET, MAIN STREET
Town: BENNINGTON
County: BENNINGTON

Test Operator Name: DAVID J FADEN
Agency: BENNINGTON PD

Obs. Period Start: 02:13

BLANK TEST	0.000	02:49
CALIBRATION CHECK	PASSED	02:49
SIMULATOR VAPOR	INTERFERENCE	02:49

INTERFERENCE DETECTED

All alcohol values reported in g/210L
Sim. Vapor Concentration 0.100
Sim. Vapor Range 0.095 - 0.105
Simulator Temp 34.0°C

RECEIVED

MAY 31 2011

Vermont Department of
Health Laboratory



DUI Subject Test

STATE OF VERMONT

DataMaster DMT: 100146
BENNINGTON PD
Certification Date: 01/20/2011
Installation Date: 02/18/2011

Date: 05/21/2011
Time: 02:44:15

Subject Name: LEE A LYKSTAD
Date of Birth: 09/29/1981 Age: 29
Gender: F
License #: 81917892 VT

Case #: 11BN03394
Time of Operation: 01:19
Location of Incident: MORGAN STREET NEAR MAIN STREET
Town: BENNINGTON
County: BENNINGTON

Test Operator Name: DAVID J FADEN
Agency: BENNINGTON PD

Obs. Period Start: 02:13

BLANK TEST	0.000	02:45
CALIBRATION CHECK	PASSED	02:45
SIMULATOR VAPOR	INTERFERENCE	02:45

INTERFERENCE DETECTED

All alcohol values reported in g/210L
Sim. Vapor Concentration 0.100
Sim. Vapor Range 0.095 - 0.105
Simulator Temp 34.1°C

RECEIVED

MAY 31 2011

Vermont Department of
Health Laboratory



DUI Subject Test

STATE OF VERMONT

DataMaster DMT: 100146
BENNINGTON PD
Certification Date: 01/20/2011
Installation Date: 02/18/2011

Date: 05/21/2011
Time: 02:39:22

Subject Name: LEE A LYKSTAD
Date of Birth: 09/29/1981 Age: 29
Gender: F
License #: 81917892 VT

Case #: 11BN03394
Time of Operation: 01:19
Location of Incident: MORGAN STREET NEAR MAIN STREET
Town: BENNINGTON
County: BENNINGTON

Test Operator Name: DAVID J FADEN
Agency: BENNINGTON PD

Obs. Period Start: 02:13

BLANK TEST	0.000	02:40
CALIBRATION CHECK	PASSED	02:40
SIMULATOR VAPOR	INTERFERENCE	02:40

INTERFERENCE DETECTED

All alcohol values reported in g/210L
Sim. Vapor Concentration 0.100
Sim. Vapor Range 0.095 - 0.105
Simulator Temp 33.9°C

RECEIVED

MAY 31 2011

Vermont Department of
Health Laboratory



DUI Subject Test

STATE OF VERMONT

DataMaster DMT: 100146
BENNINGTON PD
Certification Date: 01/20/2011
Installation Date: 02/18/2011

Date: 05/21/2011
Time: 02:32:06

Subject Name: LEE A LYKSTAD
Date of Birth: 09/29/1981 Age: 29
Gender: F
License #: 81917892 VT

Case #: 11BN03394
Time of Operation: 01:19
Location of Incident: MORGAN STREET NEAR MAIN STREET
Town: BENNINGTON
County: BENNINGTON

Test Operator Name: DAVID J FADEN
Agency: BENNINGTON PD

Obs. Period Start: 02:13

BLANK TEST	0.000	02:33
CALIBRATION CHECK	PASSED	02:33
SIMULATOR VAPOR	INTERFERENCE	02:33

INTERFERENCE DETECTED

All alcohol values reported in g/210L
Sim. Vapor Concentration 0.100
Sim. Vapor Range 0.095 - 0.105
Simulator Temp 34.1°C

RECEIVED

MAY 31 2011

Vermont Department of
Health Laboratory



DUI Subject Test

STATE OF VERMONT

DataMaster DMT: 100146
BENNINGTON PD
Certification Date: 01/20/2011
Installation Date: 02/18/2011

Date: 05/21/2011
Time: 02:29:24

Subject Name: LEE A LYKSTAD
Date of Birth: 09/29/1981 Age: 29
Gender: F
License #: 81917892 VT

Case #: 11BN03394
Time of Operation: 01:19
Location of Incident: MORGAN STREET NEAR MAIN STREET
Town: BENNINGTON
County: BENNINGTON

Test Operator Name: DAVID J FADEN
Agency: BENNINGTON PD

Obs. Period Start: 02:13

BLANK TEST	0.000	02:30
CALIBRATION CHECK	PASSED	02:30
SIMULATOR VAPOR	INTERFERENCE	02:30

INTERFERENCE DETECTED

All alcohol values reported in g/210L
Sim. Vapor Concentration 0.100
Sim. Vapor Range 0.095 - 0.105
Simulator Temp 34.0°C

RECEIVED

MAY 31 2011

Vermont Department of
Health Laboratory



DUI Subject Test

STATE OF VERMONT

DataMaster DMT: 100146
BENNINGTON PD
Certification Date: 01/20/2011
Installation Date: 02/18/2011

Date: 05/22/2011
Time: 01:40:30

Subject Name: BRANDON M MAYHEW
Date of Birth: 07/30/1987 Age: 23
Gender: M
License #: 32194842 VT

Case #: 11BN03437
Time of Operation: 01:18
Location of Incident: COUNTY STREET
Town: BENNINGTON
County: BENNINGTON

Test Operator Name: ANDREW W HUNT
Agency: BENNINGTON PD

Obs. Period Start: 01:18

BLANK TEST	0.000	01:41
CALIBRATION CHECK	PASSED	01:41
SIMULATOR VAPOR	INTERFERENCE	01:41

INTERFERENCE DETECTED

All alcohol values reported in g/210L
Sim. Vapor Concentration 0.100
Sim. Vapor Range 0.095 - 0.105
Simulator Temp 34.2°C

RECEIVED

MAY 31 2011

Vermont Department of
Health Laboratory



DUI Subject Test

STATE OF VERMONT

DataMaster DMT: 100146
BENNINGTON PD
Certification Date: 01/20/2011
Installation Date: 02/18/2011

Date: 05/22/2011
Time: 01:37:32

Subject Name: BRANDON M MAYHEW
Date of Birth: 07/30/1987 Age: 23
Gender: M
License #: 32194842 VT

Case #: 11BN03437
Time of Operation: 01:18
Location of Incident: COUNTY STREET
Town: BENNINGTON
County: BENNINGTON

Test Operator Name: ANDREW W HUNT
Agency: BENNINGTON PD

Obs. Period Start: 01:18

BLANK TEST	0.000	01:38
CALIBRATION CHECK	PASSED	01:38
SIMULATOR VAPOR	INTERFERENCE	01:38

INTERFERENCE DETECTED

All alcohol values reported in g/210L
Sim. Vapor Concentration 0.100
Sim. Vapor Range 0.095 - 0.105
Simulator Temp 34.0°C

RECEIVED

MAY 31 2011

Vermont Department of
Health Laboratory



DUI Subject Test

STATE OF VERMONT

DataMaster DMT: 100146
BENNINGTON PD
Certification Date: 01/20/2011
Installation Date: 02/18/2011

Date: 05/22/2011
Time: 01:34:54

Subject Name: BRANDON M MAYHEW
Date of Birth: 07/30/1987 Age: 23
Gender: M
License #: 32194842 VT

Case #: 11BN03437
Time of Operation: 01:18
Location of Incident: COUNTY STREET
Town: BENNINGTON
County: BENNINGTON

Test Operator Name: ANDREW W HUNT
Agency: BENNINGTON PD

Obs. Period Start: 01:18

BLANK TEST	0.000	01:35
CALIBRATION CHECK	PASSED	01:35
SIMULATOR VAPOR	INTERFERENCE	01:36

INTERFERENCE DETECTED

All alcohol values reported in g/210L
Sim. Vapor Concentration 0.100
Sim. Vapor Range 0.095 - 0.105
Simulator Temp 34.1°C

RECEIVED

MAY 31 2011

Vermont Department of
Health Laboratory

Harnois, Steven

From: Chris Lauzon [clauzon@dps.state.vt.us]
Sent: Tuesday, May 24, 2011 7:37 AM
To: Harnois, Steven
Subject: RE: [SPAM?] FW: Revised Proper Sample Acceptance test (pg 17 supervisor manual D-Alc-011 Addendum A)
stevo, just interference...no subject in the room...just the message each time they try and test

Corporal Christopher "Pappy" Lauzon
Bennington Police Department
118 South Street
Bennington, Vermont 05201
802-442-1030 voice
802-442-1067 fax

From: Harnois, Steven [mailto:Steven.Harnois@ahs.state.vt.us]
Sent: Mon 5/23/2011 1:26 PM
To: Chris Lauzon
Cc: Kimball, Kirk; Bolduc, Amanda; Merrill, Steven; Harnois, Steven
Subject: RE: [SPAM?] FW: Revised Proper Sample Acceptance test (pg 17 supervisor manual D-Alc-011 Addendum A)

Chris,

Thanks for the notification, unfortunately, there is a lot instances of interference that can happen, is this a subject breath, Radio Frequency Interference, when does this happen?

Steven Harnois
State of Vermont
Agency Of Human Services
Department of Health
Public Health Electronics Technician
195 Colchester Avenue
Burlington, Vermont 05401
Phone (802) 863 7641
Fax (802) 863 7632
E Mail: NOTE NEW EMAIL Steven.Harnois@ahs.state.vt.us

From: Chris Lauzon [mailto:clauzon@dps.state.vt.us]
Sent: Monday, May 23, 2011 11:51 AM
To: Harnois, Steven
Subject: RE: [SPAM?] FW: Revised Proper Sample Acceptance test (pg 17 supervisor manual D-Alc-011 Addendum A)

stevo, i keep getting "interference" messages on the dmt...chris

Corporal Christopher "Pappy" Lauzon
Bennington Police Department
118 South Street
Bennington, Vermont 05201

6/21/2011

802-442-1030 voice
802-442-1067 fax

From: Harnois, Steven [mailto:Steven.Harnois@ahs.state.vt.us]
Sent: Mon 5/23/2011 8:56 AM
Subject: [SPAM?] FW: Revised Proper Sample Acceptance test (pg 17 supervisor manual D-Alc-011 Addendum A)

DataMaster Supervisors,

Please see the two Attachments & Note that June is almost here and June is time to do an RPC at your site, when you have completed your RPC, **We will only except A COLOR scan of the RPC report OR Send a COLOR Copy to the lab by U.S.P.S.** , Please No Faxes, Faxes are not Color and therefore we cannot tell the black from the blue line.

Thanks, Steve

Steven Harnois
State of Vermont
Agency Of Human Services
Department of Health
Public Health Electronics Technician
195 Colchester Avenue
Burlington, Vermont 05401
Phone (802) 863 7641
Fax (802) 863 7632
E Mail: NOTE NEW EMAIL Steven.Harnois@ahs.state.vt.us

From: Kimball, Kirk
Sent: Friday, May 20, 2011 2:24 PM
To: Harnois, Steven
Cc: Bolduc, Amanda; 'Schurr, Stuart'
Subject: Revised Proper Sample Acceptance test (pg 17 supervisor manual D-Alc-011 Addendum A)

Steve,

Here is the letter and D-Alc-011 Addendum A files that need to go to all DMT DataMaster Supervisors in the state.

Kirk L. Kimball
Organic Chemistry Program Chief
VT. Dept. of Health
Burlington, VT 05402
(802) 951-1293
kirk.kimball@ahs.state.vt.us

6/21/2011

802-863-7632



DUI Subject Test

STATE OF VERMONT

DataMaster DMT: 100146
BENNINGTON PD
Certification Date: 01/20/2011
Installation Date: 02/10/2011

Date: 05/22/2011
Time: 01:40:30

Subject Name: BRANDON M MAYHEW
Date of Birth: 07/30/1987 Age: 23
Gender: M
License #: 32194042 VT

Case #: 11BN03437
Time of Operation: 01:10
Location of Incident: COUNTY STREET
Town: BENNINGTON
County: BENNINGTON

Test Operator Name: ANDREW W HUNT
Agency: BENNINGTON PD

Obs. Period Start: 01:10

BLANK TEST	0.000	01:41
CALIBRATION CHECK	PASSED	01:41
STIMULATOR VAPOR	INTERFERENCE	01:41

INTERFERENCE DETECTED

All alcohol values reported in g/210L.
Sim. Vapor Concentration 0.100
Sim. Vapor Range 0.095 - 0.105
Simulator Temp 34.2°C



Paul J. Doucette
CHIEF OF POLICE

BENNINGTON POLICE DEPARTMENT
118 SOUTH STREET
BENNINGTON, VERMONT 05201

FAX TRANSMITTAL COVER SHEET

TO: Steve Haanis
FAX #: 802-863-7632
FROM: C. Luzzo
DATE: 5/24/11

MESSAGE:

NUMBER OF PAGES: 2
(including cover sheet)

This message is intended for the use of the individual or entity to which it is addressed and may contain information that is privileged, confidential, and exempt from disclosure under applicable law. If the reader of this message is not the intended recipient or the employee agent responsible for delivering the message to the intended recipient, you are hereby notified that the dissemination, distribution or copying of this communication is strictly prohibited. If you have received this communication in error, please notify us immediately by telephone and return all portions of the communication to us at the above address via the U.S. Postal Service.