BAC DATAMASTER CHECK-UP and MAINTENANCE LOG

Manchester Pd
Department/Agency

100142
DataMaster Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY
8/31/10	ere] **DataMaster Simulator Solution** Lot # 11-04-100 $EtOH_{aq}$ Conc: 0.098 Range: 0.093– 0.103 Preparation Date: 8/19/2010 Expires: 8/19/2011 Installed Per Protocal	Steven Harrois
9/2/10	[Place Simulator Solution Label Here] Activated Sim temp and tolerance check	A. Bolden
11/2/10	[Place Simulator Solution Label Here] **DataMaster Simulator Solution** Lot # 11-14-100 $EtOH_{aq}$ Conc: 0.098 Range: 0.093– 0.103 Preparation Date: 10/8/2010 Expires: 10/8/2011 Quartzely maint	Sgt. M. [illegible]
2/28/11	[Place Simulator Solution Label Here] **DataMaster Simulator Solution** Lot # 11-04-100 $EtOH_{aq}$ Conc: 0.098 Range: 0.093– 0.103 Preparation Date: 8/19/2010 Expires: 8/19/2011 QUARTERLY MAINT.	Sgt. M. [illegible]
	[Place Simulator Solution Label Here]	
	[Place Simulator Solution Label Here]	

ALC 803 Rev. 2 06/07