Vermont Department of Health Laboratory to
National Patent Analytical Systems
DataMaster DMT Shipping Memo

SIM # G10919
NIST Temp = 34.1°C
10-12-11
10/10/11

To:
National Patent Analytical Systems
2090 Harrington Memorial Road
Mansfield, Ohio 44901

From:
State of Vermont Health Department Laboratory
195 Colchester Ave
Burlington, Vermont 05401

Date: 10/12/11

Re:
Shipment of DataMaster DMT S/N: 100156 **for the Vermont Certification Project.**

Instrument Preparation:

1. **Removed Simulator & secured simulator arms with Ty-Wrap.** 10-12-11
2. **Removed breath tube.** 10-12-11
3. **Downloaded instrument files to USB thumb drive.** 10-12-11
4. **Removed keyboard & stylus.** 10-12-11
5. **Removed security screws & replaced with manufacturer supplied screws.** 10-12-11
6. **Instrument bagged & packaged for shipment.**
7. **How many TSIs are included with instrument?** 0

Notes:

Shipment in this Box Contains One DataMaster DMT with NO accessories, and one Shipping Memo.

Packed by: 10-12-11 **Witness by:** 10/10/11

Shrinkwrap YSP

Shaftsbury VSP

Breath Record Summary

X = Interference, R = Refusal, V = Invalid, I = Incomplete

Test Date	Test Time	Agency	Operator Name	Alcohol Val. 1	Alcohol Val. 2	Sim. Conc. 1	Sim. Conc. 2	Status Message
09/04/2010	18:55	SHAFTSBURY VSP	WAYNE M GODFREY	0.040		0.094		
09/11/2010	02:01	SHAFTSBURY VSP	JOHN-PAUL SCHMIDT	0.108		0.095		
09/19/2010	21:23	SHAFTSBURY VSP	WAYNE M GODFREY	0.101	0.097	0.095	0.094	
10/02/2010	08:48	BENNINGTON PD	JOSHUA M STEMP	0.064		0.094		
10/02/2010	23:28	SHAFTSBURY VSP	WAYNE M GODFREY	0.075	0.075	0.094	0.093	
10/14/2010	21:26	BENNINGTON COUNTY SO	JUSTIN R NILES	R		0.093		
10/15/2010	18:55	SHAFTSBURY VSP	JAMES A WRIGHT	R		0.094		
10/29/2010	00:54	SHAFTSBURY VSP	WAYNE M GODFREY	0.131	0.134	0.101	0.101	
10/30/2010	02:28	VSP	WAYNE M GODFREY	0.206		0.101		
10/31/2010	22:44	SHAFTSBURY VSP	TODD L WILKINS	R		0.101		
11/04/2010	13:33	SHAFTSBURY VSP	GARY A SHUHART			0.101		RESPONSE TIMEOUT
11/04/2010	13:39	SHAFTSBURY VSP	GARY A SHUHART	I		0.101		BLANK ERROR
11/04/2010	14:00	SHAFTSBURY VSP	GARY A SHUHART	0.138		0.100		
11/04/2010	14:30	SHAFTSBURY VSP	GA4RY A SHUHART	0.127		0.101		
11/11/2010	19:58	BENNINGTON COUNTY SO	JOEL R HOWARD	0.106	0.111	0.101	0.100	
11/25/2010	01:11	SHAFTSBURY VSP	JAMES A WRIGHT	0.200		0.099		
12/08/2010	22:21	VSP SHAFTSBURY	CHRIS W BURNETT	0.000		0.099		
12/18/2010	05:08	SHAFTSBURY VSP	ROBERT C ZINK	0.220		0.098		
01/01/2011	00:53	SHAFTSBURY VSP	ROBERT C ZINK	0.209		0.098		
01/05/2011	19:16	BENNINGTON COUNTY SO	JOEL R HOWARD	V		0.098		
01/05/2011	19:41	BENNINGTON COUNTY SO	JOEL R HOWARD	0.116		0.098		
01/06/2011	21:17	SHAFTSBURY VSP	ROBERT C ZINK	0.213		0.098		
01/15/2011	01:57	BENNINGTON PD	ANDREW W HUNT	0.106		0.097		
01/21/2011	02:38	BENNINGTON PD	ROSCOE E HARRINGTON	0.201		0.098		
01/21/2011	22:38	BENNINGTON PD	THALIA J HUDSON	0.177		0.097		
01/30/2011	03:28	BENNINGTON PD	ANDREW W HUNT	0.189	0.212	0.097	0.096	
01/31/2011	23:09	BENNINGTON PD	ANTHONY SILVESTRO	0.256	0.243	0.097	0.096	
02/05/2011	01:28	SHAFTSBURY VSP	JAMES A WRIGHT	0.121	0.118	0.096	0.096	
02/06/2011	01:06	BENNINGTON PD	THALIA J HUDSON	0.179		0.096		
02/07/2011	13:59	SHAFTSBURY VSP	GARY A SHUHART	0.190	0.201	0.098	0.097	
02/13/2011	02:14	BENNINGTON PD	JOSHUA M STEMP	0.187		0.101		
02/19/2011	00:06	SHAFTSBURY VSP	TRAVIS M HESS	0.173	0.176	0.101	0.100	
02/23/2011	20:38	SHAFTSBURY VSP	JESSE W ROBSON	0.073		0.101		
02/24/2011	19:19	VSP SHAFTSBURY	CHRIS W BURNETT	0.130		0.101		
03/05/2011	01:26	SHAFTSBURY VSP	WAYNE M GODFREY	0.204		0.099		
03/12/2011	19:06	SHAFTSBURY VSP	ROBERT C ZINK	0.134		0.100		
03/19/2011	14:26	SHAFTSBURY VSP	JESSE ROBSON	0.219	0.215	0.100	0.099	
03/26/2011	01:12	SHAFTSBURY VSP	CHRIS W BURNETT	0.142		0.099		
03/30/2011	18:27	SHAFTSBURY VSP	WAYNE M GODFREY	0.077		0.100		
04/01/2011	23:02	BENNINGTON COUNTY SO	PHILIP J FRASIER	0.075		0.100		
04/15/2011	02:43	SHAFTSBURY VSP	JOHN-PAUL SCHMIDT	0.089		0.099		
05/01/2011	15:46	BENNINGTON COUNTY SO	PHILIP J FRASIER	0.137		0.100		
07/14/2011	16:09	SHAFTSBURY VSP	NICK L MAYER	0.000		0.100		
07/21/2011	09:28	SHAFTSBURY VSP	NICK L MAYER	0.000		0.101		
07/24/2011	18:43	SHAFTSBURY VSP	TRAVIS M HESS	0.151	0.152	0.100	0.100	
07/31/2011	17:31	SHAFTSBURY VSP	ROBERT C ZINK	0.274		0.101		
08/28/2011	00:17	BENNINGTON COUNTY SO	JOEL R HOWARD	0.131		0.099		
09/03/2011	22:09	SHAFTSBURY VSP	JUSTIN W WALKER	0.069	0.069	0.099	0.099	
09/04/2011	00:54	SHAFTSBURY VSP	JOHN-PAUL SCHMIDT	0.131	0.127	0.099	0.099	
09/18/2011	00:08	SHAFTSBURY VSP	MICHAEL R TIETZ	0.197		0.099		

Status Record Summary

X = Interference, R = Refusal, V = Invalid, I = Incomplete

Test Date	Test Time	Test Type	Status Message
11/04/2010	13:33	Breath Test	RESPONSE TIMEOUT
11/04/2010	13:39	Breath Test	BLANK ERROR