

VERMONT FORENSIC LABORATORY

PO BOX 47
 WATERBURY, VERMONT 05676-0047
 TEL: 802-244-8788
 FAX: 802-241-5557
<http://vfl.vermont.gov>



DataMaster DMT Site Evaluation Checklist

Agency or Site Information

Name: St. Johnsbury PD

Address: 1187 Main Street
Suit 1

Phone: 802 748 2314

Fax: 802 748 1268

Personnel Information

Head of Site: Chief Hood

Current DataMaster Supervisor (s):
Lester Cleary
Mark Birkford

Proposed DataMaster Supervisor(s):

(Proposed) DataMaster DMT Environment

Limited Controlled Access:	<input checked="" type="radio"/> Y	N
Low Activity Area:	<input checked="" type="radio"/> Y	N
Unlimited Access by Other Law Enforcement Personnel	<input checked="" type="radio"/> Y	N
Access by VFL Technical Services	<input checked="" type="radio"/> Y	N
Air Conditioned	<input checked="" type="radio"/> Y	N
Room Temperature 65° - 85° F	<input checked="" type="radio"/> Y	N
Adequate Ventilation	<input checked="" type="radio"/> Y	N
Free of Cleaning Supplies	<input checked="" type="radio"/> Y	N
Clean	<input checked="" type="radio"/> Y	N

(Proposed) DataMaster DMT Connection(s)

Power:
 Min 15 Amp Y N

Limited low power sharing: Y N

Phone Line Connection Number: - - -

Phone Line Connection Type N/A

(Proposed) DataMaster DMT Site Maintenance Agreement

Will notify VFL Alcohol Program staff of maintenance and supply needs promptly: Y N

Will notify VFL Alcohol Program staff of changes in building conditions or moves promptly: Y N

Will notify VFL Alcohol Program staff of relevant personnel changes at site promptly: Y N

VFL Representative:	<u>Robert Driscoll</u>	<u>[Signature]</u>	<u>11/8/12</u>
	Print Name	Signature	Date
Agency Representative:	<u>Lester Cleary</u>	<u>[Signature]</u>	<u>11/8/12</u>
	Print Name	Signature	Date