VERMONT FORENSIC LABORATORY

PO BOX 47
WATERBURY, VERMONT 05676-0047
TEL: 802-244-8788
FAX: 802-241-5557
http://vfl.vermont.gov

DataMaster DMT Site Evaluation Checklist

Agency or Site Information

Name: Hardwick PD

Address: 56 High Street
Hardwick, VT 05843

Phone: 802 472 5175

Fax: 802 472 - 6865

Personnel Information

Head of Site: Aaron Cochran

Current DataMaster Supervisor (s):

Aaron Cochran

Michael Gero

Proposed DataMaster Supervisor(s):

(Proposed) DataMaster DMT Environment

Limited Controlled Access:	Y	N
Low Activity Area:	Y	N
Unlimited Access by Other Law Enforcement Personnel	Y	N
Access by VFL Technical Services	Y	N
Air Conditioned	Y	N
Room Temperature 65° - 85° F	Y	N
Adequate Ventilation	Y	N
Free of Cleaning Supplies	Y	N
Clean	Y	N

(Proposed) DataMaster DMT Connection(s)

Power: Min 15 Amp	Y	N
Limited low power sharing:	Y	N
Ethernet Available:	Y	N

Distance to Ethernet

(Proposed) DataMaster DMT Site Maintenance Agreement

Will notify VFL Alcohol Program staff of maintenance and supply needs promptly:	Y	N
Will notify VFL Alcohol Program staff of changes in building conditions or moves promptly:	Y	N
Will notify VFL Alcohol Program staff of relevant personnel changes at site promptly:	Y	N

	Print Name	Signature	Date
VFL Representative:	Robert Driscoll		1/4/13
Agency Representative:	Aaron Cochran		1/4/13