VERMONT FORENSIC LABORATORY

PO BOX 47
WATERBURY, VERMONT 05676-0047
TEL: 802-244-8788
FAX: 802-241-5557
http://vfl.vermont.gov

DATAMASTER TECHNICAL SUPPORT INQUIRY

Agency/Site: St. Johnsbury PD **DataMaster Serial #:** 103609

Date: 1/22/2013 Requestor: R. Driscoll

Instrument Supervisor: Lester Cleary Phone #: (802) 748-2314

Taken By: R. Driscoll

Problem: Simulator leaking during instrument move/installation on 1/9/2013

Laboratory Response: R. Driscoll will replace simulator in advance of February RPC

Work Performed: Replaced and assigned simulator (G10189) to DMT following the procedure outlined in P-ALC 203 using NIST traceable thermometer s/n: 111814522

☒ On Site ☐ In-house Attachments: 0 pages

Performed By: R. Driscoll RCD Date: 1/22/2013
Final Review: Date: 1/24/2013