Vermont Department of Health Laboratory to National Patent Analytical Systems DataMaster DMT Shipping Memo

Sim # G8969
NIST Temp = 34.1°C
[illegible] 12-14-11
[illegible] 12/14/11

To:
National Patent Analytical Systems
2090 Harrington Memorial Road
Mansfield, Ohio 44901

From:
State of Vermont Health Department Laboratory
195 Colchester Ave
Burlington, Vermont 05401

Date: 12/13/11

Re:
Shipment of DataMaster DMT S/N: 123306 **for the Vermont Certification Project.**

Instrument Preparation:

1. **Removed Simulator & secured simulator arms with Ty-Wrap.** [illegible] 12-14-11
2. **Removed breath tube.** [illegible] 12-14-11
3. **Downloaded instrument files to USB thumb drive.** [illegible] 12-13-11
4. **Removed keyboard & stylus.** [illegible] 12-14-11
5. **Removed security screws & replaced with manufacturer supplied screws.** [illegible] 12-14-11
6. **Instrument bagged & packaged for shipment.**
7. **How many TSIs are included with instrument?** 5

Notes:

Shipment in this Box Contains One DataMaster DMT with NO accessories, and one Shipping Memo.

Packed by: [illegible] 12-14-11 **Witness by:** [illegible] 12/14/11

Shelburne PD

Breath Record Summary

X = Interference, R = Refusal, V = Invalid, I = Incomplete

Shelburne PD.

Test Date	Test Time	Agency	Operator Name	Alcohol Val. 1	Alcohol Val. 2	Sim. Conc. 1	Sim. Conc. 2	Status Message
09/28/2010	12:46	VBVGU	VTGYVGY V VBHUGVBHU	0.099		0.099		
09/28/2010	12:51	FVTUO	FTY F FUQ	0.000		0.000		
09/28/2010	12:57	FTY8F8	DF759 D DR6FT7	0.376		0.000		
10/17/2010	02:39	SHELBURNE PD	BRUCE R BEUERLEIN	R		0.099		
10/21/2010	03:36	SHELBURNE PD	JOSH J FLORE	I	0.208	0.098	0.096	STANDARD OUT OF RANGE
10/29/2010	21:50	SHELBURNE PD	MATTHEW D TATRO	X	X	0.097	0.096	
10/29/2010	22:30	SHELBURNE PD	ALLEN A FORTIN	0.000		0.097		
10/29/2010	23:41	SHELBURNE PD	MICHAEL T THOMAS			0.097		RESPONSE TIMEOUT
10/29/2010	23:46	SHELBURNE PD	MICHAEL T THOMAS	0.138	0.139	0.098	0.096	
10/30/2010	02:20	SHELBURNE PD	ALLEN A FORTIN	0.157	0.157	0.097	0.097	
11/02/2010	21:03	HINESBURG PD	CHRIS M BATAILLE	0.128	0.125	0.096	0.096	
11/20/2010	04:16	SHELBURNE PD	ALLEN A FORTIN	0.140		0.096		
12/12/2010	01:10	SHELBURNE PD	MICHAEL T THOMAS	0.101		0.095		
12/12/2010	01:33	SHELBURNE PD	MICHAEL T THOMAS	0.093		0.096		
12/12/2010	04:04	SHELBURNE PD	JOSHUA J FLORE	0.251		0.095		
12/23/2010	18:20	NEW HAVEN VSP	PETER M DEMPSEY	0.099		0.097		
12/23/2010	20:40	HINESBURG PD	CHRIS M BATAILLE	0.090	0.093	0.097	0.097	
12/30/2010	01:11	SHELBURNE PD	BRUCE R BEUERLEIN	I	0.086	0.097		
01/01/2011	08:44	SHELBURNE PD	ALLEN A FORTIN	0.087	0.082	0.097	0.096	
01/07/2011	22:49	SHELBURNE PD	BRUCE R BEUERLEIN	0.207	0.209	0.095	0.095	
01/20/2011	10:21	HINESBURG PD	BRIAN K FOX			0.095		RESPONSE TIMEOUT
01/20/2011	11:02	HINESBURG PD	BRIAN K FOX	0.232	I	0.096	0.094	RESPONSE TIMEOUT
01/22/2011	01:11	SHELBURNE PD	BRUCE R BEUERLEIN	0.097	0.098	0.095	0.095	
01/24/2011	02:30	SOUTH BURLINGTON PD	D BRUNELLE	0.049		0.095		
01/28/2011	04:03	SHELBURNE PD	JOSHUA J FLORE	0.184		0.095	0.094	
02/01/2011	02:43	SHELBURNE PD	BRUCE R BEUERLEIN	0.034		0.095		
02/05/2011	03:18	SHELBURNE PD	JOSHUA J FLORE	0.092	0.089	0.094	0.094	
02/20/2011	02:58	SHELBURNE PD	FORTIN A ALLEN	R		0.099		
02/20/2011	03:07	SHELBURNE PD	FORTIN A ALLEN	0.067		0.099		RESPONSE TIMEOUT
03/05/2011	21:20	WILLISTON PD	STEPHANIE A TANNER	R		0.098		
03/06/2011	00:20	MILTON PD	LOCKE PAUL	0.189		0.099		
03/06/2011	02:59	CHITTENDEN COUNTY SO	MICHAEL E LONGEVIN	0.125		0.100		
03/06/2011	04:23	SHELBURNE PD	MICHAEL T THOMAS	0.191	0.194	0.099	0.098	
03/12/2011	03:58	SHELBURNE PD	BRUCE R BEUERLEIN	R		0.098		
03/12/2011	23:44	SHELBURNE PD	JOSHUA J FLORE	0.068		0.098		
03/13/2011	03:10	SHELBURNE PD	FORTIN A ALLEN	0.202		0.098		
03/18/2011	02:06	SHELBURNE PD	ALLEN A FORTIN			0.099		RESPONSE TIMEOUT
03/18/2011	02:23	SHELBURNE PD	ALLEN A FORTIN	V		0.099		
03/18/2011	02:45	SHELBURNE PD	ALLEN A FORTIN	0.136	0.130	0.099	0.098	
03/18/2011	22:04	SHELBURNE PD	JOSHUA J FLORE	0.151	0.152	0.097	0.097	
03/19/2011	19:22	SHELBURNE PD	BRUCE R BEUERLEIN	R		0.098		
03/23/2011	02:46	SHELBURNE PD	JAMES B MCKNIGHT			0.097		RESPONSE TIMEOUT
03/23/2011	03:11	SHELBURNE PD	JAMES B MCKNIGHT	0.149		0.097		
04/09/2011	01:45	SHELBURNE PD	JAMES B MCKNIGHT	0.106	0.118	0.098	0.096	
04/09/2011	23:16	HINESBURG	FOX K BRIAN	0.121		0.097		
04/10/2011	23:39	SHELBURNE PD	JON M MARCOUX	I	R	0.097		
04/11/2011	03:02	SHELBURNE	CHRIS O MORRELL	0.207		0.097		
04/16/2011	01:51	SHELBURNE PD	BRUCE R BEUERLEIN	0.120	0.118	0.096	0.096	
04/16/2011	02:30	SHELBURNE PD	ALLEN A FORTIN			0.097		RESPONSE TIMEOUT
04/16/2011	03:04	SHELBURNE PD	ALLEN A FORTIN	0.079		0.097		
04/21/2011	04:21	SHELBURNE PD	BRUCE R BEUERLEIN			0.096		RESPONSE TIMEOUT
04/21/2011	04:25	SHELBURNE PD	BRUCE R BEUERLEIN	0.149		0.097	0.096	
04/26/2011	03:26	SHELBURNE PD	JOSHUA J FLORE			0.096		RESPONSE TIMEOUT
04/26/2011	03:38	SHELBURNE PD	JOSHUA J FLORE	R		0.097		

Breath Record Summary

X = Interference, R = Refusal, V = Invalid, I = Incomplete

Test Date	Test Time	Agency	Operator Name	Alcohol Val. 1	Alcohol Val. 2	Sim. Conc. 1	Sim. Conc. 2	Status Message
05/01/2011	23:29	SHELBURNE PD	MATTHEW D TATRO	0.159	0.149	0.096	0.096	
05/14/2011	03:57	SHELBURNE PD	ALLEN A FORTIN	R		0.101		
05/14/2011	04:48	SHELBURNE PD	MICHAEL T THOMAS	0.227		0.100		CALIBRATION CHECK ERROR
05/15/2011	06:11	SHELBURNE PD	ALLEN A FORTIN	0.167	0.176	0.101	0.100	
05/28/2011	01:23	SOUTH BURLINGTON PD	MATTHEW J PLUNKETT	0.118		0.100		
05/28/2011	02:17	SOUTH BURLINGTON PD	D BRUNELLE	0.097		0.099		
05/31/2011	01:00	HINESBURG PD	CHRIS M BATAILLE	0.087	0.086	0.099	0.099	
06/04/2011	01:03	SHELBURNE PD	ALLEN A FORTIN	0.073		0.099		
06/04/2011	02:35	SOUTH BURLINGTON PD	D BRUNELLE	0.099		0.099		
06/11/2011	02:45	SHELBURNE PD	BRUCE R BEUERLEIN	0.144		0.097		
06/12/2011	21:21	HINESBURG PD	GEORGE D FON	0.252	0.262	0.097	0.095	
06/15/2011	00:20	SHELBURNE PD	MATTHEW D TATRO	0.079		0.096		
06/22/2011	01:11	SHELBURNE PD	BRUCE R BEUERLEIN	0.279		0.097		
06/26/2011	02:57	SHELBURNE PD	BRUCE R BEUERLEIN	0.164		0.096		
07/05/2011	20:34	SHELBURNE PD	BRUCE BEUERLEIN	R		0.096		
07/06/2011	03:15	SOUTH BURLINGTON PD	DAVID SOLOMON	0.144	0.149	0.095	0.094	
07/11/2011	00:55	SHELBURNE PD	JOSHUA J FLORE	0.138		0.096	0.094	
07/16/2011	04:02	SHELBURNE PD	ALLEN A FORTIN	0.110	0.104	0.094	0.095	
07/24/2011	02:25	WILLISTON PD	STEPHANIE A TANNER	I	I	0.100		RESPONSE TIMEOUT
07/24/2011	03:30	SHELBURNE PD	ALLEN A FORTIN	0.096	0.094	0.099	0.098	
07/26/2011	23:01	SHELBURNE PD	BRUCE R BEUERLEIN	0.112		0.099		
07/29/2011	19:49	HINESBURG PD	CHRIS M BATAILLE			0.098		RESPONSE TIMEOUT
07/29/2011	19:54	HINESBURG PD	CHRIS M BATAILLE	0.353		0.099		
07/31/2011	01:18	SHELBURNE PD	BRUCE R BEUERLEIN	0.093	0.084	0.098	0.098	
08/16/2011	22:02	SHELBURNE PD	CHRIS O MORRELL	0.143	0.142	0.096	0.096	
08/17/2011	02:40	SHELBURNE PD	BRUCE R BEUERLEIN	0.232	0.234	0.097	0.096	
08/19/2011	00:44	SOUTH BURLINGTON PD	MATTHEW J PLUNKETT	0.172		0.095		
08/20/2011	03:33	SOUTH BURLINGTON PD	BRIANNE E WILLIAMSON	0.184		0.097		
08/21/2011	03:10	SHELBURNE PD	JOSHUA J FLORE	0.176		0.097		
08/21/2011	06:14	SHELBURNE PD	ALLEN A FORTIN			0.097		RESPONSE TIMEOUT
08/21/2011	06:34	SHELBURNE PD	ALLEN A FORTIN	0.077		0.097		
08/27/2011	03:46	SOUTH BURLINGTON PD	BRIANNE E WILLIAMSON	R		0.097		
08/27/2011	04:08	SHELBURNE PD	JASON LAWTON	0.083	0.084	0.097	0.097	
09/02/2011	21:32	HINESBURG PD	CHRIS M BATAILLE	V		0.097		
09/02/2011	21:40	HINESBURG PD	CHRIS M BATAILLE	V		0.097		
09/02/2011	21:49	HINESBURG PD	CHRIS M BATAILLE	V		0.098		
09/04/2011	00:56	SPRINGFIELD PD	MIKE T THOMAS	0.000		0.096		
09/10/2011	02:27	SOUTH BURLINGTON PD	MATTHEW PLUNKETT	R		0.095		
09/10/2011	06:21	SHELBURNE PD	JOSHUA J FLORE	0.118		0.096		
09/11/2011	03:51	SOUTH BURLINGTON PD	MATTHEW J PLUNKETT	0.112		0.095		
09/22/2011	17:01	SHELBURNE PD	JONATHAN M MARCOUX			0.093		STANDARD OUT OF RANGE
09/30/2011	00:48	SHELBURNE PD	BRUCE R BEUERLEIN	0.156	0.164	0.097	0.098	
09/30/2011	17:59	SHELBURNE PD	CHRIS O MORRELL	I	I	0.098	0.096	RESPONSE TIMEOUT
09/30/2011	18:18	SHELBURNE PD	CHRIS O MORRELL	0.218		0.099	0.096	RESPONSE TIMEOUT
10/01/2011	04:15	SHELBURNE PD	MICHAEL T THOMAS	R		0.098		
10/02/2011	03:41	SOUTH BURLINGTON PD	DAVID SOLOMON	R		0.099		
10/02/2011	04:00	SOUTH BURLINGTON	DAVID SOLOMON	0.123		0.098		
10/12/2011	15:41	SHELBURNE PD	JOSHUA J FLORE	0.143		0.099		
10/13/2011	23:55	SHELBURNE PD	BRUCE R BEUERLEIN	R		0.099		
10/16/2011	05:32	SHELBURNE PD	JASON LAWTON	0.046		0.098		
10/22/2011	02:34	SHELBURNE PD	JASON LAWTON	I	0.127	0.098		
10/29/2011	03:03	SOUTH BURLINGTON PD	DAVID SOLOMON			0.096		STANDARD OUT OF RANGE
11/07/2011	00:42	SHELBURNE PD	JAMES B MCKNIGHT	0.129		0.099	0.096	STANDARD OUT OF RANGE
11/10/2011	02:13	SHELBURNE PD	MICHAEL T THOMAS	0.184		0.099		

Breath Record Summary

X = Interference, R = Refusal, V = Invalid, I = Incomplete

Test Date	Test Time	Agency	Operator Name	Alcohol Val. 1	Alcohol Val. 2	Sim. Conc. 1	Sim. Conc. 2	Status Message
11/11/2011	22:25	MILTON PD	JASON E PORTER	R		0.098		
11/11/2011	22:44	MILTON PD	JASON E PORTER	0.089		0.099		
11/13/2011	03:32	SOUTH BURLINGTON PD	DAVID SOLOMON	0.189		0.099		
11/14/2011	04:29	SOUTH BURLINGTON PD	CHRIS M BATAILLE	0.162		0.098		
11/19/2011	22:25	SHELBURNE PD	BRUCE R BEUERLEIN	0.104		0.097		
11/24/2011	02:15	SHELBURNE PD	JONATHAN M MARCOUX	0.213		0.098		
11/24/2011	03:51	SHELBURNE PD	BRUCE R BEUERLEIN	0.170	0.170	0.098	0.097	
11/24/2011	04:36	SHELBURNE PD	JOSHUA J FLORE	0.154		0.097		
11/26/2011	05:29	SHELBURNE PD	MICHAEL T THOMAS	0.123	0.125	0.098	0.097	
11/27/2011	03:08	SHELBURNE PD	MICHAEL T THOMAS	0.149	0.159	0.098	0.097	
12/02/2011	23:59	SHELBURNE PD	ALLEN A FORTIN	0.083	0.082	0.097	0.097	
12/11/2011	02:53	NEW HAVEN VSP	PETER M DEMPSEY			0.096		STANDARD OUT OF RANGE
12/11/2011	03:02	SHELBURNE PD	MIKE T THOMAS			0.096		STANDARD OUT OF RANGE

Shelburne PD

Status Record Summary

X = Interference, R = Refusal, V = Invalid, I = Incomplete

Test Date	Test Time	Test Type	Status Message
10/21/2010	03:36	Breath Test	STANDARD OUT OF RANGE
10/29/2010	23:41	Breath Test	RESPONSE TIMEOUT
01/20/2011	10:21	Breath Test	RESPONSE TIMEOUT
01/20/2011	11:02	Breath Test	RESPONSE TIMEOUT
02/20/2011	03:07	Breath Test	RESPONSE TIMEOUT
03/18/2011	02:06	Breath Test	RESPONSE TIMEOUT
03/23/2011	02:46	Breath Test	RESPONSE TIMEOUT
04/16/2011	02:30	Breath Test	RESPONSE TIMEOUT
04/21/2011	04:21	Breath Test	RESPONSE TIMEOUT
04/26/2011	03:26	Breath Test	RESPONSE TIMEOUT
05/14/2011	04:48	Breath Test	CALIBRATION CHECK ERROR
07/24/2011	02:25	Breath Test	RESPONSE TIMEOUT
07/29/2011	19:49	Breath Test	RESPONSE TIMEOUT
08/21/2011	06:14	Breath Test	RESPONSE TIMEOUT
09/22/2011	17:01	Breath Test	STANDARD OUT OF RANGE
09/30/2011	17:59	Breath Test	RESPONSE TIMEOUT
09/30/2011	18:18	Breath Test	RESPONSE TIMEOUT
10/29/2011	03:03	Breath Test	STANDARD OUT OF RANGE
11/07/2011	00:42	Breath Test	STANDARD OUT OF RANGE
12/11/2011	02:53	Breath Test	STANDARD OUT OF RANGE