

VERMONT FORENSIC LABORATORY

PO BOX 47
 WATERBURY, VERMONT 05676-0047
 TEL: 802-244-8788
 FAX: 802-241-5557
<http://vfl.vermont.gov>



DataMaster DMT Site Evaluation Checklist

Agency or Site Information

Name: VSP-Williston
 Address: 2777 St. George Rd
Williston, VT 05495
 Phone: 802-878-7111
 Fax: " " - 2742

Personnel Information

Head of Site: LT. Marc Thomas
 Current DataMaster Supervisor (s):
SGT. Greg Campbell
S/T PAUL RAVELIN
S/T TODD AMBROZ
 Proposed DataMaster Supervisor(s):

(Proposed) DataMaster DMT Environment

Limited Controlled Access:	<input checked="" type="radio"/>	N
Low Activity Area:	<input checked="" type="radio"/>	N
Unlimited Access by Other Law Enforcement Personnel	<input checked="" type="radio"/>	N
Access by VFL Technical Services	<input checked="" type="radio"/>	N
Air Conditioned	<input checked="" type="radio"/>	N
Room Temperature 65° - 85° F	<input checked="" type="radio"/>	N
Adequate Ventilation	<input checked="" type="radio"/>	N
Free of Cleaning Supplies	<input checked="" type="radio"/>	N
Clean	<input checked="" type="radio"/>	N

(Proposed) DataMaster DMT Connection(s)

Power:
 Min 15 Amp N
 Limited low power sharing: N
 Phone Line Connection Number: _____ - _____ - _____
 Phone Line Connection Type _____

(Proposed) DataMaster DMT Site Maintenance Agreement

Will notify VFL Alcohol Program staff of maintenance and supply needs promptly: N
 Will notify VFL Alcohol Program staff of changes in building conditions or moves promptly: N
 Will notify VFL Alcohol Program staff of relevant personnel changes at site promptly: N

VFL Representative:	<u>Amanda Bolduc</u> Print Name	<u>[Signature]</u> Signature	<u>5/31/12</u> Date
Agency Representative:	<u>Todd Ambroz</u> Print Name	<u>[Signature]</u> Signature	<u>5/31/12</u> Date