VERMONT FORENSIC LABORATORY

PO BOX 47
WATERBURY, VERMONT 05676-0047
TEL: 802-244-8788
FAX: 802-241-5557
http://vfl.vermont.gov

DataMaster DMT Site Evaluation Checklist

Agency or Site Information

Name: Colchester Police Department

Address: 480 Hercules Drive

Colchester, VT

Phone: (802) 264 5555

Fax: (802) 264 5550

Personnel Information

Head of Site: Charles Kirker

Current DataMaster Supervisor (s):

James Roy, Jack Lehneman, Chris Jones

Proposed DataMaster Supervisor(s):

N/A

(Proposed) DataMaster DMT Environment

Limited Controlled Access:	Y	N
Low Activity Area:	Y	N
Unlimited Access by Other Law Enforcement Personnel	Y	N
Access by VFL Technical Services	Y	N
Air Conditioned	Y	N
Room Temperature 65° - 85° F	Y	N
Adequate Ventilation	Y	N
Free of Cleaning Supplies	Y	N
Clean	Y	N

(Proposed) DataMaster DMT Connection(s)

Power: Min 15 Amp	Y	N
Limited low power sharing:	Y	N

Phone Line Connection Number: 802-_____-_____

Phone Line Connection Type _____

(Proposed) DataMaster DMT Site Maintenance Agreement

Will notify VFL Alcohol Program staff of maintenance and supply needs promptly:	Y	N
Will notify VFL Alcohol Program staff of changes in building conditions or moves promptly:	Y	N
Will notify VFL Alcohol Program staff of relevant personnel changes at site promptly:	Y	N

	Print Name	Signature	Date
VFL Representative:	Steven Harnois		03/08/2012
Agency Representative:	DOUGLAS ALLEN		3/8/2012