DataMaster Monthly Update Check-list

Department/Agency: Winooski DM serial #: 104409

Date: 9/30/14 August 14 / September

	Provided	None this month
Operators Logbook	☒	☐
Maintenance Logbook	☒	☐
Breath Summary Report	☒	☐
Status Summary Report	☐	☐
Calibration	☐	☐
Certification	☐	☐
Installation	☐	☐
Annual Preventative Maintenance	☐	☐
Routine Performance Check	☐	☐
Simulator Solution Change	☐	☐
Other Reports	☐	☐

List other report type and date:

DATAMASTER DMT OPERATOR USE LOG

WINDOSKI
Department/Agency

DataMaster DMT Serial Number

Date	DMT Operator's Name	Subject's Name	Simulator Vapor 1	Simulator Vapor 2
06/01/14	DAVIS	EDIS MUSONAVIC	.099	.095
6/6/14	GENEST	HARKA KARKI	.096	.096
6/11/14	Adams	Glen Critchlow	.098	.099
6/22/14	Adams	Nicholas Lawrence	.097	.095
6/25/14	Adams	Grady Shea	.099	.099
7/1/14	R. Hebert	RPC	.099	
7/27/14	J. TREIER	MONICA CULINA	.096	.095
7/27/14	Borthwick	Mary Carton	.096	.095 recomp test .099
8-1-14	Chiaravalle	Gary Provost	.097	
8-11-14	Hebert	TEST	.096	
8/12/14	R. Driscoll	TEST	.096	—
8-30-14	Chiaravalle	Kristen KAMALSKY	.095	.095
9-1-14	Chiaravalle	Michael Schaden	.095	
9-22-14	CRAM	Ben Rodgers	.093	
9-27-14	"	"	.094	

ALC 603 Rev 3 08/11

DATAMASTER DMT MAINTENANCE LOG

Winooski PD
Department/Agency

104409
DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
5/31/14	DataMaster Simulator Solution Lot # 13-16-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 [For RPC or S Preparation Date: 9/26/2013 Expires: 09/26/2014 bel Here] Solution Change OPENED: 5/31/14 RMH	R Hebert 904
7/1/14	DataMaster Simulator Solution Lot # 13-16-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 [For RPC Preparation Date: 9/26/2013 Expires: 09/26/2014 e Label Here] OPENED: 7/1/14 RMH RPC	R Hebert 904
8/12/14	Performed check-in for passing [For RPC or Simulator Solution Change Place Label Here] sample acceptance.	R. Driscoll DMT TECH
	[For RPC or Simulator Solution Change Place Label Here]	

ALC 803 REV 03 08/2011

Breath Record History

Test Date	Test Time	Agency	Operator Name	Alcohol Val. 1	Alcohol Val. 2	Sim. Conc. 1	Sim. Conc. 2	Status Message
09/30/2013	01:37:58	CHITTENDEN COUNTY SO	BENJAMIN C CHIARAVALLE	0.137	0.133	0.097	0.096	
10/30/2013	23:47:15	WINOOSKI PD	RICHARD M HEBERT			0.098		RF Detected
11/22/2013	03:11:48	UVM PD	ROBERT M BAILEY	0.124		0.096		
11/23/2013	21:12:07	WINOOSKI PD	JAMES V CHARKALIS			0.095		Simulator Out of Range
11/23/2013	21:15:49	WINOOSKI PD	JAMES V CHARKALIS			0.094		Simulator Out of Range
12/04/2013	19:03:32	winooski PD	RAINER J SHEEHAN			0.094		Simulator Out of Range
12/18/2013	23:24:39	WINOOSKI PD	MICHAEL E CRAM	0.184	0.180	0.097	0.096	
01/01/2014	00:44:44	CHITTENDEN COUNTY SO	BENJAMIN C CHIARAVALLE	0.114		0.096	0.096	
01/05/2014	21:17:22	BURLINGTON PD	DWAYNE A MELLIS	0.144		0.095		
01/14/2014	18:21:19	WINOOSKI PD	MICHAEL E CRAM	0.153		0.096		
01/18/2014	02:32:07	sbpd	D BRUNELLE	0.122		0.096		
02/01/2014	00:27:30	CHITTENDEN COUNTY SO	BENJAMIN C CHIARAVALLE			0.093		Simulator Out of Range
02/01/2014	00:31:05	CHITTENDEN COUNTY SO	BENJAMIN C CHIARAVALLE	REFUSED		0.095		
02/01/2014	16:36:25	WINOOSKI PD	MICHAEL E CRAM	REFUSED		0.095		
02/03/2014	01:29:32	ADDISON COUNTY SO	OSCAR J GARDNER	0.078		0.094		RESPONSE TIMEOUT
02/19/2014	15:15:57	WINOOSKI PD	MACHAVERN CHRIS			0.093		Simulator Out of Range
03/01/2014	21:08:04	WINOOSKI PD	MICHAEL E CRAM	0.079		0.096		
03/02/2014	08:49:32	winooski	BRENDA J DAVIS	0.093		0.096		
03/02/2014	23:31:06	winooski	RAINER J SHEEHAN	0.138		0.096		
03/03/2014	13:15:55	WINOOSKI PD	CHRIS MACHAVERN	0.128		0.096		
03/04/2014	10:27:37	WINOOSKI PD	CHRIS MACHAVERN	0.051		0.096		
03/10/2014	10:03:30	WINOOSKI PD	CHRIS MACHAVERN			0.093		Simulator Out of Range
03/10/2014	10:43:25	WINOOSKI PD	RICHARD M HEBERT	0.000		0.096		
03/11/2014	10:54:04	WINOOSKI PD	SCOTT D MCGIVERN	0.046		0.096		
03/15/2014	04:57:39	WINOOSKI PD	BENJAMIN J ADAMS	0.186		0.095		
03/28/2014	22:28:22	CHITTENDEN COUNTY SO	BENJAMIN C CHIARAVALLE	0.211		0.095		
04/18/2014	01:27:43	WINOOSKI PD	MICHAEL E CRAM	0.147		0.094		
04/20/2014	02:40:21	CHITTENDEN COUNTY SO	BENJAMIN C CHIARAVALLE			0.093		Simulator Out of Range
04/20/2014	02:43:55	CHITTENDEN COUNTY SO	BENJAMIN C CHIARAVALLE	0.089	0.086	0.094	0.094	
04/25/2014	00:21:09	WINOOSKI PD	MICHAEL E CRAM	0.146		0.094		
05/03/2014	03:51:48	WINOOSKI PD	MICHAEL E CRAM			0.093		Simulator Out of Range
05/12/2014	20:23:01	SOUTH BURLINGTON PD	CHRIS M BATAILLE	0.142		0.094	0.093	Simulator Out of Range
05/12/2014	20:29:59	SOUTH BURLINGTON PD	CHRIS M BATAILLE			0.093		Simulator Out of Range
05/24/2014	09:29:34	WINOOSKI PD	WILLIAM J ELOVIRTA			0.093		Simulator Out of Range
05/24/2014	09:34:23	WINOOSKI PD	WILLIAM J ELOVIRTA	0.112		0.094		
05/25/2014	12:09:17	Colchester	KEITH A SCHAFFER			0.092		Simulator Out of Range
06/01/2014	04:33:45	winooski	BRENDA J DAVIS	INCOMPLETE	INCOMPLETE	0.098	0.097	
06/05/2014	19:28:30	WINOOSKI PD	JAMES V CHARKALIS	REFUSED		0.097		
06/06/2014	00:06:10	WINOOSKI PD	SKYLER W GENEST	INCOMPLETE	0.136	0.096	0.096	
06/11/2014	22:00:13	WINOOSKI PD	BENJAMIN ADAMS	0.199	0.196	0.098	0.096	
06/22/2014	01:18:01	WINOOSKI PD	BENJAMIN ADAMS	0.122	0.123	0.097	0.095	
06/25/2014	03:38:55	WINOOSKI PD	BENJAMIN J ADAMS	0.134	0.145	0.098	0.098	
07/01/2014	15:55:42	WINOOSKI PD	RICHARD HEBERT	0.090		0.098		
07/27/2014	06:19:35	COLCHESTER PD	JESSE L TREIER	0.127	0.125	0.096	0.095	
07/27/2014	21:35:11	WINOOSKI PD	ROBERT D BORTHWICK	0.171	INCOMPLETE	0.096	0.095	
08/01/2014	22:19:06	CHITTENDEN COUNTY SO	BENJAMIN C CHIARAVALLE	0.055		0.097		
08/11/2014	13:31:18	WINOOSKI PD	RICHARD M HEBERT	0.000		0.096		
08/12/2014	06:45:09	VFL	ROBERT DRISCOLL	0.000		0.096		
08/30/2014	03:07:52	CHITTENDEN COUNTY SO	BENJAMIN C CHIARAVALLE	0.145	0.155	0.095	0.095	
09/01/2014	01:19:24	CHITTENDEN COUNTY SO	BENJAMIN C CHIARAVALLE	REFUSED		0.096		
09/27/2014	04:03:15	WINOOSKI PD	MICHAEL E CRAM			0.093		Simulator Out of Range
09/27/2014	04:07:46	WINOOSKI PD	MICHAEL E CRAM	0.207		0.094		