

VERMONT FORENSIC LABORATORY

PO BOX 47
 WATERBURY, VERMONT 05676-0047
 TEL: 802-244-8788
 FAX: 802-241-5557
<http://vfl.vermont.gov>



DataMaster DMT Site Evaluation Checklist

Agency or Site Information

Name: Essex PD

Address: 145 Maple Street
Essex Junction, VT 05452

Phone: (802) 878-8331

Fax: (802) 878-1340

Personnel Information

Head of Site: Brad Larose

Current DataMaster Supervisor (s):
Rob Kissinger
Steve Dunning, John Dunn

Proposed DataMaster Supervisor(s):

(Proposed) DataMaster DMT Environment

Limited Controlled Access:	<input checked="" type="radio"/>	N
Low Activity Area:	<input checked="" type="radio"/>	N
Unlimited Access by Other Law Enforcement Personnel	<input checked="" type="radio"/>	N
Access by VFL Technical Services	<input checked="" type="radio"/>	N
Air Conditioned	<input checked="" type="radio"/>	N
Room Temperature 65° - 85° F	<input checked="" type="radio"/>	N
Adequate Ventilation	<input checked="" type="radio"/>	N
Free of Cleaning Supplies	<input checked="" type="radio"/>	N
Clean	<input checked="" type="radio"/>	N

(Proposed) DataMaster DMT Connection(s)

Power:		
Min 15 Amp	<input checked="" type="radio"/>	N
Limited low power sharing:	<input checked="" type="radio"/>	N
Ethernet Available:	<input checked="" type="radio"/>	N
Distance to Ethernet	<u>2'</u>	

(Proposed) DataMaster DMT Site Maintenance Agreement

Will notify VFL Alcohol Program staff of maintenance and supply needs promptly:	<input checked="" type="radio"/>	N
Will notify VFL Alcohol Program staff of changes in building conditions or moves promptly:	<input checked="" type="radio"/>	N
Will notify VFL Alcohol Program staff of relevant personnel changes at site promptly:	<input checked="" type="radio"/>	N

VFL Representative:	<u>Robert Darsoll</u>	<u>[Signature]</u>	<u>10/13/14</u>
	Print Name	Signature	Date
Agency Representative:	<u>Robert Kissinger</u>	<u>[Signature]</u>	<u>10-13-14</u>
	Print Name	Signature	Date