2012 DataMaster DMT APM and Software Update

Date	6/26/2012
Agency	CANAAN PD
DMT Serial Number	100170
Technician Name	TLC
Software Updated	YES → TLC 6/26/12 NO
Lock Changed	YES NO
DPS Property Tag Number	20173
APM Protocol Performed	YES → TLC 6/26/12 NO
Other/Notes	↳ SIMULATOR CHECKED: PRESSURE OK O-RINGS CHANGED * APM PROTOCOL NOT PERFORMED → NO HAND HELD RADIO *

thermometer s/n 111814522

NOTE: LOGGED ON AS KIRK KIMBALL SINCE OLD VERSION OF SOFTWARE WAS ON INSTRUMENT

A. Bolden
6/28/12