Vermont Department of Health Laboratory to National Patent Analytical Systems DataMaster DMT Shipping Memo

Sim# G11308
NIST Temp = 34.0°C
[illegible] 1-30-12
[illegible] 01/30/12

To:
National Patent Analytical Systems
2090 Harrington Memorial Road
Mansfield, Ohio 44901

From:
State of Vermont Health Department Laboratory
195 Colchester Ave
Burlington, Vermont 05401

Date: 1-30-12

Re:
Shipment of DataMaster DMT S/N: 103006 **for the Vermont Certification Project.**

Instrument Preparation:

1. **Removed Simulator & secured simulator arms with Ty-Wrap.**
2. **Removed breath tube.** [illegible] 1-30-12
3. **Downloaded instrument files to USB thumb drive.** [illegible] 1-30-12
4. **Removed keyboard & stylus.** [illegible] 1-30-12
5. **Removed security screws & replaced with manufacturer supplied screws.** [illegible] 1-30-12
6. **Instrument bagged & packaged for shipment.**
7. **How many TSIs are included with instrument?** 4

Notes:

Shipment in this Box Contains One DataMaster DMT with NO accessories, and one Shipping Memo.

Packed by: [illegible] 1-30-12 **Witness by:** [illegible] 01/30/12

ST Albans PD / Swanton PD

Swanton PD / ST A VSP

Breath Record Summary

X = Interference, R = Refusal, V = Invalid, I = Incomplete

Test Date	Test Time	Agency	Operator Name	Alcohol Val. 1	Alcohol Val. 2	Sim. Conc. 1	Sim. Conc. 2	Status Message
09/08/2010	07:03	ADDISON COUNTY SO	FU G YUFGYU			0.130		RESPONSE TIMEOUT
09/08/2010	07:07	ADDISON COUNTY SO	FU G YUFGYU	0.129	0.130	0.131	0.131	
09/08/2010	07:17	CGUK	BHJK V VGJK	0.000	0.000			
09/21/2010	23:33	GRAND ISLE COUNTY SO	BLAKE R ALLEN	0.118	0.119	0.103	0.102	
10/09/2010	23:24	SWANTON PD	ALEXIS R DUBOIS II	I	R	0.102		
10/09/2010	23:51	SWANTON PD	ALEXIS R DUBOIS II	I	0.113	0.102	0.102	RFI DETECTED
10/10/2010	00:05	SWANTON PD	ALEXIS R DUBOIS II	R		0.101		
10/31/2010	21:32	SWANTON PD	DEAN M SCOTT	0.206		0.100		
11/20/2010	01:22	SWANTON PD	DONALD M GILBERT	0.111		0.099		
03/31/2011	16:46	SWANTON PD	DONALD M GILBERT	0.000		0.096		
04/03/2011	20:24	SWANTON PD	DEAN M SCOTT	0.137		0.100	0.100	RESPONSE TIMEOUT
05/05/2011	21:05	GRAND ISLE COUNTY SO	DUSTIN J ABELL	0.237		0.098		
05/16/2011	00:14	SWANTON PD	DEAN M SCOTT	R		0.097		
05/16/2011	00:33	SWANTON PD	DEAN M SCOTT	0.188	0.193	0.097	0.096	
06/03/2011	22:24	SWANTON PD	CHAD B PARAH	0.143		0.096		
06/19/2011	02:13	SWANTON PD	MICHAEL S AHER	0.090	0.097	0.095	0.096	
06/29/2011	16:01	SWANTON PD	DONALD M GILBERT	0.000		0.105		
07/09/2011	23:56	SWANTON PD	CHAD B PARAH	0.071		0.104		
07/24/2011	19:18	SWANTON PD	DEAN M SCOTT	0.114	0.112	0.103	0.102	
08/26/2011	00:52	SWANTON PD	CHAD B PARAH	0.104		0.101		
09/15/2011	14:49	SWANTON PD	EUGENE G RICH	0.264		0.100		
09/23/2011	10:25	VDHL	STEVEN E HARNOIS	0.000		0.103		
09/24/2011	01:14	ST. ALBANS VSP	JACOB A METAYER	0.226		0.103		
10/06/2011	21:58	SWANTON PD	CHAD B PARAH	0.091		0.103		
10/15/2011	00:07	ST. ALBANS VSP	JACOB A METAYER	0.115	0.112	0.102	0.102	
10/19/2011	19:16	ST. ALBANS VSP	RICHARD C STEPIEN	0.125		0.102		
10/22/2011	00:30	SWANTON PD	CHAD B PARAH	I		0.103		RESPONSE TIMEOUT
10/23/2011	02:39	ST. ALBANS VSP	D ROLANDINI	0.198	0.192	0.102	0.101	
11/01/2011	23:40	ST. ALBANS VSP	JACOB A METAYER	0.128	0.127	0.101	0.101	
11/11/2011	16:53	ST. ALBANS VSP	MATTHEW HILL	0.074		0.101		RESPONSE TIMEOUT
11/15/2011	21:48	ST. ALBANS VSP	ASHLEY A FARMER	R		0.100		
11/20/2011	01:09	ST. ALBANS VSP	JAY RIGGEN	I	R	0.100		
11/20/2011	01:31	ST. ALBANS VSP	JAY RIGGEN	0.204	0.212	0.100	0.099	
11/23/2011	19:48	ST. ALBANS VSP	MICHAEL R ANDERSON	0.165		0.099		
11/26/2011	10:14	ST. ALBANS VSP	MATTHEW R JOHNSON	0.087		0.099		RESPONSE TIMEOUT
12/03/2011	23:38	ST. ALBANS VSP	RICHARD C STEPIEN	0.283	0.257	0.098	0.098	
12/08/2011	19:19	ST. ALBANS VSP	D ROLANDINI	R		0.098		
12/11/2011	04:33	ST. ALBANS VSP	JACOB A METAYER	R		0.098		
12/13/2011	18:38	ST. ALBANS VSP	JAY RIGGEN	I	I	0.099	0.099	
12/13/2011	21:23	ST. ALBANS VSP	D ROLANDINI	0.159		0.099		
12/19/2011	23:49	FRANKLIN COUNTY SO	ANDREW G UNDERWOOD	0.139	0.142	0.098	0.099	
12/21/2011	21:58	ST. ALBANS PD	WILLIAM R FURNARI	0.157		0.098		
12/22/2011	01:57	ST. ALBANS PD	SEAN E POPE	R		0.099		
12/22/2011	21:38	ST. ALBANS PD	PAUL A TALLEY	0.140	0.149	0.099	0.098	
12/23/2011	02:00	ST. ALBANS PD	PAUL MORITS	0.226		0.098		
12/28/2011	12:22	ST. ALBANS VSP	MATTHEW R JOHNSON	0.278		0.098		
12/30/2011	01:09	ST. ALBANS PD	SEAN E POPE	R		0.097		
12/31/2011	02:02	WILLISTON VSP	ANGELA S BAKER	0.110	0.109	0.098	0.098	
01/05/2012	23:40	ST. ALBANS VSP	MATTHEW HILL	0.124		0.097		
01/12/2012	00:19	SWANTON PD	CHAD B PARAH	0.123		0.097		
01/15/2012	09:16	ST. ALBANS VSP	JOHN L BRUZZI	V		0.097		
01/15/2012	09:36	ST. ALBANS VSP	JOHN L BRUZZI	0.147	0.132	0.098	0.097	
01/21/2012	03:46	ST. ALBANS PD	SEAN E POPE	0.175		0.097		
01/22/2012	20:07	ST. ALBANS VSP	MATTHEW JOHNSON			0.096		RESPONSE TIMEOUT

Breath Record Summary

X = Interference, R = Refusal, V = Invalid, I = Incomplete

Test Date	Test Time	Agency	Operator Name	Alcohol Val. 1	Alcohol Val. 2	Sim. Conc. 1	Sim. Conc. 2	Status Message
01/27/2012	01:49	ST. ALBANS VSP	JACOB A METAYER	0.048		0.097		

Swanton PD / St Albans VSP

Status Record Summary

X = Interference, R = Refusal, V = Invalid, I = Incomplete

Test Date	Test Time	Test Type	Status Message
09/08/2010	07:03	Breath Test	RESPONSE TIMEOUT
10/09/2010	23:51	Breath Test	RFI DETECTED
04/03/2011	20:24	Breath Test	RESPONSE TIMEOUT
10/22/2011	00:30	Breath Test	RESPONSE TIMEOUT
11/11/2011	16:53	Breath Test	RESPONSE TIMEOUT
11/26/2011	10:14	Breath Test	RESPONSE TIMEOUT
01/22/2012	20:07	Breath Test	RESPONSE TIMEOUT