Vermont Department of Health Laboratory to National Patent Analytical Systems DataMaster DMT Shipping Memo

Sim # G7452
NIST Temp: 34.2°C
KLK 1-17-12
SH 01/17/12

To:
National Patent Analytical Systems
2090 Harrington Memorial Road
Mansfield, Ohio 44901

From:
State of Vermont Health Department Laboratory
195 Colchester Ave
Burlington, Vermont 05401

Date: 1-17-12

Re:
Shipment of DataMaster DMT S/N: 103206 **for the Vermont Certification Project.**

Instrument Preparation:

1. **Removed Simulator & secured simulator arms with Ty-Wrap.** KLK 1-18-12
2. **Removed breath tube.** KLK 1-18-12
3. **Downloaded instrument files to USB thumb drive.** KLK 1-17-12
4. **Removed keyboard & stylus.** KLK 1-18-12
5. **Removed security screws & replaced with manufacturer supplied screws.** KLK 1-18-12
6. **Instrument bagged & packaged for shipment.** KLK 1-18-12
7. **How many TSIs are included with instrument?** 8

Notes:

Shipment in this Box Contains One DataMaster DMT with NO accessories, and one Shipping Memo.

Packed by: KLK 1-18-12 **Witness by:** SH 01/18/12

St Albans PD

Status Record Summary

ST Albans PD

X = Interference, R = Refusal, V = Invalid, I = Incomplete

Test Date	Test Time	Test Type	Status Message
10/15/2010	09:46	Breath Test	RESPONSE TIMEOUT
10/24/2010	05:18	Breath Test	INTERFERENCE DETECTED
11/14/2010	04:26	Breath Test	INTERFERENCE DETECTED
11/14/2010	04:29	Breath Test	INTERFERENCE DETECTED
11/14/2010	04:31	Breath Test	INTERFERENCE DETECTED
11/14/2010	04:36	Breath Test	INTERFERENCE DETECTED
11/17/2010	01:48	Breath Test	STANDARD OUT OF RANGE
11/19/2010	01:01	Breath Test	INTERFERENCE DETECTED
11/19/2010	01:05	Breath Test	STANDARD OUT OF RANGE
11/19/2010	01:15	Breath Test	RESPONSE TIMEOUT
11/19/2010	01:19	Breath Test	INTERFERENCE DETECTED
11/19/2010	01:22	Breath Test	INTERFERENCE DETECTED
11/19/2010	01:26	Breath Test	INTERFERENCE DETECTED
11/19/2010	01:29	Breath Test	INTERFERENCE DETECTED
11/19/2010	01:33	Breath Test	INTERFERENCE DETECTED
11/19/2010	01:36	Breath Test	INTERFERENCE DETECTED
12/23/2010	06:01	Breath Test	INTERFERENCE DETECTED
01/21/2011	03:46	Breath Test	INTERFERENCE DETECTED
01/21/2011	03:49	Breath Test	INTERFERENCE DETECTED
01/24/2011	16:13	Breath Test	INTERFERENCE DETECTED
01/24/2011	16:32	Breath Test	STANDARD OUT OF RANGE
02/06/2011	03:09	Breath Test	INTERFERENCE DETECTED
02/27/2011	08:45	Breath Test	INTERFERENCE DETECTED
02/27/2011	09:18	Breath Test	INTERFERENCE DETECTED
03/04/2011	19:57	Breath Test	INTERFERENCE DETECTED
03/05/2011	02:31	Breath Test	STANDARD OUT OF RANGE
03/05/2011	22:12	Breath Test	INTERFERENCE DETECTED
03/05/2011	22:15	Breath Test	INTERFERENCE DETECTED
05/01/2011	03:38	Breath Test	STANDARD OUT OF RANGE
06/16/2011	13:20	Breath Test	STANDARD OUT OF RANGE
06/21/2011	01:21	Breath Test	RESPONSE TIMEOUT
06/24/2011	23:40	Breath Test	RFI DETECTED
07/14/2011	03:14	Breath Test	STANDARD OUT OF RANGE
08/06/2011	01:13	Breath Test	STANDARD OUT OF RANGE
08/06/2011	01:16	Breath Test	STANDARD OUT OF RANGE
08/06/2011	22:01	Breath Test	STANDARD OUT OF RANGE
09/10/2011	03:52	Breath Test	RESPONSE TIMEOUT
09/10/2011	03:59	Breath Test	RESPONSE TIMEOUT
11/08/2011	11:46	Breath Test	RESPONSE TIMEOUT
11/08/2011	11:55	Breath Test	STANDARD OUT OF RANGE
11/19/2011	03:45	Breath Test	STANDARD OUT OF RANGE
12/04/2011	04:19	Breath Test	STANDARD OUT OF RANGE
12/04/2011	04:23	Breath Test	STANDARD OUT OF RANGE

Breath Record Summary

X = Interference, R = Refusal, V = Invalid, I = Incomplete

Test Date	Test Time	Agency	Operator Name	Alcohol Val. 1	Alcohol Val. 2	Sim. Conc. 1	Sim. Conc. 2	Status Message
10/15/2010	09:40			0.100		0.102		
10/15/2010	09:46			0.000				RESPONSE TIMEOUT
10/15/2010	11:41			0.000	0.381			
10/24/2010	05:18	MILTON PD	WILLIAM M BOSWORTH			X		INTERFERENCE DETECTED
10/24/2010	05:23	MILTON PD	WILLIAM M BOSWORTH	0.076		0.098		
10/24/2010	05:47	MILTON PD	WILLIAM M BOSWORTH	0.071		0.099		
11/14/2010	04:26	ST. ALBANS PD	KEITH P MCMAHON			X		INTERFERENCE DETECTED
11/14/2010	04:29	ST. ALBANS PD	KEITH P MCMAHON			X		INTERFERENCE DETECTED
11/14/2010	04:31	ST. ALBANS PD	KEITH P MCMAHON			X		INTERFERENCE DETECTED
11/14/2010	04:36	ST. ALBANS PD	KEITH P MCMAHON			X		INTERFERENCE DETECTED
11/14/2010	04:39	ST. ALBANS PD	KEITH P MCMAHON	0.196		0.098		
11/17/2010	01:48	ST. ALBANS VSP	ADAM P MARCHAND			0.095		STANDARD OUT OF RANGE
11/19/2010	01:01	ST. ALBANS PD	SEAN E POPE			X		INTERFERENCE DETECTED
11/19/2010	01:05	ST. ALBANS PD	SEAN E POPE	I	0.096	0.097	0.096	STANDARD OUT OF RANGE
11/19/2010	01:15	ST. ALBANS PD	SEAN E POPE			0.097		RESPONSE TIMEOUT
11/19/2010	01:19	ST. ALBANS PD	SEAN E POPE			X		INTERFERENCE DETECTED
11/19/2010	01:22	ST. ALBANS PD	SEAN E POPE			X		INTERFERENCE DETECTED
11/19/2010	01:26	ST. ALBANS PD	SEAN E POPE			X		INTERFERENCE DETECTED
11/19/2010	01:29	ST. ALBANS PD	SEAN E POPE			X		INTERFERENCE DETECTED
11/19/2010	01:33	ST. ALBANS PD	SEAN E POPE			X		INTERFERENCE DETECTED
11/19/2010	01:36	ST. ALBANS PD	SEAN E POPE			X		INTERFERENCE DETECTED
11/25/2010	04:09	ST. ALBANS PD	SEAN E POPE	0.193		0.100		
12/02/2010	00:33	ST. ALBANS PD	HOLLY B MOORE	0.129	0.132	0.101	0.100	
12/04/2010	03:43	ST. ALBANS PD	SEAN E POPE	0.113	0.112	0.101	0.099	
12/11/2010	03:14	ST. ALBANS PD	SEAN E POPE	0.135	0.131	0.099	0.099	
12/12/2010	05:28	ST. ALBANS PD	SEAN E POPE	R		0.101		
12/23/2010	06:01	MILTON PD	WILLIAM M BOSWORTH			X		INTERFERENCE DETECTED
12/23/2010	06:04	MILTON PD	WILLIAM M BOSWORTH	0.100		0.101		
12/26/2010	00:55	ST. ALBANS PD	SEAN E POPE	0.095		0.099		
01/17/2011	23:45	ST. ALBANS PD	MARIE L BELAND	0.116	0.118	0.098	0.098	
01/21/2011	03:46	FRANKLIN COUNTY SO	ANDREW G UNDERWOOD			X		INTERFERENCE DETECTED
01/21/2011	03:49	FRANKLIN COUNTY SO	ANDREW G UNDERWOOD			X		INTERFERENCE DETECTED
01/21/2011	03:52	FRANKLIN COUNTY SO	ANDREW G UNDERWOOD	0.210		0.100		
01/24/2011	16:13	DERBY VSP	OWEN D BALLINGER			X		INTERFERENCE DETECTED
01/24/2011	16:32	DERBY VSP	OWEN D BALLINGER	0.164		0.099	0.096	STANDARD OUT OF RANGE
02/06/2011	03:09	ST. ALBANS PD	SEAN E POPE			X		INTERFERENCE DETECTED
02/06/2011	03:13	ST. ALBANS PD	SEAN E POPE	0.170		0.099		
02/27/2011	08:45	ST. ALBANS PD	KENNETH G TARDIE			X		INTERFERENCE DETECTED
02/27/2011	09:18	ST. ALBANS PD	KENNETH G TARDIE			X		INTERFERENCE DETECTED
03/04/2011	19:57	ST. ALBANS PD	MARIE L BELAND			X		INTERFERENCE DETECTED
03/04/2011	20:00	ST. ALBANS PD	MARIE L BELAND	0.183		0.098		
03/05/2011	02:31	ST. ALBANS PD	SEAN E POPE	0.120		0.098	0.093	STANDARD OUT OF RANGE
03/05/2011	02:38	ST. ALBANS PD	SEAN E POPE	0.124		0.098	0.097	
03/05/2011	22:12	ST. ALBANS PD	MARIE L BELAND			X		INTERFERENCE DETECTED
03/05/2011	22:15	ST. ALBANS PD	MARIE L BELAND			X		INTERFERENCE DETECTED
03/05/2011	22:18	ST. ALBANS PD	MARIE L BELAND	0.000		0.100		
03/28/2011	19:32	ST. ALBANS VSP	DREW E COTA	0.111		0.097		
04/01/2011	02:23	ST. ALBANS PD	SEAN E POPE	0.100		0.097		
04/03/2011	02:52	ST. ALBANS PD	SEAN E POPE	0.086	0.089	0.098	0.098	
04/16/2011	23:13	ST. ALBANS PD	MARIE L BELAND	0.131	0.139	0.097	0.097	
04/24/2011	04:52	ST. ALBANS PD	SEAN E POPE	0.169		0.098		
04/30/2011	02:08	ST. ALBANS PD	SEAN E POPE	0.127		0.097		
05/01/2011	03:38	ST. ALBANS PD	SEAN E POPE	0.136		0.096	0.095	STANDARD OUT OF RANGE
05/01/2011	03:46	ST. ALBANS PD	SEAN E POPE	I	0.134	0.097		

Breath Record Summary

St Albans PD

X = Interference, R = Refusal, V = Invalid, I = Incomplete

Test Date	Test Time	Agency	Operator Name	Alcohol Val. 1	Alcohol Val. 2	Sim. Conc. 1	Sim. Conc. 2	Status Message
05/08/2011	02:08	FRANKLIN COUNTY SO	ANDREW G UNDERWOOD	0.095		0.099		
05/08/2011	03:02	FRANKLIN COUNTY SO	DEAN M SCOTT	0.156		0.098		
05/13/2011	15:33	ST. ALBANS PD	JOSEPH E THOMAS	0.268		0.099		
05/26/2011	01:02	ST. ALBANS PD	SEAN E POPE	0.076		0.097		
05/28/2011	01:45	ST. ALBANS PD	SEAN E POPE	0.146		0.097		
06/04/2011	03:11	ST. ALBANS PD	SEAN E POPE	R		0.097		
06/10/2011	23:01	ST. ALBANS PD	PAUL MORITS	0.095		0.096		
06/11/2011	04:21	ST. ALBANS PD	SEAN E POPE	R		0.097		
06/12/2011	01:47	ST. ALBANS PD	HOWARD P MCGUIRE	R		0.097		
06/16/2011	13:20	ST. ALBANS PD	PAUL A TALLEY			0.081		STANDARD OUT OF RANGE
06/21/2011	01:21	ST. ALBANS PD	SEAN E POPE			0.098		RESPONSE TIMEOUT
06/24/2011	23:40	ST. ALBANS PD	MARIE L BELAND					RFI DETECTED
06/25/2011	00:00	ST. ALBANS PD	MARIE L BELAND	0.121	0.131	0.098	0.095	
07/12/2011	12:13	ST. ALBANS PD	PAUL A TALLEY	0.143		0.096		
07/14/2011	03:14	ST. ALBANS PD	MICHAEL P BEGIN			0.094		STANDARD OUT OF RANGE
07/14/2011	03:33	ST. ALBANS PD	MICHAEL P BEGIN	0.130		0.096		
08/06/2011	01:13	ST. ALBANS PD	TREVOR A SARGENT			0.091		STANDARD OUT OF RANGE
08/06/2011	01:16	ST. ALBANS PD	TREVOR A SARGENT			0.094		STANDARD OUT OF RANGE
08/06/2011	22:01	ST. ALBANS PD	SUSAN A HEAD			0.093		STANDARD OUT OF RANGE
08/11/2011	00:45	ST. ALBANS PD	NICHOLAS D PALMIERJ	0.088	0.079	0.099	0.099	
08/12/2011	01:29	ST. ALBANS PD	MICHAEL P BEGIN	0.194	0.194	0.099	0.099	
08/20/2011	01:55	ST. ALBANS PD	HOWARD P MCGUIRE	R		0.095		
08/20/2011	20:20	ST. ALBANS PD	GEORGE HRUBOVCAK	R		0.099		
08/29/2011	04:48	ST. ALBANS PD	HOLLY B MOORE	0.198		0.098		
09/10/2011	03:52	ST. ALBANS PD	NICHOLAS D PALMIER			0.099		RESPONSE TIMEOUT
09/10/2011	03:59	ST. ALBANS PD	NICHOLAS D PALMIER			0.098		RESPONSE TIMEOUT
09/10/2011	04:06	ST. ALBANS PD	NICHOLAS D PALMIER	0.146		0.099		
09/18/2011	15:23	ST. ALBANS PD	SEAN K BOURGEOIS	0.166		0.098		
09/27/2011	16:12	ST. ALBANS PD	SEAN E POPE	0.091		0.100		
10/02/2011	17:05	ST. ALBANS PD	MICHAEL P BEGIN	0.132		0.096		
10/18/2011	20:21	ST. ALBANS PD	NICHOLAS D PALMIER	0.142	0.150	0.096	0.095	
10/30/2011	01:52	FRANKLIN COUNTY SO	CHAD A MILES	0.173		0.097		
11/08/2011	11:46	ST. ALBANS PD	FRANCIS J MCCARTY			0.097		RESPONSE TIMEOUT
11/08/2011	11:55	ST. ALBANS PD	FRANCIS J MCCARTY			0.096		STANDARD OUT OF RANGE
11/08/2011	12:03	ST. ALBANS PD	FRANCIS J MCCARTY	R		0.097		
11/19/2011	03:45	ST. ALBANS PD	MICHAEL P BEGIN			0.093		STANDARD OUT OF RANGE
12/04/2011	04:19	ST. ALBANS PD	SEAN E POPE			0.093		STANDARD OUT OF RANGE
12/04/2011	04:23	ST. ALBANS PD	SEAN E POPE			0.093		STANDARD OUT OF RANGE