1-7-13

Swanton PD Datamaster # 121506

Franklin County

Breath Record History

Test Date	Test Time	Agency	Operator Name	Alcohol Val. 1	Alcohol Val. 2	Sim. Conc. 1	Sim. Conc. 2	Status Message
12/09/2012	20:23:49	SWANTON PD	DEAN M SCOTT	REFUSED		0.099		
12/10/2012	19:38:25	FRANKLIN COUNTY SO	CHAD A MILES	0.217	0.216	0.098	0.098	

Status Record History

Status Date	Status Time	Test Type	Status Message

Log On Record History

Test Date	Test Time	Name

DATAMASTER DMT OPERATOR USE LOG

Department/Agency

DataMaster DMT Serial Number

Date	DMT Operator's Name	Subject's Name	Simulator Vapor 1	Simulator Vapor 2
11/4/12	D. Scott	Whipple	.100	
12/9/12	D. Scott	VLAD Intoxicated	.100	
12/10/12	C. Miles	Allen Paradee	.098	

ALC 603 Rev 3 08/11