VSP - St. Albans 100171

Breath Record History

Test Date	Test Time	Agency	Operator Name	Alcohol Val. 1	Alcohol Val. 2	Sim. Conc. 1	Sim. Conc. 2	Status Message
03/02/2013	04:22:42	ST. ALBANS VSP	JOSEPH PAQUETTE	0.245	0.233	0.097	0.095	
03/02/2013	13:25:48	ST. ALBANS VSP	LUCAS J HALL	0.291		0.097		
03/03/2013	01:45:36	ST. ALBANS VSP	JAY RIGGEN	0.219		0.097		
03/07/2013	14:49:14	vsp st. albans	CORY L LOZIER	REFUSED		0.097		
03/09/2013	02:16:11	ST. ALBANS VSP	MATTHEW CANNON	0.073		0.098		
03/09/2013	03:30:37	ST. ALBANS VSP	JAY RIGGEN	0.123		0.096		
03/10/2013	01:24:24	ST. ALBANS VSP	D ROLANDINI	0.095		0.097		
03/14/2013	04:47:57	ST. ALBANS VSP	JAY RIGGEN	REFUSED		0.097		
03/23/2013	01:25:44	ST. ALBANS VSP	MATTHEW CANNON	0.113	0.108	0.097	0.097	
03/23/2013	02:25:58	ST. ALBANS VSP	JAY RIGGEN	0.162		0.097		
03/23/2013	02:52:58	ST. ALBANS VSP	BENJAMIN M PATNODE	0.070		0.096		
03/25/2013	00:35:14	ST. ALBANS VSP	RICHARD C STEPIEN	0.125	0.124	0.097	0.096	
03/30/2013	01:56:06	ST. ALBANS VSP	JAY RIGGEN	0.108	0.101	0.097	0.097	
03/30/2013	02:51:06	ST. ALBANS VSP	MATTHEW CANNON	0.161	0.159	0.097	0.097	
03/30/2013	18:08:29	ST. ALBANS VSP	JOHN P RAFTERY	0.208	0.209	0.096	0.097	
04/01/2013	15:17:24	vsp st. albans	CORY L LOZIER	0.163		0.097		
04/01/2013	20:18:09	ST. ALBANS PD	PAUL MORITS	0.213	0.212	0.096	0.096	
04/03/2013	10:44:15	ST. ALBANS VSP	DUSTIN P ROBINSON	0.147		0.096		

VSP- St Albans 100171

Status Record History

Status Date	Status Time	Test Type	Status Message

DATAMASTER DMT OPERATOR USE LOG

VSP St Albans
Department/Agency

100171
DataMaster DMT Serial Number

Date	DMT Operator's Name	Subject's Name	Simulator Vapor 1	Simulator Vapor 2
3/2/13	Paquette	Benjamin Percy	.098	
3/3/13	Hall	Christensen, Adam	.098	
3/03/13	RIGGEN	HAZY, MICHAEL	.098	
3/7/13	LOZIER	Perley, Tim	.098	
3/9/13	CANNON	DINUNZIO, FELICIA	.094	
03/09/13	RIGGEN	JACOB AINSWORTH	.096	
3/10/13	Polandini	Hogue, Amanda	.097	
3/14/13	Riggen	Sartwell, Daniel	.097	
3/20/13	CANNON	QUINOZ, ALMANDO	.097	.098
3/23/13	RIGGEN	KRISTINA KING	.097	
03/23/13	PATNODE		.096	
03/25/13	Stepien	Joseph Strouse	.097	.096
03/30/13	RIGGEN	JIGGY TESSICA	.097	
3/30/13	CANNON	DANIEL GEISWEIT	.097	.097
3/30/13	Hall	John Raffery	.208	.209
4/1/13	Lozier	Lowe, Kimberly	.098	
4/1/13	Morits	JUSTIN BARNES	.096	.096
4/2/13	~~Ritchie David~~ Robinson	Ritchie David		

ALC 603 Rev 3 08/11

DATAMASTER DMT MAINTENANCE LOG

VSP ST Albans
Department/Agency

100171
DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
1/11/13	APM protocol completed [For RPC or Simulator Solution Change Place Label Here]	R. Driscoll DMT TECH
02-19-13	[For RPC or Simulator Solution Change Place Label Here] **DataMaster Simulator Solution** Lot # 12-73-100 EtOHaq Conc: 0.098 Range: 0.093 - 0.103 Preparation Date: 04/27/2012 Expires: 04/27/2013 OPENED: 02-19-13 P. Whitney RPC	P. Whitney
3-11-13	[For RPC or Simulator Solution Change Place Label Here] Reset time 1 hour ahead	P. Whitney
	[For RPC or Simulator Solution Change Place Label Here]	

ALC 803 REV 03 08/2011