Vermont Department of Health Laboratory to National Patent Analytical Systems DataMaster DMT Shipping Memo

SIM#G7439
NIST Temp: 34.1°C
KRK 1-27-12
01/27/12

To:
National Patent Analytical Systems
2090 Harrington Memorial Road
Mansfield, Ohio 44901

From:
State of Vermont Health Department Laboratory
195 Colchester Ave
Burlington, Vermont 05401

Date: 1-27-12

Re:
Shipment of DataMaster DMT S/N: 100144 **for the Vermont Certification Project.**

Instrument Preparation:

1. **Removed Simulator & secured simulator arms with Ty-Wrap.** KRK 1-27-12
2. **Removed breath tube.** KRK 1-27-12
3. **Downloaded instrument files to USB thumb drive.** KRK 1-27-12
4. **Removed keyboard & stylus.** KRK 1-27-12
5. **Removed security screws & replaced with manufacturer supplied screws.** KRK 1-27-12
6. **Instrument bagged & packaged for shipment.** KRK 1-27-12
7. **How many TSIs are included with instrument?** 1

Notes:

Shipment in this Box Contains One DataMaster DMT with NO accessories, and one Shipping Memo.

Packed by: KRK 1-27-12 **Witness by:** 01/27/12

Stowe PD

Stowe PD

Breath Record Summary

X = Interference, R = Refusal, V = Invalid, I = Incomplete

Test Date	Test Time	Agency	Operator Name	Alcohol Val. 1	Alcohol Val. 2	Sim. Conc. 1	Sim. Conc. 2	Status Message
05/24/2011	22:18	STOWE PD	KYLE WALKER	0.120		0.104		
05/25/2011	02:24	STOWE PD	DAVID M KNIGHT	R		0.104		
05/31/2011	18:00	STOWE PD	FREDERICK G WHITCOMB	R		0.103		
06/04/2011	23:50	STOWE PD	MICHAEL J DOUGHERTY	0.089	0.090	0.104	0.104	
06/11/2011	02:14	STOWE PD	KYLE WALKER	0.138	0.130	0.103	0.103	
06/15/2011	23:57	WILLISTON VSP	TIMOTHY M CARLSON	R		0.104		
06/16/2011	00:16	WILLISTON VSP	TIMOTHY M CARLSON	0.092	0.084	0.104	0.103	
06/18/2011	07:43	STOWE PD	STEVEN C STEWART	0.051		0.104		
06/20/2011	09:07	STOWE PD	BRUCE W EMERSON	0.039		0.103		
06/24/2011	01:35	STOWE PD	MICHAEL J DOUGHERTY	R		0.103		
06/25/2011	03:38	STOWE PD	KYLE WALKER	0.190	0.192	0.102	0.103	
06/30/2011	13:15	STOWE PD	DAVID M KNIGHT	0.000		0.106		
07/01/2011	06:08	STOWE PD	KYLE WALKER	R		0.103		
07/10/2011	03:23	STOWE PD	DAVID M KNIGHT	0.200		0.102		
07/19/2011	18:05	STOWE PD	DAVID M KNIGHT	0.266	0.251	0.102	0.101	
07/25/2011	01:22	STOWE PD	DAVID M KNIGHT	R		0.101		
07/31/2011	02:23	STOWE PD	EDWIN W WEBSTER	R		0.102		
08/06/2011	00:25	STOWE PD	DAVID M KNIGHT	0.120		0.102		
08/31/2011	13:16	VDHL	STEVEN E HARNOIS	0.000		0.100		
09/24/2011	00:08	STOWE PD	JON K ZYGMUNTOWICZ	0.121	0.123	0.099	0.099	
09/25/2011	02:38	STOWE PD	MICHAEL J DOUGHERTY	I		0.099		RESPONSE TIMEOUT
09/25/2011	03:01	STOWE PD	MICHAEL J DOUGHERTY	0.185	0.183	0.099	0.099	
10/01/2011	06:00	STOWE PD	KYLE WALKER	0.192	0.197	0.099	0.100	
10/02/2011	01:10	STOWE PD	KYLE WALKER	0.176		0.099		
10/22/2011	20:01	STOWE PD	MICHAEL J DOUGHERTY	0.155		0.102		
10/22/2011	20:26	STOWE PD	MICHAEL J DOUGHERTY	0.150		0.102		
10/29/2011	02:08	STOWE PD	JON K ZYGMUNTOWICZ	R		0.101		
10/29/2011	04:08	STOWE PD	KYLE WALKER	0.151	0.146	0.102	0.102	
10/29/2011	04:50	STOWE PD	JON K ZYGMUNTOWICZ	R		0.101		
11/09/2011	03:00	STOWE PD	DAVID M KNIGHT	0.102		0.101		
11/27/2011	02:50	STOWE PD	KYLE WALKER	0.151	0.171	0.100	0.098	
12/05/2011	23:17	STOWE PD	KYLE WALKER	0.087		0.099		
12/11/2011	03:45	STOWE PD	CHRIS B ROGERS	0.181		0.098		
12/13/2011	15:10	STOWE PD	FREDERICK WHITCOMB	R		0.099		
12/18/2011	19:08	STOWE PD	FREDERICK WHITCOMB	V		0.099		
12/18/2011	19:30	STOWE PD	FREDERICK WHITCOMB	0.196		0.099		
12/24/2011	07:31	STOWE PD	BRUCE W EMERSON	0.146		0.098		
12/28/2011	00:50	STOWE PD	MICHAEL J DOUGHERTY	0.106	0.102	0.099	0.099	
01/01/2012	03:03	STOWE PD	JON K ZYGMUNTOWICZ	0.231		0.098		RESPONSE TIMEOUT
01/01/2012	03:28	STOWE PD	JON K ZYGMUNTOWICZ	0.227		0.099		
01/06/2012	14:33	STOWE PD	CHRIS B ROGERS	0.077		0.099		RESPONSE TIMEOUT
01/13/2012	19:50	STOWE PD	FREDERICK WHITCOMB	0.281	0.294	0.098	0.097	
01/21/2012	23:15	WILLISTON VSP	TIMOTHY M CARLSON			0.098		RESPONSE TIMEOUT
01/21/2012	23:34	WILLISTON VSP	TIMOTHY M CARLSON	I		0.098		RESPONSE TIMEOUT
01/23/2012	18:36	STOWE PD	FREDERICK WHITCOMB	0.182		0.098		

Breath Record Summary

X = Interference, R = Refusal, V = Invalid, I = Incomplete

Test Date	Test Time	Agency	Operator Name	Alcohol Val. 1	Alcohol Val. 2	Sim. Conc. 1	Sim. Conc. 2	Status Message
08/30/2010	20:29	STOWE PD	CHRIS B ROGERS	0.073		0.102		
09/04/2010	23:38	STOWE PD	CHRISTOPHER P MCHUGH	0.170		0.102		RESPONSE TIMEOUT
09/05/2010	02:08	STOWE PD	FREDERICK G WHITCOMB	0.150	0.142	0.102	0.102	
09/06/2010	22:04	STOWE PD	MICHAEL J DOUGHERTY	0.000		0.103		
09/06/2010	23:15	STOWE POLICE DEPARTMENT	MICHAEL J DOUGHERTY	0.168		0.102		
09/10/2010	20:00	STOWE PD	DAVID M KNIGHT	0.047		0.102		
09/15/2010	01:57	STOWE PD	KYLE R WALKER1	0.084		0.102		
09/19/2010	03:17	STOWE PD	FREDERICK G WHITCOMB	R		0.102		
09/19/2010	03:54	STOWE PD	KYLE R WALKER	R		0.102		
10/03/2010	08:52	STOWE PD	BRUCE W EMERSON	0.009		0.101		
10/25/2010	00:40	WATERBURY PD	ADAM J HUBACZ	0.105		0.099		
10/30/2010	02:31	STOWE PD	KYLE WALKER	R		0.097		
10/30/2010	20:48	STOWE PD	CHRISTOPHER P MCHUGH	0.182		0.098		
11/15/2010	02:10	ST. ALBANS PD	KYLE WALKER	0.137	0.144	0.097	0.096	
11/27/2010	01:41	STOWE PD	KYLE WALKER	0.201		0.096		
11/28/2010	02:27	STOWE PD	KYLE WALKER	0.080	0.083	0.096	0.096	
11/29/2010	23:44	STOWE PD	DAVID M KNIGHT	0.101	0.100	0.096	0.097	
12/05/2010	03:13	STOWE	MICHAEL J DOUGHERTY	0.080	0.085	0.096	0.097	
12/10/2010	18:36	STOWE PD	FREDERICK G WHITCOMB	0.206		0.096		
12/11/2010	03:20	STOWE PD	KYLE WALKER	V		0.096		
12/11/2010	03:39	STOWE PD	KYLE R WALKER	0.190	0.183	0.096	0.096	
12/12/2010	02:17	STOWE PD	KYLE WALKER	0.110		0.096		
12/24/2010	02:42	STOWE PD	JON K ZYGMUNTOWICZ			0.096		RESPONSE TIMEOUT
12/24/2010	02:46	STOWE PD	JON K ZYGMUNTOWICZ	0.076		0.097		
12/30/2010	00:10	STOWE PD	CHRISTOPHER P MCHUGH	0.072		0.095		
01/01/2011	01:48	STOWE POLICE DEPT	MICHAEL J DOUGHERTY	R		0.095		
01/08/2011	03:50	STOWE PD	KYLE WALKER	0.121		0.094		
01/22/2011	02:24	STOWE PD	KYLE WALKER	0.097		0.094		
01/23/2011	01:05	STOWE PD	KYLE WALKER	0.133		0.094		
01/30/2011	03:59	STOWE PD	EDWIN W WEBSTER	0.187	0.196	0.093	0.094	
02/05/2011	00:36	STOWE PD	KYLE WALKER			0.098		RESPONSE TIMEOUT
02/05/2011	00:41	STOWE PD	KYLE WALKER	0.225	0.244	0.099	0.098	
02/05/2011	02:38	STOWE PD	KYLE WALKER	0.129		0.099		
02/06/2011	02:58	STOWE PD	KYLE WALKER	0.188	0.188	0.098	0.098	
02/11/2011	01:48	STOWE PD	KYLE WALKER	0.129		0.099		
02/21/2011	01:35	STOWE PD	KYLE WALKER	R		0.096		
03/02/2011	14:19	STOWE PD	FREDERICK G WHITCOMB			0.096		RESPONSE TIMEOUT
03/02/2011	14:24	STOWE PD	FREDERICK G WHITCOMB	R		0.096		
03/06/2011	03:31	STOWE PD	KYLE WAQLKER	0.103		0.096		
03/06/2011	03:50	STOWE PD	KYLE WALKER	0.095		0.097		
03/17/2011	23:43	STOWE PD	FREDERICK G WHITCOMB	0.194		0.095		
03/27/2011	03:44	STOWE PD	JON K ZYGMUNTOWICZ	R		0.095		
04/02/2011	01:04	STOWE PD	KYLE WALKER	0.100	0.100	0.094	0.094	
04/05/2011	11:09	STOWE PD	STEVEN C STEWART	0.282	V	0.094	0.094	BLANK ERROR
04/10/2011	03:19	STOWE PD	EDWIN W WEBSTER	R		0.094		
04/16/2011	01:37	STOWE PD	KYLE WALKER	0.123		0.093	0.094	
04/16/2011	02:01	STOWE PD	KYLE WALKER	0.117		0.094		
04/17/2011	02:42	STOWE PD	KYLE WALKER	R		0.094		
04/22/2011	01:18	STOWE PD	KYLE WALKER	0.085	0.085	0.093	0.094	
04/30/2011	01:32	STOWE PD	KYLE WALKER	0.151	0.147	0.105	0.105	
04/30/2011	23:48	STOWE PD	KYLE WALKER	0.116	0.117	0.104	0.105	
05/01/2011	02:37	STOWE PD	KYLE WALKER	0.220	0.217	0.105	0.104	
05/06/2011	01:51	STOWE PD	KYLE WALKER	R		0.104		
05/24/2011	21:59	STOWE PD	KYLE WALKER	R		0.105		

Stowe PD

Status Record Summary

X = Interference, R = Refusal, V = Invalid, I = Incomplete

Test Date	Test Time	Test Type	Status Message
09/04/2010	23:38	Breath Test	RESPONSE TIMEOUT
12/24/2010	02:42	Breath Test	RESPONSE TIMEOUT
02/05/2011	00:36	Breath Test	RESPONSE TIMEOUT
03/02/2011	14:19	Breath Test	RESPONSE TIMEOUT
04/05/2011	11:09	Breath Test	BLANK ERROR
09/25/2011	02:38	Breath Test	RESPONSE TIMEOUT
01/01/2012	03:03	Breath Test	RESPONSE TIMEOUT
01/06/2012	14:33	Breath Test	RESPONSE TIMEOUT
01/21/2012	23:15	Breath Test	RESPONSE TIMEOUT
01/21/2012	23:34	Breath Test	RESPONSE TIMEOUT