Vermont Department of Health Laboratory to National Patent Analytical Systems DataMaster DMT Shipping Memo

Lamoille County S.O.

SIM # G6505
NIST Temp = 34.1°C
01/30/12
1-30-12

To:
National Patent Analytical Systems
2090 Harrington Memorial Road
Mansfield, Ohio 44901

From:
State of Vermont Health Department Laboratory
195 Colchester Ave
Burlington, Vermont 05401

Date: 1-30-12

Re:
Shipment of DataMaster DMT S/N: 100147 **for the Vermont Certification Project.**

Instrument Preparation:

1. **Removed Simulator & secured simulator arms with Ty-Wrap.** 1-30-12
2. **Removed breath tube.** 1-30-12
3. **Downloaded instrument files to USB thumb drive.** 1-30-12
4. **Removed keyboard & stylus.** 1-30-12
5. **Removed security screws & replaced with manufacturer supplied screws.** 1-30-12
6. **Instrument bagged & packaged for shipment.** 1-30-12
7. **How many TSIs are included with instrument?** 0

Notes:

Shipment in this Box Contains One DataMaster DMT with NO accessories, and one Shipping Memo.

Packed by: 1-30-12 **Witness by:** 01/30/12

Lamoille County S.O.

Breath Record Summary

X = Interference, R = Refusal, V = Invalid, I = Incomplete

Test Date	Test Time	Agency	Operator Name	Alcohol Val. 1	Alcohol Val. 2	Sim. Conc. 1	Sim. Conc. 2	Status Message
08/25/2010	18:54	MORRISTOWN PD	JASON P LUNEAU	0.084	0.084	0.103	0.101	
08/27/2010	01:04	LAMOILLE COUNTY SO	MICHAEL J STENGEL	R		0.103		
08/28/2010	01:57	LAMOILLE COUNTY SO	ADAM AUDET	0.195	0.209	0.103	0.101	
08/30/2010	22:02	LAMOILLE COUNTY SO	CHRISTIAN T WATSON	0.114		0.102		
09/11/2010	22:43	LAMOILLE COUNTY SO	ADAM C AUDET	0.130		0.102	0.101	
09/12/2010	02:34	LAMOILLE COUNTY SO	ADAM AUDET	0.183		0.102		
09/19/2010	23:15	LAMOILLE COUNTY SO	CHRISTIAN T WATSON	0.088		0.101		
09/24/2010	16:16	LAMOILLE COUNTY SO	ANDREW E PLANTE	0.042		0.101		
09/25/2010	08:14	LAMOILLE COUNTY SO	ANDREW E PLANTE	0.100		0.100		RESPONSE TIMEOUT
09/25/2010	08:21	LAMOILLE COUNTY SO	ANDREW E PLANTE	0.101		0.100		
10/02/2010	20:17	LAMOILLE COUNTY SO	CHRISTIAN T WATSON	0.153	0.146	0.100	0.097	
10/04/2010	02:31	LAMOILLE COUNTY SO	CHRISTIAN WATSON	I	R	0.101		
10/09/2010	02:22	LAMOILLE COUNTY SO	KEVIN R MAXFIELD	0.143		0.100		
10/24/2010	05:43	WILLISTON VSP	CARI L CRICK	0.215		0.099	0.098	
10/31/2010	03:14	WILLISTON VSP	MICHELLE A LEBLANC	R		0.098		
12/10/2010	03:29	LAMOILLE COUNTY SHERIFF	RONALD R AUDET	0.192	0.197	0.097	0.096	
01/05/2011	00:41	LAMOILLE COUNTY SO	JAMES L SAWYER	0.209		0.095		RESPONSE TIMEOUT
01/05/2011	00:46	LAMOILLE COUNTY SO	JAMES L SAWYER	0.208		0.096		
01/12/2011	02:52	WILLISTON VSP	PHILLIP W GLANZ	R		0.096		
01/12/2011	03:40	WILLISTON VSP	PHILLIP W GLANZ	I	0.123	0.095		
01/17/2011	16:29	LAMOILLE COUNTY SO	BRIANA R CLARK	0.149		0.096		
01/30/2011	00:17	MORRISTOWN PD	AARON COCHRAN	0.123		0.094		
02/01/2011	01:26	MORRISTOWN PD	TANNER C ATWOOD	0.098	I	0.093	0.093	
02/01/2011	23:05	STOWE PD	KYLE R WALKER	0.138		0.094		
02/03/2011	18:11	MORRISTOWN PD	GARTH R CHRISTENSEN			0.093		RESPONSE TIMEOUT
02/03/2011	18:16	MORRISTOWN PD	GARTH R CHRISTENSEN	I		0.094		RESPONSE TIMEOUT
02/03/2011	18:40	MORRISTOWN PD	GARTH R CHRISTENSEN	R		0.093		
02/05/2011	01:25	LAMOILLE COUNTY SO	RONALD R AUDET	0.188	I	0.094	0.093	
02/16/2011	23:42	LAMOILLE COUNTY SO	ADAM AUDET	0.165		0.093		
02/23/2011	19:42	LAMOILLE COUNTY SO	BRIANA R CLARK	0.189		0.093		RESPONSE TIMEOUT
02/26/2011	08:08	LAMOILLE COUNTY SO	JAMES L SAWYER	0.133		0.105		
03/07/2011	04:20	LAMOILLE COUNTY SO	CHRISTIAN T WATSON			0.105		RESPONSE TIMEOUT
03/07/2011	04:40	LAMOILLE COUNTY SO	CHRISTIAN T WATSON			0.105		RESPONSE TIMEOUT
03/07/2011	04:46	LAMOILLE COUNTY SO	CHRISTIAN T WATSON			0.105		RESPONSE TIMEOUT
03/07/2011	04:50	LAMOILLE COUNTY SO	CHRISTIAN T WATSON			0.105		RESPONSE TIMEOUT
03/07/2011	04:55	LAMOILLE COUNTY SO	CHRISTIAN T WATSON	R		0.105		
03/20/2011	00:06	LAMOILLE COUNTY SO	CHRISTIAN T WATSON	0.130		0.105		
03/27/2011	16:03	HARDWICK-GREENSBORO PD	KEVIN P LEHOE	0.115		0.104	0.104	RESPONSE TIMEOUT
03/27/2011	16:11	HARDWICK-GREENSBORO PD	KEVIN P LEHOE	0.115		0.104		
03/28/2011	20:02	LAMOILLE COUNTY SO	BRIANA R CLARK	0.000		0.104		
04/01/2011	22:46	LAMOILLE COUNTY SO	CHRISTIAN T WATSON	0.226		0.104		RESPONSE TIMEOUT
04/27/2011	21:41	MORRISTOWN PD	WILLIAM MORLEY	0.221	0.216	0.103	0.101	
05/06/2011	00:30	LAMOILLE COUNTY SO	CHRISTIAN T WATSON	0.162	0.167	0.102	0.100	
05/07/2011	01:54	LAMOILLE COUNTY SO	ADAM AUDET	0.148	0.140	0.101	0.100	
05/07/2011	18:40	LAMOILLE COUNTY SO	ANDREW E PLANTE	0.177		0.101		
05/14/2011	18:03	LAMOILLE COUNTY SO	RONALD R AUDET	0.204		0.101		RESPONSE TIMEOUT
05/22/2011	09:34	LAMOILLE COUNTY SO	ANDREW E PLANTE	0.122		0.102		RESPONSE TIMEOUT
05/22/2011	20:39	LAMOILLE COUNTY SO	JAMES L SAWYER	0.122	0.140	0.103	0.101	
05/28/2011	18:35	LAMOILLE COUNTY SO	CHRISTIAN T WATSON	R		0.103		
06/12/2011	22:22	WILLISTON VSP	PHILLIP W GLANZ	0.118	0.115	0.101	0.101	
06/19/2011	00:34	LAMOILLE COUNTY SO	JACQUES R MARCOUX			0.096		STANDARD OUT OF RANGE
07/03/2011	00:02	WILLISTON VSP	PHILLIP GLANZ	0.068		0.103		
07/08/2011	07:49	LAMOILLE COUNTY SO	RONALD R AUDET	0.029		0.103		
07/09/2011	00:57	LAMOILLE COUNTY SO	CHRISTIAN T WATSON	0.143	0.143	0.103	0.103	

Breath Record Summary

X = Interference, R = Refusal, V = Invalid, I = Incomplete

Test Date	Test Time	Agency	Operator Name	Alcohol Val. 1	Alcohol Val. 2	Sim. Conc. 1	Sim. Conc. 2	Status Message
07/09/2011	03:09	LAMOILLE COUNTY SO	JACQUES R MARCOUX	0.110	0.109	0.104	0.102	
07/17/2011	19:24	LAMOILLE COUNTY SO	ANDREW E PLANTE	0.156		0.102		
08/05/2011	21:11	LAMOILLE COUNTY SO	CHRISTIAN T WATSON	R		0.102		
08/05/2011	21:45	LAMOILLE COUNTY SO	CHRISTIAN T WATSON	0.144	0.150	0.102	0.101	
08/20/2011	02:22	LAMOILLE COUNTY SO	CHRISTIAN T WATSON	0.125	0.116	0.102	0.102	
08/31/2011	10:47	VDHL	STEVEN E HARNOIS	0.000		0.100		
09/02/2011	23:01	LAMOILLE COUNTY SO	CHRISTIAN T WATSON	0.142	0.141	0.101	0.100	
09/05/2011	00:16	LAMOILLE COUNTY SO	CHRISTIAN T WATSON	0.147		0.100		
09/09/2011	01:36	LAMOILLE COUNTY SO	CHRISTIAN T WATSON	0.177		0.100		
09/22/2011	01:31	LAMOILLE COUNTY SO	CHRISTIAN T WATSON	0.112		0.100		
09/25/2011	02:34	LAMOILLE COUNTY SO	ADAM AUDET	R		0.100		
10/26/2011	14:18	LAMOILLE COUNTY SO	ANDREW E PLANTE	0.000		0.100		
11/06/2011	00:49	LAMOILLE COUNTY SO	ADAM AUDET	0.165	0.174	0.100	0.099	
11/07/2011	18:58	WILLISTON VSP	PHILLIP GLANZ	0.281		0.100		
11/17/2011	23:01	LAMOILLE COUNTY SO	KEVIN M LAPLANTE	0.119		0.099		PUMP ERROR
11/17/2011	23:23	LAMOILLE COUNTY SO	KEVIN M LAPLANTE	0.112		0.099		
11/18/2011	23:54	LAMOILLE COUNTY SO	RONALD R AUDET	0.143	0.145	0.100	0.098	
11/23/2011	19:46	LAMOILLE COUNTY SO	ANDREW E PLANTE	0.152		0.099		
11/30/2011	19:41	LAMOILLE COUNTY SO	RONALD R AUDET	V		0.099		
11/30/2011	20:02	LAMOILLE COUNTY SO	RONALD R AUDET	I	0.181	0.098	0.097	
12/05/2011	19:00	MORRISTOWN PD	MICHAEL W REEVE	R		0.098		
12/17/2011	03:52	LAMOILLE COUNTY SO	JAMES L SAWYER	R		0.099		
12/18/2011	22:39	LAMOILLE COUNTY SO	JAMES L SAWYER	R		0.098		
01/03/2012	16:57	HARDWICK-GREENSBORO PD	AARON COCHRAN	0.064	0.063	0.096	0.096	
01/19/2012	00:16	LAMOILLE COUNTY SO	ADAM AUDET	R		0.097		
01/19/2012	00:34	LAMOILLE COUNTY SO	ADAM AUDET	0.107		0.098		
01/21/2012	03:09	MORRISTOWN PD	SCOTT N MCCULLOUGH	0.059		0.097		

Lamoille County S.O.

Status Record Summary

X = Interference, R = Refusal, V = Invalid, I = Incomplete

Test Date	Test Time	Test Type	Status Message
09/25/2010	08:14	Breath Test	RESPONSE TIMEOUT
01/05/2011	00:41	Breath Test	RESPONSE TIMEOUT
02/03/2011	18:11	Breath Test	RESPONSE TIMEOUT
02/03/2011	18:16	Breath Test	RESPONSE TIMEOUT
02/23/2011	19:42	Breath Test	RESPONSE TIMEOUT
03/07/2011	04:20	Breath Test	RESPONSE TIMEOUT
03/07/2011	04:40	Breath Test	RESPONSE TIMEOUT
03/07/2011	04:46	Breath Test	RESPONSE TIMEOUT
03/07/2011	04:50	Breath Test	RESPONSE TIMEOUT
03/27/2011	16:03	Breath Test	RESPONSE TIMEOUT
04/01/2011	22:46	Breath Test	RESPONSE TIMEOUT
05/14/2011	18:03	Breath Test	RESPONSE TIMEOUT
05/22/2011	09:34	Breath Test	RESPONSE TIMEOUT
06/19/2011	00:34	Breath Test	STANDARD OUT OF RANGE
11/17/2011	23:01	Breath Test	PUMP ERROR