Vermont Department of Health Laboratory to National Patent Analytical Systems DataMaster DMT Shipping Memo

STM # G9280
NIST Temp = 34.1°C
1-30-12
01/30/12

To:
National Patent Analytical Systems
2090 Harrington Memorial Road
Mansfield, Ohio 44901

From:
State of Vermont Health Department Laboratory
195 Colchester Ave
Burlington, Vermont 05401

Date: 1-30-12

Re:
Shipment of DataMaster DMT S/N: 100150 **for the Vermont Certification Project.**

Instrument Preparation:

1. **Removed Simulator & secured simulator arms with Ty-Wrap.** 1-30-12
2. **Removed breath tube.** 1-30-12
3. **Downloaded instrument files to USB thumb drive.** 1-30-12
4. **Removed keyboard & stylus.** 1-30-12
5. **Removed security screws & replaced with manufacturer supplied screws.** 1-30-12
6. **Instrument bagged & packaged for shipment.** 1-30-12
7. **How many TSIs are included with instrument?** 0

Notes:

Shipment in this Box Contains One DataMaster DMT with NO accessories, and one Shipping Memo.

Packed by: 1-30-12 **Witness by:** 01/30/12

Morristown PD

Breath Record Summary

X = Interference, R = Refusal, V = Invalid, I = Incomplete

Test Date	Test Time	Agency	Operator Name	Alcohol Val. 1	Alcohol Val. 2	Sim. Conc. 1	Sim. Conc. 2	Status Message
09/21/2011	09:15	MORRISTOWN PD	RYAN T BJERKE	0.052		0.103		
09/24/2011	00:22	WILLISTON VSP	TIMOTHY M CARLSON	0.091	0.087	0.102	0.101	
09/30/2011	16:10	WILLISTON VSP	CHRISTIAN HUNT	0.139		0.101		
10/07/2011	08:30	MORRISTOWN PD	GARTH R CHRISTENSEN	0.015		0.098		
10/15/2011	07:22	MORRISTOWN PD	RYAN T BJERKE	0.016		0.100		
10/15/2011	20:24	MORRISTOWN PD	MICHAEL W REEVE	0.268	0.280	0.099	0.099	
10/16/2011	07:19	MORRISTOWN PD	RYAN T BJERKE	0.060		0.100		
10/22/2011	07:15	MORRISTOWN PD	GARTH R CHRISTENSEN	0.010		0.100		
10/23/2011	02:02	MORRISTOWN PD	ERIC J DODGE	0.231	0.228	0.100	0.099	
10/29/2011	03:25	MORRISTOWN PD	SCOTT N MCCULLOUGH	0.151		0.098		
10/31/2011	09:19	MORRISTOWN PD	MICHAEL W REEVE	0.048		0.102		
11/03/2011	14:31	MORRISTOWN PD	MICHAEL W REEVE	0.000		0.102		
11/04/2011	00:18	MORRISTOWN PD	MICHAEL W REEVE	0.120		0.101		
11/07/2011	18:28	WILLISTON VSP	PHILLIP GLANZ			0.100		BLANK ERROR
11/07/2011	18:32	WILLISTON VSP	PHILLIP GLANZ			0.095		STANDARD OUT OF RANGE
11/07/2011	18:37	MORRISTOWN PD	MICHAEL W REEVE	0.000		0.101		
11/08/2011	23:36	MIDDLESEX VSP	JAMES V FOX	0.085		0.103		
11/11/2011	15:34	MORRISTOWN PD	RYAN T BJERKE	R		0.103		
11/13/2011	02:21	MORRISTOWN PD	SCOTT N MCCULLOUGH	0.104	0.102	0.103	0.102	
11/13/2011	04:26	MORRISTOWN PD	MICHAEL W REEVE	R		0.101		
11/19/2011	06:52	MORRISTOWN PD	GARTH R CHRISTENSEN	0.088		0.102		
11/20/2011	18:29	MORRISTOWN PD	ERIC J DODGE	0.209	0.173	0.101	0.100	
11/26/2011	18:38	WILLISTON VSP	DAVID R ROBILLARD	R		0.101		
11/27/2011	03:41	MORRISTOWN PD	SCOTT N MCCULLOUGH			0.101		RESPONSE TIMEOUT
11/27/2011	04:00	MORRISTOWN PD	SCOTT N MCCULLOUGH	0.091	0.091	0.102	0.101	
12/03/2011	01:22	MORRISTOWN PD	TANNER C ATWOOD			0.101		RESPONSE TIMEOUT
12/03/2011	01:42	MORRISTOWN PD	TANNER C ATWOOD	0.073		0.100		
12/06/2011	01:35	MORRISTOWN PD	SCOTT N MCCULLOUGH	0.202	V	0.101	0.101	
12/06/2011	21:47	MORRISTOWN PD	SCOTT N MCCULLOUGH	0.195	0.195	0.100	0.101	
12/10/2011	07:13	MORRISTOWN PD	RYAN T BJERKE	0.043		0.100		
12/11/2011	03:22	MORRISTOWN PD	MICHAEL W REEVE			0.099		RESPONSE TIMEOUT
12/11/2011	03:26	MORRISTOWN PD	MICHAEL W REEVE	0.209		0.100		
12/13/2011	09:42	MORRISTOWN PD	GARTH R CHRISTENSEN	0.025		0.099		
12/17/2011	05:01	MORRISTOWN PD	TANNER C ATWOOD	0.139	0.145	0.099	0.101	
12/30/2011	16:03	MORRISTOWN PD	DAMIEN A LAROSE	0.165	0.164	0.097	0.097	
01/03/2012	14:47	MORRISTOWN PD	RYAN T BJERKE	0.179		0.097		
01/04/2012	23:38	WILLISTON VSP	TIMOTHY M CARLSON	0.083	I	0.099	0.098	RESPONSE TIMEOUT
01/06/2012	17:39	MORRISTOWN PD	MICHAEL W REEVE	R		0.097		
01/07/2012	01:57	MORRISTOWN PD	SCOTT N MCCULLOUGH	0.107	0.104	0.098	0.099	
01/11/2012	09:16	MORRISTOWN PD	RYAN T BJERKE			0.097		RESPONSE TIMEOUT
01/13/2012	17:29	MORRISTOWN PD	JAY T LAMELL	0.207		0.098		
01/20/2012	00:12	MORRISTOWN PD	JAY T LAMELL	R		0.098		
01/21/2012	02:45	MORRISTOWN PD	SCOTT N MCCULLOUGH			0.096		STANDARD OUT OF RANGE
01/22/2012	02:18	MORRISTOWN PD	DAMIEN A LAROSE	I		0.099		RESPONSE TIMEOUT

Breath Record Summary

X = Interference, R = Refusal, V = Invalid, I = Incomplete

Test Date	Test Time	Agency	Operator Name	Alcohol Val. 1	Alcohol Val. 2	Sim. Conc. 1	Sim. Conc. 2	Status Message
04/18/2011	14:52	MORRISTOWN PD	WILLIAM MORLEY	0.168		0.102		
04/21/2011	09:29	MORRISTOWN PD	RYAN T BJERKE	0.013		0.101		
04/27/2011	21:54	MORRISTOWN PD	SCOTT N MCCULLOUGH	0.138	0.138	0.099	0.099	
05/01/2011	01:03	WILLISTON VSP	TIMOTHY M CARLSON	0.085		0.099		
05/06/2011	01:00	MORRISTOWN PD	MICHAEL W REEVE	0.177		0.099		
05/06/2011	02:22	MORRISTOWN PD	SCOTT N MCCULLOUGH	0.108	0.101	0.100	0.100	
05/06/2011	20:13	MORRISTOWN PD	GARTH R CHRISTENSEN	R		0.100		
05/20/2011	03:40	MORRISTOWN PD	SCOTT N MCCULLOUGH	0.232		0.100		
05/20/2011	23:07	MORRISTOWN PD	GARTH R CHRISTENSEN	I		0.099		RESPONSE TIMEOUT
05/20/2011	23:30	MORRISTOWN PD	GARTH R CHRISTENSEN	0.134	0.135	0.100	0.100	
06/04/2011	01:47	WILLISTON VSP	TIMOTHY M CARLSON	0.119	0.119	0.099	0.098	
06/08/2011	02:31	MORRISTOWN PD	SCOTT N MCCULLOUGH	0.180		0.098		
06/12/2011	21:46	WILLISTON VSP	PHILLIP W GLANZ			0.097		STANDARD OUT OF RANGE
06/17/2011	23:58	MORRISTOWN PD	WILLIAM MORLEY	V		0.103		
06/18/2011	00:22	MORRISTOWN PD	WILLIAM MORLEY	0.157	0.159	0.104	0.103	
06/18/2011	07:22	MORRISTOWN PD	ANDREW GLOVER	0.020		0.104		
06/19/2011	01:07	LAMOILLE COUNTY SO	JACQUES R MARCOUX	0.143	0.156	0.103	0.103	
06/21/2011	02:14	MORRISTOWN PD	MICHAEL W REEVE	0.166	0.170	0.103	0.102	
06/25/2011	06:29	LAMOILLE COUNTY SO	CHRISTIAN T WATSON	0.153	0.154	0.105	0.103	
06/29/2011	22:15	MORRISTOWN PD	MICHAEL W REEVE	R		0.103		
07/02/2011	04:44	MORRISTOWN PD	MICHAEL W REEVE	0.143		0.103		
07/02/2011	07:17	MORRISTOWN PD	GASRTH R CHRISTENSEN	0.071		0.103		
07/04/2011	21:10	WILLISTON VSP	MIKE KAMERLING	0.159		0.103		
07/04/2011	22:59	WILLISTON VSP	PHILLIP GLANZ	0.090	0.087	0.104	0.104	
07/09/2011	00:36	MORRISTOWN PD	SCOTT N MCCULLOUGH	0.084		0.102		
07/10/2011	02:33	MORRISTOWN PD	SCOTT N MCCULLOUGH	0.156		0.103		
07/10/2011	23:08	WILLISTON VSP	TIMOTHY M CARLSON	0.087	0.089	0.102	0.102	
07/15/2011	03:59	MORRISTOWN PD	SCOTT N MCCULLOUGH	0.178		0.104		RESPONSE TIMEOUT
07/19/2011	20:02	MORRISTOWN PD	ERIC J DODGE	0.054		0.102		
07/20/2011	04:38	WILLISTON VSP	TIMOTHY M CARLSON	0.089		0.104		
07/21/2011	03:13	MORRISTOWN PD	TANNER C ATWOOD	0.141		0.102		RESPONSE TIMEOUT
07/21/2011	03:34	MORRISTOWN PD	TANNER C ATWOOD	0.137	0.134	0.102	0.102	
07/23/2011	03:51	MORRISTOWN PD	SCOTT N MCCULLOUGH	0.156		0.102		
07/23/2011	19:34	WILLISTON VSP	TIMOTHY M CARLSON	0.111		0.102		
07/23/2011	23:05	LAMOILLE COUNTY SO	JACQUES R MARCOUX	R		0.103		
07/28/2011	01:07	MORRISTOWN PD	MICHAEL W REEVE	0.124	0.119	0.103	0.101	
07/30/2011	02:51	MORRISTOWN PD	ATWOOD C ATWOOD	0.092	0.092	0.102	0.101	
07/31/2011	00:07	MORRISTOWN PD	TANNER C ATWOOD	0.227		0.102	0.101	
07/31/2011	00:29	MORRISTOWN PD	TANNER C ATWOOD	0.228		0.102		
08/03/2011	00:42	WILLISTON VSP	TIMOTHY M CARLSON	0.100		0.100		
08/06/2011	01:18	WILLISTON VSP	PHILLIP GLANZ	I	R	0.101		
08/06/2011	01:41	WILLISTON VSP	PHILLIP GLANZ	0.154	0.161	0.101	0.101	
08/10/2011	20:45	MORRISTOWN PD	MICHAEL W REEVE	0.190	0.188	0.101	0.099	
08/18/2011	08:53	MORRISTOWN PD	GARTH R CHRISTENSEN	0.014		0.100		
08/20/2011	01:30	MORRISTOWN PD	SCOTT N MCCULLOUGH	R		0.102		
08/20/2011	07:11	MORRISTOWN PD	TANNER C ATWOOD	0.027		0.100		
08/31/2011	11:54	VDHL	STEVEN E HARNOIS	0.000		0.099		
09/07/2011	18:14	MORRISTOWN PD	MICHAEL W REEVE	0.115	0.114	0.099	0.099	
09/08/2011	20:04	MORRISTOWN PD	MICHAEL W REEVE	0.255		0.099		
09/09/2011	23:59	WILLISTON VSP	TIMOTHY M CARLSON	0.081	0.079	0.099	0.099	
09/10/2011	07:12	MORRISTOWN PD	GARTH R CHRISTENSEN	0.026		0.100		
09/13/2011	04:34	MORRISTOWN PD	SCOTT N MCCULLOUGH	0.105		0.100		
09/15/2011	18:56	MORRISTOWN PD	ERIC J DODGE	0.066		0.101		
09/17/2011	03:01	MORRISTOWN PD	TANNER C ATWOOD	0.173	0.162	0.101	0.100	

Breath Record Summary

X = Interference, R = Refusal, V = Invalid, I = Incomplete

Test Date	Test Time	Agency	Operator Name	Alcohol Val. 1	Alcohol Val. 2	Sim. Conc. 1	Sim. Conc. 2	Status Message
08/25/2010	23:29	MORRISTOWN PD	MICHAEL W REEVE	0.110	0.108	0.100	0.100	
08/27/2010	22:45	MORRISTOWN PD	GARTH R CHRISTENSEN	0.144	0.148	0.100	0.100	
08/29/2010	01:54	WILLISTON VSP	MICHAEL KAMERLING	R		0.099		
08/30/2010	01:04	MORRISTOWN PD	TANNER C ATWOOD	0.151	0.154	0.098	0.098	
09/04/2010	23:35	MORRISTOWN PD	LUNEAU JASON	0.104		0.098		
09/07/2010	16:17	LAMOILLE VSP	MARK C LUCAS	0.139		0.100		
09/09/2010	22:55	MORRISTOWN PD	JASON P LUNEAU	0.125	0.126	0.098	0.099	
09/12/2010	03:42	MORRISTOWN PD	MICHAEL W REEVE	0.231	0.223	0.100	0.099	
09/12/2010	20:58	WILLISTON VSP	GARRY M SCOTT	0.110	0.106	0.098	0.100	
09/17/2010	17:13	MORRISTOWN PD	MICHAEL W REEVE	0.249		0.097		
09/17/2010	19:06	MORRISTOWN PD	MICHAEL W REEVE	0.260		0.098		
09/18/2010	20:59	WILLISTON VSP	PHILLIP W GLANZ	0.147		0.098		
09/18/2010	22:22	MORRISTOWN PD	JASON P LUNEAU	0.077		0.098		
09/19/2010	03:14	MORRISTOWN PD	JASON P LUNEAU	0.202		0.098		
10/01/2010	23:36	WILLISTON VSP	GARRY M SCOTT	R		0.098		
10/02/2010	07:25	MORRISTOWN PD	RYAN T BJERKE	0.010		0.099		
10/03/2010	02:03	MORRISTOWN PD	MICHAEL W REEVE	R		0.099		
10/09/2010	19:32	ST. ALBANS VSP	DUSTIN P ROBINSON	0.153		0.099		
10/24/2010	05:04	WILLISTON VSP	CARI L CRICK			0.094		STANDARD OUT OF RANGE
10/24/2010	05:06	WILLISTON VSP	CARI L CRICK			0.096		STANDARD OUT OF RANGE
11/14/2010	01:07	MORRISTOWN PD	MICHAEL W REEVE	0.162		0.096		
12/11/2010	00:23	MORRISTOWN PD	MICHAEL W REEVE	R		0.095		
12/15/2010	04:27	MORRISTOWN PD	ATWOOD C TANNER	0.098	0.100	0.094	0.094	
12/27/2010	18:30	MORRISTOWN PD	AARON L COCHRAN	0.122	0.125	0.095	0.094	
01/01/2011	02:19	MORRISTOWN PD	WILLIAM MORLEY	0.082	0.083	0.095	0.094	
01/01/2011	08:36	WILLISTON VSP	JOSEPH A PAQUETTE	0.182	0.172	0.093	0.093	
01/06/2011	19:28	MORRISTOWN PD	MICHAEL W REEVE	R		0.094		
01/19/2011	00:30	MORRISTOWN PD	TANNER C ATWOOD	0.191		0.093		
01/19/2011	20:11	MORRISTOWN PD	MICHAEL W REEVE	0.178		0.093		
01/29/2011	23:50	MORRISTOWN PD	AARON COCHRAN			0.092		STANDARD OUT OF RANGE
02/03/2011	17:43	MORRISTOWN PD	GARTH R CHRISTENSEN			0.092		STANDARD OUT OF RANGE
02/05/2011	01:42	MORRISTOWN PD	TANNER C ATWOOD	0.180		0.106		RESPONSE TIMEOUT
02/07/2011	01:30	MORRISTOWN PD	WILLIAM F MORLEY	0.156		0.107		
02/13/2011	02:37	MORRISTOWN PD	MICHAEL W REEVE	0.141	0.137	0.105	0.105	
02/20/2011	22:56	MORRISTOWN PD	TANNER C ATWOOD	0.200	0.198	0.106	0.105	
02/24/2011	23:18	MORRISTOWN PD	TANNER C ATWOOD			0.105		RESPONSE TIMEOUT
02/24/2011	23:37	MORRISTOWN PD	TANNER C ATWOOD	0.104	0.114	0.105	0.104	
02/28/2011	22:26	WILLISTON VSP	PHILLIP W GLANZ	0.343		0.104		
03/01/2011	19:38	MORRISTOWN PD	TANNER C ATWOOD	I	R	0.105		
03/04/2011	20:39	WILLISTON VSP	TIMOTHY M CARLSON	0.094	0.092	0.105	0.104	
03/05/2011	01:30	MORRISTOWN PD	WILLIAM F MORLEY	0.101	0.123	0.104	0.104	
03/06/2011	07:11	WILLISTON VSP	CHRISTIAN HUNT			0.105		RESPONSE TIMEOUT
03/06/2011	07:31	WILLISTON VSP	CHRISTIAN HUNT	0.087	0.080	0.104	0.104	
03/12/2011	03:37	MORRISTOWN PD	WILLIAM F MORLEY	0.138	0.143	0.104	0.103	
03/15/2011	20:53	WILLISTON VSP	TIMOTHY M CARLSON	0.068		0.105		
03/17/2011	23:51	WILLISTON VSP	PHILLIP W GLANZ	0.137	0.135	0.105	0.104	
03/19/2011	22:10	WILLISTON VSP	TIMOTHY M CARLSON	R		0.104		
03/31/2011	18:23	MORRISTOWN PD	MICHAEL W REEVE	0.296		0.103		
04/07/2011	02:35	MORRISTOWN PD	WILLIAM MORLEY	R		0.102		
04/10/2011	02:35	MORRISTOWN PD	SCOTT N MCCULLOUGH			0.103		RESPONSE TIMEOUT
04/10/2011	02:55	MORRISTOWN PD	SCOTT N MCCULLOUGH	0.179	0.185	0.102	0.101	
04/14/2011	00:57	WILLISTON VSP	TIMOTHY M CARLSON	0.068		0.105		
04/14/2011	19:52	MORRISTOWN PD	SCOTT N MCCULLOUGH	R		0.103		
04/16/2011	03:53	MORRISTOWN PD	TANNER C ATWOOD	R		0.103		

Morristown PD

Status Record Summary

X = Interference, R = Refusal, V = Invalid, I = Incomplete

Test Date	Test Time	Test Type	Status Message
10/24/2010	05:04	Breath Test	STANDARD OUT OF RANGE
10/24/2010	05:06	Breath Test	STANDARD OUT OF RANGE
01/29/2011	23:50	Breath Test	STANDARD OUT OF RANGE
02/03/2011	17:43	Breath Test	STANDARD OUT OF RANGE
02/05/2011	01:42	Breath Test	RESPONSE TIMEOUT
02/24/2011	23:18	Breath Test	RESPONSE TIMEOUT
03/06/2011	07:11	Breath Test	RESPONSE TIMEOUT
04/10/2011	02:35	Breath Test	RESPONSE TIMEOUT
05/20/2011	23:07	Breath Test	RESPONSE TIMEOUT
06/12/2011	21:46	Breath Test	STANDARD OUT OF RANGE
07/15/2011	03:59	Breath Test	RESPONSE TIMEOUT
07/21/2011	03:13	Breath Test	RESPONSE TIMEOUT
11/07/2011	18:28	Breath Test	BLANK ERROR
11/07/2011	18:32	Breath Test	STANDARD OUT OF RANGE
11/27/2011	03:41	Breath Test	RESPONSE TIMEOUT
12/03/2011	01:22	Breath Test	RESPONSE TIMEOUT
12/11/2011	03:22	Breath Test	RESPONSE TIMEOUT
01/04/2012	23:38	Breath Test	RESPONSE TIMEOUT
01/11/2012	09:16	Breath Test	RESPONSE TIMEOUT
01/21/2012	02:45	Breath Test	STANDARD OUT OF RANGE
01/22/2012	02:18	Breath Test	RESPONSE TIMEOUT