



## Driscoll, Rob

---

**From:** Flowers, Heather  
**Sent:** Monday, August 18, 2014 11:15 AM  
**To:** Driscoll, Rob  
**Subject:** RE: DMT

I am going to get some solution from St. J. I would appreciate you looking at it tomorrow though and bringing solution for us and to replace St. J's. Night shift has been getting DUIs and they are trying to get stats for the labor day campaign so I don't want to take it out of service. Thank you!

Trooper Heather Flowers  
Vermont State Police  
Bradford Barracks

Phone: 802-222-4680  
Fax: 802-222-4028  
E-mail: [heather.flowers@state.vt.us](mailto:heather.flowers@state.vt.us)

---

**From:** Driscoll, Rob  
**Sent:** Monday, August 18, 2014 11:13 AM  
**To:** Flowers, Heather  
**Subject:** Re: DMT

Heather,  
You can pick some solution up at the forensics lab, or wait for me to come take a look tomorrow. I am on the road and in Bennington all day today. If you wait until I'm available tomorrow just put the instrument out of service.

Regards,  
Rob

Sent from my iPhone

On Aug 18, 2014, at 10:11 AM, "Flowers, Heather" <[Heather.Flowers@state.vt.us](mailto:Heather.Flowers@state.vt.us)> wrote:

I also am in emergency need of simulator solution. Ours is out of range and I apparently don't have any.

Trooper Heather Flowers  
Vermont State Police  
Bradford Barracks

Phone: 802-222-4680  
Fax: 802-222-4028  
E-mail: [heather.flowers@state.vt.us](mailto:heather.flowers@state.vt.us)

---

**From:** Flowers, Heather  
**Sent:** Monday, August 18, 2014 8:43 AM  
**To:** Driscoll, Rob  
**Subject:** DMT

Hello,

Our simulator solution is going out of range every month. Nothing appears to be wrong with the machine but Sgt. White is wondering if you might be able to check it. Also, I need more solution. Thank you.

Trooper Heather Flowers  
Vermont State Police  
Bradford Barracks

Phone: 802-222-4680

Fax: 802-222-4028

E-mail: [heather.flowers@state.vt.us](mailto:heather.flowers@state.vt.us)



### ACCURACY & PRECISION REPORT

#### STATE OF VERMONT

DataMeter DMT: 100140  
Location: Bradford VSP

Date: 08/19/2014  
Time: 10:01:49

SUPERVISOR NAME:  
ROBERT DRISCOLL

SOLUTION LOT #: 14-05-100  
SOLUTION CONCENTRATION: 0.098

BLANK TEST	0.000	10:02
CALIBRATION CHECK	PASSED	10:02
SIMULATOR VAPOR 34.0°C	0.097	10:03
SIMULATOR VAPOR 34.1°C	0.097	10:04
SIMULATOR VAPOR 34.0°C	0.097	10:05
SIMULATOR VAPOR 34.1°C	0.098	10:06
SIMULATOR VAPOR 34.0°C	0.098	10:07
SIMULATOR VAPOR 34.1°C	0.098	10:08
SIMULATOR VAPOR 34.1°C	0.098	10:09
SIMULATOR VAPOR 34.1°C	0.098	10:10
SIMULATOR VAPOR 34.0°C	0.098	10:11
SIMULATOR VAPOR 34.0°C	0.098	10:12
BLANK TEST	0.000	10:13

Average = 0.097  
Std Dev = 0.000

