Vermont Department of Health Laboratory to National Patent Analytical Systems DataMaster DMT Shipping Memo

Sim# G8224
NIST TEMP=34.0°C
11-3-11
11/03/11

To:
National Patent Analytical Systems
2090 Harrington Memorial Road
Mansfield, Ohio 44901

From:
State of Vermont Health Department Laboratory
195 Colchester Ave
Burlington, Vermont 05401

Date: 11-3-11

Re:
Shipment of DataMaster DMT S/N: 100159 **for the Vermont Certification Project.**

Instrument Preparation:

1. **Removed Simulator & secured simulator arms with Ty-Wrap.** 11-3-11
2. **Removed breath tube.** 11-3-11
3. **Downloaded instrument files to USB thumb drive.** 11-3-11
4. **Removed keyboard & stylus.** 11-3-11
5. **Removed security screws & replaced with manufacturer supplied screws.** 11-3-11
6. **Instrument bagged & packaged for shipment.**
7. **How many TSIs are included with instrument?** 3

Notes:

Shipment in this Box Contains One DataMaster DMT with NO accessories, and one Shipping Memo.

Packed by: 11-3-11 **Witness by:** 11/03/11

Rutland PD.

Rutland PD

Breath Record Summary

X = Interference, R = Refusal, V = Invalid, I = Incomplete

Test Date	Test Time	Agency	Operator Name	Alcohol Val. 1	Alcohol Val. 2	Sim. Conc. 1	Sim. Conc. 2	Status Message
09/20/2010	23:53	RUTLAND CITY PD	JOSEPH J BARTLETT	R		0.098		
09/21/2010	01:53	RUTLAND	ELIZABETH R RYAN	0.081		0.097		RESPONSE TIMEOUT
09/21/2010	02:03	RUTLAND CITY PD	ELIZABETH R RYAN	0.096		0.097		
10/02/2010	02:30	RUTLAND COUNTY SO	JEFFREY W STEPHENSON	0.107	0.107	0.096	0.095	
10/06/2010	02:14	RUTLAND CITY PD	TED R WASHBURN	0.165		0.096		
10/07/2010	00:54	RUTLAND COUNTY SO	RYAN BRADY	0.203		0.096		
10/07/2010	01:39	RUTLAND CITY PD	TED R WASHBURN	R		0.096		
10/08/2010	15:48	RUTLAND COUNTY SHERIFFS	ANDREW S CROSS	R		0.097		
10/16/2010	00:58	RUTLAND COUNTY SO	JEFFREY W STEPHENSON	0.235		0.095		
10/25/2010	02:25	RUTLAND CITY PD	EARL F POST	R		0.095		
10/25/2010	02:53	RUTLAND CITY PD	EARL F POST	0.229	0.234	0.095	0.094	
10/26/2010	22:43	RUTLAND CITY PD	NICOLE A CHAPLEAU	R		0.094		
10/31/2010	01:12	RUTLAND CITY PD	NICOLE A CHAPLEAU	0.077		0.095		
11/01/2010	18:33	RUTLAND CITY	JUSTIN D SOUZA	0.185		0.094		
11/06/2010	23:50	RUTLAND CITY PD	ANDREW N TODD	0.232		0.095		
11/07/2010	05:16	RUTLAND CITY PD	ANDREW N TODD	R		0.095		
11/13/2010	21:34	RUTLAND CITY PD	JOSEPH J BARTLETT	R		0.095		
11/14/2010	01:38	RUTLAND VSP	ARON C MCNEIL	0.255		0.094		
11/14/2010	02:36	RUTLAND CITY PD	MATTHEW A PROUTY	0.259		0.095		
11/15/2010	02:55	RUTLAND CITY PD	JOSEPH J BARTLETT	I	R	0.094		
11/15/2010	03:22	RUTLAND CITY PD	JOSEPH J BARTLETT	0.106	0.108	0.094	0.094	
11/19/2010	19:26	RUTLAND CITY PD	NICOLE A CHAPLEAU	0.150		0.094	0.093	
11/19/2010	19:35	RUTLAND CITY PD	NICOLE A CHAPLEAU	0.158		0.094		
11/21/2010	02:39	RUTLAND VSP	ROBERT A RIDER	0.134		0.093		
11/25/2010	00:42	RUTLAND CITY PD	JOSEPH J BARTLETT	0.109	0.114	0.093	0.093	
11/26/2010	16:32	RUTLAND CITY PD	GREG S SHELDON	0.184		0.094	0.093	
12/03/2010	20:58	RUTLAND COUNTY SO	RYAN BRADY	0.141		0.094		
12/06/2010	21:52	RUTLAND CITY PD	TED R WASHBURN	0.162		0.093	0.091	STANDARD OUT OF RANGE
12/08/2010	03:07	RUTLAND VSP	ARON C MCNEIL	0.160	0.161	0.093	0.093	
12/08/2010	21:39	RUTLAND CITY PD	MATTYHEW A PROUTY	0.145	0.147	0.093	0.093	
12/10/2010	09:15	RUTLAND CITY PD	JUSTIN D SOUZA			0.092		STANDARD OUT OF RANGE
12/18/2010	03:57	RUTLAND COUNTY SO	JEFFREY W STEPHENSON	0.089	0.092	0.099	0.099	
12/24/2010	23:07	RUTLAND CITY PD	TED R WASHBURN	R		0.097		
12/24/2010	23:27	RUTLAND CITY PD	TED R WASHBURN	0.191		0.098		
12/29/2010	07:55	RUTLAND COUNTY SO	RICK B PUTNAM	0.151		0.096		
12/29/2010	23:01	RUTLAND COUNTY SO	MICHAEL F MILLER	0.166		0.097		RESPONSE TIMEOUT
01/02/2011	21:53	RUTLAND CITY PD	EARL F POST	0.146		0.097		
01/08/2011	21:22	RUTLAND CITY PD	ANDREW N TODD	0.124	0.129	0.096	0.096	
01/21/2011	04:17	RUTLAND CITY PD	TED R WASHBURN	0.171		0.095		
01/22/2011	18:13	RUTLAND COUNTY SO	MICHAEL F MILLER	0.101		0.095		
01/27/2011	07:58	RUTLAND COUNTY SO	RICHARD B PUTNAM	0.138		0.096		
01/30/2011	02:28	RUTLAND COUNTY SO	MICHAEL F MILLER	0.146	0.138	0.096	0.095	
02/05/2011	03:41	RUTLAND CITY PD	EARL F POST	0.082	0.085	0.095	0.094	
02/14/2011	21:12	RUTLAND CITY PD	TED R WASHBURN	0.126	0.135	0.094	0.095	
02/19/2011	15:53	RUTLAND CITY PD	CHARLES K HALL			0.095		RESPONSE TIMEOUT
02/19/2011	15:58	RUTLAND CITY PD	CHARLES K HALL			0.095		RESPONSE TIMEOUT
02/19/2011	16:03	RUTLAND CITY PD	CHARLES K HALL	0.203		0.095		
02/23/2011	18:13	RUTLAND COUNTY SO	MICHAEL J TARBELL	0.000		0.103		
02/26/2011	03:17	RUTLAND CITY PD	ANDREW N TODD	R		0.103		
02/27/2011	02:06	RUTLAND COUNTY SO	JEFFREY W STEPHENSON	0.172		0.102		RESPONSE TIMEOUT
02/27/2011	02:14	RUTLAND COUNTY SO	JEFFREY W STEPHENSON	0.175		0.103		
03/07/2011	04:32	RUTLAND CITY PD	EDWARD E DUMAS II	0.270		0.102		RESPONSE TIMEOUT
03/09/2011	17:42	RUTLAND CITY PD	NICOLE A CHAPLEAU	0.154		0.102		
03/11/2011	23:56	RUTLAND COUNTY SO	LAWRENCE J YNESTA	0.149	0.137	0.102	0.102	

Breath Record Summary

X = Interference, R = Refusal, V = Invalid, I = Incomplete

Test Date	Test Time	Agency	Operator Name	Alcohol Val. 1	Alcohol Val. 2	Sim. Conc. 1	Sim. Conc. 2	Status Message
03/12/2011	03:40	RUTLAND COUNTY SO	JEFFREY W STEPHENSON	0.149		0.102	0.101	
03/13/2011	04:45	RUTLAND CITY PD	TED R WASHBURN	I	I	0.103		PUMP ERROR
03/13/2011	05:03	RUTLAND CITY PD	TED R WASHBURN	I		0.101		PUMP ERROR
03/18/2011	08:12	RUTLAND COUNTY SO	RICHARD B PUTNAM	0.052		0.102		
03/19/2011	00:31	RUTLAND COUNTY SO	DAVID J FOX	R		0.102		
03/19/2011	02:08	RUTLAND COUNTY SO	RYAN BRADY	0.130		0.103		
03/20/2011	02:23	RUTLAND COUNTY SO	JEFFREY W STEPHENSON	0.134		0.102		RESPONSE TIMEOUT
03/20/2011	02:31	RUTLAND COUNTY SO	JEFFREY W STEPHENSON	R		0.102		
03/20/2011	02:52	RUTLAND CITY PD	ED E DUMAS II	0.097	0.096	0.101	0.101	
03/21/2011	20:10	RUTLAND VSP	DAN HALL	0.161		0.101		
03/21/2011	20:51	RUTLAND VSP	ARON MCNEIL	0.169		0.102		
03/22/2011	18:50	RUTLAND CITY PD	THOMAS W FULLER	0.290	0.286	0.101	0.100	
03/24/2011	10:50	RUTLAND COUNTY SO	DAVID J FOX	R		0.101		
03/24/2011	11:11	RUTLAND COUNTY SO	DAVID J FOX	0.083	0.083	0.101	0.100	
03/25/2011	05:14	RUTLAND CITY PD	SETH A RICHARDSON	0.098		0.101		
03/25/2011	08:16	RUTLAND COUNTY SO	RICHARD B PUTNAM	0.055		0.101		
03/27/2011	21:17	RUTLAND CITY PD	JENNIFER L CZACHOR	0.173		0.100		RESPONSE TIMEOUT
03/30/2011	08:10	RUTLAND COUNTY SO	FRANK L WILK	0.033		0.101		
04/01/2011	09:37	RUTLAND CITY PD	NICOLE A CHAPLEAU	R		0.101		
04/01/2011	23:02	RUTLAND CITY PD	JENNIFER L CZACHOR	0.164		0.100		
04/03/2011	18:45	RUTLAND COUNTY SO	JAMES M BENNICK	0.066		0.100		
05/06/2011	23:05	RUTLAND COUNTY SO	JEFFREY W STEPHENSON	0.077		0.098		
05/08/2011	02:07	ROYALTON VSP	ARON MCNEIL	R		0.097		
05/08/2011	02:48	RUTLAND VSP	ROBERT A RIDER	0.115	0.112	0.097	0.097	
05/08/2011	21:17	RUTLAND VSP	MICHAEL R TIETZ	0.210		0.098		RESPONSE TIMEOUT
05/11/2011	02:07	RUTLAND VSP	ROBERT A RIDER	0.149	0.150	0.097	0.096	
05/12/2011	00:33	VSP-RUTLAND	CHRISTOPHER G LOYZELLE	0.170		0.098		
05/12/2011	03:24	RUTLAND VSP	SGT T MOZZER	0.161		0.097		
05/14/2011	01:39	RUTLAND COUNTY SO	JEFFREY W STEPHENSON	0.070	I	0.097	0.098	
05/15/2011	02:55	RUTLAND VSP	SGT MOZZER	0.080		0.098		
05/15/2011	03:46	RUTLAND CITY PD	ANDREW N TODD	0.173		0.097		
06/05/2011	01:59	RUTLAND COUNTY SO	JEFFREY W STEPHENSON			0.095		STANDARD OUT OF RANGE
06/17/2011	20:37	RUTLAND COUNTY SO	JEFFREY W STEPHENSON	0.197		0.098		
06/18/2011	02:49	RUTLAND COUNTY SO	JEFFREY W STEPHENSON	R		0.098		
06/19/2011	00:02	RUTLAND COUNTY SO	JEFFREY W STEPHENSON					RFI DETECTED
06/19/2011	00:03	RUTLAND COUNTY SO	JEFFREY W STEPHENSON	I	R	0.098		
06/19/2011	00:25	RUTLAND COUNTY SO	JEFFREY W STEPHENSON	0.210		0.098		RESPONSE TIMEOUT
07/03/2011	01:18	RUTLAND COUNTY SO	LISA J FAVRO	I	R	0.096		
07/03/2011	19:05	RUTLAND CITY PD	JUSTIN D SOUZA	0.115		0.096		
07/09/2011	23:38	RUTLAND CITY PD	JOSEPH J BARTLETT	0.173		0.096		
07/10/2011	18:47	RUTLAND COUNTY SO	MILICIA A LYNDS	0.039		0.097		
07/10/2011	18:52	RUTLAND COUNTY SO	MILICIA A LYNDS	0.040		0.096		
07/16/2011	20:29	RUTLAND COUNTY SO	JASON R ALLEN	0.044		0.096		
07/17/2011	02:47	VSP-RUTLAND	CHRISTOPHER G LOYZELLE	0.206		0.096		
07/17/2011	08:21	RUTLAND COUNTY SO	MILICIA A LYNDS	0.118		0.097		
07/18/2011	21:54	RUTLAND CITY PD	JOSEPH J BARTLETT	0.097		0.096		
07/21/2011	21:37	RUTLAND COUNTY SO	LAWRENCE J YNESTA	0.123		0.097		
07/23/2011	05:13	RUTLAND CITY PD	JUSTIN D SOUZA	0.074		0.097		
07/24/2011	00:52	RUTLAND COUNTY SO	JEFFREY W STEPHENSON	0.117		0.097		
07/29/2011	22:50	RUTLAND COUNTY SO	JEFFREY W STEPHENSON	0.070		0.096		
07/31/2011	03:43	RUTLAND CITY PD	JUSTIN D SOUZA			0.095		STANDARD OUT OF RANGE
08/24/2011	23:07	RUTLAND CITY PD	EARL F POST			0.095		STANDARD OUT OF RANGE
08/27/2011	01:47	RUTLAND CITY PD	ED E DUMAS			0.095		STANDARD OUT OF RANGE
09/03/2011	23:30	RUTLAND CITY PD	MATTHEW A PROUTY			0.095		STANDARD OUT OF RANGE

Rutland PP

Breath Record Summary

X = Interference, R = Refusal, V = Invalid, I = Incomplete

Test Date	Test Time	Agency	Operator Name	Alcohol Val. 1	Alcohol Val. 2	Sim. Conc. 1	Sim. Conc. 2	Status Message
09/08/2011	07:40	RUTLAND COUNTY SO	RICHARD B PUTNAM			0.093		STANDARD OUT OF RANGE
10/10/2011	03:00	RUTLAND CITY PD	EARL F POST			0.093		STANDARD OUT OF RANGE
10/27/2011	19:55	RUTLAND COUNTY SO	MISTY L POWELL			0.092		STANDARD OUT OF RANGE

Rutland PD

Status Record Summary

X = Interference, R = Refusal, V = Invalid, I = Incomplete

Test Date	Test Time	Test Type	Status Message
09/21/2010	01:53	Breath Test	RESPONSE TIMEOUT
12/06/2010	21:52	Breath Test	STANDARD OUT OF RANGE
12/10/2010	09:15	Breath Test	STANDARD OUT OF RANGE
12/29/2010	23:01	Breath Test	RESPONSE TIMEOUT
02/19/2011	15:53	Breath Test	RESPONSE TIMEOUT
02/19/2011	15:58	Breath Test	RESPONSE TIMEOUT
02/27/2011	02:06	Breath Test	RESPONSE TIMEOUT
03/07/2011	04:32	Breath Test	RESPONSE TIMEOUT
03/13/2011	04:45	Breath Test	PUMP ERROR
03/13/2011	05:03	Breath Test	PUMP ERROR
03/20/2011	02:23	Breath Test	RESPONSE TIMEOUT
03/27/2011	21:17	Breath Test	RESPONSE TIMEOUT
05/08/2011	21:17	Breath Test	RESPONSE TIMEOUT
06/05/2011	01:59	Breath Test	STANDARD OUT OF RANGE
06/19/2011	00:02	Breath Test	RFI DETECTED
06/19/2011	00:25	Breath Test	RESPONSE TIMEOUT
07/31/2011	03:43	Breath Test	STANDARD OUT OF RANGE
08/24/2011	23:07	Breath Test	STANDARD OUT OF RANGE
10/10/2011	03:00	Breath Test	STANDARD OUT OF RANGE