VERMONT FORENSIC LABORATORY

PO BOX 47
WATERBURY, VERMONT 05676-0047
TEL: 802-244-8788
FAX: 802-241-5557
http://vfl.vermont.gov

Certificate of Analysis of DataMaster Interference Solution

On 5/22/2013, an interference solution was prepared by Rebecca Mead for use as a DataMaster Interference Check solution. The solution was prepared per standard laboratory practice and assigned the lot number 13-05-08M4 and an expiration date of 5/10/2014.

On 5/13/2013 and 5/15/2013, the solution was analyzed against calibration VFL05132013 and SIM05152013. Analysis of the solution was performed using headspace gas chromatography. The instrument was calibrated prior to analyses. It was found to have met required quality assurance performance specifications and was operating properly. Quality checks throughout the analytical process demonstrated that the instrument continued to operate properly.

This solution is certified as presenting 0.080 g/210L simulated breath ethanol 0.04% vol/vol methanol concentration and approved for use as a DataMaster Interference Check solution.

Dated at Waterbury, Vermont on Wednesday, May 22, 2013.

Rebecca Mead

Rebecca Mead