

SOLUTION CHANGE

DataMaster DMT: 100147
Location: VFL
Date: 09/13/2013
Performed by: REBECCA MEAD



Accuracy and Precision Check

Concentration = 0.100 g/210L
Lot # = 13-13-100
Range = 0.095 - 0.105
Average = 0.096 g/210L
Std Dev = 0.0000

Simulator Temperature: 34°C

Performed by

Rebecca Mead

Date

09/13/2013