Digital Thermometer Performance Check

Reference Item	Reference Temperature	Test Temperature	Difference	Pass/Fail*
NIST Traceable QC Thermometer	33.9	34.0	0.1	Pass

* +/- 1 degree Celsius acceptable range

Certified Thermometer Serial Number 91A-4704

Thermometer Serial Number 111814367

Date 9/4/2013

Test Perfomed by Rebecca Mead

Reviewer

VFL Thermometer blank template: public/lab1/QC/QAQCfolder/thermometers