

VERMONT FORENSIC LABORATORY

PO BOX 47  
WATERBURY, VERMONT 05676-0047  
TEL: 802-244-8788  
FAX: 802-241-5557  
<http://vfl.vermont.gov>



**Certificate of Analysis of DataMaster Interference Solution**

On 10/22/2013, an interference solution was prepared by Amanda Bolduc for use as a DataMaster Interference Check solution. The solution was prepared per standard laboratory practice and assigned the lot number 13-05-08A1 and an expiration date of 5/10/2014.

On 5/13/2013 and 5/15/2013, the solution was analyzed against calibration VFL05132013 and SIM05152013. Analysis of the solution was performed using headspace gas chromatography. The instrument was calibrated prior to analyses. It was found to have met required quality assurance performance specifications and was operating properly. Quality checks throughout the analytical process demonstrated that the instrument continued to operate properly.

This solution is certified as presenting 0.080 g/210L simulated breath ethanol and 0.01% (vol/vol) acetone concentration and approved for use as a DataMaster Interference Check solution.

Dated at Waterbury, Vermont on Tuesday, October 22, 2013.

A handwritten signature in black ink, appearing to be "AB", is written over a horizontal line. The signature is stylized and cursive.

Amanda Bolduc