

VERMONT FORENSIC LABORATORY

PO BOX 47
WATERBURY, VERMONT 05676-0047
TEL: 802-244-8788
FAX: 802-241-5557
<http://vfl.vermont.gov>



Certificate of Analysis of DataMaster Interference Solution

On 11/6/2013, an interference solution was prepared by Rebecca Mead for use as a DataMaster Interference Check solution. The solution was prepared per standard laboratory practice and assigned the lot number 13-18-08A1 and an expiration date of 10/25/2014.

On 10/28/2013 and 10/31/2013, the solution was analyzed against calibration VFL10282013 and VFL10312013. Analysis of the solution was performed using headspace gas chromatography. The instrument was calibrated prior to analyses. It was found to have met required quality assurance performance specifications and was operating properly. Quality checks throughout the analytical process demonstrated that the instrument continued to operate properly.

This solution is certified as presenting 0.080 g/210L simulated breath ethanol 0.01% vol/vol acetone concentration and approved for use as a DataMaster Interference Check solution.

Dated at Waterbury, Vermont on Wednesday, November 06, 2013.

A handwritten signature in cursive script, appearing to read "Rebecca Mead", is written over a horizontal line.

Rebecca Mead