



Report of Calibration

Bio-Tek Services, Inc. 5310 South Laburnum Avenue,
 Richmond, VA. 23231 (1-800-792-3625)
 ISO/IEC 17025:2005 Accredited

Calibration Type: As Found

**Vermont Forensic Laboratory -
 Department of Public Safety**

Contact Name: Rebecca L. Mead 103 South Main Street
Location/Dep't: Waterbury, 05671 VT
Phone: (802) 244-8788

Device ID: JJ03213 **Serial#:** JJ03213 **Cal Date:** 21 Jan 2014 **Next Due Date:** Jul 2014
Manufacturer: Finnpiquette, fixed, 250, 1-ch **Size:** 250 ul **Tip used:** Generic Pipette Tip
Method: Level I **Test Plan:** ISO 17025, fixed, 250, 1-ch **Status:** **Passed**
Operator: Al Gunduz **Balance ID:** AD4212B-101 S/N # T0100294 **Certificate#:** 25934-006
Environmental Factors **Weight SN#** BTS-WT-039 S/N 1012016 **Comment:**

Temperature: 19.90 ° C Z Factor: 1.00283 Bar Pressure 752.60 mmHg Density: 0.9982 Humidity: 11.70 % Air Density: 0.00119

SUMMARY STATISTICS

Target Volume (uL)	Channel	Mean Volume	SD	Accuracy %		Precision %		Status	Sample	Sample Wt. (mg)	Actual Volume (uL)
				Actual	Target	Actual	Target				
250.00	1	254.2170	1.1260	1.69	2.00	0.44	2.00	Passed			
Uncertainty = 2.252											
									1	252.4000	253.1140
									2	252.7000	253.4150
									3	254.2000	254.9190
									4	254.7000	255.4200

Signature:

BTS-SOP-002

Template Rev 20 April 2012

Page 1 of 1