



# Report of Calibration

Bio-Tek Services, Inc. 5310 South Laburnum Avenue,  
Richmond, VA. 23231 (1-800-792-3625)  
ISO/IEC 17025:2005 Accredited

Calibration Type: **As Found**

Vermont Forensic Laboratory -  
Department of Public Safety

Contact Name: Rebecca L. Mead 103 South Main Street  
Location/Dep't: Waterbury, 05671  
Phone: (802) 244-8788 VT

Device ID: **HH94382** Serial#: **HH94382** Cal Date: **21 Jan 2014** Next Due Date: **Jul 2014**  
Manufacturer: Finnpiquette, fixed, 500, Size: 500 ul Tip used: Generic Pipette Tip  
1-ch Test Plan: ISO 17025, 500ul, fixed, 1-ch Status: **Passed**  
Method: Level I Balance ID: AD4212B-101 S/N # T0100294 Certificate#: 25934-007  
Operator: Al Gunduz Comment:  
Environmental Factors Weight SN# BTS-WT-039 S/N 1012016

Temperature: 19.90 ° C Z Factor: 1.00283 Bar Pressure 752.60 mmHg Density: 0.9982 Humidity: 11.70 % Air Density: 0.00119

### SUMMARY STATISTICS

Target Volume (uL)	Channel	Mean Volume	SD	Accuracy %		Precision %		Status	Sample	Sample Wt. (mg)	Actual Volume (uL)
				Actual	Target	Actual	Target				
500.00	1	504.8490	1.4400	0.97	2.00	0.29	2.00	Passed	1	505.3000	506.7290
									2	503.3000	504.7230
									3	503.3000	504.7230
									4	501.8000	503.2190

Uncertainty = 2.88

Signature:

BTS-SOP-002

Template Rev 20 April 2012

Page 1 of 1