Vermont Department of Health Laboratory to National Patent Analytical Systems DataMaster DMT Shipping Memo

Sim S/N # G-4536
NIST Temp = 34.0°C
11-30-11
MCS 11/30/11

To:
National Patent Analytical Systems
2090 Harrington Memorial Road
Mansfield, Ohio 44901

From:
State of Vermont Health Department Laboratory
195 Colchester Ave
Burlington, Vermont 05401

Date: 11-29-11

Re:
Shipment of DataMaster DMT S/N: 104009 **for the Vermont Certification Project.**

Instrument Preparation:

1. **Removed Simulator & secured simulator arms with Ty-Wrap.** 11-29-11
2. **Removed breath tube.** 11-30-11
3. **Downloaded instrument files to USB thumb drive.** 11-29-11
4. **Removed keyboard & stylus.** 11-30-11
5. **Removed security screws & replaced with manufacturer supplied screws.** 11-30-11
6. **Instrument bagged & packaged for shipment.**
7. **How many TSIs are included with instrument?** 2

Notes: _____

Shipment in this Box Contains One DataMaster DMT with NO accessories, and one Shipping Memo.

Packed by: _____ 11-30-11 **Witness by:** MCS 11/30/11

Northfield PD

removed 11/29

Breath Record Summary

X = Interference, R = Refusal, V = Invalid, I = Incomplete

Test Date	Test Time	Agency	Operator Name	Alcohol Val. 1	Alcohol Val. 2	Sim. Conc. 1	Sim. Conc. 2	Status Message
12/28/2009	01:17	MIDDLESEX VSP	BENJAMIN J BARTON	0.075		0.096		
03/18/2010	02:17	NORTHFIELD PD	STEPHEN J CLARK	0.094		0.099		
03/19/2010	00:57	NORTHFIELD PD	STEPHEN J CLARK	0.073		0.099		
03/28/2010	03:14	NORTHFIELD PD	STEPHEN J CLARK	0.180	0.178	0.098	0.098	
04/02/2010	23:34	NORTHFIELD PD	STEPHEN J CLARK	0.080	0.090	0.098	0.098	
04/03/2010	05:02	NORTHFIELD PD	STEPHEN J CLARK	0.067		0.099		
04/10/2010	01:52	NORTHFIELD PD	BRIAN W HOAR	0.072		0.099		
04/23/2010	06:43	NORTHFIELD PD	JEFFREY L SHAW	I	0.172	0.098	0.098	
04/25/2010	00:49	NORTHFIELD PD	BRIAN W HOAR	0.079		0.098		
04/25/2010	03:28	WASHINGTON COUNTY SO	CHARLES F SATTERFIELD	R		0.098		
05/09/2010	02:45	NORTHFIELD PD	STEPHEN J CLARK	I	I	0.098	0.097	
05/26/2010	22:22	NORTHFIELD PD	STEPHEN J CLARK	0.174		0.097		
06/05/2010	23:36	NORTHFIELD PD	BRIAN W HOAR	R		0.097		
06/13/2010	22:58	NORTHFIELD PD	BRIAN W HOAR	0.239		0.097		
06/26/2010	00:56	NORTHFIELD PD	STEPHEN J CLARK	0.203		0.097		
07/06/2010	03:33	NORTHFIELD PD	BRIAN W HOAR	0.098		0.097		
07/10/2010	06:31	NORTHFIELD PD	STEPHEN J CLARK	0.115		0.096		
07/16/2010	06:19	NORTHFIELD PD	STEPHEN J CLARK	0.063	0.063	0.096	0.097	
07/23/2010	23:17	NORTHFIELD PD	STEPHEN J CLARK	0.092		0.097		
07/24/2010	01:55	NORTHFIELD PD	STEPHEN J CLARK	0.114		0.096		RESPONSE TIMEOUT
07/24/2010	02:02	NORTHFIELD PD	STEPHEN J CLARK	0.115		0.096		
07/31/2010	03:17	NORTHFIELD PD	STEPHEN J CLARK	0.083	0.083	0.097	0.096	
08/13/2010	02:14	NORTHFIELD PD	STEPHEN J CLARK	0.061		0.096		
08/13/2010	19:27	NORTHFIELD PD	CHARLES F SATTERFIELD	0.224		0.096		
08/19/2010	21:24	NORTHFIELD PD	STEPHEN J CLARK	0.105	0.107	0.096	0.096	
08/28/2010	00:21	NORTHFIELD PD	STEPHEN J CLARK			0.096		RESPONSE TIMEOUT
08/28/2010	00:41	NORTHFIELD PD	STEPHEN J CLARK	I	R	0.095		
08/28/2010	01:08	NORTHFIELD PD	STEPHEN J CLARK	R		0.095		
08/28/2010	01:38	NORTHFIELD PD	STEPHEN J CLARK	0.155	0.158	0.096	0.095	
08/29/2010	03:01	WASHINGTON COUNTY SO	CHARLES F SATTERFIELD	0.128	0.118	0.095	0.095	
09/04/2010	03:23	NORTHFIELD PD	STEPHEN J CLARK	0.102		0.096		
09/25/2010	03:49	NORTHFIELD PD	CHARLES F SATTERFIELD	0.078		0.095		
09/26/2010	02:12	NORTHFIELD PD	CHARLES F SATTERFIELD	0.110	0.110	0.095	0.094	
10/01/2010	16:20	NORTHFIELD PD	JAMES C DZIOBEK	0.034		0.094		
10/03/2010	05:06	NORTHFIELD PD	MICHAEL A ARMANTROUT	R		0.094		
10/16/2010	04:24	NORTHFIELD PD	MICHAEL A ARMANTROUT	0.090	0.089	0.097	0.096	
10/24/2010	01:32	NORTHFIELD PD	CHARLES F SATTERFIELD	R		0.095		
10/29/2010	01:58	NORTHFIELD PD	MICHAEL A ARMANTROUT	R		0.096		
10/30/2010	01:37	NORTHFIELD PD	BRIAN W HOAR	0.153		0.096		
10/31/2010	04:01	NORTHFIELD PD	BRIAN W HOAR	0.064	0.065	0.096	0.096	
11/05/2010	01:08	NORTHFIELD PD	CHARLES F SATTERFIELD	0.067	0.066	0.096	0.096	
11/06/2010	03:49	NORTHFIELD PD	MICHAEL A ARMANTROUT	R		0.095		
11/28/2010	01:45	NORTHFIELD PD	CHARLES F SATTERFIELD	0.182		0.095		
12/04/2010	03:06	NORTHFIELD PD	CHARLES F SATTERFIELD	0.160	I	0.095	0.094	
12/05/2010	02:32	NORTHFIELD PD	MICHAEL A ARMANTROUT	0.127		0.095		
12/18/2010	04:11	NORTHFIELD PD	CHARLES F SATTERFIELD	0.074	0.081	0.095	0.095	
12/18/2010	06:39	NORTHFIELD PD	CHARLES F SATTERFIELD	0.093		0.094		
01/01/2011	03:12	NORTHFIELD PD	CHARLES F SATTERFIELD	0.067	0.068	0.095	0.095	
01/08/2011	20:49	NORTHFIELD PD	CHARLES F SATTERFIELD	0.083		0.095		
01/17/2011	22:11	NORTHFIELD PD	BRIAN W HOAR	V		0.093		
01/17/2011	22:32	NORTHFIELD PD	BRIAN W HOAR					CALIBRATION CHECK ERROR
01/18/2011	08:39	NORTHFIELD PD	CHRISTOPHER M QUESNEL	0.000		0.094		
01/22/2011	02:43	NORTHFIELD PD	CHARLES F SATTERFIELD	0.131	0.128	0.094	0.094	
01/23/2011	02:27	NORTHFIELD PD	MICHAEL A ARMANTROUT	R		0.094		

Northfield PD

Breath Record Summary

X = Interference, R = Refusal, V = Invalid, I = Incomplete

Test Date	Test Time	Agency	Operator Name	Alcohol Val. 1	Alcohol Val. 2	Sim. Conc. 1	Sim. Conc. 2	Status Message
01/28/2011	02:28	NORTHFIELD PD	MICHAEL A ARMANTROUT			0.094		RESPONSE TIMEOUT
01/28/2011	02:38	NORTHFIELD PD	MICHAEL A ARMANTROUT			0.094		RESPONSE TIMEOUT
01/28/2011	02:44	NORTHFIELD PD	MICHAEL A ARMANTROUT	R		0.093		
02/21/2011	02:29	NORTHFIELD PD	BRIAN W HOAR	0.161		0.093		
02/24/2011	08:19	VDHL	STEVEN E HARNOIS	0.000		0.093		
03/12/2011	01:16	NORTHFIELD PD	JEAN-MIGUEL R BARITEAU	0.131	0.135	0.103	0.103	
03/18/2011	02:12	NORTHFIELD PD	JEAN-MIGUEL R BARITEAU	0.143	0.152	0.103	0.102	
04/02/2011	04:18	NORTHFIELD PD	STEPHEN J CLARK	0.071		0.102		
05/06/2011	18:24	NORTHFIELD PD	MICHAEL A ARMANTROUT	0.160	0.164	0.101	0.100	
05/07/2011	03:28	NORTHFIELD PD	CHARLES F SATTERFIELD	0.193		0.101		
05/09/2011	23:16	NORTHFIELD PD	BRIAN W HOAR	R		0.101		
05/15/2011	01:40	NORTHFIELD PD	MICHAEL A ARMANTROUT	0.120		0.101		
05/17/2011	01:37	NORTHFIELD PD	BRIAN W HOAR	0.125	0.130	0.101	0.101	
05/19/2011	21:54	NORTHFIELD PD	STEPHEN J CLARK	0.108		0.100		
05/27/2011	23:45	NORTHFIELD PD	STEPHEN J CLARK	R		0.100		
05/28/2011	00:35	NORTHFIELD PD	MICHAEL A ARMANTROUT	R		0.101		
05/29/2011	00:27	NORTHFIELD PD	MICHAEL A ARMANTROUT	0.199		0.101		
05/30/2011	16:36	NORTHFIELD PD	CHRISTOPHER M QUESNEL	0.092		0.100		
05/31/2011	21:16	NORTHFIELD PD	BRIAN W HOAR	0.112		0.101		
06/15/2011	17:13	NORTHFIELD PD	CHARLES F SATTERFIELD	0.145	0.146	0.101	0.100	
06/18/2011	08:16	NORTHFIELD PD	CHRISTOPHER QUESNEL	0.037		0.100		
06/24/2011	22:06	NORTHFIELD PD	STEPHEN J CLARK	V		0.100		
06/24/2011	22:27	NORTHFIELD PD	STEPHEN J CLARK	0.065		0.099		
07/02/2011	01:35	NORTHFIELD PD	MARK A POULIN	0.108		0.103		
07/03/2011	20:44	NORTHFIELD PD	CHARLES F SATTERFIELD	I	I	0.103		
07/10/2011	20:05	NORTHFIELD PD	MICHAEL A ARMANTROUT	0.107		0.103		
07/17/2011	01:35	NORTHFIELD PD	MARK A POULIN	0.000		0.103		
07/17/2011	03:56	NORTHFIELD PD	CHRIS L WADE	0.158	0.159	0.103	0.102	
07/23/2011	03:11	NORTHFIELD PD	CHRISTOPHER L WADE	0.081	0.075	0.103	0.103	
07/31/2011	02:37	NORTHFIELD PD	STEPHEN J CLARK	0.153		0.102		
07/31/2011	04:59	NORTHFIELD PD	STEPHEN J CLARK	0.064		0.103		RESPONSE TIMEOUT
08/09/2011	01:42	NORTHFIELD PD	STEPHEN J CLARK	0.123		0.102	0.102	RESPONSE TIMEOUT
08/11/2011	04:46	NORTHFIELD PD	BRIAN W HOAR	0.078		0.103		
08/13/2011	21:55	NORTHFIELD PD	MARK A POULIN	0.232		0.102		
08/25/2011	09:08	NORTHFIELD PD	MICHAEL A ARMANTROUT	0.034		0.101		
08/26/2011	20:00	NORTHFIELD PD	MICHAEL A ARMANTROUT	0.183		0.102		
09/25/2011	01:33	NORTHFIELD PD	MARK A POULIN	0.113	0.111	0.100	0.100	
10/01/2011	20:31	MIDDLESEX VSP	BRANDON L DOLL	0.227		0.100		
10/30/2011	16:36	BERLIN PD	STEPHEN J CLARK	0.085		0.099		
11/04/2011	21:39	NORTHFIELD PD	CHARLES F SATTERFIELD	0.068	0.060	0.100	0.100	
11/11/2011	01:12	NORTHFIELD PD	CHRISTOPHER L WADE			0.099		RESPONSE TIMEOUT
11/11/2011	01:43	NORTHFIELD PD	CHRISTOPHER L WADE	0.079		0.099		
11/19/2011	16:45	NORTHFIELD PD	CHRISTOPHER M QUESNEL	0.140		0.098		
11/26/2011	01:20	NORTHFIELD PD	MARK A POULIN	0.088		0.098		
11/26/2011	23:14	NORTHFIELD PD	MICHAEL A ARMANTROUT	0.134		0.099		

Status Record Summary

X = Interference, R = Refusal, V = Invalid, I = Incomplete

Test Date	Test Time	Test Type	Status Message
07/24/2010	01:55	Breath Test	RESPONSE TIMEOUT
08/28/2010	00:21	Breath Test	RESPONSE TIMEOUT
01/17/2011	22:32	Breath Test	CALIBRATION CHECK ERROR
01/28/2011	02:28	Breath Test	RESPONSE TIMEOUT
01/28/2011	02:38	Breath Test	RESPONSE TIMEOUT
07/31/2011	04:59	Breath Test	RESPONSE TIMEOUT
08/09/2011	01:42	Breath Test	RESPONSE TIMEOUT
11/11/2011	01:12	Breath Test	RESPONSE TIMEOUT

Status Record Summary

X = Interference, R = Refusal, V = Invalid, I = Incomplete

Test Date	Test Time	Test Type	Status Message
07/24/2010	01:55	Breath Test	RESPONSE TIMEOUT
08/28/2010	00:21	Breath Test	RESPONSE TIMEOUT
01/17/2011	22:32	Breath Test	CALIBRATION CHECK ERROR
01/28/2011	02:28	Breath Test	RESPONSE TIMEOUT
01/28/2011	02:38	Breath Test	RESPONSE TIMEOUT
07/31/2011	04:59	Breath Test	RESPONSE TIMEOUT
08/09/2011	01:42	Breath Test	RESPONSE TIMEOUT
11/11/2011	01:12	Breath Test	RESPONSE TIMEOUT