Vermont Department of Health Laboratory to National Patent Analytical Systems DataMaster DMT Shipping Memo

Sim# G4545
NIST Temp =
34.0°C
KHK 12-7-11
12/07/11

To:
National Patent Analytical Systems
2090 Harrington Memorial Road
Mansfield, Ohio 44901

From:
State of Vermont Health Department Laboratory
195 Colchester Ave
Burlington, Vermont 05401

Date: 12-7-11

Re:
Shipment of DataMaster DMT S/N: 104409 **for the Vermont Certification Project.**

Instrument Preparation:

1. **Removed Simulator & secured simulator arms with Ty-Wrap.** KHK 12-7-11
2. **Removed breath tube.** KHK 12-7-11
3. **Downloaded instrument files to USB thumb drive.** KHK 12-7-11
4. **Removed keyboard & stylus.** KHK 12-7-11
5. **Removed security screws & replaced with manufacturer supplied screws.** KHK 12-7-11
6. **Instrument bagged & packaged for shipment.** KHK 12-7-11
7. **How many TSIs are included with instrument?** 4

Notes:

Shipment in this Box Contains One DataMaster DMT with NO accessories, and one Shipping Memo.

Packed by: KHK 12-7-11 **Witness by:** SH 12/07/11

Middlesex VSP

Breath Record Summary

X = Interference, R = Refusal, V = Invalid, I = Incomplete

Test Date	Test Time	Agency	Operator Name	Alcohol Val. 1	Alcohol Val. 2	Sim. Conc. 1	Sim. Conc. 2	Status Message
12/23/2009	23:10	MIDDLESEX VSP	BRANDON L DOLL	R		0.096		
12/25/2009	01:29	MIDDLESEX VSP	BENJAMIN J BARTON	0.134		0.096		
12/29/2009	08:35	MIDDLESEX VSP	DANIEL J SCHNEIDER	0.160		0.094	0.090	STANDARD OUT OF RANGE
12/30/2009	20:10	WATERBURY PD	ANTHONY J MAZZILLI			0.096		RESPONSE TIMEOUT
12/30/2009	20:15	WATERBURY PD	ANTHONY J MAZZILLI	0.111		0.096		
01/01/2010	02:49	MIDDLESEX VSP.	BRANDON L DOLL	R		0.095		
01/07/2010	23:42	MIDDLESEX VSP	BRANDON L DOLL	0.186	0.178	0.094	0.093	
01/08/2010	21:04	MIDDLESEX VSP	ROBERT J LUCAS	0.061	0.064	0.094	0.095	
01/11/2010	23:21	MIDDLESEX VSP	CRISTA MAURICE	0.071	0.069	0.095	0.095	
01/15/2010	00:02	MIDDLESEX VSP	BENJAMIN BARTON	0.152	0.153	0.094	0.093	
01/16/2010	19:45	MIDDLESEX VSP	BENJAMIN J BARTON	0.210	0.213	0.095	0.094	
01/16/2010	21:14	WASHINGTON COUNTY SO	CHARLES F SATTERFIELD	0.045		0.094		
01/25/2010	02:10	MIDDLESEX VSP	BRANDON L DOLL	0.096	0.093	0.094	0.094	
01/29/2010	01:05	WATERBURY PD	ANTHONY J MAZZILLI	0.144	0.159	0.093	0.093	
02/20/2010	21:41	MIDDLESEX VSP	ROBERT J LUCAS	0.249	0.246	0.097	0.096	
02/20/2010	22:07	MIDDLESEX VSP	BENJAMIN J BARTON	I	I	0.096		RESPONSE TIMEOUT
02/21/2010	09:08	MIDDLESEX VSP	DANIEL J SCHNEIDER	R		0.097		
03/20/2010	00:35	MIDDLESEX VSP	CRISTA MAURICE	V		0.097		
03/20/2010	00:57	MIDDLESEX VSP	CRISTA MAURICE	0.175	0.173	0.096	0.095	
03/28/2010	00:29	MIDDLESEX VSP	JAMES V FOX	0.154		0.095	0.095	
03/28/2010	01:13	MIDDLESEX VSP	DARRYL M CREMO	0.104		0.096	0.094	STANDARD OUT OF RANGE
04/04/2010	19:39	MIDDLESEX VSP	RAYMOND J RODRIGUEZ	0.065	0.067	0.097	0.096	
04/07/2010	22:47	WATERBURY PD	ANTHONY J MAZZILLI	0.222		0.096		
04/09/2010	05:29	MIDDLESEX VSP	CRISTA MAURICE	0.149		0.095		
04/14/2010	22:50	MIDDLESEX VSP	CRISTA MAURICE	0.261		0.095	0.093	STANDARD OUT OF RANGE
04/18/2010	22:55	MIDDLESEX VSP	JAMES V FOX			0.093		STANDARD OUT OF RANGE
04/18/2010	22:58	MIDDLESEX VSP	JAMES V FOX	0.204		0.095		
04/24/2010	00:47	MIDDLESEX VSP	BENJAMIN J BARTON	0.029		0.095		
04/25/2010	23:41	MIDDLESEX VSP	BRANDON L DOLL	0.140		0.095	0.094	STANDARD OUT OF RANGE
05/02/2010	10:10	MIDDLESEX VSP	DANIEL J SCHNEIDER	0.147		0.095	0.094	STANDARD OUT OF RANGE
05/31/2010	23:25	WATERBURY PD	ADAM J HUBACZ	0.229		0.096		
06/02/2010	02:54	MIDDLESEX VSP	JAMES V FOX	R		0.096		
06/12/2010	01:38	MIDDLESEX VSP	BRANDON L DOLL	0.161		0.096		
06/12/2010	05:16	WATERBURY PD	ADAM J HUBACZ	0.174		0.096		
06/25/2010	18:20	MIDDLESEX VSP	CRISTA MAURICE	0.104		0.094		RESPONSE TIMEOUT
06/26/2010	02:13	MIDDLESEX VSP	CRISTA MAURICE	0.129		0.096		
06/27/2010	22:39	WATERBURY PD	ADAM J HUBACZ	0.100		0.095		
07/03/2010	01:12	MIDDLESEX VSP	DARRYL M CREMO	R		0.095		
07/10/2010	23:16	MIDDLESEX VSP	RYAN J GREEN	0.045		0.094		
07/16/2010	17:32	MIDDLESEX VSP	TROOPER D TROTTIER	0.287		0.094		RESPONSE TIMEOUT
08/03/2010	21:37	MIDDLESEX VSP	RAYMOND G LEBLANC			0.093		RESPONSE TIMEOUT
08/03/2010	22:03	MIDDLESEX VSP	RAYMOND G LEBLANC	0.158		0.093		
08/07/2010	02:30	MIDDLESEX VSP	BRANDON L DOLL	0.074		0.094		
08/07/2010	23:34	MIDDLESEX VSP	BRANDON L DOLL			0.092		STANDARD OUT OF RANGE
08/07/2010	23:36	MIDDLESEX VSP	BRANDON L DOLL	0.189		0.093	0.092	STANDARD OUT OF RANGE
08/12/2010	00:04	MIDDLESEX VSP	DANIEL A TROTTIER	0.094		0.099		
08/12/2010	10:24	VDHL	STEVEN E HARNOIS	0.000		0.097		
08/14/2010	00:20	MIDDLESEX VSP	JAMES V FOX	0.128		0.099		
08/14/2010	08:21	MIDDLESEX VSP	DARRYL CREMO	R		0.101		
08/14/2010	17:55	MIDDLESEX VSP	JAMES V FOX	R		0.100		
08/15/2010	00:45	MIDDLESEX VSP	BRANDON L DOLL	0.082	0.082	0.100	0.100	
08/18/2010	20:47	WATERBURY PD	ANTHONY J MAZZILLI	0.116		0.100		
09/04/2010	22:25	WATERBURY PD	ANTHONY J MAZZILLI	0.194		0.100		
09/17/2010	01:08	MIDDLESEX VSP	BRANDON L DOLL	0.191		0.099		

Breath Record Summary

X = Interference, R = Refusal, V = Invalid, I = Incomplete

Test Date	Test Time	Agency	Operator Name	Alcohol Val. 1	Alcohol Val. 2	Sim. Conc. 1	Sim. Conc. 2	Status Message
09/27/2010	22:01	MIDDLESEX VSP	CRISTA MAURICE	0.178		0.098		
10/03/2010	00:42	MIDDLESEX VSP	JESSE J LAWLISS	0.081		0.098		RESPONSE TIMEOUT
10/03/2010	21:31	MIDDLESEX VSP	BENJAMIN J BARTON	0.166	0.163	0.097	0.097	
10/07/2010	01:23	MIDDLESEX VSP	BRANDON L DOLLQ	0.049		0.098		
10/10/2010	06:55	MIDDLESEX VSP	BENJAMIN J BARTON	0.097		0.100		RESPONSE TIMEOUT
10/19/2010	16:00	VDHL	STEVEN E HARNOIS	0.000		0.098		
10/19/2010	16:19	VDHL	STEVEN E HARNOIS	0.000		0.099		
10/28/2010	23:49	WATERBURY PD	ADAM J HUBACZ	0.079		0.099		
10/29/2010	23:54	WATERBURY PD	ADAM J HUBACZ	0.066		0.099		
10/31/2010	00:53	MIDDLESEX VSP	BRANDON L DOLL	R		0.098		
11/05/2010	19:17	MIDDLESEX VSP	RYAN J GREEN	0.096	0.096	0.098	0.098	
11/18/2010	02:03	MIDDLESEX VSP	MATTHEW H SWEITZER	0.147		0.098		
12/01/2010	01:07	MIDDLESEX VSP	RICHARD W CONLON	0.072		0.098		
12/11/2010	00:41	MIDDLESEX VSP	ROBERT SNETSINGER	0.133		0.097		
12/12/2010	13:03	MIDDLESEX VSP	JAMES V FOX	0.139		0.096		
12/14/2010	05:35	MIDDLESEX VSP	RYAN J GREEN	I	I	0.097		
12/15/2010	21:55	WATERBURY PD	ANTHONY J MAZZILLI	0.185		0.097		
12/17/2010	12:30	VDHL	STEVEN E HARNOIS	0.000		0.096		
12/18/2010	00:18	MIDDLESEX VSP	BRANDON L DOLL	0.169	0.167	0.097	0.097	
12/25/2010	02:25	MIDDLESEX VSP	CRISTA MAURICE	0.130	0.134	0.096	0.095	
12/30/2010	01:06	MIDDLESEX VSP	MATTHEW H SWEITZER	0.010	0.009	0.096	0.097	
01/07/2011	00:39	MIDDLESEX VSP	RICHARD W CONLON			0.097		RESPONSE TIMEOUT
01/07/2011	00:59	MIDDLESEX VSP	RICHARD W CONLON	0.163	0.168	0.096	0.096	
01/07/2011	21:25	MIDDLESEX VSP	BRANDON L DOLL	0.087		0.096		
01/17/2011	15:51	MIDDLESEX VSP	RYAN J GREEN	0.000	0.000	0.096	0.096	
01/24/2011	19:40	MIDDLESEX VSP	AMBER M POULIOT	0.184		0.096		RESPONSE TIMEOUT
01/28/2011	00:24	MIDDLESEX VSP	BRANDON L DOLL	R		0.096		
01/29/2011	02:07	MIDDLESEX VSP	RICHARD W CONLON	0.132	0.135	0.096	0.095	
02/02/2011	19:23	MIDDLESEX VSP	RYAN J GREEN			0.100		RFI DETECTED
02/03/2011	22:46	MIDDLESEX VSP	RYAN J GREEN	0.074		0.098		
02/14/2011	16:47	MIDDLESEX VSP	DARRYL M CREMO	0.049		0.099		
02/15/2011	17:12	MIDDLESEX VSP	DANIEL A TROTTIER	0.015		0.100		
02/20/2011	13:20	MIDDLESEX VSP	BENJAMIN J BARTON	0.143	0.137	0.098	0.098	
02/28/2011	22:28	MIDDLESEX VSP	RICHARD W CONLON	0.126	0.127	0.098	0.098	
03/01/2011	21:37	MIDDLESEX VSP	RYAN J GREEN	0.155		0.098		
03/04/2011	17:55	MIDDLESEX VSP	DANIEL A TROTTIER			0.098		RESPONSE TIMEOUT
03/04/2011	17:59	MIDDLESEX VSP	DANIEL A TROTTIER	R		0.098		
03/07/2011	17:31	MIDDLESEX VSP	DANIEL A TROTTIER	R		0.098		
03/12/2011	00:02	MIDDLESEX VSP	ALEX A COMTOIS	0.089	0.089	0.098	0.097	
03/18/2011	15:26	MIDDLESEX VSP	JERRY PARTIN	R		0.098		
03/19/2011	00:54	MIDDLESEX VSP	BENJAMIN J BARTON	0.096		0.098		
03/26/2011	23:54	MIDDLESEX VSP	BRANDON L DOLL	0.082	0.079	0.097	0.096	
04/03/2011	02:47	MIDDLESEX VSP	BRANDON L DOLL	0.126		0.097		
04/08/2011	23:40	MIDDLESEX VSP	BRANDON L DOLL	0.257		0.097		
04/12/2011	23:45	MIDDLESEX VSP	RICHARD W CONLON	0.166	0.158	0.097	0.097	
04/22/2011	16:45	MIDDLESEX VSP	DANIEL A TROTTIER			0.097		RESPONSE TIMEOUT
04/26/2011	14:13	MIDDLESEX VSP	D CREMO	R		0.096		
04/26/2011	14:33	MIDDLESEX VSP	DARRYL M CREMO	0.309	0.318	0.096	0.096	
04/29/2011	08:47	MIDDLESEX VSP	AMBER POULIOT	0.125	0.124	0.095	0.095	
05/17/2011	01:23	MIDDLESEX VSP	ALEX A COMTOIS	0.143		0.096		
05/21/2011	10:56	VSP MIDDLESEX	DANIEL J SCHNEIDER	0.182		0.096		
06/09/2011	23:57	MIDDLESEX VSP	CRISTA MAURICE	0.060		0.098		
06/16/2011	14:39	MIDDLESEX VSP	DANIEL TROTTIER	0.133		0.096		
06/16/2011	23:49	MIDDLESEX VSP	BRANDON L DOLL	0.100		0.098		

Breath Record Summary

X = Interference, R = Refusal, V = Invalid, I = Incomplete

Test Date	Test Time	Agency	Operator Name	Alcohol Val. 1	Alcohol Val. 2	Sim. Conc. 1	Sim. Conc. 2	Status Message
06/18/2011	00:55	MIDDLESEX VSP	BENJAMIN BARTON	R		0.098		
06/21/2011	09:12	MIDDLESEX VSP	DANIEL A TROTTIER	0.165		0.098		
06/22/2011	23:05	MIDDLESEX VSP	BRANDON L DOLL	0.094		0.098		
06/23/2011	17:09	MIDDLESEX VSP	BENJAMIN J BARTON	R		0.096		
06/23/2011	17:59	MIDDLESEX VSP	BENJAMIN J BARTON	0.040	0.041	0.097	0.098	
06/25/2011	01:15	VERMONT STATE POLICE	ALEX A COMOTIS	0.138	0.139	0.098	0.097	
07/02/2011	17:46	MIDDLESEX VSP	JERRY PARTIN	0.239	0.226	0.101	0.100	
07/05/2011	17:06	VERMONT STATE POLICE	ALEX A COMTOIS			0.100		RESPONSE TIMEOUT
07/05/2011	17:29	VERMONT STATE POLICE	ALEX A COMTOIS	R		0.100		
07/05/2011	18:00	VERMONT STATE POLICE	ALEX A COMTOIS	0.194		0.100		
07/05/2011	23:35	MIDDLESEX VSP	BRANDON L DOLL	0.138		0.102		
07/09/2011	15:28	MIDDLESEX VSP	RICHARD W CONLON	0.161		0.099		
07/09/2011	22:37	MIDDLESEX VSP	MICHAEL S BRUCE	R		0.099		
07/09/2011	22:58	MIDDLESEX VSP	RYAN J GREEN			0.099		RESPONSE TIMEOUT
07/15/2011	02:06	MIDDLESEX VSP	BENJAMIN J BARTON	0.142		0.103		
07/24/2011	00:04	VERMONT STATE POLICE	ALEX A COMTOIS	0.068		0.100		
07/24/2011	23:56	MIDDLESEX VSP	BRANDON L DOLL	0.094		0.102		
08/01/2011	21:28	MIDDLESEX VSP	BENJAMIN J BARTON	0.094		0.101		
08/03/2011	23:51	MIDDLESEX VSP	RYAN J GREEN	0.145	0.139	0.100	0.100	
08/05/2011	22:14	MIDDLESEX VSP	BENJAMIN J BARTON	0.177		0.100		
08/10/2011	18:43	MIDDLESEX VSP	BRANDON L DOLL	0.190	0.196	0.099	0.098	
08/19/2011	00:03	WATERBURY PD	ANTHONY J MAZZILLI			0.097		RESPONSE TIMEOUT
08/19/2011	00:42	WATERBURY PD	ANTHONY J MAZZILLI	0.067		0.099		
08/31/2011	00:58	MIDDLESEX VSP	TREVOR CARBO	0.155	0.149	0.100	0.098	
09/01/2011	17:28	MIDDLESEX VSP	RAYMOND G LEBLANC	0.251		0.098		
09/08/2011	00:06	MIDDLESEX VSP	BRANDON L DOLL	0.077		0.099		
09/29/2011	10:33	VDHL	STEVEN E HARNOIS	0.000		0.099		
10/02/2011	08:42	MIDDLESEX VSP	DANIEL A TROTTIER	0.071		0.098		
10/05/2011	14:57	MIDDLESEX VSP	RICHARD W CONLON	0.068	0.071	0.099	0.099	
10/12/2011	20:06	VERMONT STATE POLICE	ALEX A COMTOIS	0.198		0.104		
10/14/2011	23:42	MIDDLESEX VSP	SAMUEL W LUND	R		0.105		
10/17/2011	20:59	MIDDLESEX VSP	CRISTA MAURICE			0.104		RESPONSE TIMEOUT
10/17/2011	21:02	MIDDLESEX VSP	CRISTA MAURICE	R		0.104		
10/21/2011	00:47	WILLISTON VSP	DUSTIN P ROBINSON	0.113	0.115	0.104	0.104	
10/22/2011	20:41	VERMONT STATE POLICE	ALEX A COMTOIS	0.165		0.103		
10/30/2011	07:49	MIDDLESEX VSP	DANIEL A TROTTIER	0.100		0.102		
11/03/2011	20:37	MIDDLESEX VSP	MICHAEL T MUISE	0.101		0.103		
11/05/2011	20:09	MIDDLESEX VSP	MATTHEW H SWEITZER	0.072		0.102		
11/11/2011	23:35	MIDDLESEX VSP	TREVOR J CARBO	0.167		0.102		RESPONSE TIMEOUT
12/04/2011	00:06	WASHINGTON COUNTY SO	ANDREW R BENT	0.094		0.101		

Status Record Summary

X = Interference, R = Refusal, V = Invalid, I = Incomplete

Test Date	Test Time	Test Type	Status Message
12/29/2009	08:35	Breath Test	STANDARD OUT OF RANGE
12/30/2009	20:10	Breath Test	RESPONSE TIMEOUT
02/20/2010	22:07	Breath Test	RESPONSE TIMEOUT
03/28/2010	01:13	Breath Test	STANDARD OUT OF RANGE
04/14/2010	22:50	Breath Test	STANDARD OUT OF RANGE
04/18/2010	22:55	Breath Test	STANDARD OUT OF RANGE
04/25/2010	23:41	Breath Test	STANDARD OUT OF RANGE
05/02/2010	10:10	Breath Test	STANDARD OUT OF RANGE
06/25/2010	18:20	Breath Test	RESPONSE TIMEOUT
07/16/2010	17:32	Breath Test	RESPONSE TIMEOUT
08/03/2010	21:37	Breath Test	RESPONSE TIMEOUT
08/07/2010	23:34	Breath Test	STANDARD OUT OF RANGE
08/07/2010	23:36	Breath Test	STANDARD OUT OF RANGE
10/03/2010	00:42	Breath Test	RESPONSE TIMEOUT
10/10/2010	06:55	Breath Test	RESPONSE TIMEOUT
01/07/2011	00:39	Breath Test	RESPONSE TIMEOUT
01/24/2011	19:40	Breath Test	RESPONSE TIMEOUT
03/04/2011	17:55	Breath Test	RESPONSE TIMEOUT
04/22/2011	16:45	Breath Test	RESPONSE TIMEOUT
07/05/2011	17:06	Breath Test	RESPONSE TIMEOUT
07/09/2011	22:58	Breath Test	RESPONSE TIMEOUT
08/19/2011	00:03	Breath Test	RESPONSE TIMEOUT
10/17/2011	20:59	Breath Test	RESPONSE TIMEOUT
11/11/2011	23:35	Breath Test	RESPONSE TIMEOUT