ROUTINE PERFORMANCE CHECK REPORT

DataMaster DMT: 300151
Location: Wilmington PD
Calibration Date: 12/07/2011
Certification Date: 12/08/2011
Installation Date: 08/10/2012
RPC Date: 11/01/2012
Supervisor Name: MARK J DENAULT

Diagnostic Results

```
VERSIONS
DMT: 1.01
PIC: 2.06
Modem: 2.4
Questions: 2.1

TEMPERATURES

Sample Chamber = 49.0°C
Breath Tube    = 41.6°C
Digital Sim    = 34.0°C

SETTINGS
Lamp Voltage   = 1.71 V
Cooler Voltage = 1.74 V
Bias Voltage   = 80 V
Chopper Freq   = 544 Hz

PUMP INFO
Flow Rate = 4.786 L/M

DETECTOR INFO
PUMP     ON        OFF
MAX(V)   0.3594    0.3620
MIN(V)   0.3579    0.3604

FILTER INFO
Filter 1   0.361   Zero = true
Filter 2   1.090   Zero = true
Filter 3   0.646   Zero = true

CALIBRATION CHECK
Xq  = 0.088  0.79%
```

RCD

Routine Performance Check Passed

Accuracy and Precision Check

Concentration	= 0.099 g/210L
Lot #	= 12-80-100
Range	= 0.094 - 0.104
Average	= 0.098 g/210L
Std Dev	= 0.0000

RF Detection Test

Passed

Sample Acceptance Test

Passed



Performed by _____ Date 11/01/2012

Reviewed by _____ Date 11/5/12

11/7/2012