

SOLUTION CHANGE

DataMaster DMT: 100169
Location: Bellows Falls PD
Date: 04/29/2013
Performed by: MICHAEL J KEEFE



Accuracy and Precision Check

Concentration = 0.099 g/210L
Lot # = 12-77-100
Range = 0.094 - 0.104
Average = 0.102 g/210L
Std Dev = 0.0000

Simulator Temperature: 34.1°C

Performed by

M. J. Keefe

Date

06/11/2013