DATAMASTER DMT MAINTENANCE LOG

Bellows Falls
Department/Agency

100169
DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
9/7/11	**DataMaster Ext. Std. Solution** Lot # 11-41-100 $EtOH_{aq}$ Conc: 0.101 Range: 0.096 - 0.106 Preparation Date: 06/29/2011 Expires: 06/29/2012 [nge Place Label Here] OPENED: 9/7/11 Installed Instrument Per Protocol SBM 09/07/11	Steven Hornous
9/30/11	[For RPC or Simulator Solution Change Place Label Here] Updated Software	A. Belle Gadon
11/7/11	Software update Installed 11/7 [For RPC or Simulator Solution Change Place Label Here]	Prog. Chief Org Chem VDH&L
11/11/11	RPC All OK [For RPC or Simulator Solution Change Place Label Here] Ext .098	[illegible]

ALC 803 REV 03 08/2011

DATAMASTER DMT MAINTENANCE LOG

Department/Agency

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
02-01-12	**DataMaster Simulator Solution** Lot # 12-32-100 $EtOH_{aq}$ Conc: 0.100 Range: 0.095 - 0.105 Preparation Date: 10/26/2011 Expires: 10/26/2012 hange Place Label Here] OPENED: 02-01-12 RPC 02-12 EXT. 101 SIM SOLUTION CHANGE LAST DUI TEST FAILED FOR SOLUTION OUT OF RANGE	OFC. M. KEEFE
04-16-12 ~~06-18-12~~	**DataMaster Ext. Std. Solution** Lot # 11-40-100 $EtOH_{aq}$ Conc: 0.100 Range: 0.095 - 0.105 Preparation Date: 06/15/2011 Expires: 06/15/2012 iange Place Label Here] SIM SOLUTION CHANGE EXT. 102 LAST CHECK IN TEST FAILED FOR SOLUTION OUT OF RANGE	OFC. M. KEEFE
06-19-12	**DataMaster Simulator Solution** Lot # 12-72-100 EtOHaq Conc: 0.100 [Range: 0.095 - 0.105 Preparation Date: 04/24/2012 Expires: 04/24/2013 nge Place Label Here] OPENED: 06-19-12 RPC, SIM SOLUTION CHANGE EXT. 102 SOLUTION OUT OF RANGE ON 6/15/12 PUMP ERROR ON 06-13-12 - ALL OK	OFC. M. KEEFE
8/23/12	APM Completed + Software update [For RPC or Simulator Solution Change Place Label Here]	R. Driscoll

ALC 803 REV 03 08/2011

DATAMASTER DMT MAINTENANCE LOG

Department/Agency

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
10-02-12	**DataMaster Simulator Solution** Lot # 12-58-100 EtOH$_{aq}$ Conc: 0.100 Range: 0.095 - 0.105 Preparation Date: 02/09/2012 Expires: 02/09/2013 ange Place Label Here] OPENED: 10-02-2012 RPC, SIM SOLUTION CHANGE FIRST RPC FAILED - SOLUTION JAR NOT TIGHT ENOUGH SECOND RPC - OK. EXT. 102	OFC. M. KEEFE
01-28-13	**DataMaster Simulator Solution** Lot # 12-72-100 EtOHaq Conc: 0.100 Range: 0.095 - 0.105 Preparation Date: 04/24/2012 Expires: 04/24/2013 hange Place Label Here] OPENED: 01-28-13 SIM SOLUTION CHANGE EXT. 102 LAST DATE TEST FAILED SIM SOLUTION OUT OF RANGE ON 01/26/13	OFC. M. KEEFE
02-04-13	**DataMaster Simulator Solution** Lot # 12-84-100 EtOHaq Conc: 0.101 Range: 0.096 - 0.106 Preparation Date: 9/24/2012 Expires: 9/24/2013 ange Place Label Here] OPENED: 02-04-13 RPC, SIM SOLUTION CHANGE ALL OK EXT. 101	OFC. M. KEEFE
04-29-13	**DataMaster Simulator Solution** Lot # 12-77-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 05/08/2012 Expires: 05/08/2013 hange Place Label Here] OPENED: 04-29-13 SIM SOLUTION CHANGE EXT. 102 SOLUTION WAS @ LOW END OF RANGE PRIOR TO CHANGE.	OFC. M. KEEFE

ALC 803 REV 03 08/2011

DATAMASTER DMT MAINTENANCE LOG

Bellows Falls PD
Department/Agency

100169
DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
06-11-13	**DataMaster Simulator Solution** Lot # 12-84-100 EtOHaq Conc: 0.101 Range: 0.096 - 0.106 Preparation Date: 9/24/2012 Expires: 9/24/2013 hange Place Label Here] OPENED: 06-11-13 Solution Change Ext .102 June RPC All ok	Ofc. M. Keefe
7/10/13	Completed APM [For RPC or Simulator Solution Change Place Label Here]	R. Donnell DMT Tech.
10/02/13	**DataMaster Simulator Solution** Lot # 12-89-100 EtOHaq Conc: 0.097 Range: 0.092 - 0.102 Preparation Date: 11/21/2012 Expires: 11/21/2013 hange Place Label Here] OPENED: 10/02/13 Ext .102 Solution Change / Sim Solution was reported out of range on 9/30/13 Oct RPC All ok	Ofc M. Keefe
02/05/14	**DataMaster Simulator Solution** Lot # 13-19-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 10/25/2013 Expires: 10/25/2014 hange Place Label Here] OPENED: 02-05-14 Feb RPC / Ext .101 Solution Change All ok	Ofc. M. Keefe

ALC 803 REV 03 08/2011

DATAMASTER DMT MAINTENANCE LOG

Bellows Falls PD
Department/Agency

100169
DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
06/06/14	**DataMaster Simulator Solution** Lot # 13-19-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 10/25/2013 Expires: 10/25/2014 hange Place Label Here] OPENED: 06-06-14 EXT. 103 SOLUTION CHANGE JUNE RPC ALL OK	OFC. M. Keefe
7/21/14	Completed ARM Protocol [For RPC or Simulator Solution Change Place Label Here]	R. Driscoll DMT TECH.
10/05/14	**DataMaster Simulator Solution** Lot # 13-21-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 12/20/2013 Expires: 12/20/2014 Change Place Label Here] OPENED: 10/05/14 SOLUTION CHANGE OCT RPC - FIRST TEST FAILED - TESTER ERROR ON SHALLOW BREATH TEST - SECOND TEST ALL OK EXT. 101	OFC. M. Keefe
01/28/15	**DataMaster Simulator Solution** Lot # 14-10-100 EtOHaq Conc: 0.097 Range: 0.092 - 0.102 Preparation Date: 7/14/2014 Expires: 7/14/2015 hange Place Label Here] OPENED: 01-28-15 FEB RPC ALL OK SOLUTION CHANGE EXT. 101	OFC. M. Keefe

ALC 803 REV 03 08/2011

DATAMASTER DMT MAINTENANCE LOG

Department/Agency

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
06/03/15	**DataMaster Simulator Solution** Lot # 14-10-100 EtOHaq Conc: 0.097 Range: 0.092 - 0.102 Preparation Date: 7/14/2014 Expires: 7/14/2015 hange Place Label Here] OPENED: 06-03-15 JUNE 2015 RPC ALL OK SOLUTION CHANGE EXT .102	CPL M. KEEFE
	[For RPC or Simulator Solution Change Place Label Here]	
	[For RPC or Simulator Solution Change Place Label Here]	
	[For RPC or Simulator Solution Change Place Label Here]	

ALC 803 REV 03 08/2011