Breath Record History

Test Date	Test Time	Agency	Operator Name	Alcohol Val. 1	Alcohol Val. 2	Sim. Conc. 1	Sim. Conc. 2	Status Message
10/09/2011	02:52:12	HARTFORD PD	ERIC J CLIFFORD	0.119	0.116	0.097	0.098	
10/09/2011	13:01:08	WOODSTOCK PD	LISA B WILDING	0.078		0.097		
10/13/2011	02:36:53	HARTFORD PD	FREDERICK C PEYTON	0.058		0.097		
10/21/2011	22:48:23	ROYALTON VSP	ANDREW COLLIER			0.095		Simulator Out of Range
10/21/2011	22:51:49	ROYALTON VSP	ANDREW COLLIER	0.154	0.154	0.096	0.097	
10/23/2011	23:34:57	HARTFORD PD	ERIC J CLIFFORD			0.095		Simulator Out of Range
10/23/2011	23:37:54	HARTFORD PD	ERIC J CLIFFORD	0.111		0.096	0.095	Simulator Out of Range
10/23/2011	23:44:21	HARTFORD PD	ERIC J CLIFFORD	0.108		0.096		
11/10/2011	22:28:22	ROYALTON VSP	STACY CORLISS	0.133	0.146	0.096	0.096	
01/14/2012	01:43:45	WOODSTOCK PD	JENNIFER HUTCHINS			0.086		Simulator Out of Range
01/14/2012	01:46:08	WOODSTOCK PD	JENNIFER HUTCHINS			0.086		Simulator Out of Range
01/14/2012	01:48:52	WOODSTOCK PD	JENNIFER HUTCHINS			0.086		Simulator Out of Range
01/14/2012	02:17:39	woodstock	JENNIFER HUTCHINS			0.085		Simulator Out of Range
01/14/2012	21:57:38	WOODSTOCK PD	JENNIFER HUTCHINS			0.084		Simulator Out of Range
01/16/2012	20:44:17	ROYALTON VSP	ANDREW COLLIER	0.069	0.068	0.097	0.097	
01/16/2012	21:18:45	ROYALTON VSP	ANDREW COLLIER			0.094		Simulator Out of Range
01/16/2012	21:21:29	ROYALTON VSP	ANDREW COLLIER	INTERFERENCE		0.095		
01/16/2012	21:44:30	ROYALTON VSP	ANDREW COLLIER	0.134		0.096		
02/03/2012	03:19:03	HARTFORD PD	JON S KUSTAFIK	0.086	0.086	0.097	0.096	
02/03/2012	15:25:15	HARTFORD PD	CHRISTOPHER C OKEEFFE	INTERFERENCE		0.096		
02/03/2012	15:53:04	HARTFORD PD	CHRISTOPHER C OKEEFFE	INTERFERENCE		0.095		
02/17/2012	17:31:33	WOODSTOCK PD	JOSEPH B SWANSON	0.000		0.097		
03/17/2012	17:37:47	WOODSTOCK PD	MARK J FIORE					RF Detected
03/17/2012	17:40:46	WOODSTOCK PD	MARK J FIORE	0.115		0.099		
04/01/2012	00:20:52	Woostock PD	PETER J MANTELLO	0.117		0.097		
04/05/2012	04:17:47	WOODSTOCK PD	JOSEPH B SWANSON	INCOMPLETE	0.138	0.097	0.095	
04/07/2012	06:47:10	WOODSTOCK PD	MARK J FIORE			0.094		Simulator Out of Range
04/07/2012	06:50:16	WOODSTOCK PD	MARK J FIORE	0.097	0.096	0.096	0.095	
04/24/2012	14:34:47	WOODSTOCK PD	JENNIFER HUTCHINS	0.000		0.096		
06/09/2012	02:22:12	ROYALTON VSP	ANDREW COLLIER	0.111	0.120	0.097	0.097	
06/20/2012	15:00:33	WOODSTOCK PD	JOSEPH B SWANSON	0.243	0.251	0.098	0.097	
06/22/2012	22:03:51	WINDSOR COUNTY SO	THOMAS P BATTISTA	0.082		0.096		RESPONSE TIMEOUT
06/24/2012	00:49:19	WOODSTOCK PD	PETER J MANTELLO	0.109	0.110	0.095	0.096	
08/13/2012	21:09:35	WOODSTOCK PD	PETER J MANTELLO			0.092		Simulator Out of Range
08/17/2012	20:09:10	WOODSTOCK PD	JOSEPH B SWANSON	0.125		0.096		
08/26/2012	05:45:35	woodstock police dept	PETER J MANTELLO	0.129		0.096		
09/20/2012	15:47:25	WOODSTOCK PD	JENNIFER HUTCHINS	0.000		0.096		
09/27/2012	16:08:42	a2978	JENNIFER HUTCHINS	0.000		0.096		
09/27/2012	18:10:15	WINDSOR COUNTY SO	ERIK P ROBINSON	REFUSED		0.095		
09/27/2012	18:32:06	WINDSOR COUNTY SO	ERIK P ROBINSON	0.150	0.155	0.094	0.095	Pump Error
09/28/2012	10:43:15	WOODSTOCK PD	MARK J FIORE	0.112		0.095		
10/05/2012	23:39:19	WOODSTOCK PD	JENNIFER HUTCHINS	0.153		0.096	0.095	
10/06/2012	12:51:03	WINDSOR COUNTY SO	TYLER J TROMBLEY			0.096		RESPONSE TIMEOUT
10/06/2012	12:55:43	WINDSOR COUNTY SO	TYLER J TROMBLEY	REFUSED		0.096		
10/07/2012	06:59:06	HARTFORD PD	DANIEL C SOLOMITA	0.120	0.114	0.094	0.094	
10/09/2012	23:56:04	WOODSTOCK PD	JENNIFER HUTCHINS			0.092		Simulator Out of Range
10/09/2012	23:59:02	WOODSTOCK PD	JENNIFER HUTCHINS	0.207	INCOMPLETE	0.093	0.093	
10/18/2012	18:31:18	WOODSTOCK PD	JOSEPH B SWANSON	0.083		0.099		
10/30/2012	22:13:49	woodstock	JOSEPH B SWANSON	0.111		0.099		CALIBRATION CHECK ERROR
10/30/2012	22:27:42	woodstock	JOSEPH B SWANSON			0.099		RESPONSE TIMEOUT
10/30/2012	23:02:21	woodstock	JOSEPH B SWANSON	0.111		0.099		
11/03/2012	17:46:01	WINDSOR COUNTY SO	JONATHAN G NORTON			0.096		RESPONSE TIMEOUT
11/10/2012	01:43:24	WOODSTOCK PD	WILLIAM E HANSON			0.096		RESPONSE TIMEOUT
11/10/2012	01:55:42	WOODSTOCK PD	WILLIAM E HANSON	0.134		0.098		RESPONSE TIMEOUT
11/10/2012	02:02:55	WOODSTOCK PD	WILLIAM E HANSON	0.137		0.096		
11/18/2012	01:17:25	Woosdstock PD	CYNTHIA P PARKER			0.097		RESPONSE TIMEOUT
01/11/2013	03:44:35	WOODSTOCK PD	JOSEPH B SWANSON	REFUSED		0.094		
02/17/2013	00:48:18	WOODSTOCK PD	PETER J MANTELLO			0.095		Pump Error
02/17/2013	00:51:32	WOODSTOCK PD	PETER J MANTELLO	INCOMPLETE	0.081	0.094		RESPONSE TIMEOUT
03/09/2013	08:24:30	WINDSOR COUNTY SO	TYLER J TROMBLEY			0.094		RESPONSE TIMEOUT
03/09/2013	08:53:34	WINDSOR COUNTY SO	TYLER J TROMBLEY	REFUSED		0.093		
03/15/2013	04:23:32	hartford	LAURA S TARBELL	0.209	0.197	0.095	0.094	
03/17/2013	02:30:23	HARTFORD PD	DANIEL C SOLOMITA	0.099		0.095		
03/17/2013	02:53:12	HARTFORD PD	DANIEL C SOLOMITA	0.091		0.095		

Status Record History

Status Date	Status Time	Test Type	Status Message
09/14/2011	14:04:42	Accuracy & Precision Records	Simulator Temperature Too Low
09/16/2011	11:44:32	Accuracy & Precision Records	Simulator Timed Out
10/21/2011	22:49:51	DUI Subject Record	Simulator Out of Range
10/23/2011	23:36:24	DUI Subject Record	Simulator Out of Range
10/23/2011	23:42:51	DUI Subject Record	Simulator Out of Range
01/14/2012	01:45:14	DUI Subject Record	Simulator Out of Range
01/14/2012	01:47:39	DUI Subject Record	Simulator Out of Range
01/14/2012	01:50:24	DUI Subject Record	Simulator Out of Range
01/14/2012	02:19:12	DUI Subject Record	Simulator Out of Range
01/14/2012	21:59:10	DUI Subject Record	Simulator Out of Range
01/16/2012	21:20:13	DUI Subject Record	Simulator Out of Range
02/05/2012	17:50:56	RPC Record	Diagnostic Check Failed
02/16/2012	12:15:28	RPC Record	RF NOT DETECTED
03/17/2012	17:39:08	DUI Subject Record	RF Detected
04/07/2012	06:48:38	DUI Subject Record	Simulator Out of Range
06/02/2012	16:37:57	RPC Record	RF NOT DETECTED
06/22/2012	22:08:44	DUI Subject Record	RESPONSE TIMEOUT
08/13/2012	21:11:04	DUI Subject Record	Simulator Out of Range
08/13/2012	22:15:04	Solution Change Record	ACCURACY AND PRECISION CHECK FAILED
09/27/2012	18:38:29	DUI Subject Record	Pump Error
10/06/2012	12:54:15	DUI Subject Record	RESPONSE TIMEOUT
10/09/2012	23:57:32	DUI Subject Record	Simulator Out of Range
10/30/2012	22:19:40	DUI Subject Record	CALIBRATION CHECK ERROR
10/30/2012	22:30:54	DUI Subject Record	RESPONSE TIMEOUT
11/03/2012	17:49:13	DUI Subject Record	RESPONSE TIMEOUT
11/10/2012	01:46:36	DUI Subject Record	RESPONSE TIMEOUT
11/10/2012	02:00:32	DUI Subject Record	RESPONSE TIMEOUT
11/18/2012	01:20:37	DUI Subject Record	RESPONSE TIMEOUT
02/04/2013	04:57:58	RPC Record	SAMPLE ACCEPTANCE TEST FAILED
02/04/2013	18:17:53	RPC Record	SAMPLE ACCEPTANCE TEST FAILED
02/06/2013	14:00:18	RPC Record	ACCURACY AND PRECISION CHECK FAILED
02/17/2013	00:49:59	DUI Subject Record	Pump Error
02/17/2013	01:00:20	DUI Subject Record	RESPONSE TIMEOUT
03/09/2013	08:27:42	DUI Subject Record	RESPONSE TIMEOUT

Breath Record History

Test Date	Test Time	Agency	Operator Name	Alcohol Val. 1	Alcohol Val. 2	Sim. Conc. 1	Sim. Conc. 2	Status Message
04/06/2013	20:06:21	WOODSTOCK PD	MARK J FIORE	0.126		0.094		
05/11/2013	06:25:49	woodstock	PETER J MANTELLO	0.067		0.093		