

- Add a Simulator test (1 shot) to APM protocol
- Help menu which will contain our training manual
- Aborted RPC does not put unit out of service
- Ability to use previous timer on protocols
- Averaging glitch
- Standard deviation to 4 decimals
- Change background color when a password is in
- Update rev # for each update
- Training mode overrides timers
- Options reset does NOT change sim solution info
- Suck-back error does NOT abort sample acceptance test
- Temperature glitch
- Save certifications
- New passwords!!!
- Suck back error should not fail a sample acceptance test
- Ability to print to USB when making summery reports
- Option for allowing card swipe in the future
- Please add the following police agencies:
  - Capitol Police
  - Dept. of Liquor Control
  - Brighton Police
- Add the following towns:
  - Bellows Falls in Windham County
  - Brighton in Essex County
  - Island Pond in Essex County